



Agency Edgewater PD

S/N 80-001243

Date In 01/19/2022 DI Completion Date 01-31-2022

☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Calibration Adjustment

By _____

☐ Post Calibration Adjustment Stability Checks

Notes/Suggested Service: Department Inspection was started after changing breath hose. While doing Minimum Volume test it was noticed that air was leaking through breath hose. Inspection was aborted, a new breath hose was put on, and the Department Inspection restarted and completed with nominal values.

Department Inspection

By IS

Attachments

☒ Instrument Complies with Chapter 11D-8, FAC
☐ Instrument Does Not Comply with Chapter 11D-8, FAC

☒ Return to/Place into Evidentiary Use
☐ Remain Out of Evidentiary Use

☒ Conduct an Agency Inspection Before Evidentiary Use

Taylor
Gutschow

Digitally signed by Taylor
Gutschow
Date: 2022.02.03 09:01:07
-05'00'

Tech Review / Date

Admin Review / Date

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: EDGEWATER P.D.

Time of Inspection: 13:52

Date of Inspection: 01/31/2022

Serial Number: 80-001243

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202010A Exp: 10/05/2022	0.08g/210L Test (g/210L) Lot#:202010B Exp: 10/05/2022	0.20g/210L Test (g/210L) Lot#:202010D Exp: 10/06/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG113403 Exp: 05/14/2023
0.000	0.048	0.078	0.199	0.078
0.000	0.048	0.078	0.199	0.078
0.000	0.049	0.079	0.198	0.079
0.000	0.049	0.078	0.198	0.079
0.000	0.049	0.078	0.199	0.079
0.000	0.049	0.079	0.198	0.079
0.000	0.049	0.079	0.199	0.079
0.000	0.049	0.079	0.198	0.079
0.000	0.049	0.079	0.199	0.079
0.000	0.049	0.079	0.198	0.079


Standard Deviations	0.0004	0.0005	0.0005	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



ISRAEL SOTO

Signature and Printed Name

01/31/2022
Date

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: EDGEWATER P.D.

Time of Inspection: 12:15

Date of Inspection: 01/31/2022

Serial Number: 80-001243

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK		No	Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK		No
Alcohol Free Subject Test: 0.000		No	Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No	Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Standard Deviations				
---------------------	--	--	--	--

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: _____ Number of Simulators Used: _____

Remarks:


REPLACED HOSE, HOSE IS LEAKING AIR Non-compliance:HOSE SOUNDED LIKE IT WAS LEAKING.

Replaced breath hose prior to inspection, hose was leaking air during minimum volume test, stopped inspection and will repeat after replacing breath hose.



The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



ISRAEL SOTO

Signature and Printed Name

01/31/2022
Date

EDGEWATER P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001243
01/19/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:35
Control Test	0.049	13:36
Air Blank	0.000	13:37
Control Test	0.049	13:37
Air Blank	0.000	13:38
Control Test	0.050	13:39
Air Blank	0.000	13:39
Control Test Stats		
Average	0.0493	
Std Dev	0.0006	
Rel Std Dev(%)	1.1703	

.05

RAW

Operator's Signature

EDGEWATER P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001243
01/19/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:45
Control Test	0.078	13:46
Air Blank	0.000	13:47
Control Test	0.078	13:47
Air Blank	0.000	13:48
Control Test	0.077	13:49
Air Blank	0.000	13:49
Control Test Stats		
Average	0.0777	
Std Dev	0.0006	
Rel Std Dev(%)	0.7434	

.08
wct

RAW

Operator's Signature

EDGEWATER P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001243
01/19/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:51
Control Test	0.199	13:52
Air Blank	0.000	13:52
Control Test	0.198	13:53
Air Blank	0.000	13:53
Control Test	0.197	13:54
Air Blank	0.000	13:55
Control Test Stats		
Average	0.1980	
Std Dev	0.0010	
Rel Std Dev(%)	0.5051	

.20

RAW

Operator's Signature

EDGEWATER P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001243
01/19/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:05
Control Test	0.079	14:05
Air Blank	0.000	14:06
Control Test	0.078	14:06
Air Blank	0.000	14:07
Control Test	0.079	14:07
Air Blank	0.000	14:07
Control Test Stats		
Average	0.0787	
Std Dev	0.0006	
Rel Std Dev(%)	0.7339	

.08
DGS

RAW

Operator's Signature

Stability Checks

80-001243

01/19/22

RAW



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2729 Fort Knox Blvd.
Bldg. 2, Suite 1300
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001243, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001243</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>EDGEWATER P.D.</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>01/31/2022</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>13:52</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Israel Soto
Israel Soto
Date: 2022.01.31
14:24:42 -05'00'

01/31/2022

Date

ISRAEL SOTO,

Department Inspector

FDLE/ATP Form 69 January 2022

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Page 1 of 1