



INSTRUMENT PROCESSING SHEET

Agency Indian Harbour Beach Police Department

S/N 80-001263

Florida Department of
Law Enforcement

Date In 9/21/2022 DI Completion Date 9/22/2022

☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake By <u>DERR</u> ☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: <u>Instrument arrived in an FDLE loaner shipping box. DERR</u>	Quality Checks By <u>DERR</u> Date <u>9/22/2022</u> ☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>213</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP101</u> 32 mm <u>0.152</u> (.139 - .169) 36 mm <u>0.167</u> (.156 - .190) 53 mm <u>0.238</u> (.228 - .278) 103 mm <u>0.503</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28199</u> ☒ Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> <tr> <td>0.050</td> <td>MP6286</td> <td>202201C 01/11/2024</td> </tr> <tr> <td>0.080</td> <td>MP6287</td> <td>202201D 01/18/2024</td> </tr> <tr> <td>0.200</td> <td>MP6288</td> <td>202201E 01/18/2024</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>00521080A2 02/05/2023</td> </tr> </table>	Simulator	Serial #	Lot #/Exp	0.050	MP6286	202201C 01/11/2024	0.080	MP6287	202201D 01/18/2024	0.200	MP6288	202201E 01/18/2024	0.080 DGS	N/A	00521080A2 02/05/2023	Flow Calibration By _____ Date _____ Flow Column # _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																														
Simulator	Serial #	Lot #/Exp																																																													
0.050	MP6286	202201C 01/11/2024																																																													
0.080	MP6287	202201D 01/18/2024																																																													
0.200	MP6288	202201E 01/18/2024																																																													
0.080 DGS	N/A	00521080A2 02/05/2023																																																													
		Maintenance By <u>DERR</u> ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☒ Other <u>Gas tank tare before doing quality checks. DERR</u>																																																													
Calibration Adjustment By _____ Barometric Pressure Gauge _____ ID # _____ <table border="1" style="width:100%; border-collapse: collapse;"> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </table> ☐ Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </table>		Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Department Inspection By <u>DERR</u> Barometric Pressure ID# <u>26932</u> Gauge <u>1010</u> Instrument <u>1009</u> Mouth Alcohol Solution Lot # <u>2021-D</u> Acetone Stock Solution Lot # <u>2021-C</u> <table border="1" style="width:100%; border-collapse: collapse;"> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> <tr> <td>0.000</td> <td>MP6284</td> </tr> <tr> <td>Interferent</td> <td>MP6285</td> </tr> <tr> <td>0.050</td> <td>MP6286</td> </tr> <tr> <td>0.080</td> <td>MP6287</td> </tr> <tr> <td>0.200</td> <td>MP6288</td> </tr> </table>		Simulator	Serial Number	0.000	MP6284	Interferent	MP6285	0.050	MP6286	0.080	MP6287	0.200	MP6288
Simulator	Serial #	Lot #	Expiration																																																												
0.000		N/A	N/A																																																												
0.040																																																															
0.100																																																															
0.200																																																															
0.300																																																															
0.080 DGS	N/A																																																														
Simulator	Serial #	Lot #	Expiration																																																												
0.050																																																															
0.080																																																															
0.200																																																															
0.080 DGS	N/A																																																														
Simulator	Serial Number																																																														
0.000	MP6284																																																														
Interferent	MP6285																																																														
0.050	MP6286																																																														
0.080	MP6287																																																														
0.200	MP6288																																																														
Notes/Suggested Service: <u>Conducted Gas Tank Sensor Calibration Adjustment IAW Technical SOP 11.2 DERR.</u>		Attachments ☒ Form 41 ☐ Post-Stability Checks ☒ Stability Checks ☐ Flow Calibration ☒ Calibration Certificate ☐ Form 40 ☐ Calibration Adjustment ☐ Other _____																																																													
		☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																																																													
		Israel Soto Digitally signed by Israel Soto Date: 2022.09.22 09:32:30 -04'00' Tech Review / Date _____ Admin Review / Date <u>09/22/2022</u>																																																													



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001263, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001263</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>INDIAN HARBOR PD</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>09/22/2022</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>08:46</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.
This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

09/22/2022

Date


DAVID E REYES-RIVERA,

Department Inspector

FDLE/ATP Form 69 December 2021
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Page 1 of 1

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: INDIAN HARBOUR PD
Time of Inspection: 08:46

Date of Inspection: 09/22/2022

Serial Number: 80-001263
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:00521080A2 Exp: 02/05/2023
0.000	0.049	0.078	0.200	0.079
0.000	0.049	0.078	0.200	0.079
0.000	0.049	0.078	0.199	0.079
0.000	0.049	0.078	0.199	0.079
0.000	0.049	0.078	0.200	0.079
0.000	0.050	0.078	0.199	0.079
0.000	0.049	0.079	0.200	0.079
0.000	0.049	0.079	0.200	0.079
0.000	0.049	0.078	0.200	0.079
0.000	0.050	0.079	0.200	0.079

Standard Deviations	0.0004	0.0004	0.0004	0.0000
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

David E Reyes-Rivera
Signature and Printed Name

09/22/2022
Date

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-001263	Indian Harbour Beach Police Department	9/22/2022	DERR <i>[Signature]</i>

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																																
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083																																																																																																																																																
INDIAN HARBOUR PD Intoxilizer - Alcohol Analyzer Model 8000 SN 80-001263 09/22/2022 Software: 8100.27 <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>07:05</td></tr><tr><td>Control Test</td><td>0.049</td><td>07:05</td></tr><tr><td>Air Blank</td><td>0.000</td><td>07:06</td></tr><tr><td>Control Test</td><td>0.049</td><td>07:06</td></tr><tr><td>Air Blank</td><td>0.000</td><td>07:07</td></tr><tr><td>Control Test</td><td>0.049</td><td>07:08</td></tr><tr><td>Air Blank</td><td>0.000</td><td>07:08</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0490</td><td></td></tr><tr><td>Std Dev</td><td>0.0000</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr></table> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	07:05	Control Test	0.049	07:05	Air Blank	0.000	07:06	Control Test	0.049	07:06	Air Blank	0.000	07:07	Control Test	0.049	07:08	Air Blank	0.000	07:08	Control Test Stats			Average	0.0490		Std Dev	0.0000		Rel Std Dev(%)	0.0000		INDIAN HARBOUR PD Intoxilizer - Alcohol Analyzer Model 8000 SN 80-001263 09/22/2022 Software: 8100.27 <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>07:10</td></tr><tr><td>Control Test</td><td>0.078</td><td>07:10</td></tr><tr><td>Air Blank</td><td>0.000</td><td>07:11</td></tr><tr><td>Control Test</td><td>0.079</td><td>07:11</td></tr><tr><td>Air Blank</td><td>0.000</td><td>07:12</td></tr><tr><td>Control Test</td><td>0.078</td><td>07:12</td></tr><tr><td>Air Blank</td><td>0.000</td><td>07:13</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0783</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.7370</td><td></td></tr></table> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	07:10	Control Test	0.078	07:10	Air Blank	0.000	07:11	Control Test	0.079	07:11	Air Blank	0.000	07:12	Control Test	0.078	07:12	Air Blank	0.000	07:13	Control Test Stats			Average	0.0783		Std Dev	0.0006		Rel Std Dev(%)	0.7370		INDIAN HARBOUR PD Intoxilizer - Alcohol Analyzer Model 8000 SN 80-001263 09/22/2022 Software: 8100.27 <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>07:14</td></tr><tr><td>Control Test</td><td>0.199</td><td>07:15</td></tr><tr><td>Air Blank</td><td>0.000</td><td>07:15</td></tr><tr><td>Control Test</td><td>0.199</td><td>07:16</td></tr><tr><td>Air Blank</td><td>0.000</td><td>07:16</td></tr><tr><td>Control Test</td><td>0.199</td><td>07:17</td></tr><tr><td>Air Blank</td><td>0.000</td><td>07:18</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.1990</td><td></td></tr><tr><td>Std Dev</td><td>0.0000</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr></table> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	07:14	Control Test	0.199	07:15	Air Blank	0.000	07:15	Control Test	0.199	07:16	Air Blank	0.000	07:16	Control Test	0.199	07:17	Air Blank	0.000	07:18	Control Test Stats			Average	0.1990		Std Dev	0.0000		Rel Std Dev(%)	0.0000		INDIAN HARBOUR PD Intoxilizer - Alcohol Analyzer Model 8000 SN 80-001263 09/22/2022 Software: 8100.27 <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>07:19</td></tr><tr><td>Control Test</td><td>0.079</td><td>07:19</td></tr><tr><td>Air Blank</td><td>0.000</td><td>07:20</td></tr><tr><td>Control Test</td><td>0.080</td><td>07:20</td></tr><tr><td>Air Blank</td><td>0.000</td><td>07:20</td></tr><tr><td>Control Test</td><td>0.080</td><td>07:21</td></tr><tr><td>Air Blank</td><td>0.000</td><td>07:21</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0797</td><td></td></tr><tr><td>Std Dev</td><td>0.0005</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.7247</td><td></td></tr></table> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	07:19	Control Test	0.079	07:19	Air Blank	0.000	07:20	Control Test	0.080	07:20	Air Blank	0.000	07:20	Control Test	0.080	07:21	Air Blank	0.000	07:21	Control Test Stats			Average	0.0797		Std Dev	0.0005		Rel Std Dev(%)	0.7247	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	07:05																																																																																																																																																	
Control Test	0.049	07:05																																																																																																																																																	
Air Blank	0.000	07:06																																																																																																																																																	
Control Test	0.049	07:06																																																																																																																																																	
Air Blank	0.000	07:07																																																																																																																																																	
Control Test	0.049	07:08																																																																																																																																																	
Air Blank	0.000	07:08																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0490																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	07:10																																																																																																																																																	
Control Test	0.078	07:10																																																																																																																																																	
Air Blank	0.000	07:11																																																																																																																																																	
Control Test	0.079	07:11																																																																																																																																																	
Air Blank	0.000	07:12																																																																																																																																																	
Control Test	0.078	07:12																																																																																																																																																	
Air Blank	0.000	07:13																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0783																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7370																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	07:14																																																																																																																																																	
Control Test	0.199	07:15																																																																																																																																																	
Air Blank	0.000	07:15																																																																																																																																																	
Control Test	0.199	07:16																																																																																																																																																	
Air Blank	0.000	07:16																																																																																																																																																	
Control Test	0.199	07:17																																																																																																																																																	
Air Blank	0.000	07:18																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1990																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	07:19																																																																																																																																																	
Control Test	0.079	07:19																																																																																																																																																	
Air Blank	0.000	07:20																																																																																																																																																	
Control Test	0.080	07:20																																																																																																																																																	
Air Blank	0.000	07:20																																																																																																																																																	
Control Test	0.080	07:21																																																																																																																																																	
Air Blank	0.000	07:21																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0797																																																																																																																																																		
Std Dev	0.0005																																																																																																																																																		
Rel Std Dev(%)	0.7247																																																																																																																																																		