

INSTRUMENT PROCESSING SHEET

Agency Pinellas CSO

S/N 80-001274

Florida Department of
Law Enforcement

Date In 5/2/2022DI Completion Date 5/3/2022☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By TDG	Quality Checks	By TDG	Date 5/3/2022	Flow Calibration	By	Date																																																												
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes: _____ _____ _____ _____ _____ _____ _____ _____ _____ _____		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value 150 ☒ Flow Verification (L/s) Flow Column # ATP104 32 mm 0.144 (.139 - .169) 36 mm 0.164 (.156 - .190) 53 mm 0.230 (.228 - .278) 103 mm 0.496 (.447 - .547) ☒ Barometric Pressure Check Gauge ID # 68639 ☒ Stability Checks			Flow Column # _____ ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																																														
		<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP5092</td> <td>202010A 10/05/2022</td> </tr> <tr> <td>0.080</td> <td>MP5093</td> <td>202010B 10/05/2022</td> </tr> <tr> <td>0.200</td> <td>MP5094</td> <td>202010D 10/06/2022</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG115904 06/08/2023</td> </tr> </tbody> </table>			Simulator	Serial #	Lot #/Exp	0.050	MP5092	202010A 10/05/2022	0.080	MP5093	202010B 10/05/2022	0.200	MP5094	202010D 10/06/2022	0.080 DGS	N/A	AG115904 06/08/2023	<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Maintenance</th> <th>By</th> </tr> </thead> <tbody> <tr> <td>☐ Battery Replacement</td> <td></td> </tr> <tr> <td>☐ Dry Gas Regulator Replacement</td> <td></td> </tr> <tr> <td>☐ Breath Tube Replacement</td> <td></td> </tr> <tr> <td>☐ Other _____</td> <td></td> </tr> <tr> <td>_____</td> <td></td> </tr> <tr> <td>_____</td> <td></td> </tr> <tr> <td>_____</td> <td></td> </tr> </tbody> </table>			Maintenance	By	☐ Battery Replacement		☐ Dry Gas Regulator Replacement		☐ Breath Tube Replacement		☐ Other _____		_____		_____		_____																														
Simulator	Serial #	Lot #/Exp																																																																	
0.050	MP5092	202010A 10/05/2022																																																																	
0.080	MP5093	202010B 10/05/2022																																																																	
0.200	MP5094	202010D 10/06/2022																																																																	
0.080 DGS	N/A	AG115904 06/08/2023																																																																	
Maintenance	By																																																																		
☐ Battery Replacement																																																																			
☐ Dry Gas Regulator Replacement																																																																			
☐ Breath Tube Replacement																																																																			
☐ Other _____																																																																			

Calibration Adjustment By _____				Department Inspection By TDG _____																																																															
Barometric Pressure Gauge _____ ID # _____ <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> ☐ Post Calibration Adjustment Stability Checks <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>				Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Barometric Pressure ID# 28199 Gauge 1018 Instrument 1011 Mouth Alcohol Solution Lot # 2021-B Acetone Stock Solution Lot # 2021-C <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>SD3963</td> </tr> <tr> <td>Interferent</td> <td>SD3966</td> </tr> <tr> <td>0.050</td> <td>MP5092</td> </tr> <tr> <td>0.080</td> <td>MP5093</td> </tr> <tr> <td>0.200</td> <td>MP5094</td> </tr> </tbody> </table>				Simulator	Serial Number	0.000	SD3963	Interferent	SD3966	0.050	MP5092	0.080	MP5093	0.200	MP5094
Simulator	Serial #	Lot #	Expiration																																																																
0.000		N/A	N/A																																																																
0.040																																																																			
0.100																																																																			
0.200																																																																			
0.300																																																																			
0.080 DGS	N/A																																																																		
Simulator	Serial #	Lot #	Expiration																																																																
0.050																																																																			
0.080																																																																			
0.200																																																																			
0.080 DGS	N/A																																																																		
Simulator	Serial Number																																																																		
0.000	SD3963																																																																		
Interferent	SD3966																																																																		
0.050	MP5092																																																																		
0.080	MP5093																																																																		
0.200	MP5094																																																																		
Notes/Suggested Service: Checked breath tube screen and replaced o-rings on 5/2. (TDG) _____ _____ _____ _____ _____ _____ _____ _____ _____				Attachments ☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Calibration Adjustment ☐ Post-Stability Checks ☐ Flow Calibration ☐ Form 40 ☐ Other _____																																																															
				☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																																																															
David Eliezer Reyes-Rivera				Digitally signed by David Eliezer Reyes-Rivera Date: 2022.05.04 07:54:12 -04'00' 14:14:0																																																															
Tech Review / Date _____				Admin Review / Date _____																																																															

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-001274	Pineallas CSO	05/03/2022	TDG <i>ML</i>

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083
PINELLAS COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 05/03/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 10:40 Control Test 0.049 10:41 Air Blank 0.000 10:42 Control Test 0.049 10:42 Air Blank 0.000 10:43 Control Test 0.049 10:43 Air Blank 0.000 10:44 Control Test Stats Average 0.0490 Std Dev 0.0000 Rel Std Dev(%) 0.0000 <i>ML</i> Operator's Signature	PINELLAS COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 05/03/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 10:47 Control Test 0.079 10:48 Air Blank 0.000 10:49 Control Test 0.079 10:49 Air Blank 0.000 10:50 Control Test 0.078 10:50 Air Blank 0.000 10:51 Control Test Stats Average 0.0787 Std Dev 0.0006 Rel Std Dev(%) 0.7339 <i>ML</i> Operator's Signature	PINELLAS COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 05/03/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 10:54 Control Test 0.200 10:55 Air Blank 0.000 10:55 Control Test 0.199 10:56 Air Blank 0.000 10:56 Control Test 0.198 10:57 Air Blank 0.000 10:58 Control Test Stats Average 0.1990 Std Dev 0.0010 Rel Std Dev(%) 0.5025 <i>ML</i> Operator's Signature	<i>065</i> PINELLAS COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 05/03/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 10:59 Control Test 0.079 10:59 Air Blank 0.000 11:00 Control Test 0.080 11:00 Air Blank 0.000 11:01 Control Test 0.079 11:01 Air Blank 0.000 11:02 Control Test Stats Average 0.0793 Std Dev 0.0006 Rel Std Dev(%) 0.7277 <i>ML</i> Operator's Signature

Comments:

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: PINELLAS COUNTY SO
Time of Inspection: 13:46

Date of Inspection: 05/03/2022

Serial Number: 80-001274
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202010A Exp: 10/05/2022	0.08g/210L Test (g/210L) Lot#:202010B Exp: 10/05/2022	0.20g/210L Test (g/210L) Lot#:202010D Exp: 10/06/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG115904 Exp: 06/08/2023
0.000	0.049	0.079	0.199	0.078
0.000	0.049	0.079	0.200	0.079
0.000	0.049	0.079	0.199	0.078
0.000	0.049	0.079	0.199	0.079
0.000	0.049	0.079	0.200	0.079
0.000	0.049	0.079	0.199	0.079
0.000	0.049	0.079	0.199	0.079
0.000	0.049	0.079	0.200	0.078
0.000	0.049	0.080	0.199	0.078
0.000	0.049	0.079	0.199	0.078


Standard Deviations	0.0000	0.0003	0.0004	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

05/03/2022
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001274, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	80-001274	UNCERTAINTY* $\pm$
Owning Agency:	PINELLAS COUNTY SO	0.050 g/ 210 L 0.004
Calibration Date:	05/03/2022	0.080 g/ 210 L 0.004
Calibration Time:	13:46	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Taylor D Gutschow

05/03/2022

Date

TAYLOR D GUTSCHOW,

Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality