



Agency Jacksonville SO

S/N 80-001279

Date In 02/01/22

DI Completion Date 02/09/2022

Ship

☐ P/U

☐ H/D CMI

EE

Calibration Adjustment

By

☐ Post Calibration Adjustment Stability Checks

Notes/Suggested Service:

Department Inspection

By RAW

Attachments

- ☒ Instrument Complies with Chapter 11D-8, FAC
- ☐ Instrument Does Not Comply with Chapter 11D-8, FAC
- ☒ Return to/Place into Evidentiary Use
- ☐ Remain Out of Evidentiary Use
- ☒ Conduct an Agency Inspection Before Evidentiary Use

Taylor Gutschow Digitally signed by Taylor Gutschow
Date: 2022.02.11 09:26:00 -0500

Tech Review / Date

Admin Review Date 05/00

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: JACKSONVILLE SO
Time of Inspection: 14:12

Date of Inspection: 02/09/2022

Serial Number: 80-001279
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202010A Exp: 10/05/2022	0.08g/210L Test (g/210L) Lot#:202010B Exp: 10/05/2022	0.20g/210L Test (g/210L) Lot#:202010D Exp: 10/06/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG113403 Exp: 05/14/2023
0.000	0.050	0.081	0.202	0.078
0.000	0.050	0.081	0.203	0.078
0.000	0.050	0.081	0.203	0.077
0.000	0.050	0.081	0.203	0.078
0.000	0.050	0.081	0.202	0.077
0.000	0.050	0.081	0.202	0.077
0.000	0.051	0.081	0.202	0.077
0.000	0.051	0.081	0.203	0.077
0.000	0.051	0.082	0.202	0.077
0.000	0.051	0.082	0.202	0.077

Standard Deviations	0.0005	0.0004	0.0005	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 RICHARD A WILLIAMS
Signature and Printed Name

02/09/2022
Date

JACKSONVILLE SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001279
02/09/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:40
Control Test	0.049	10:40
Air Blank	0.000	10:41
Control Test	0.050	10:41
Air Blank	0.000	10:42
Control Test	0.050	10:43
Air Blank	0.000	10:43
Control Test Stats		
Average	0.0497	
Std Dev	0.0006	
Rel Std Dev(%)	1.1625	

.05


Operator's Signature

JACKSONVILLE SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001279
02/09/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:01
Control Test	0.080	11:01
Air Blank	0.000	11:02
Control Test	0.080	11:03
Air Blank	0.000	11:03
Control Test	0.080	11:04
Air Blank	0.000	11:05
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

.08


Operator's Signature

JACKSONVILLE SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001279
02/09/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:27
Control Test	0.202	11:27
Air Blank	0.000	11:28
Control Test	0.201	11:29
Air Blank	0.000	11:29
Control Test	0.200	11:30
Air Blank	0.000	11:30
Control Test Stats		
Average	0.2010	
Std Dev	0.0010	
Rel Std Dev(%)	0.4975	

.20


Operator's Signature

JACKSONVILLE SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001279
02/09/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:34
Control Test	0.078	11:34
Air Blank	0.000	11:34
Control Test	0.078	11:35
Air Blank	0.000	11:35
Control Test	0.079	11:36
Air Blank	0.000	11:36
Control Test Stats		
Average	0.0783	
Std Dev	0.0006	
Rel Std Dev(%)	0.7370	

.08
DGS


Operator's Signature

Stability Checks
80-001279
02/09/22




Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2729 Fort Knox Blvd.
Bldg. 2, Suite 1300
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001279, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001279</u>	Serial Number:	<u>80-001279</u>
Owning Agency:	<u>JACKSONVILLE SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>02/09/2022</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>14:12</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Richard A
Williams
Digitally signed by Richard
A Williams
Date: 2022.02.10 17:23:59
-05'00'

02/09/2022

Date

RICHARD A WILLIAMS,
Department Inspector

FDLE/ATP Form 69 January 2022

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality