

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: UNION COUNTY SO
Time of Inspection: 10:52

Date of Inspection: 08/18/2022

Serial Number: 80-001288
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:08121080A1 Exp: 05/05/2023
0.000	0.050	0.078	0.199	0.077
0.000	0.050	0.078	0.198	0.077
0.000	0.050	0.079	0.198	0.077
0.000	0.050	0.078	0.198	0.077
0.000	0.050	0.078	0.199	0.077
0.000	0.050	0.078	0.199	0.077
0.000	0.050	0.078	0.198	0.077
0.000	0.050	0.078	0.198	0.077
0.000	0.050	0.078	0.198	0.077
0.000	0.050	0.078	0.199	0.077


Standard Deviations	0.0000	0.0003	0.0005	0.0000
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0002 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



PHIL NICODEMO

Signature and Printed Name

08/18/2022
Date

Type of Test	Serial Number	Agency	Date	Performed By
Stability Checks	80-001288	Union County Sheriff's Office	8/18/2022	PN
0.05g/210L	0.047 to 0.053	0.08g/210L	0.20g/210L	DGS 0.08g/210L
✓	✓	✓	✓	✓
UNION COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001288 08/18/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 08:33 Control Test 0.050 08:33 Air Blank 0.000 08:34 Control Test 0.050 08:35 Air Blank 0.000 08:35 Control Test 0.049 08:36 Air Blank 0.000 08:36 Control Test Stats Average 0.0497 Std Dev 0.0006 Rel Std Dev(%) 1.1625 UNION COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001288 08/18/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 08:44 Control Test 0.199 08:45 Air Blank 0.000 08:45 Control Test 0.198 08:46 Air Blank 0.000 08:46 Control Test 0.199 08:47 Air Blank 0.000 08:48 Control Test Stats Average 0.1987 Std Dev 0.0006 Rel Std Dev(%) 0.2906 UNION COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001288 08/18/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 08:38 Control Test 0.079 08:39 Air Blank 0.000 08:40 Control Test 0.079 08:40 Air Blank 0.000 08:41 Control Test 0.079 08:41 Air Blank 0.000 08:42 Control Test Stats Average 0.0790 Std Dev 0.0000 Rel Std Dev(%) 0.0000 UNION COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001288 08/18/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 08:49 Control Test 0.077 08:50 Air Blank 0.000 08:50 Control Test 0.078 08:51 Air Blank 0.000 08:51 Control Test 0.077 08:51 Air Blank 0.000 08:52 Control Test Stats Average 0.0773 Std Dev 0.0006 Rel Std Dev(%) 0.7466 Operator's Signature	UNION COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001288 08/18/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 08:49 Control Test 0.077 08:50 Air Blank 0.000 08:50 Control Test 0.078 08:51 Air Blank 0.000 08:51 Control Test 0.077 08:51 Air Blank 0.000 08:52 Control Test Stats Average 0.0773 Std Dev 0.0006 Rel Std Dev(%) 0.7466 Operator's Signature	UNION COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001288 08/18/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 08:49 Control Test 0.077 08:50 Air Blank 0.000 08:50 Control Test 0.078 08:51 Air Blank 0.000 08:51 Control Test 0.077 08:51 Air Blank 0.000 08:52 Control Test Stats Average 0.0773 Std Dev 0.0006 Rel Std Dev(%) 0.7466 Operator's Signature	UNION COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001288 08/18/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 08:49 Control Test 0.077 08:50 Air Blank 0.000 08:50 Control Test 0.078 08:51 Air Blank 0.000 08:51 Control Test 0.077 08:51 Air Blank 0.000 08:52 Control Test Stats Average 0.0773 Std Dev 0.0006 Rel Std Dev(%) 0.7466 Operator's Signature	UNION COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001288 08/18/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 08:49 Control Test 0.077 08:50 Air Blank 0.000 08:50 Control Test 0.078 08:51 Air Blank 0.000 08:51 Control Test 0.077 08:51 Air Blank 0.000 08:52 Control Test Stats Average 0.0773 Std Dev 0.0006 Rel Std Dev(%) 0.7466 Operator's Signature
Comments: N/A				



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road.
Suite B1032
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001288, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001288</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>UNION COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>08/18/2022</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>10:52</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

FDLE/ATP Form 69 March 2022

Issuing Authority: Alcohol Testing Program

08/18/2022

Date

PHIL NICODEMO,

Department Inspector

Service • Integrity • Respect • Quality

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: UNION COUNTY SO
Time of Inspection: 08:26

Date of Inspection: 08/18/2022

Serial Number: 80-001288
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____

Number of Simulators Used: _____

Remarks:

BYPASSED AI TO OPERATE INSTRUMENT

COMPLIANCE NOT DETERMINED

PN 9/18/2022

The above instrument complies (☒) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

PHIL

PHIL NICODEMO

Signature and Printed Name

08/18/2022
Date