

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: LEVY COUNTY SO
Time of Inspection: 11:22

Date of Inspection: 10/05/2022

Serial Number: 80-001292
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:08121080A1 Exp: 05/05/2023
0.000	0.049	0.078	0.199	0.079
0.000	0.049	0.078	0.199	0.079
0.000	0.049	0.078	0.200	0.079
0.000	0.049	0.078	0.200	0.080
0.000	0.049	0.078	0.200	0.079
0.000	0.049	0.078	0.200	0.079
0.000	0.049	0.078	0.200	0.079
0.000	0.049	0.078	0.199	0.079
0.000	0.049	0.078	0.200	0.079
0.000	0.049	0.078	0.200	0.079

Standard Deviations	0.0000	0.0000	0.0004	0.0003
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0001 Number of Simulators Used: 5

Remarks:

*Middle Initial Character accidentally typed During inspection,
Corrected on printed FORM 41. P.N 10/5/2022*

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

PHIL

PHIL / NICODEMO

Signature and Printed Name

10/05/2022
Date

P.N 10/5/2022

Type of Test	Serial Number	Agency	Date	Performed By
Stability Checks	80-001292	Levy County Sheriff's Office	10/05/2022	PN

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083
LEUV COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 10/05/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 09:21 Control Test 0.049 09:22 Air Blank 0.000 09:22 Control Test 0.049 09:23 Air Blank 0.000 09:23 Control Test 0.049 09:24 Air Blank 0.000 09:25 Control Test Stats Average 0.0490 Std Dev 0.0000 Rel Std Dev(%) 0.0000 Operator's Signature 	LEUV COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 10/05/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 09:26 Control Test 0.078 09:27 Air Blank 0.000 09:27 Control Test 0.078 09:28 Air Blank 0.000 09:29 Control Test 0.078 09:29 Air Blank 0.000 09:30 Control Test Stats Average 0.0780 Std Dev 0.0000 Rel Std Dev(%) 0.0000 Operator's Signature 	LEUV COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 10/05/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 09:31 Control Test 0.199 09:32 Air Blank 0.000 09:32 Control Test 0.200 09:33 Air Blank 0.000 09:34 Control Test 0.200 09:34 Air Blank 0.000 09:35 Control Test Stats Average 0.1997 Std Dev 0.0006 Rel Std Dev(%) 0.2892 Operator's Signature 	LEUV COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 10/05/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 09:39 Control Test 0.078 09:39 Air Blank 0.000 09:40 Control Test 0.079 09:40 Air Blank 0.000 09:41 Control Test 0.079 09:41 Air Blank 0.000 09:41 Control Test Stats Average 0.0787 Std Dev 0.0006 Rel Std Dev(%) 0.7339 Operator's Signature 

Comments: N/A



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road.
Suite B1032
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001292, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001292</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>LEVY COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>10/05/2022</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>11:22</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Phil Nicodemo
Digitally signed by Phil Nicodemo
Date: 2022.10.05 12:08:31 -04'00'

10/05/2022

Date

PHIL - NICODEMO,
Department Inspector

FDLE/ATP Form 69 March 2022
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: LEVY COUNTY SO

Time of Inspection: 11:42

Date of Inspection: 01/04/2022

Serial Number: 80-001292

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202010A Exp: 10/05/2022	0.08g/210L Test (g/210L) Lot#:202010B Exp: 10/05/2022	0.20g/210L Test (g/210L) Lot#:202010D Exp: 10/06/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG113403 Exp: 05/14/2023
0.000	0.050	0.080	0.200	0.079
0.000	0.050	0.080	0.201	0.080
0.000	0.050	0.079	0.201	0.080
0.000	0.050	0.080	0.200	0.080
0.000	0.050	0.079	0.201	0.080
0.000	0.050	0.080	0.200	0.080
0.000	0.050	0.080	0.200	0.080
0.000	0.051	0.080	0.201	0.080
0.000	0.050	0.080	0.200	0.080
0.000	0.051	0.080	0.201	0.080


Standard Deviations	0.0004	0.0004	0.0005	0.0003
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



Signature and Printed Name

RICHARD A WILLIAMS

01/04/2022
Date

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: LEVY COUNTY SO
Time of Inspection: 13:35

Date of Inspection: 01/03/2022

Serial Number: 80-001292
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202010A Exp: 10/05/2022	0.08g/210L Test (g/210L) Lot#:202010B Exp: 10/05/2022	0.20g/210L Test (g/210L) Lot#:202010D Exp: 10/06/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG113403 Exp: 05/14/2023
0.000	0.047	/ 0.079	0.200	0.079
0.000	0.047	/ 0.079	0.200	0.079
0.000	0.047	/ 0.079	0.200	0.083
0.000	0.048	/ 0.079	0.200	0.083
0.000	0.047	/ 0.079	0.200	0.083
0.000	0.047	/ 0.079	0.201	0.083
0.000	0.047	/ 0.079	0.201	0.083
0.000	0.047	/ 0.079	0.201	0.083
0.000	0.048	/ 0.079	0.201	0.083
0.000	0.048	/ 0.079	0.201	0.083
Standard Deviations	0.0004	/ 0.0000	0.0005	0.0016


Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0006 Number of Simulators Used: 5

Remarks:

08: Ambient Fail, TOO MANY INSTRUMENTS IN ROOM.

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



Signature and Printed Name

RICHARD A WILLIAMS

01/03/2022
Date

LEUY COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001292
01/03/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	AMB*	10:01
Air Blank	0.000	10:01

*Ambient Fail

Operator's Signature

LEUY COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001292
01/03/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:02
Control Test	0.079	10:03
Air Blank	0.000	10:03
Control Test	0.080	10:03
Air Blank	0.000	10:04
Control Test	0.080	10:04
Air Blank	0.000	10:05
Control Test Stats		
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	

pre
08
wct

Operator's Signature

Pre-Cal Stability
Checks

80-001292

01/03/22

RAW

LEUY COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001292
01/03/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:11
Control Test	0.046	10:12
Air Blank	0.000	10:12
Control Test	0.047	10:13
Air Blank	0.000	10:13
Control Test	0.048	10:14
Air Blank	0.000	10:15
Control Test Stats		
Average	0.0470	
Std Dev	0.0010	
Rel Std Dev(%)	2.1277	

05
#1
pre

Operator's Signature

LEUY COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001292
01/03/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:25
Control Test	0.201	10:26
Air Blank	0.000	10:26
Control Test	0.201	10:27
Air Blank	0.000	10:27
Control Test	0.200	10:28
Air Blank	0.000	10:28
Control Test Stats		
Average	0.2007	
Std Dev	0.0006	
Rel Std Dev(%)	0.2877	

020
pre

Operator's Signature

LEUY COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001292
01/03/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:15
Control Test	0.048	10:16
Air Blank	0.000	10:16
Control Test	0.047	10:17
Air Blank	0.000	10:18
Control Test	0.048	10:18
Air Blank	0.000	10:19
Control Test Stats		
Average	0.0477	
Std Dev	0.0006	
Rel Std Dev(%)	1.2112	

pre
05
#2

Operator's Signature

LEUY COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001292
01/03/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:20
Control Test	0.079	10:21
Air Blank	0.000	10:21
Control Test	0.079	10:22
Air Blank	0.000	10:23
Control Test	0.078	10:23
Air Blank	0.000	10:24
Control Test Stats		
Average	0.0787	
Std Dev	0.0006	
Rel Std Dev(%)	0.7339	

08
DGS
pre

Operator's Signature

LEVY COUNT: 50
Intoxilyzer - Alcohol Analyzer
Model 8100 SN 80-001292
01/03/2022 15:06:09

Auto Calibration
Max Power Res Value = 35
Auto Range Res Value = 22

Sol Value = 0.000 g/210L ***
Fit value = 0.0000 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12944, Sum Io = 13752

Channel 1 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 0.1680 (-0.0220)
Sample #2 = 0.1780 (0.0070)
Sample #3 = 0.1450 (0.0190)
Sample #4 = 0.1670 (0.0070)
Avg % Abs = 0.1633 (0.0110)
STD DEV = 0.0168 (0.0069)
REL STD DEV = 10.287 (62.984)

Channel 2 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 0.1390 (-0.0090)
Sample #2 = 0.1330 (0.0050)
Sample #3 = 0.1380 (0.0210)
Sample #4 = 0.1370 (0.0120)
Avg % Abs = 0.1360 (0.0127)
STD DEV = 0.0026 (0.0080)
REL STD DEV = 1.945 (63.322)

Sol Value = 0.040 g/210L ***
Fit value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12938, Sum Io = 13748
Channel 1 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 0.9560 (-0.0290)
Sample #2 = 0.9630 (-0.0130)
Sample #3 = 0.9270 (-0.0080)
Sample #4 = 0.9190 (0.0250)
Avg % Abs = 0.9363 (0.0013)
STD DEV = 0.0234 (0.0206)
REL STD DEV = 2.503 (1548.588)

Channel 2 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 1.6310 (-0.0190)
Sample #2 = 1.5920 (0.0000)
Sample #3 = 1.5880 (0.0010)
Sample #4 = 1.5940 (0.0000)
Avg % Abs = 1.5913 (0.0003)
STD DEV = 0.0031 (0.0006)
REL STD DEV = 0.192 (173.205)

Sol Value = 0.100 g/210L ***
Fit value = 0.4762 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12934, Sum Io = 13745
Channel 1 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 2.0320 (-0.0070)
Sample #2 = 2.0390 (-0.0060)
Sample #3 = 2.0410 (0.0250)
Sample #4 = 2.0390 (0.0400)
Avg % Abs = 2.0397 (0.0197)
STD DEV = 0.0012 (0.0235)
REL STD DEV = 0.057 (119.284)

Channel 2 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 3.7300 (-0.0200)
Sample #2 = 3.7370 (-0.0160)
Sample #3 = 3.7550 (0.0010)
Sample #4 = 3.7450 (0.0160)
Avg % Abs = 3.7457 (0.0003)
STD DEV = 0.0090 (0.0160)
REL STD DEV = 0.241 (4803.124)

Sol Value = 0.200 g/210L ***
Fit value = 0.9524 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12928, Sum Io = 13744
Channel 1 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 3.8250 (-0.0120)
Sample #2 = 3.7980 (0.0170)
Sample #3 = 3.8090 (0.0370)
Sample #4 = 3.8330 (0.0190)
Avg % Abs = 3.8133 (0.0243)
STD DEV = 0.0179 (0.0110)
REL STD DEV = 0.469 (45.268)

Channel 2 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 7.1180 (-0.0060)
Sample #2 = 7.0610 (0.0620)
Sample #3 = 7.0710 (0.0650)
Sample #4 = 7.0880 (0.0560)
Avg % Abs = 7.0733 (0.0613)
STD DEV = 0.0137 (0.0050)
REL STD DEV = 0.193 (8.206)

Sol Value = 0.300 g/210L ***
Fit value = 1.4286 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12926, Sum Io = 13739
Channel 1 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 5.5330 (-0.0100)
Sample #2 = 5.5430 (-0.0100)
Sample #3 = 5.5210 (0.0270)
Sample #4 = 5.5480 (0.0140)
Avg % Abs = 5.5373 (0.0103)
STD DEV = 0.0144 (0.0188)
REL STD DEV = 0.259 (181.650)

Channel 2 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 10.2520 (0.0000)
Sample #2 = 10.2590 (-0.0090)
Sample #3 = 10.2600 (0.0170)
Sample #4 = 10.2380 (0.0170)
Avg % Abs = 10.2523 (0.0083)
STD DEV = 0.0124 (0.0150)
REL STD DEV = 0.121 (180.133)

**** AUTO CAL DATA ****

Channel 1 Data:
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.163
Std Dev = 0.02 Rel Std Dev = 10.29
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.935
Std Dev = 0.02 Rel Std Dev = 2.50
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 2.040
Std Dev = 0.00 Rel Std Dev = 0.06
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 3.813
Std Dev = 0.02 Rel Std Dev = 0.47
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 5.537
Std Dev = 0.01 Rel Std Dev = 0.25
Zero Order Coef = -424.12
First Order Coef = 2481.67
Second Order Coef = 31.84
Standard Deviation = 18.940121

Channel 2 Data:
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.136
Std Dev = 0.00 Rel Std Dev = 1.95
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 1.591
Std Dev = 0.00 Rel Std Dev = 0.19
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 3.746
Std Dev = 0.01 Rel Std Dev = 0.24
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 7.073
Std Dev = 0.01 Rel Std Dev = 0.19
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 10.252
Std Dev = 0.01 Rel Std Dev = 0.12
Zero Order Coef = -170.80
First Order Coef = 1273.43
Second Order Coef = 13.38
Standard Deviation = 17.344715

Optical Bench Calibration
(pt.1)

80-001292

01/03/22

RAW

Solution Stats Quadratic Fit Chan 1

Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	-0.000	0.0004
0.040	0.040	-0.0005
0.100	0.100	-0.0002
0.200	0.200	0.0005
0.300	0.300	-0.0002

Solution Stats Quadratic Fit Chan 2

Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	0.000	-0.0001
0.040	0.040	0.0003
0.100	0.101	-0.0005
0.200	0.200	0.0004
0.300	0.300	-0.0001

Sol Value = 0.060 g/210L ***
 Fit value = 0.3810 mg/l %%%
 Samples Taken = 4, Discarded = 1

***** CHANNEL 1

Sample #1 = 2782.00
 Sample #2 = 2935.00
 Sample #3 = 2910.00
 Sample #4 = 2806.00
 Average Result = 2883.6667
 STD DEV = 68.4130
 REL STD DEV = 2.372

***** CHANNEL 2

Sample #1 = 3251.00
 Sample #2 = 3304.00
 Sample #3 = 3350.00
 Sample #4 = 3311.00
 Average Result = 3321.6667
 STD DEV = 24.7857
 REL STD DEV = 0.746

Dry Gas H2O Adjust Results *****

Barometric Pressure = 1021
 3 um H2O Adjust (mg/l*10,000) = 926
 9 um H2O Adjust (mg/l*10,000) = 488

*** AUTO CAL PASS

LEUW COUNTY SC
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001292
 01/03/2022
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:39
Control Test	0.050	13:39
Air Blank	0.000	13:40
Control Test	0.047	13:41
Air Blank	0.000	13:41
Control Test	0.046	13:42
Air Blank	0.000	13:42
Control Test	0.047	13:43
Air Blank	0.000	13:44
Control Test	0.046	13:44
Air Blank	0.000	13:45
Control Test	0.046	13:46
Air Blank	0.000	13:46
Control Test	0.046	13:47
Air Blank	0.000	13:47
Control Test	0.046	13:48
Air Blank	0.000	13:49
Control Test	0.046	13:49
Air Blank	0.010	13:50
Control Test	0.046	13:50
Air Blank	0.000	13:51
Control Test	0.046	13:52
Air Blank	0.000	13:52
Control Test	0.046	13:53
Air Blank	0.000	13:53
Control Test	0.044	13:54
Air Blank	0.000	13:55
Control Test	0.045	13:55
Air Blank	0.000	13:56
Control Test	0.046	13:57
Air Blank	0.000	13:57
Control Test	0.047	13:58
Air Blank	0.000	13:58
Control Test	0.046	13:59
Air Blank	0.000	14:00
Control Test	0.046	14:00
Air Blank	0.000	14:01
Control Test	0.046	14:02
Air Blank	0.000	14:02
Control Test	0.048	14:03
Air Blank	0.000	14:03

Control Test Stats

Average 0.0463
 Std Dev 0.0012
 Rel Std Dev(%) 2.5363

105

RAW

Optical Bench Calibration

+

Extra Stability Checks

80-001292

01/03/22

RAW

Operator's Signature

LEUY COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001292
01/03/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	16:16
Control Test	0.049	16:17
Air Blank	0.000	16:18
Control Test	0.049	16:18
Air Blank	0.000	16:19
Control Test	0.048	16:20
Air Blank	0.000	16:20
Control Test Stats		
Average	0.0407	
Std Dev	0.0006	
Rel Std Dev(%)	1.1863	

0.05
post

RAW

Operator's Signature

LEUY COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001292
01/03/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	16:34
Control Test	0.079	16:35
Air Blank	0.000	16:35
Control Test	0.079	16:36
Air Blank	0.000	16:37
Control Test	0.079	16:37
Air Blank	0.000	16:38
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

0.08
wet
post

RAW

Operator's Signature

LEUY COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001292
01/03/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	16:06
Control Test	0.201	16:07
Air Blank	0.000	16:07
Control Test	0.198	16:08
Air Blank	0.000	16:09
Control Test	0.197	16:09
Air Blank	0.000	16:10
Control Test Stats		
Average	0.1987	
Std Dev	0.0021	
Rel Std Dev(%)	1.0478	

0.20
post

RAW

Operator's Signature

LEUY COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001292
01/03/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	15:59
Control Test	0.080	15:59
Air Blank	0.000	16:00
Control Test	0.080	16:00
Air Blank	0.000	16:01
Control Test	0.080	16:01
Air Blank	0.000	16:02
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

0.08
DGS
post

RAW

Operator's Signature

Post-Cal Stabilities

01/03/22

80-001292

RAW

LEVY COUNTY
Intoxilyzer
Model 8000
01/05/2022

Auto Calibration
Max Power Res Value = 56
Auto Range Res Value = 62

Sol Value = 0.000 g/210L ***
Fit value = 0.0000 mg/l %%%
Samples Taken = 4, Discarded =
Sum Io = 12407, Sum Io = 12407

Sample #1 = 0.1310 (0.1310)
Sample #2 = 2.1710 (0.1310)
Sample #3 = 0.3150 (0.1310)
Sample #4 = 0.3910 (0.1310)
Avg % Abs = 0.9590 (1.4410)
STD DEV = 1.0503 (1.1383)
REL STD DEV = 109.521 (79.028)

Sample #1 = 4.8670 (0.5680)
Sample #2 = 0.5880 (0.5880)
Sample #3 = -3.0810 (10.7520)
Sample #4 = 0.0420 (12.5520)
Avg % Abs = -0.8170 (9.9613)
STD DEV = 1.9796 (3.0635)
REL STD DEV = 242.301 (30.754)

Sol Value = 0.040 g/210L ***
Fit value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded =
Sum Io = 10972, Sum Io = 12763
Sample #1 = -7.7980 (5.9420)
Sample #2 = 2.2620 (-0.1730)
Sample #3 = 4.3050 (3.4110)
Sample #4 = -7.2210 (14.5110)
Avg % Abs = -0.2180 (5.9163)
STD DEV = 6.1502 (7.6559)
REL STD DEV = 2821.196 (129.402)

Sample #1 = -2.8710 (3.6110)
Sample #2 = 2.2530 (0.4110)
Sample #3 = 3.3030 (2.1750)
Sample #4 = -2.5850 (7.8510)
Avg % Abs = 0.9903 (3.4737)
STD DEV = 3.1405 (3.8796)
REL STD DEV = 317.118 (111.686)

**** AUTO CAL FAIL

LEVY COUNTY
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001292
01/05/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:27
Air Blank	0.000	12:28

**** AUTO CAL FAIL

Operator's Signature

LEVY COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001292
01/05/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:28
Control Test	0.000	12:29
Air Blank	0.000	12:30
Control Test	0.000	12:30
Air Blank	0.000	12:31
Control Test	0.000	12:31
Air Blank	0.000	12:32
Control Test Stats		
Average	0.0000	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature

Avg % Abs = 1.6110 (0.0000)
STD DEV = 1.0510 (0.0000)
REL STD DEV = 1.929 (65.808)

**** AUTO CAL FAIL

LEVY COUNTY
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001292
01/05/2022 11:59:49

Auto Calibration
Max Power Res Value = 56
Auto Range Res Value = 62

Sol Value = 0.000 g/210L ***
Fit value = 0.0000 mg/l %%%
Samples Taken = 4, Discarded =
Sum Io = 12407, Sum Io = 12407
Sample #1 = 0.1310 (-0.1310)
Sample #2 = 2.1710 (0.1310)
Sample #3 = 0.3150 (0.1310)
Sample #4 = 0.3910 (0.1310)
Avg % Abs = 0.9590 (1.4410)
STD DEV = 1.0503 (1.1383)
REL STD DEV = 109.521 (79.028)

Sample #1 = 0.1210 (-0.0160)
Sample #2 = 1.1230 (0.0720)
Sample #3 = 0.1720 (1.0580)
Sample #4 = 0.2400 (1.0560)
Avg % Abs = 0.5117 (0.7287)
STD DEV = 0.5305 (0.5687)
REL STD DEV = 103.665 (78.045)

Sol Value = 0.040 g/210L ***
Fit value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 11946, Sum Io = 13241
Sample #1 = 1.4530 (-0.1480)
Sample #2 = 0.7640 (0.5860)
Sample #3 = 0.5990 (0.4770)
Sample #4 = 1.1020 (1.3000)
Avg % Abs = 0.8217 (0.7877)
STD DEV = 0.2564 (0.4470)
REL STD DEV = 31.206 (56.753)

Sample #1 = 1.8400 (-0.0730)
Sample #2 = 1.4990 (0.3020)
Sample #3 = 1.4020 (0.2460)
Sample #4 = 1.6830 (0.6430)
Avg % Abs = 1.5280 (0.3970)
STD DEV = 0.1427 (0.2149)
REL STD DEV = 9.341 (34.125)

**** AUTO CAL FAIL

RAW 1/5/22
Richard Williams Optical Bench Cal
80-001292
P+1

Sol Value = 0.040 g/210L ***
Fit value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12077, Sum Io = 13311

<<<< CHANNEL 1 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1	0.9310	(-0.1010)
Sample #2	0.9430	(0.1350)
Sample #3	1.1030	(0.3040)
Sample #4	0.7830	(0.3510)
Avg % Abs	0.9430	(0.2633)
STD DEV	0.1600	(0.1136)
REL STD DEV	16.967	(43.138)

<<<< CHANNEL 2 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1	1.5560	(-0.0460)
Sample #2	1.5870	(0.0600)
Sample #3	1.6460	(0.1430)
Sample #4	1.5030	(0.1610)
Avg % Abs	1.5787	(0.1213)
STD DEV	0.0719	(0.0539)
REL STD DEV	4.552	(44.401)

Sol Value = 0.040 g/210L ***
Fit value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12639, Sum Io = 13716

<<<< CHANNEL 1 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1	0.0540	(0.0000)
Sample #2	0.0490	(0.2770)
Sample #3	0.1360	(0.3360)
Sample #4	0.1840	(0.5450)
Avg % Abs	0.1230	(0.3860)
STD DEV	0.0684	(0.1408)
REL STD DEV	55.636	(36.483)

<<<< CHANNEL 2 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1	0.0390	(-0.0180)
Sample #2	0.0260	(0.1160)
Sample #3	0.0820	(0.1410)
Sample #4	0.1080	(0.2410)
Avg % Abs	0.0720	(0.1660)
STD DEV	0.0419	(0.0661)
REL STD DEV	58.201	(39.846)

Sol Value = 0.040 g/210L ***
Fit value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12605, Sum Io = 13591

<<<< CHANNEL 1 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1	0.1520	(0.0430)
Sample #2	-0.0150	(0.0270)
Sample #3	0.1790	(-0.0480)
Sample #4	-0.0200	(0.0960)
Avg % Abs	0.0480	(0.0250)
STD DEV	0.1135	(0.0720)
REL STD DEV	236.410	(238.083)

<<<< CHANNEL 2 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1	0.0670	(0.0130)
Sample #2	-0.0240	(0.0020)
Sample #3	0.0700	(-0.0390)
Sample #4	0.0090	(0.0200)
Avg % Abs	0.0183	(-0.0057)
STD DEV	0.0477	(0.0382)
REL STD DEV	260.127	(533.611)

Intoximeter Alconol Analyzer
Model: 8500 SN 80-001292
01/05/2022 10:33:00

Auto Calibration
Max Lower Res Value = 51
Auto Range Res Value = 58

Sol Value = 0.040 g/210L ***
Fit value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12971, Sum Io = 13791

<<<< CHANNEL 1 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1	0.2200	(0.0000)
Sample #2	0.3000	(0.1350)
Sample #3	0.2940	(0.3730)
Sample #4	0.2950	(0.5730)
Avg % Abs	0.2963	(0.3803)
STD DEV	0.0032	(0.1891)
REL STD DEV	1.085	(49.721)

<<<< CHANNEL 2 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1	0.1220	(0.0010)
Sample #2	0.1930	(0.1150)
Sample #3	0.1910	(0.1900)
Sample #4	0.1950	(0.2870)
Avg % Abs	0.1930	(0.1973)
STD DEV	0.0020	(0.0862)
REL STD DEV	1.036	(43.700)

Sol Value = 0.040 g/210L ***
Fit value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12674, Sum Io = 13629

<<<< CHANNEL 1 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1	0.0400	(0.0150)
Sample #2	0.0250	(0.0580)
Sample #3	0.1750	(0.1160)
Sample #4	0.0740	(0.2570)
Avg % Abs	0.0913	(0.1437)
STD DEV	0.0765	(0.1023)
REL STD DEV	83.745	(71.237)

<<<< CHANNEL 2 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1	0.0210	(0.0150)
Sample #2	0.0160	(0.0130)
Sample #3	0.1020	(0.0440)
Sample #4	0.0380	(0.1230)
Avg % Abs	0.0520	(0.0600)
STD DEV	0.0447	(0.0567)
REL STD DEV	85.917	(94.531)

Sol Value = 0.040 g/210L ***
Fit value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12531, Sum Io = 13557

<<<< CHANNEL 1 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1	0.9630	(-0.0170)
Sample #2	0.9370	(0.1000)
Sample #3	1.0560	(0.1530)
Sample #4	1.0520	(0.2640)
Avg % Abs	1.0150	(0.1723)
STD DEV	0.0676	(0.0837)
REL STD DEV	6.658	(48.564)

<<<< CHANNEL 2 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1	1.6060	(-0.0060)
Sample #2	1.6010	(0.0580)
Sample #3	1.6500	(0.0780)
Sample #4	1.6240	(0.1630)
Avg % Abs	1.6250	(0.0997)
STD DEV	0.0245	(0.0558)
REL STD DEV	1.509	(55.939)

Optical Bench Cal

80-001292

1/5/22

RAW

Pt. 2

Sol Value = 0.100 g/210L ***
Fit value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12435, Sum Io = 13504

<<<< CHANNEL 1 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1	0.9170	(-0.0310)
Sample #2	1.0470	(0.0990)
Sample #3	0.9460	(0.3260)
Sample #4	0.9730	(0.5110)
Avg % Abs	0.9887	(0.3120)
STD DEV	0.0523	(0.2064)
REL STD DEV	5.289	(66.140)

<<<< CHANNEL 2 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1	1.5530	(0.0160)
Sample #2	1.6260	(0.0810)
Sample #3	1.5810	(0.1710)
Sample #4	1.5850	(0.2810)
Avg % Abs	1.5973	(0.1777)
STD DEV	0.0249	(0.1002)
REL STD DEV	1.559	(56.379)

Sol Value = 0.040 g/210L ***
Fit value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12283, Sum Io = 13419

<<<< CHANNEL 1 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1	1.0410	(-0.0160)
Sample #2	0.9380	(0.1710)
Sample #3	0.9360	(0.2110)
Sample #4	1.0230	(0.5170)
Avg % Abs	0.9657	(0.2997)
STD DEV	0.0497	(0.1893)
REL STD DEV	5.143	(63.162)

<<<< CHANNEL 2 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1	1.6170	(-0.0130)
Sample #2	1.6080	(0.0730)
Sample #3	1.5760	(0.0970)
Sample #4	1.6780	(0.2390)

LEVY COUNTY
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001292
01/05/2022 09:28:47

Auto Calibration
Max Power Res Value = 51
Auto Range Res Val

Sol Value = 0.100 g/210L ***
Fit value = 0.0000 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 13014, Sum Io = 13824

<<<< CHANNEL 1 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1	0.1970	(-0.0210)
Sample #2	0.1900	(0.1240)
Sample #3	0.1540	(0.1610)
Sample #4	0.2980	(0.0520)
Avg % Abs	0.2140	(0.1123)
STD DEV	0.0749	(0.0554)
REL STD DEV	35.019	(49.343)

<<<< CHANNEL 2 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1	0.1420	(-0.0080)
Sample #2	0.1610	(0.0770)
Sample #3	0.1450	(0.0900)
Sample #4	0.2130	(0.0660)
Avg % Abs	0.1730	(0.0777)
STD DEV	0.0356	(0.0120)
REL STD DEV	20.551	(15.469)

Sol Value = 0.040 g/210L ***
Fit value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12985, Sum Io = 13802

<<<< CHANNEL 1 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1	-0.0160	(0.0540)
Sample #2	-0.1710	(0.1250)
Sample #3	0.0280	(0.0270)
Sample #4	0.1140	(0.0090)
Avg % Abs	-0.0097	(0.0537)
STD DEV	0.1462	(0.0624)
REL STD DEV	1512.269	(116.327)

<<<< CHANNEL 2 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1	0.0300	(0.0170)
Sample #2	-0.0590	(0.0630)
Sample #3	0.0100	(0.0350)
Sample #4	0.0800	(0.0180)
Avg % Abs	0.0103	(0.0387)
STD DEV	0.0695	(0.0227)
REL STD DEV	672.586	(58.766)

Sol Value = 0.100 g/210L ***
Fit value = 0.4762 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12971, Sum Io = 13791

<<<< CHANNEL 1 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1	2.1160	(-0.0120)
Sample #2	2.0990	(0.1460)
Sample #3	1.9900	(0.2450)
Sample #4	2.0770	(0.2510)
Avg % Abs	2.0553	(0.2140)
STD DEV	0.0576	(0.0590)
REL STD DEV	2.804	(27.554)

<<<< CHANNEL 2 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1	3.7650	(-0.0020)
Sample #2	3.7710	(0.0470)
Sample #3	3.7190	(0.0890)
Sample #4	3.7990	(0.0730)
Avg % Abs	3.7630	(0.0697)
STD DEV	0.0406	(0.0212)
REL STD DEV	1.079	(28.477)

Sol Value = 0.200 g/210L ***
Fit value = 0.9524 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12923, Sum Io = 13771

<<<< CHANNEL 1 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1	3.9330	(-0.0100)
Sample #2	3.8660	(0.0720)
Sample #3	3.9210	(0.1510)
Sample #4	3.9390	(0.3320)
Avg % Abs	3.9087	(0.1850)
STD DEV	0.0380	(0.1333)
REL STD DEV	0.973	(72.050)

<<<< CHANNEL 2 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1	7.1890	(0.0070)
Sample #2	7.1370	(0.0340)
Sample #3	7.1320	(0.0730)
Sample #4	7.1870	(0.1760)
Avg % Abs	7.1520	(0.0943)
STD DEV	0.0304	(0.0734)
REL STD DEV	0.425	(77.771)

Optical Bench Cal.

80-001292

1/5/22

RAW
pt. 3

Sol Value = 0.300 g/210L ***
 Fit value = 1.4286 mg/l %%%
 Samples Taken = 4, Discarded = 1
 Sum Io = 12849, Sum Io = 13727

<<<< CHANNEL 1 >>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 5.6260 (0.0000)
 Sample #2 = 5.5780 (-0.0090)
 Sample #3 = 5.5330 (0.1810)
 Sample #4 = 5.8250 (0.2370)
 Avg % Abs = 5.6453 (0.1363)
 STD DEV = 0.1572 (0.1289)
 REL STD DEV = 2.785 (94.576)

<<<< CHANNEL 2 >>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 10.3560 (-0.0030)
 Sample #2 = 10.2380 (0.0040)
 Sample #3 = 10.3060 (0.0890)
 Sample #4 = 10.4230 (0.1180)
 Avg % Abs = 10.3223 (0.0703)
 STD DEV = 0.0936 (0.0592)
 REL STD DEV = 0.907 (84.239)

Optical Bench Cal

80-001292

1/5/22

PAW
 Pt. 9

 <<<< CHANNEL 1 >>>>
 Sol Val = 0.0000 mg/l or 0.000 g/210L
 % Abs = 0.214
 Std Dev = 0.07 Rel Std Dev = 35.02
 Sol Val = 0.1905 mg/l or 0.040 g/210L
 % Abs = -0.010
 Std Dev = 0.15 Rel Std Dev = 1512.27
 Sol Val = 0.4762 mg/l or 0.100 g/210L
 % Abs = 2.055
 Std Dev = 0.06 Rel Std Dev = 2.80
 Sol Val = 0.9524 mg/l or 0.200 g/210L
 % Abs = 3.909
 Std Dev = 0.04 Rel Std Dev = 0.97
 Sol Val = 1.4286 mg/l or 0.300 g/210L
 % Abs = 5.645
 Std Dev = 0.16 Rel Std Dev = 2.78
 Zero Order Coef = 837.07
 First Order Coef = 1635.57
 Second Order Coef = 135.53
 Standard Deviation = 815.266357

<<<< CHANNEL 2 >>>>
 Sol Val = 0.0000 mg/l or 0.000 g/210L
 % Abs = 0.173
 Std Dev = 0.04 Rel Std Dev = 20.55
 Sol Val = 0.1905 mg/l or 0.040 g/210L
 % Abs = 0.010
 Std Dev = 0.07 Rel Std Dev = 672.59
 Sol Val = 0.4762 mg/l or 0.100 g/210L
 % Abs = 3.763
 Std Dev = 0.04 Rel Std Dev = 1.08
 Sol Val = 0.9524 mg/l or 0.200 g/210L
 % Abs = 7.152
 Std Dev = 0.03 Rel Std Dev = 0.43
 Sol Val = 1.4286 mg/l or 0.300 g/210L
 % Abs = 10.322
 Std Dev = 0.09 Rel Std Dev = 0.91
 Zero Order Coef = 674.52
 First Order Coef = 913.30
 Second Order Coef = 38.04
 Standard Deviation = 733.567322

Solution Stats Quadratic Fit Chan 1

Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	0.025	-0.0251
0.040	0.017	0.0228
0.100	0.100	-0.0002
0.200	0.195	0.0047
0.300	0.302	-0.0022

Solution Stats Quadratic Fit Chan 2

g/210L	Fit	Residual
g/210L	g/210L	g/210L
0.000	0.022	-0.0227
0.040	0.019	0.0214
0.100	0.102	-0.0018
0.200	0.196	0.0036
0.300	0.301	-0.0015

**** AUTO END ****



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2729 Fort Knox Blvd.
Bldg. 2, Suite 1300
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001292, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001292</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>LEVY COUNTY SO</u>	0.050 g/ 210 L 0.005
Calibration Date:	<u>01/04/2022</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>11:42</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Richard A
Williams
Digitally signed by Richard A
Williams
Date: 2022.02.17 11:38:38
-05'00'

01/04/2022

Date

RICHARD A WILLIAMS,
Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Return Material Authorization

Ship to: ☒ CMI, Inc.

☐ Enforcement Electronics

Shipment to repair facility authorized by: Levy Co. SO on 1/5/22

Items Returned: Instrument ☒ Supplies ☐ Other ☐ Describe: _____

Instrument Model: Intoxilyzer 8000 Serial Number: 80-001292

Bill To Address:

Levy Co. SO

Ship to Address:

FDLE-ATP Tallahassee

Tallahassee, FL

Reason for Return:

Instrument source issues. Optical Bench Calibration fail.

Please choose one of the following options:

☐ 1. I _____, authorize all repairs.

☐ 2. I _____, authorize repairs up to \$ _____.

☒ 3. I require an estimate **BEFORE** any repairs will be authorized and/ or conducted.

Please contact: Name: Jason Levy

Phone #: 352-339-0143

Email: motorman190@gmail.com

ATP Contact Name: Richard Williams

ATP Email: richardwilliams@fdle.state.fl.us