



# INSTRUMENT PROCESSING SHEET

Agency Madison County Sheriff's OfficeS/N 80-001307

Florida Department of Law Enforcement

Date In 12/28/2021 DI Completion Date 1/4/2022 Ship  P/U  H/D  CMI  EE

<b>Intake</b> By <u>IS</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE  Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight  Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable  Notes: _____ _____ _____ _____ _____ _____ _____ _____ _____ _____	<b>Quality Checks</b> By <u>IS</u> Date <u>01-03-2022</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>153</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP-102</u> 32 mm <u>0.144</u> (.139 - .169) 36 mm <u>0.164</u> (.156 - .190) 53 mm <u>0.234</u> (.228 - .278) 103 mm <u>0.507</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>30793</u> <input checked="" type="checkbox"/> Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td rowspan="2">0.050</td> <td rowspan="2">MP5088</td> <td>202010A</td> </tr> <tr> <td>10-05-2022</td> </tr> <tr> <td rowspan="2">0.080</td> <td rowspan="2">MP5089</td> <td>202010B</td> </tr> <tr> <td>10-05-2022</td> </tr> <tr> <td rowspan="2">0.200</td> <td rowspan="2">MP5090</td> <td>202010D</td> </tr> <tr> <td>10-06-2022</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG113403</td> </tr> <tr> <td></td> <td></td> <td>05-14-2023</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	MP5088	202010A	10-05-2022	0.080	MP5089	202010B	10-05-2022	0.200	MP5090	202010D	10-06-2022	0.080 DGS	N/A	AG113403			05-14-2023	<b>Flow Calibration</b> By _____ Date _____ Flow Column # _____ <input type="checkbox"/> 5L/min – 17mm <input type="checkbox"/> 15L/min – 53mm <input type="checkbox"/> 30L/min – 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)  <b>Maintenance</b> By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____  <b>DI Temp. Checks</b> By <u>IS</u> <input checked="" type="checkbox"/> Lab Temp °C <u>20.96</u> External Digital Therm. ID#: <u>381189</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>MP5088</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>MP5089</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>MP5090</u>
Simulator	Serial #	Lot #/Exp																					
0.050	MP5088	202010A																					
		10-05-2022																					
0.080	MP5089	202010B																					
		10-05-2022																					
0.200	MP5090	202010D																					
		10-06-2022																					
0.080 DGS	N/A	AG113403																					
		05-14-2023																					

<b>Calibration Adjustment</b> By <u>IS</u> Barometric Pressure Gauge <u>1026</u> ID # <u>28421</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP6294</td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td>MP6295</td> <td>21070</td> <td>03-01-2023</td> </tr> <tr> <td>0.100</td> <td>MP6296</td> <td>21080</td> <td>03-08-2023</td> </tr> <tr> <td>0.200</td> <td>MP6297</td> <td>20510</td> <td>12-03-2022</td> </tr> <tr> <td>0.300</td> <td>MP6298</td> <td>21030</td> <td>02-02-2023</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>08121080A1</td> <td>05-05-2023</td> </tr> </tbody> </table> <input checked="" type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP5088</td> <td>202010A</td> <td>10-05-2022</td> </tr> <tr> <td>0.080</td> <td>MP5089</td> <td>202010B</td> <td>10-05-2022</td> </tr> <tr> <td>0.200</td> <td>MP5090</td> <td>202010D</td> <td>10-06-2022</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG113403</td> <td>05-14-2023</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #	Expiration	0.000	MP6294	N/A	N/A	0.040	MP6295	21070	03-01-2023	0.100	MP6296	21080	03-08-2023	0.200	MP6297	20510	12-03-2022	0.300	MP6298	21030	02-02-2023	0.080 DGS	N/A	08121080A1	05-05-2023	Simulator	Serial #	Lot #	Expiration	0.050	MP5088	202010A	10-05-2022	0.080	MP5089	202010B	10-05-2022	0.200	MP5090	202010D	10-06-2022	0.080 DGS	N/A	AG113403	05-14-2023	<b>Department Inspection</b> By <u>IS</u> Barometric Pressure ID# <u>30793</u> Gauge <u>1026</u> Instrument <u>1026</u> Mouth Alcohol Solution Lot # <u>2021-B</u> Acetone Stock Solution Lot # <u>2021-C</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP5086</td> </tr> <tr> <td>Interferent</td> <td>MP5087</td> </tr> <tr> <td>0.050</td> <td>MP5088</td> </tr> <tr> <td>0.080</td> <td>MP5089</td> </tr> <tr> <td>0.200</td> <td>MP5090</td> </tr> </tbody> </table>	Simulator	Serial Number	0.000	MP5086	Interferent	MP5087	0.050	MP5088	0.080	MP5089	0.200	MP5090
Simulator	Serial #	Lot #	Expiration																																																										
0.000	MP6294	N/A	N/A																																																										
0.040	MP6295	21070	03-01-2023																																																										
0.100	MP6296	21080	03-08-2023																																																										
0.200	MP6297	20510	12-03-2022																																																										
0.300	MP6298	21030	02-02-2023																																																										
0.080 DGS	N/A	08121080A1	05-05-2023																																																										
Simulator	Serial #	Lot #	Expiration																																																										
0.050	MP5088	202010A	10-05-2022																																																										
0.080	MP5089	202010B	10-05-2022																																																										
0.200	MP5090	202010D	10-06-2022																																																										
0.080 DGS	N/A	AG113403	05-14-2023																																																										
Simulator	Serial Number																																																												
0.000	MP5086																																																												
Interferent	MP5087																																																												
0.050	MP5088																																																												
0.080	MP5089																																																												
0.200	MP5090																																																												

Notes/Suggested Service: <u>More than 1% difference in pressure between gauge and instrument during Barometric Pressure Check.</u> <u>Performed Optical Bench Calibration Adjust.</u> <u>IS 01-04-2022</u> _____ _____ _____	<b>Attachments</b> <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Stability Checks <input checked="" type="checkbox"/> Calibration Certificate <input checked="" type="checkbox"/> Calibration Adjustment <input checked="" type="checkbox"/> Post-Stability Checks <input type="checkbox"/> Flow Calibration <input type="checkbox"/> Form 40 <input type="checkbox"/> Other _____  <input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use <div style="text-align: right;">  2022.01.0          6 12:13:39          05'00'          Tech Review / Date                      Admin Review / Date       </div>
--	--

# Florida Department of Law Enforcement Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: MADISON COUNTY SO  
Time of Inspection: 11:36

Date of Inspection: 01/04/2022

Serial Number: 80-001307  
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202010A Exp: 10/05/2022	0.08g/210L Test (g/210L) Lot#:202010B Exp: 10/05/2022	0.20g/210L Test (g/210L) Lot#:202010D Exp: 10/06/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG113403 Exp: 05/14/2023
0.000	0.049	0.079	0.198	0.080
0.000	0.049	0.078	0.197	0.079
0.000	0.049	0.078	0.197	0.080
0.000	0.048	0.078	0.197	0.080
0.000	0.049	0.078	0.197	0.080
0.000	0.050	0.078	0.197	0.080
0.000	0.049	0.079	0.197	0.080
0.000	0.049	0.078	0.196	0.080
0.000	0.049	0.078	0.197	0.080
0.000	0.049	0.079	0.197	0.080

Standard Deviations	0.0004	0.0004	0.0004	0.0003
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

1

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Israel Soto*

ISRAEL SOTO

Signature and Printed Name

01/04/2022  
Date

# stability checks

MADISON COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001307  
01/03/2022  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	15:31
Control Test	0.049	15:32
Air Blank	0.000	15:33
Control Test	0.049	15:33
Air Blank	0.000	15:34
Control Test	0.048	15:34
Air Blank	0.000	15:35
Control Test Stats		
Average	0.0487	
Std Dev	0.0006	
Rel Std Dev(%)	1.1853	

  
-----  
Operator's Signature

MADISON COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001307  
01/03/2022  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	15:36
Control Test	0.079	15:36
Air Blank	0.000	15:37
Control Test	0.078	15:38
Air Blank	0.000	15:38
Control Test	0.078	15:39
Air Blank	0.000	15:39
Control Test Stats		
Average	0.0783	
Std Dev	0.0006	
Rel Std Dev(%)	0.7370	

wet

  
-----  
Operator's Signature

MADISON COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001307  
01/03/2022  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	15:42
Control Test	0.197	15:42
Air Blank	0.000	15:43
Control Test	0.198	15:43
Air Blank	0.000	15:44
Control Test	0.197	15:45
Air Blank	0.000	15:45
Control Test Stats		
Average	0.1973	
Std Dev	0.0006	
Rel Std Dev(%)	0.2926	

  
-----  
Operator's Signature

MADISON COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001307  
01/03/2022  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	15:50
Control Test	0.080	15:50
Air Blank	0.000	15:50
Control Test	0.080	15:51
Air Blank	0.000	15:51
Control Test	0.081	15:52
Air Blank	0.000	15:52
Control Test Stats		
Average	0.0803	
Std Dev	0.0006	
Rel Std Dev(%)	0.7187	

Dry

  
-----  
Operator's Signature



# Calibration Certificate

Florida Department of Law Enforcement  
Alcohol Testing Program  
2729 Fort Knox Blvd.  
Bldg. 2, Suite 1300  
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001307, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001307</u>	UNCERTAINTY* ±	
Owning Agency:	<u>MADISON COUNTY SO</u>	0.050 g/ 210 L	0.005
Calibration Date:	<u>01/04/2022</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>11:36</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

\*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

## TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

**Israel Soto**  
Soto  
Date: 2022.01.04 12:00:38  
-05'00'

Digitally signed by Israel Soto

01/04/2022

Date

ISRAEL SOTO,

Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

# Optical Bench Cal. Adj.

MADISON COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001307  
01/04/2022 08:27:17

Auto Calibration  
Max Power Res Value = 39  
Auto Range Res Value = 33

Sol Value = 0.000 g/210L \*\*\*  
Fit value = 0.0000 mg/l %%%  
Samples Taken = 4, Discarded = 1  
Sum Io = 12849, Sum Io = 12730

<<<< CHANNEL 1 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1	0.1920	(-0.0110)
Sample #2	0.1540	(0.0110)
Sample #3	0.1550	(0.0410)
Sample #4	0.1660	(0.0130)
Avg % Abs	0.1583	(0.0217)
STD DEU	0.0067	(0.0168)
REL STD DEU	4.205	(77.414)

<<<< CHANNEL 2 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1	0.1310	(-0.0140)
Sample #2	0.1300	(-0.0190)
Sample #3	0.1200	(-0.0030)
Sample #4	0.1310	(-0.0120)
Avg % Abs	0.1270	(-0.0113)
STD DEU	0.0061	(0.0080)
REL STD DEU	4.790	(70.772)

Sol Value = 0.040 g/210L \*\*\*  
Fit value = 0.1905 mg/l %%%  
Samples Taken = 4, Discarded = 1  
Sum Io = 12846, Sum Io = 12730

<<<< CHANNEL 1 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1	0.9290	(-0.0110)
Sample #2	0.8900	(0.0050)
Sample #3	0.9180	(0.0050)
Sample #4	0.8900	(0.0130)
Avg % Abs	0.8993	(0.0077)
STD DEU	0.0162	(0.0046)
REL STD DEU	1.796	(60.245)

<<<< CHANNEL 2 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1	1.5680	(-0.0080)
Sample #2	1.5470	(-0.0060)
Sample #3	1.5470	(0.0000)
Sample #4	1.5290	(0.0100)
Avg % Abs	1.5410	(0.0013)
STD DEU	0.0104	(0.0081)
REL STD DEU	0.674	(605.218)

Sol Value = 0.100 g/210L \*\*\*  
Fit value = 0.4762 mg/l %%%  
Samples Taken = 4, Discarded = 1  
Sum Io = 12843, Sum Io = 12727

<<<< CHANNEL 1 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1	2.0100	(-0.0200)
Sample #2	1.9760	(-0.0140)
Sample #3	1.9830	(0.0070)
Sample #4	1.9650	(0.0220)
Avg % Abs	1.9747	(0.0050)
STD DEU	0.0091	(0.0181)
REL STD DEU	0.460	(361.663)

<<<< CHANNEL 2 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1	3.5990	(-0.0010)
Sample #2	3.5840	(-0.0020)
Sample #3	3.5880	(0.0100)
Sample #4	3.5700	(0.0120)
Avg % Abs	3.5807	(0.0067)
STD DEU	0.0095	(0.0076)
REL STD DEU	0.264	(113.578)

Sol Value = 0.200 g/210L \*\*\*  
Fit value = 0.9524 mg/l %%%  
Samples Taken = 4, Discarded = 1  
Sum Io = 12843, Sum Io = 12727

<<<< CHANNEL 1 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1	3.7690	(-0.0160)
Sample #2	3.7150	(0.0080)
Sample #3	3.7260	(0.0270)
Sample #4	3.7130	(0.0370)
Avg % Abs	3.7180	(0.0240)
STD DEU	0.0070	(0.0147)
REL STD DEU	0.188	(61.379)

<<<< CHANNEL 2 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1	6.8530	(-0.0020)
Sample #2	6.8410	(0.0020)
Sample #3	6.8330	(0.0180)
Sample #4	6.8200	(0.0200)
Avg % Abs	6.8313	(0.0133)
STD DEU	0.0105	(0.0099)
REL STD DEU	0.155	(73.993)

Sol Value = 0.300 g/210L \*\*\*  
Fit value = 1.4286 mg/l %%%  
Samples Taken = 4, Discarded = 1  
Sum Io = 12842, Sum Io = 12726

<<<< CHANNEL 1 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1	5.4400	(-0.0130)
Sample #2	5.4480	(0.0100)
Sample #3	5.4280	(0.0070)
Sample #4	5.4200	(0.0150)
Avg % Abs	5.4320	(0.0107)
STD DEU	0.0144	(0.0040)
REL STD DEU	0.266	(37.889)

<<<< CHANNEL 2 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1	9.9730	(-0.0250)
Sample #2	9.9460	(0.0110)
Sample #3	9.9380	(0.0120)
Sample #4	9.9250	(0.0150)
Avg % Abs	9.9363	(0.0127)
STD DEU	0.0106	(0.0021)
REL STD DEU	0.107	(16.434)

\*\*\*\*\* AUTO CAL DATA \*\*\*\*\*

<<<< CHANNEL 1 >>>>

Sol Val = 0.0010 mg/l or 0.000 g/210L  
% Abs = 0.158  
Std Dev = 0.01 Rel Std Dev = 4.21  
Sol Val = 0.1905 mg/l or 0.040 g/210L  
% Abs = 0.899  
Std Dev = 0.02 Rel Std Dev = 1.89  
Sol Val = 0.4762 mg/l or 0.100 g/210L  
% Abs = 1.975  
Std Dev = 0.01 Rel Std Dev = 0.46  
Sol Val = 0.9524 mg/l or 0.200 g/210L  
% Abs = 3.718  
Std Dev = 0.01 Rel Std Dev = 0.19  
Sol Val = 1.4286 mg/l or 0.300 g/210L  
% Abs = 5.432  
Std Dev = 0.01 Rel Std Dev = 0.27  
Zero Order Coef = -423.93  
First Order Coef = 2585.51  
Second Order Coef = 22.59  
Standard Deviation = 15.281347

<<<< CHANNEL 2 >>>>

Sol Val = 0.0000 mg/l or 0.000 g/210L  
% Abs = 0.127  
Std Dev = 0.01 Rel Std Dev = 4.79  
Sol Val = 0.1905 mg/l or 0.040 g/210L  
% Abs = 1.541  
Std Dev = 0.01 Rel Std Dev = 0.67  
Sol Val = 0.4762 mg/l or 0.100 g/210L  
% Abs = 3.581  
Std Dev = 0.01 Rel Std Dev = 0.26  
Sol Val = 0.9524 mg/l or 0.200 g/210L  
% Abs = 6.831  
Std Dev = 0.01 Rel Std Dev = 0.16  
Sol Val = 1.4286 mg/l or 0.300 g/210L  
% Abs = 9.936  
Std Dev = 0.01 Rel Std Dev = 0.11  
Zero Order Coef = -176.46  
First Order Coef = 1337.40  
Second Order Coef = 11.91  
Standard Deviation = 6.870618



Optical Bench  
Cal. Adj-  
*sd*

Post-stability checks

Solution Stats Quadratic Fit Chan 1		
Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	-0.000	0.0003
0.040	0.040	-0.0003
0.100	0.100	-0.0002
0.200	0.200	0.0004
0.300	0.300	-0.0001

Solution Stats Quadratic Fit Chan 2		
Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	-0.000	0.0001
0.040	0.040	-0.0002
0.100	0.100	-0.0001
0.200	0.200	0.0002
0.300	0.300	-0.0001

Sci Value = 0.080 g/210L \*\*\*  
 Fit value = 0.3810 mg/l %%%  
 Samples Taken = 4, Discarded = 1  
 \*\*\*\*\* CHANNEL 1  
 Sample #1 = 3007.00  
 Sample #2 = 3002.00  
 Sample #3 = 3001.00  
 Sample #4 = 3008.00  
 Average Result = 3003.6667  
 STD DEV = 3.7859  
 REL STD DEV = 0.126

\*\*\*\*\*  
 \*\*\*\*\* CHANNEL 2  
 Sample #1 = 3352.00  
 Sample #2 = 3383.00  
 Sample #3 = 3363.00  
 Sample #4 = 3379.00  
 Average Result = 3375.0000  
 STD DEV = 10.5830  
 REL STD DEV = 0.314  
 \*\*\*\*\*  
 Dry Gas H2O Adjust Results \*\*\*\*\*  
 Barometric Pressure = 1026  
 3 um H2O Adjust (mg/l\*10,000) = 806  
 9 um H2O Adjust (mg/l\*10,000) = 434  
 \*\*\* AUTO CAL PASS

MADISON COUNTY SO  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-001307  
 01/04/2022  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:40
Control Test	0.049	09:41
Air Blank	0.000	09:41
Control Test	0.049	09:42
Air Blank	0.000	09:42
Control Test	0.049	09:43
Air Blank	0.000	09:44
Control Test Stats		
Average	0.0490	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

*sd*  
 -----  
 Operator's Signature

MADISON COUNTY SO  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-001307  
 01/04/2022  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:45
Control Test	0.078	09:45
Air Blank	0.000	09:46
Control Test	0.078	09:46
Air Blank	0.000	09:47
Control Test	0.079	09:48
Air Blank	0.000	09:48
Control Test Stats		
Average	0.0783	
Std Dev	0.0006	
Rel Std Dev(%)	0.7370	

wet

*sd*  
 -----  
 Operator's Signature

MADISON COUNTY SO  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-001307  
 01/04/2022  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:49
Control Test	0.197	09:50
Air Blank	0.000	09:51
Control Test	0.198	09:51
Air Blank	0.000	09:52
Control Test	0.198	09:52
Air Blank	0.000	09:53
Control Test Stats		
Average	0.1977	
Std Dev	0.0006	
Rel Std Dev(%)	0.2921	

*sd*  
 -----  
 Operator's Signature

MADISON COUNTY SO  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-001307  
 01/04/2022  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:54
Control Test	0.080	09:54
Air Blank	0.000	09:55
Control Test	0.079	09:55
Air Blank	0.000	09:56
Control Test	0.080	09:56
Air Blank	0.000	09:56
Control Test Stats		
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	

Dry

*sd*  
 -----  
 Operator's Signature