



INSTRUMENT PROCESSING SHEET

Agency Okeechobee CSOS/N 80-001320Florida Department of
Law EnforcementDate In 3/23/2022 DI Completion Date 4/5/2022☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By TDG	Quality Checks	By TDG	Date <u>3/29/2022</u>	Flow Calibration	By	Date																																							
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: _____ _____ _____ _____ _____ _____ _____ _____		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>107</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP104</u> 32 mm <u>0.164</u> (.139 - .169) 36 mm <u>0.183</u> (.156 - .190) 53 mm <u>0.250</u> (.228 - .278) 103 mm <u>0.515</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>68639</u> ☒ Stability Checks			Flow Column # _____ ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																									
		<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP6286</td><td>202010A 10/05/2022</td></tr><tr><td>0.080</td><td>MP6287</td><td>202010B 10/05/2022</td></tr><tr><td>0.200</td><td>MP6288</td><td>202010D 10/06/2022</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG115904 06/08/2023</td></tr></tbody></table>			Simulator	Serial #	Lot #/Exp	0.050	MP6286	202010A 10/05/2022	0.080	MP6287	202010B 10/05/2022	0.200	MP6288	202010D 10/06/2022	0.080 DGS	N/A	AG115904 06/08/2023	Maintenance By _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____ _____ _____ _____ _____																										
Simulator	Serial #	Lot #/Exp																																												
0.050	MP6286	202010A 10/05/2022																																												
0.080	MP6287	202010B 10/05/2022																																												
0.200	MP6288	202010D 10/06/2022																																												
0.080 DGS	N/A	AG115904 06/08/2023																																												
Calibration Adjustment By _____		Department Inspection By <u>TDG</u>																																												
Barometric Pressure Gauge _____ ID # _____		Barometric Pressure ID# <u>28663</u> Gauge <u>1014</u> Instrument <u>1011</u> Mouth Alcohol Solution Lot # <u>2021-B</u> Acetone Stock Solution Lot # <u>2021-C</u>																																												
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP6284</td></tr><tr><td>Interferent</td><td>MP6285</td></tr><tr><td>0.050</td><td>MP6286</td></tr><tr><td>0.080</td><td>MP6287</td></tr><tr><td>0.200</td><td>MP6288</td></tr></tbody></table>					Simulator	Serial Number	0.000	MP6284	Interferent	MP6285	0.050	MP6286	0.080	MP6287	0.200	MP6288
Simulator	Serial #	Lot #	Expiration																																											
0.000		N/A	N/A																																											
0.040																																														
0.100																																														
0.200																																														
0.300																																														
0.080 DGS	N/A																																													
Simulator	Serial Number																																													
0.000	MP6284																																													
Interferent	MP6285																																													
0.050	MP6286																																													
0.080	MP6287																																													
0.200	MP6288																																													
☐ Post Calibration Adjustment Stability Checks		Attachments ☒ Form 41 ☐ Post-Stability Checks ☒ Stability Checks ☐ Flow Calibration ☒ Calibration Certificate ☐ Form 40 ☐ Calibration Adjustment ☐ Other _____																																												
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																								
Simulator	Serial #	Lot #	Expiration																																											
0.050																																														
0.080																																														
0.200																																														
0.080 DGS	N/A																																													
Notes/Suggested Service: _____ _____ _____ _____ _____ _____ _____ _____		Israel Soto Digitally signed by Israel Soto Date: 2022.04.05 13:08:23 +0400 14:53:33 Tech Review / Date _____ Admin Review / Date <u>04/06/2022</u>																																												

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-00 1320	Okeechobee CSD	03/29/2022	TDG <i>ML</i>

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083
✓	✓	✓	✓
OKEECHOBEE CO SO Intoxilyzer - Alcohol Analyzer Model 8000 03/29/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 14:12 Control Test 0.049 14:12 Air Blank 0.000 14:13 Control Test 0.049 14:13 Air Blank 0.000 14:14 Control Test 0.049 14:15 Air Blank 0.000 14:15 Control Test Stats Average 0.0490 Std Dev 0.0000 Rel Std Dev(%) 0.0000	OKEECHOBEE CO SO Intoxilyzer - Alcohol Analyzer Model 8000 03/29/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 14:28 Control Test 0.080 14:29 Air Blank 0.000 14:30 Control Test 0.079 14:30 Air Blank 0.000 14:31 Control Test 0.079 14:32 Air Blank 0.000 14:32 Control Test Stats Average 0.0793 Std Dev 0.0006 Rel Std Dev(%) 0.7277	OKEECHOBEE CO SO Intoxilyzer - Alcohol Analyzer Model 8000 03/29/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 14:53 Control Test 0.202 14:54 Air Blank 0.000 14:55 Control Test 0.202 14:55 Air Blank 0.000 14:56 Control Test 0.201 14:57 Air Blank 0.000 14:57 Control Test Stats Average 0.2017 Std Dev 0.0006 Rel Std Dev(%) 0.2863	DGS OKEECHOBEE CO SO Intoxilyzer - Alcohol Analyzer Model 8000 03/29/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 14:17 Control Test 0.078 14:17 Air Blank 0.000 14:18 Control Test 0.078 14:18 Air Blank 0.000 14:19 Control Test 0.078 14:19 Air Blank 0.000 14:19 Control Test Stats Average 0.0780 Std Dev 0.0000 Rel Std Dev(%) 0.0000
<i>ML</i> Operator's Signature	<i>ML</i> Operator's Signature	<i>ML</i> Operator's Signature	<i>ML</i> Operator's Signature

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: OKEECHOBEE CO SO
Time of Inspection: 12:51

Date of Inspection: 04/05/2022

Serial Number: 80-001320
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG115904 Exp: 06/08/2023
0.000	0.049	0.079	0.202	0.077
0.000	0.049	0.079	0.202	0.078
0.000	0.050	0.079	0.202	0.078
0.000	0.049	0.079	0.202	0.077
0.000	0.049	0.079	0.202	0.077
0.000	0.049	0.079	0.202	0.077
0.000	0.050	0.079	0.202	0.078
0.000	0.050	0.079	0.202	0.077
0.000	0.050	0.079	0.202	0.077
0.000	0.050	0.079	0.202	0.077
0.000	0.050	0.079	0.202	0.077
Standard Deviations	0.0005	0.0000	0.0000	0.0004

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0002 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

04/05/2022
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001320, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	80-001320	UNCERTAINTY* $\pm$
Owning Agency:	OKEECHOBEE CO SO	0.050 g/ 210 L 0.004
Calibration Date:	04/05/2022	0.080 g/ 210 L 0.004
Calibration Time:	12:51	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

04/05/2022

Date



TAYLOR D GUTSCHOW,

Department Inspector

Service • Integrity • Respect • Quality