



## INSTRUMENT PROCESSING SHEET

Agency Patrick Air Force BaseS/N 80-001471Florida Department of  
Law EnforcementDate In 12/9/2021 DI Completion Date 1/12/2022☒ Ship ☐ P/U ☐ H/D ☒ CMI ☐ EE

Intake	By <u>DERR</u>	Quality Checks	By <u>DERR</u>	Date <u>1/12/2022</u>	Flow Calibration	By <u>DERR</u>	Date <u>1/12/2022</u>
<input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE  Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight  Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable  Notes: _____ _____ _____ _____ _____ _____ _____ _____		<input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>172</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP106</u> 32 mm <u>0.125</u> (.139 - .169) 36 mm <u>0.140</u> (.156 - .190) 53 mm <u>0.207</u> (.228 - .278) 103 mm <u>0.488</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28663</u> <input checked="" type="checkbox"/> Stability Checks			Flow Column # <u>ATP104</u> <input checked="" type="checkbox"/> 5L/min - 17mm <input checked="" type="checkbox"/> 15L/min - 53mm <input checked="" type="checkbox"/> 30L/min - 103mm <input checked="" type="checkbox"/> R-Value <u>165</u> <input checked="" type="checkbox"/> Post Calibration Verification (L/s) Flow Column # <u>ATP106</u> 32 mm <u>0.140</u> (.139 - .169) 36 mm <u>0.160</u> (.156 - .190) 53 mm <u>0.230</u> (.228 - .278) 103 mm <u>0.503</u> (.447 - .547)		
					<b>Maintenance</b> By _____		
					<input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ _____ _____ _____ _____		

Calibration Adjustment				By _____	Department Inspection		By <u>DERR</u>	
Barometric Pressure Gauge _____ ID # _____					Barometric Pressure ID# <u>28199</u> Gauge <u>1024</u> Instrument <u>1022</u> Mouth Alcohol Solution Lot # <u>2021-D</u> Acetone Stock Solution Lot # <u>2021-C</u>			
Simulator	Serial #	Lot #	Expiration		Simulator	Serial Number		
0.000		N/A	N/A		0.000	MP6284		
0.040					Interferent	MP6285		
0.100					0.050	MP6286		
0.200					0.080	MP6287		
0.300					0.200	MP6288		
0.080 DGS	N/A							
<input type="checkbox"/> Post Calibration Adjustment Stability Checks					<b>Attachments</b>			
Simulator	Serial #	Lot #	Expiration		<input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Stability Checks <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Calibration Adjustment			<input type="checkbox"/> Post-Stability Checks <input checked="" type="checkbox"/> Flow Calibration <input type="checkbox"/> Form 40 <input type="checkbox"/> Other _____
0.050								
0.080								
0.200								
0.080 DGS	N/A							

Notes/Suggested Service: Calibrated flow to bring values closer to nominal. Instrument had an interferent detect during stabilities, unknown source, re-ran the solution on another stabilities test. During the Department Inspection, the instrument did not identify the interferent test. DERR

\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

- ☐ Instrument Complies with Chapter 11D-8, FAC  
☒ Instrument Does Not Comply with Chapter 11D-8, FAC  
☐ Return to/Place into Evidentiary Use  
☒ Remain Out of Evidentiary Use  
☐ Conduct an Agency Inspection Before Evidentiary Use

Tech Review / Date \_\_\_\_\_

Admin Review / Date \_\_\_\_\_

# Florida Department of Law Enforcement Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: PATRICK AFB SFS  
Time of Inspection: 10:35

Date of Inspection: 01/12/2022

Serial Number: 80-001471  
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect		No	Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:
0.000				
0.000				
0.000				
0.000				
0.000				
0.000				
0.000				
0.000				
0.000				
0.000				

Standard Deviations				
---------------------	--	--	--	--


Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: \_\_\_\_\_ Number of Simulators Used: 5

**Remarks:**

Int Det:DID NOT IDENTIFY THE INTERFERENT.Non-compliance:FAILED TO IDENTIFY INTERFERENT DETECT TEST..

The above instrument complies ( ) does not comply ( X ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 \_\_\_\_\_  
Signature and Printed Name

DAVID E REYES-RIVERA

01/12/2022  
Date




Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-001471	Patrick Air Force Base	1/12/2022	DERR <i>DELL</i>

0.05g/210L 0.047 to 0.053 <input checked="" type="checkbox"/>	0.08g/210L 0.077 to 0.083 <input checked="" type="checkbox"/>	0.20g/210L 0.194 to 0.206 <input checked="" type="checkbox"/>	DGS 0.08g/210L 0.077 to 0.083 <input checked="" type="checkbox"/>																																																																																																																																																
PATRICK AFB SFS Intoxiliizer - Alcohol Analyzer Model 8000 SN 80-001471 01/12/2022 Software: 8100.27	PATRICK AFB SFS Intoxiliizer - Alcohol Analyzer Model 8000 SN 80-001471 01/12/2022 Software: 8100.27	PATRICK AFB SFS Intoxiliizer - Alcohol Analyzer Model 8000 SN 80-001471 01/12/2022 Software: 8100.27	PATRICK AFB SFS Intoxiliizer - Alcohol Analyzer Model 8000 SN 80-001471 01/12/2022 Software: 8100.27																																																																																																																																																
<table border="1"> <thead> <tr> <th>Test</th><th>g/210L</th><th>Time</th></tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>08:53</td></tr> <tr><td>Control Test</td><td>0.053</td><td>08:54</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:54</td></tr> <tr><td>Control Test</td><td>0.052</td><td>08:55</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:55</td></tr> <tr><td>Control Test</td><td>0.053</td><td>08:56</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:57</td></tr> <tr><td>Control Test</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0527</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.0962</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	08:53	Control Test	0.053	08:54	Air Blank	0.000	08:54	Control Test	0.052	08:55	Air Blank	0.000	08:55	Control Test	0.053	08:56	Air Blank	0.000	08:57	Control Test			Average	0.0527		Std Dev	0.0006		Rel Std Dev(%)	1.0962		<table border="1"> <thead> <tr> <th>Test</th><th>g/210L</th><th>Time</th></tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>08:58</td></tr> <tr><td>Control Test</td><td>0.081</td><td>08:59</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:59</td></tr> <tr><td>Control Test</td><td>0.082</td><td>09:00</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:00</td></tr> <tr><td>Control Test</td><td>0.082</td><td>09:01</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:02</td></tr> <tr><td>Control Test</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0817</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7070</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	08:58	Control Test	0.081	08:59	Air Blank	0.000	08:59	Control Test	0.082	09:00	Air Blank	0.000	09:00	Control Test	0.082	09:01	Air Blank	0.000	09:02	Control Test			Average	0.0817		Std Dev	0.0006		Rel Std Dev(%)	0.7070		<table border="1"> <thead> <tr> <th>Test</th><th>g/210L</th><th>Time</th></tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>09:03</td></tr> <tr><td>Control Test</td><td>INT*</td><td>09:04</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:04</td></tr> <tr><td>Control Test</td><td>0.202</td><td>09:05</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:05</td></tr> <tr><td>Control Test</td><td>0.202</td><td>09:06</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:07</td></tr> <tr><td>Control Test</td><td></td><td></td></tr> <tr><td>Average</td><td>0.2003</td><td></td></tr> <tr><td>Std Dev</td><td>0.0029</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.4410</td><td></td></tr> </tbody> </table> <p>*Interferent Detect</p>	Test	g/210L	Time	Air Blank	0.000	09:03	Control Test	INT*	09:04	Air Blank	0.000	09:04	Control Test	0.202	09:05	Air Blank	0.000	09:05	Control Test	0.202	09:06	Air Blank	0.000	09:07	Control Test			Average	0.2003		Std Dev	0.0029		Rel Std Dev(%)	1.4410		<table border="1"> <thead> <tr> <th>Test</th><th>g/210L</th><th>Time</th></tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>09:09</td></tr> <tr><td>Control Test</td><td>0.084</td><td>09:09</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:10</td></tr> <tr><td>Control Test</td><td>0.085</td><td>09:10</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:11</td></tr> <tr><td>Control Test</td><td>0.084</td><td>09:11</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:12</td></tr> <tr><td>Control Test</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0843</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.5846</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	09:09	Control Test	0.084	09:09	Air Blank	0.000	09:10	Control Test	0.085	09:10	Air Blank	0.000	09:11	Control Test	0.084	09:11	Air Blank	0.000	09:12	Control Test			Average	0.0843		Std Dev	0.0006		Rel Std Dev(%)	0.5846	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	08:53																																																																																																																																																	
Control Test	0.053	08:54																																																																																																																																																	
Air Blank	0.000	08:54																																																																																																																																																	
Control Test	0.052	08:55																																																																																																																																																	
Air Blank	0.000	08:55																																																																																																																																																	
Control Test	0.053	08:56																																																																																																																																																	
Air Blank	0.000	08:57																																																																																																																																																	
Control Test																																																																																																																																																			
Average	0.0527																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	1.0962																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	08:58																																																																																																																																																	
Control Test	0.081	08:59																																																																																																																																																	
Air Blank	0.000	08:59																																																																																																																																																	
Control Test	0.082	09:00																																																																																																																																																	
Air Blank	0.000	09:00																																																																																																																																																	
Control Test	0.082	09:01																																																																																																																																																	
Air Blank	0.000	09:02																																																																																																																																																	
Control Test																																																																																																																																																			
Average	0.0817																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7070																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	09:03																																																																																																																																																	
Control Test	INT*	09:04																																																																																																																																																	
Air Blank	0.000	09:04																																																																																																																																																	
Control Test	0.202	09:05																																																																																																																																																	
Air Blank	0.000	09:05																																																																																																																																																	
Control Test	0.202	09:06																																																																																																																																																	
Air Blank	0.000	09:07																																																																																																																																																	
Control Test																																																																																																																																																			
Average	0.2003																																																																																																																																																		
Std Dev	0.0029																																																																																																																																																		
Rel Std Dev(%)	1.4410																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	09:09																																																																																																																																																	
Control Test	0.084	09:09																																																																																																																																																	
Air Blank	0.000	09:10																																																																																																																																																	
Control Test	0.085	09:10																																																																																																																																																	
Air Blank	0.000	09:11																																																																																																																																																	
Control Test	0.084	09:11																																																																																																																																																	
Air Blank	0.000	09:12																																																																																																																																																	
Control Test																																																																																																																																																			
Average	0.0843																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.5846																																																																																																																																																		
Operator's Signature <i>DELL</i>	Operator's Signature <i>DELL</i>	Operator's Signature <i>DELL</i>	Operator's Signature <i>DELL</i>																																																																																																																																																

PATRICK 9FB SFS  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001471  
01/12/2022  
Software: 8100.27


Flow Rate Calibration\*\*\*\*\*  
1: Rate (Liters/min) = 5  
SQRT(Diff) ) = 6.926  
2: Rate (Liters/min) = 15  
SQRT(Diff) ) = 11.746  
3: Rate (Liters/min) = 30  
SQRT(Diff) ) = 21.328  
Dependent Data Scale Factor = 100000 L/min  
Independent Data Scale Factor = 256  
Rounded Slope = 669  
Rounded Intercept = -615625  
Correlation = 0.99729

Flow Calibration	
SN:	80-001471
Agency:	Patrick Air Force Base
Date:	1/12/2022
By:	DERR 

PATRICK AFB SFS  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001471  
01/12/2022  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:00
Control Test	0.000	11:01
Air Blank	0.000	11:01
Control Test	INT*	11:02
Air Blank	0.000	11:02
Control Test	INT*	11:03
Air Blank	0.000	11:03
Control Test Stats		
Average	0.0000	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

\*Interferent Detect

  
Operator's Signature

Stabilities test of interferent detect post inspection failed to identify the first sample and recognized the following two samples.

Instrument will be kept out of evidentiary use and the agency will be notified.

DEK 1/12/2022

## Return Material Authorization

**Ship to:** ☒ CMI, Inc.

☐ Enforcement Electronics

Shipment to repair facility authorized by: TJ Graham on 3/8/2022

**Items Returned:** Instrument ☒ Supplies ☐ Other ☐ Describe: \_\_\_\_\_

Instrument Model: Intoxilyzer 8000 Serial Number: 80-001471

**Bill To Address:**

Sumter County Sheriff's Office

(PLEASE SEE NOTE BELOW)

**Ship to Address:**

Florida Department of Law Enforcement

Fort Myers Regional Operations Center

Attn: Alcohol Testing Program

4700 Terminal Drive, Suite 1

Fort Myers, FL 33907

**Reason for Return:**

Instrument intermittently fails to properly detect the presence of alcohol or incorrectly detects  
interferents when there are none. Please send quote to both FDLE and Sumter CSO.

**Please choose one of the following options:**

☐ 1. I \_\_\_\_\_, authorize all repairs.

☐ 2. I \_\_\_\_\_, authorize repairs up to \$\_\_\_\_\_.

☒ 3. I require an estimate **BEFORE** any repairs will be authorized and/ or conducted.

Please contact: Name: Terence Hollis

Phone #: 352-569-1710

Email: thollis@sumtercountysheriff.org

ATP Contact Name: Taylor Gutschow

ATP Email: TaylorGutschow@fdle.state.fl.us









Florida Department of  
Law Enforcement

## REQUEST FOR REGISTRATION

MAKE AND MODEL OF INSTRUMENT: Intoxilyzer 8000

SERIAL NUMBER: 80-001471

OWNING AGENCY: Sumter CSO

DATE OF DEPARTMENT INSPECTION: 11/15/2022

AGENCY INSPECTOR: Terence Hollis

ADDRESS: 219 E. Anderson Ave

CITY, STATE, ZIP: Bushnell, FL 33513

TELEPHONE NUMBER: 352-569-1710

FAX NUMBER: n/a

EMAIL ADDRESS (if available): thollis@sumtercountysheriff.org

**For Program Office Use Only:**

- ☐ Registration Issued
- ☒ Instrument Added to Evidentiary Instrument Database
- ☐ Instrument Added to Monthly Statistics Database
- ☒ Contact Information Added to Instrument Database

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-00 1471	Sumter CSO	11/08/2022	TDG MB

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053 ✓	0.077 to 0.083 ✓	0.194 to 0.206 ✓	0.077 to 0.083 <del>X</del> ≤0.003 of Wet ✓
SUMTER CSO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001471 11/08/2022 Software: 8100.27	SUMTER CSO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001471 11/08/2022 Software: 8100.27	SUMTER CSO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001471 11/08/2022 Software: 8100.27	DGS SUMTER CSO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001471 11/08/2022 Software: 8100.27
Test g/210L Time	Test g/210L Time	Test g/210L Time	Test g/210L Time
Air Blank 0.000 14:55	Air Blank 0.000 15:09	Air Blank 0.000 15:17	Air Blank 0.000 14:49
Control Test 0.049 14:56	Control Test 0.080 15:10	Control Test 0.198 15:17	Control Test 0.076 14:49
Air Blank 0.000 14:56	Air Blank 0.000 15:11	Air Blank 0.000 15:18	Air Blank 0.000 14:49
Control Test 0.048 14:57	Control Test 0.079 15:11	Control Test 0.196 15:19	Control Test 0.077 14:50
Air Blank 0.000 14:57	Air Blank 0.000 15:12	Air Blank 0.000 15:19	Air Blank 0.000 14:50
Control Test 0.048 14:58	Control Test 0.078 15:12	Control Test 0.197 15:20	Control Test 0.079 14:51
Air Blank 0.000 14:59	Air Blank 0.000 15:13	Air Blank 0.000 15:20	Air Blank 0.000 14:51
Control Test Stats	Control Test Stats	Control Test Stats	Control Test Stats
Average 0.0483	Average 0.0790	Average 0.1970	Average 0.0773
Std Dev 0.0006	Std Dev 0.0010	Std Dev 0.0010	Std Dev 0.0015
Rel Std Dev(%) 1.1945	Rel Std Dev(%) 1.2658	Rel Std Dev(%) 0.5076	Rel Std Dev(%) 1.9753
ML Operator's Signature	ML Operator's Signature	ML Operator's Signature	ML Operator's Signature

Comments: DGS Test Failed the Stability Checks. Will perform an optical cal adjust. MB 11/08/2022

SUMTER CSO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001471  
11/15/2022 09:15:45

Auto Calibration  
Max Power Res Value = 53  
Auto Range Res Value = 22

Sol Value = 0.000 g/210L \*\*\*  
Fit value = 0.0000 mg/l %%%  
Samples Taken = 4, Discarded = 1  
Sum Io = 12564, Sum Io = 12941

Channel 1  
Sample % Abs (% Abs Ref)  
Sample #1 = 0.2460 (0.0000)  
Sample #2 = 0.2180 (0.1240)  
Sample #3 = 0.2020 (0.2130)  
Sample #4 = 0.2060 (0.3190)  
Avg % Abs = 0.2087 (0.2187)  
STD DEV = 0.0083 (0.0976)  
REL STD DEV = 3.990 (44.645)

Channel 2  
Sample % Abs (% Abs Ref)  
Sample #1 = 0.3710 (-0.0220)  
Sample #2 = 0.3450 (0.0380)  
Sample #3 = 0.3530 (0.0560)  
Sample #4 = 0.3420 (0.0980)  
Avg % Abs = 0.3467 (0.0640)  
STD DEV = 0.0057 (0.0308)  
REL STD DEV = 1.640 (48.109)

Sol Value = 0.040 g/210L \*\*\*  
Fit value = 0.1905 mg/l %%%  
Samples Taken = 4, Discarded = 1  
Sum Io = 12495, Sum Io = 12915

Channel 1  
Sample % Abs (% Abs Ref)  
Sample #1 = 0.9320 (-0.0040)  
Sample #2 = 0.8990 (0.0550)  
Sample #3 = 0.8820 (0.1100)  
Sample #4 = 0.9090 (0.1280)  
Avg % Abs = 0.8967 (0.0977)  
STD DEV = 0.0137 (0.0380)  
REL STD DEV = 1.522 (38.939)

Channel 2  
Sample % Abs (% Abs Ref)  
Sample #1 = 1.7610 (-0.0120)  
Sample #2 = 1.7430 (0.0080)  
Sample #3 = 1.7210 (0.0610)  
Sample #4 = 1.7160 (0.0690)  
Avg % Abs = 1.7267 (0.0460)  
STD DEV = 0.0144 (0.0332)  
REL STD DEV = 0.832 (72.068)

Sol Value = 0.100 g/210L \*\*\*  
Fit value = 0.4762 mg/l %%%  
Samples Taken = 4, Discarded = 1  
Sum Io = 12460, Sum Io = 12901

Channel 1  
Sample % Abs (% Abs Ref)  
Sample #1 = 1.9900 (-0.0240)  
Sample #2 = 1.9550 (0.0190)  
Sample #3 = 1.9680 (0.0560)  
Sample #4 = 1.9450 (0.1010)  
Avg % Abs = 1.9560 (0.0587)  
STD DEV = 0.0115 (0.0411)  
REL STD DEV = 0.590 (69.997)

Channel 2  
Sample % Abs (% Abs Ref)  
Sample #1 = 3.8230 (-0.0130)  
Sample #2 = 3.7430 (0.0590)  
Sample #3 = 3.7830 (0.0550)  
Sample #4 = 3.7570 (0.0810)  
Avg % Abs = 3.7610 (0.0650)  
STD DEV = 0.0203 (0.0140)  
REL STD DEV = 0.540 (21.538)

Sol Value = 0.200 g/210L \*\*\*  
Fit value = 0.9524 mg/l %%%  
Samples Taken = 4, Discarded = 1  
Sum Io = 12424, Sum Io = 12878

Channel 1  
Sample % Abs (% Abs Ref)  
Sample #1 = 3.6630 (-0.0170)  
Sample #2 = 3.6670 (0.0080)  
Sample #3 = 3.6290 (0.0540)  
Sample #4 = 3.6530 (0.0750)  
Avg % Abs = 3.6497 (0.0457)  
STD DEV = 0.0192 (0.0343)  
REL STD DEV = 0.527 (75.041)

Channel 2  
Sample % Abs (% Abs Ref)  
Sample #1 = 7.0420 (-0.0150)  
Sample #2 = 6.9870 (0.0340)  
Sample #3 = 6.9760 (0.0590)  
Sample #4 = 6.9780 (0.0510)  
Avg % Abs = 6.9803 (0.0480)  
STD DEV = 0.0059 (0.0128)  
REL STD DEV = 0.084 (26.598)

Sol Value = 0.300 g/210L \*\*\*  
Fit value = 1.4286 mg/l %%%  
Samples Taken = 4, Discarded = 1  
Sum Io = 12395, Sum Io = 12868

Channel 1  
Sample % Abs (% Abs Ref)  
Sample #1 = 5.3360 (-0.0200)  
Sample #2 = 5.2800 (0.0440)  
Sample #3 = 5.3140 (0.0420)  
Sample #4 = 5.2830 (0.0620)  
Avg % Abs = 5.2923 (0.0493)  
STD DEV = 0.0188 (0.0110)  
REL STD DEV = 0.356 (22.328)

Channel 2  
Sample % Abs (% Abs Ref)  
Sample #1 = 10.1300 (-0.0120)  
Sample #2 = 10.0190 (0.0880)  
Sample #3 = 10.0530 (0.0850)  
Sample #4 = 10.0150 (0.1030)  
Avg % Abs = 10.0290 (0.0920)  
STD DEV = 0.0209 (0.0096)  
REL STD DEV = 0.208 (10.482)

Optical Calibration #1	
SN:	80-00 1471
Agency:	Sumter CSO
Date:	11/15/2022
Quadratic Fit:	+/- 0.002g/210L ✓
By:	TDG ML

AUTO CAL DATA  
Channel 1  
Sol Val = 0.0000 mg/l or 0.000 g/210L  
% Abs = 0.209  
Std Dev = 0.01 Rel Std Dev = 3.99  
Sol Val = 0.1905 mg/l or 0.040 g/210L  
% Abs = 0.897  
Std Dev = 0.01 Rel Std Dev = 1.52  
Sol Val = 0.4762 mg/l or 0.100 g/210L  
% Abs = 1.956  
Std Dev = 0.01 Rel Std Dev = 0.59  
Sol Val = 0.9524 mg/l or 0.200 g/210L  
% Abs = 3.650  
Std Dev = 0.02 Rel Std Dev = 0.53  
Sol Val = 1.4286 mg/l or 0.300 g/210L  
% Abs = 5.292  
Std Dev = 0.02 Rel Std Dev = 0.36  
Zero Order Coef = -538.51  
First Order Coef = 2665.80  
Second Order Coef = 25.43  
Standard Deviation = 19.888802

Channel 2  
Sol Val = 0.0000 mg/l or 0.000 g/210L  
% Abs = 0.347  
Std Dev = 0.01 Rel Std Dev = 1.64  
Sol Val = 0.1905 mg/l or 0.040 g/210L  
% Abs = 1.727  
Std Dev = 0.01 Rel Std Dev = 0.83  
Sol Val = 0.4762 mg/l or 0.100 g/210L  
% Abs = 3.761  
Std Dev = 0.02 Rel Std Dev = 0.54  
Sol Val = 0.9524 mg/l or 0.200 g/210L  
% Abs = 6.980  
Std Dev = 0.01 Rel Std Dev = 0.08  
Sol Val = 1.4286 mg/l or 0.300 g/210L  
% Abs = 10.029  
Std Dev = 0.02 Rel Std Dev = 0.21  
Zero Order Coef = -459.52  
First Order Coef = 1340.45  
Second Order Coef = 12.92  
Standard Deviation = 6.952269

Solution Stats Quadratic Fit Chan 1		
Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	0.000	-0.0004
0.040	0.039	0.0007
0.100	0.100	-0.0002
0.200	0.200	-0.0001
0.300	0.300	0.0001

Solution Stats Quadratic Fit Chan 2		
Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	0.000	-0.0001
0.040	0.040	0.0002
0.100	0.100	-0.0001
0.200	0.200	-0.0001
0.300	0.300	0.0000

Sol Value = 0.080 g/210L \*\*\*  
Fit value = 0.3810 mg/l %%%  
Samples Taken = 4, Discarded = 1  
Channel 1  
Sample #1 = 2829.00  
Sample #2 = 2804.00  
Sample #3 = 2772.00  
Sample #4 = 2871.00  
Average Result = 2815.6667  
STD DEV = 50.5206  
REL STD DEV = 1.794

Channel 2  
Sample #1 = 2928.00  
Sample #2 = 2880.00  
Sample #3 = 2885.00  
Sample #4 = 2927.00  
Average Result = 2897.3333  
STD DEV = 25.8134  
REL STD DEV = 0.891  
\*\*\*\*\*  
Dry Gas H2O Adjust Results \*\*\*\*\*  
Barometric Pressure = 1019  
3 um H2O Adjust (mg/l\*10,000) = 994  
9 um H2O Adjust (mg/l\*10,000) = 912  
\*\*\*\* AUTO CAL PASS



Type of Test	Serial Number	Agency	Date	Performed By
Stabilities (Post-Cal) #1	80-00 1471	Sumter CSO	11/15/2022	TDG MG

0.05g/210L			0.08g/210L			0.20g/210L			DGS 0.08g/210L		
0.047 to 0.053			0.077 to 0.083			0.194 to 0.206			0.077 to 0.083		
									≤0.003 of Wet		
SUMTER CSO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001471 11/15/2022 Software: 8100.27			SUMTER CSO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001471 11/15/2022 Software: 8100.27			SUMTER CSO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001471 11/15/2022 Software: 8100.27			SUMTER CSO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001471 11/15/2022 Software: 8100.27		
Test	g/210L	Time	Test	g/210L	Time	Test	g/210L	Time	Test	g/210L	Time
Air Blank	0.000	10:42	Air Blank	0.000	10:58	Air Blank	0.000	11:04	Air Blank	0.000	10:38
Control Test	0.049	10:42	Control Test	0.079	10:59	Control Test	0.199	11:05	Control Test	0.077	10:38
Air Blank	0.000	10:43	Air Blank	0.000	10:59	Air Blank	0.000	11:06	Air Blank	0.000	10:39
Control Test	0.047	10:44	Control Test	0.078	11:00	Control Test	0.198	11:06	Control Test	0.076	10:39
Air Blank	0.000	10:44	Air Blank	0.000	11:00	Air Blank	0.000	11:07	Air Blank	0.000	10:40
Control Test	0.048	10:45	Control Test	0.077	11:01	Control Test	0.198	11:08	Control Test	0.077	10:40
Air Blank	0.000	10:45	Air Blank	0.000	11:02	Air Blank	0.000	11:08	Air Blank	0.000	10:41
Control Test Stats			Control Test Stats			Control Test Stats			Control Test Stats		
Average	0.0480		Average	0.0780		Average	0.1983		Average	0.0767	
Std Dev	0.0010		Std Dev	0.0010		Std Dev	0.0006		Std Dev	0.0006	
Rel Std Dev(%)	2.0833		Rel Std Dev(%)	1.2821		Rel Std Dev(%)	0.2911		Rel Std Dev(%)	0.7531	
Operator's Signature			Operator's Signature			Operator's Signature			Operator's Signature		

Comments: DGS Test Failed the Stability Checks. Will repeat the optical cal adjust.  
MG 11/15/2022

Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	0.001	-0.0006
0.040	0.039	0.0007
0.100	0.100	0.0002
0.200	0.201	-0.0005
0.300	0.300	0.0002



Type of Test	Serial Number	Agency	Date	Performed By
Stabilities (Post-Cal) #2	80-00 1471	Sumter CSO	11/15/2022	TDG <i>ML</i>

0.05g/210L			0.08g/210L			0.20g/210L			DGS 0.08g/210L		
0.047 to 0.053			0.077 to 0.083			0.194 to 0.206			0.077 to 0.083		
									≤0.003 of Wet		
SUMTER CSO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001471 11/15/2022 Software: 8100.27			SUMTER CSO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001471 11/15/2022 Software: 8100.27			SUMTER CSO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001471 11/15/2022 Software: 8100.27			SUMTER CSO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001471 11/15/2022 Software: 8100.27		
Test	g/210L	Time	Test	g/210L	Time	Test	g/210L	Time	Test	g/210L	Time
Air Blank	0.000	12:13	Air Blank	0.000	12:20	Air Blank	0.000	12:27	Air Blank	0.000	12:32
Control Test	0.049	12:14	Control Test	0.079	12:21	Control Test	0.199	12:28	Control Test	0.079	12:32
Air Blank	0.000	12:15	Air Blank	0.000	12:22	Air Blank	0.000	12:29	Air Blank	0.000	12:33
Control Test	0.048	12:15	Control Test	0.078	12:23	Control Test	0.198	12:29	Control Test	0.080	12:33
Air Blank	0.000	12:16	Air Blank	0.000	12:23	Air Blank	0.000	12:30	Air Blank	0.000	12:34
Control Test	0.047	12:16	Control Test	0.078	12:24	Control Test	0.198	12:31	Control Test	0.080	12:34
Air Blank	0.000	12:17	Air Blank	0.000	12:24	Air Blank	0.000	12:31	Air Blank	0.000	12:34
Control Test Stats			Control Test Stats			Control Test Stats			Control Test Stats		
Average	0.0480		Average	0.0783		Average	0.1983		Average	0.0797	
Std Dev	0.0010		Std Dev	0.0006		Std Dev	0.0006		Std Dev	0.0006	
Rel Std Dev(%)	2.0833		Rel Std Dev(%)	0.7370		Rel Std Dev(%)	0.2911		Rel Std Dev(%)	0.7247	
Operator's Signature			Operator's Signature			Operator's Signature			Operator's Signature		

Comments:

# Florida Department of Law Enforcement Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: SUMTER CSO  
Time of Inspection: 14:54

Date of Inspection: 11/15/2022

Serial Number: 80-001471  
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:00521080A2 Exp: 02/05/2023
0.000	0.049	0.078	0.199	0.080
0.000	0.049	0.078	0.199	0.079
0.000	0.049	0.079	0.199	0.079
0.000	0.050	0.079	0.198	0.080
0.000	0.049	0.078	0.199	0.079
0.000	0.049	0.079	0.199	0.080
0.000	0.049	0.078	0.199	0.079
0.000	0.049	0.078	0.199	0.080
0.000	0.049	0.078	0.199	0.079
0.000	0.049	0.078	0.199	0.080
0.000	0.049	0.078	0.199	0.080
0.000	0.049	0.078	0.199	0.080
0.000	0.049	0.078	0.199	0.080
0.000	0.049	0.078	0.199	0.080
0.000	0.050	0.079	0.199	0.079
Standard Deviations	0.0004	0.0005	0.0003	0.0005

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies ( ☒ ) does not comply ( ☐ ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Taylor D Gutschow TAYLOR D GUTSCHOW  
Signature and Printed Name

11/15/2022  
Date





## Calibration Certificate

Florida Department of Law Enforcement  
Alcohol Testing Program  
4700 Terminal Drive, Suite 1  
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001471, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001471</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>SUMTER CSO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>11/15/2022</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>14:54</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within  $\pm 0.005$  or 5%, whichever is greater, of the target alcohol concentration.

\*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ( $k=3$ ).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

### TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

11/15/2022

Date

  
TAYLOR D GUTSCHOW,  
Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Page 1 of 1