



INSTRUMENT PROCESSING SHEET

Agency Cape Coral Police Department

S/N 80-001647

Florida Department of
Law Enforcement

Date In 3/21/2022 DI Completion Date 3/23/2022

☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By DERR	Quality Checks	By DERR	Date 3/23/2022	Flow Calibration	By	Date														
☒ Annual ☐ Registration ☒ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: <u>No form 51 on record</u> <u>DERR</u>		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>199</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP106</u> 32 mm <u>0.144</u> (.139 - .169) 36 mm <u>0.164</u> (.156 - .190) 53 mm <u>0.234</u> (.228 - .278) 103 mm <u>0.480</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28663</u> ☒ Stability Checks			Flow Column # _____ ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																
<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP6286</td> <td>202201C 01/11/2024</td> </tr> <tr> <td>0.080</td> <td>MP6287</td> <td>202201D 01/18/2024</td> </tr> <tr> <td>0.200</td> <td>MP6288</td> <td>202201E 01/18/2024</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG115904 06/08/2023</td> </tr> </tbody> </table>				Simulator	Serial #	Lot #/Exp	0.050	MP6286	202201C 01/11/2024	0.080	MP6287	202201D 01/18/2024	0.200	MP6288	202201E 01/18/2024	0.080 DGS	N/A	AG115904 06/08/2023	Maintenance By _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____		
Simulator	Serial #	Lot #/Exp																			
0.050	MP6286	202201C 01/11/2024																			
0.080	MP6287	202201D 01/18/2024																			
0.200	MP6288	202201E 01/18/2024																			
0.080 DGS	N/A	AG115904 06/08/2023																			

Calibration Adjustment	By DERR	Department Inspection	By DERR																																																												
Barometric Pressure Gauge <u>1014</u> ID # <u>68639</u> <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP5095</td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td>MP5096</td> <td>21070</td> <td>03/01/2023</td> </tr> <tr> <td>0.100</td> <td>MP5098</td> <td>21080</td> <td>03/08/2023</td> </tr> <tr> <td>0.200</td> <td>MP5100</td> <td>20510</td> <td>12/03/2022</td> </tr> <tr> <td>0.300</td> <td>MP5101</td> <td>21030</td> <td>02/02/2023</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>0052108A2</td> <td>06/08/2023</td> </tr> </tbody> </table> ☒ Post Calibration Adjustment Stability Checks <u>DERR 3/23/22</u> <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP6286</td> <td>202201C</td> <td>01/11/2024</td> </tr> <tr> <td>0.080</td> <td>MP6287</td> <td>202201D</td> <td>01/18/2024</td> </tr> <tr> <td>0.200</td> <td>MP6288</td> <td>202201E</td> <td>01/18/2024</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG115904</td> <td>06/08/2023</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #	Expiration	0.000	MP5095	N/A	N/A	0.040	MP5096	21070	03/01/2023	0.100	MP5098	21080	03/08/2023	0.200	MP5100	20510	12/03/2022	0.300	MP5101	21030	02/02/2023	0.080 DGS	N/A	0052108A2	06/08/2023	Simulator	Serial #	Lot #	Expiration	0.050	MP6286	202201C	01/11/2024	0.080	MP6287	202201D	01/18/2024	0.200	MP6288	202201E	01/18/2024	0.080 DGS	N/A	AG115904	06/08/2023		Barometric Pressure ID# <u>28199</u> Gauge <u>1015</u> Instrument <u>1014</u> Mouth Alcohol Solution Lot # <u>2021-D</u> Acetone Stock Solution Lot # <u>2021-C</u> <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP6284</td> </tr> <tr> <td>Interferent</td> <td>MP6285</td> </tr> <tr> <td>0.050</td> <td>MP6286</td> </tr> <tr> <td>0.080</td> <td>MP6287</td> </tr> <tr> <td>0.200</td> <td>MP6288</td> </tr> </tbody> </table>	Simulator	Serial Number	0.000	MP6284	Interferent	MP6285	0.050	MP6286	0.080	MP6287	0.200	MP6288	
Simulator	Serial #	Lot #	Expiration																																																												
0.000	MP5095	N/A	N/A																																																												
0.040	MP5096	21070	03/01/2023																																																												
0.100	MP5098	21080	03/08/2023																																																												
0.200	MP5100	20510	12/03/2022																																																												
0.300	MP5101	21030	02/02/2023																																																												
0.080 DGS	N/A	0052108A2	06/08/2023																																																												
Simulator	Serial #	Lot #	Expiration																																																												
0.050	MP6286	202201C	01/11/2024																																																												
0.080	MP6287	202201D	01/18/2024																																																												
0.200	MP6288	202201E	01/18/2024																																																												
0.080 DGS	N/A	AG115904	06/08/2023																																																												
Simulator	Serial Number																																																														
0.000	MP6284																																																														
Interferent	MP6285																																																														
0.050	MP6286																																																														
0.080	MP6287																																																														
0.200	MP6288																																																														
Notes/Suggested Service: <u>Instrument was calibrated to bring values closer to nominal. DERR</u> <u>Tech Review identified wrong lot # on cal DGS 00521080A2; expiration date 02/05/2023 DERR 3/23/2022</u>		<table border="1"> <thead> <tr> <th>Attachments</th> <th></th> </tr> </thead> <tbody> <tr> <td> ☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☒ Calibration Adjustment </td> <td> ☒ Post-Stability Checks ☐ Flow Calibration ☐ Form 40 ☐ Other _____ </td> </tr> </tbody> </table> ☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use <u>25</u> <u>10:43:50</u> <u>04:00</u>		Attachments		☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☒ Calibration Adjustment	☒ Post-Stability Checks ☐ Flow Calibration ☐ Form 40 ☐ Other _____																																																								
Attachments																																																															
☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☒ Calibration Adjustment	☒ Post-Stability Checks ☐ Flow Calibration ☐ Form 40 ☐ Other _____																																																														
Taylor Gutschow Digitally signed by Taylor Gutschow Date: 2022.03.24 08:41:49 -0400		Tech Review / Date _____ Admin. Review / Date _____																																																													



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001647, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001647</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>CAPE CORAL PD</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>03/23/2022</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>11:26</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.
This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

03/23/2022

Date

David E Reyes-Rivera
DAVID E REYES-RIVERA,

Department Inspector

Service • Integrity • Respect • Quality

Florida Department of Law Enforcement

Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: CAPE CORAL PD
Time of Inspection: 11:26

Date of Inspection: 03/23/2022

Serial Number: 80-001647
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG115904 Exp: 06/08/2023
0.000	0.050	0.079	0.199	0.080
0.000	0.049	0.079	0.199	0.079
0.000	0.050	0.079	0.199	0.079
0.000	0.049	0.080	0.200	0.079
0.000	0.049	0.079	0.199	0.080
0.000	0.049	0.080	0.199	0.079
0.000	0.050	0.080	0.200	0.079
0.000	0.050	0.079	0.199	0.079
0.000	0.050	0.079	0.199	0.079
0.000	0.049	0.079	0.199	0.078

Standard Deviations	0.0005	0.0004	0.0004	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

David E Reyes-Rivera

DAVID E REYES-RIVERA

Signature and Printed Name

03/23/2022
Date

Type of Test	Serial Number	Agency	Date	Performed By
Post Stabilities	80-001647	Cape Coral Police Department	3/23/2022	DERR <i>[Signature]</i>

0.05g/210L 0.047 to 0.053	0.08g/210L 0.077 to 0.083	0.20g/210L 0.194 to 0.206	DGS 0.08g/210L 0.077 to 0.083
CAPE CORAL PC Intoxilyzer - Alcohol Analyzer Model 8000 03/23/2022 Software: 8100.27 Test Time Air Blank 0.000 09:16 Control Test 0.049 09:17 Air Blank 0.000 09:17 Control Test 0.049 09:18 Air Blank 0.000 09:19 Control Test 0.049 09:19 Air Blank 0.000 09:20 Control Test 0.049 09:20 Average 0.0490 Std Dev 0.0000 Rel. Std Dev(%) 0.0000	CAPE CORAL PC Intoxilyzer - Alcohol Analyzer Model 8000 03/23/2022 Software: 8100.27 Test Time Air Blank 0.000 09:11 Control Test 0.079 09:12 Air Blank 0.000 09:12 Control Test 0.079 09:13 Air Blank 0.000 09:14 Control Test 0.079 09:15 Average 0.0790 Std Dev 0.0000 Rel. Std Dev(%) 0.0000	CAPE CORAL PC Intoxilyzer - Alcohol Analyzer Model 8000 03/23/2022 Software: 8100.27 Test Time Air Blank 0.000 09:16 Control Test 0.200 09:16 Air Blank 0.000 09:17 Control Test 0.199 09:18 Air Blank 0.000 09:18 Control Test 0.199 09:19 Air Blank 0.000 09:19 Control Test 0.199 09:19 Average 0.1993 Std Dev 0.0000 Rel. Std Dev(%) 0.2996	CAPE CORAL PC Intoxilyzer - Alcohol Analyzer Model 8000 03/23/2022 Software: 8100.27 Test Time Air Blank 0.000 09:21 Control Test 0.080 09:21 Air Blank 0.000 09:21 Control Test 0.080 09:22 Air Blank 0.000 09:22 Control Test 0.080 09:22 Air Blank 0.000 09:23 Control Test 0.080 09:23 Average 0.0800 Std Dev 0.0000 Rel. Std Dev(%) 0.0000
<i>[Signature]</i> Operator's Signature	<i>[Signature]</i> Operator's Signature	<i>[Signature]</i> Operator's Signature	<i>[Signature]</i> Operator's Signature

Solution Stats Quadratic Fit Chan 2
Act Fit Residual
g/210L g/210L g/210L
0.000 0.000 -0.0002
0.040 0.040 0.0003
0.430 0.100 -0.0000
0.200 0.200 -0.0002
0.300 0.300 0.0001

***** AUTO CAL DATA *****
<<<<< CHANNEL 1 >>>>>
Sol Val = 0.300 mg/l or 0.300 g/210L
% RDS = 0.147
Std Dev = 0.01 Rel Std Dev = 7.65
Sol Val = 0.105 mg/l or 0.105 g/210L
% RDS = 0.841
Std Dev = 0.02 Rel Std Dev = 2.07
% RDS = 1.910
Std Dev = 0.01 Rel Std Dev = 0.41
Sol Val = 0.352 mg/l or 0.352 g/210L
% RDS = 3.605
Std Dev = 0.01 Rel Std Dev = 0.34
Sol Val = 1.426 mg/l or 0.330 g/210L
% RDS = 5.265
Std Dev = 0.00 Rel Std Dev = 0.09
Zero Order Coef = -377.77
First Order Coef = 2556.59
Second Order Coef = 24.42
Standard Deviation = 20.55753

***** CHANNEL 2 *****
Sample % RDS (% Abs Ref)
Sample #1 = 5.8920 (-0.0130)
Sample #2 = 6.8390 (0.0090)
Sample #3 = 6.8920 (0.0080)
Sample #4 = 6.8220 (0.0080)
Avg % RDS = 6.8543 (0.0250)
STD DEV = 0.0328 (0.0154)
REL STD DEV = 0.479 (61.579)

***** CHANNEL 2 *****
Sample % RDS (% Abs Ref)
Sample #1 = 1.5750 (0.0000)
Sample #2 = 1.5810 (0.0190)
Sample #3 = 1.5640 (0.0170)
Sample #4 = 1.5660 (0.0280)
Avg % RDS = 1.5683 (0.0213)
STD DEV = 0.0112 (0.0350)
REL STD DEV = 0.711 (27.466)

CAPE CORAL PD
Intoxilyzer - 4, Control Analyzer
Model 8000
SN 82-011647
3/23/2022 07:47:36

Sol Value = 0.380 g/210L ***
Fit Value = 0.3810 mg/l ****
Samples Taken = 4, Discarded = 1
***** CHANNEL 1 *****
Sample #1 = 2727.00
Sample #2 = 2776.00
Sample #3 = 2647.00
Sample #4 = 2708.00
Average Result = 2710.3333
STD DEV = 64.5316
REL STD DEV = 2.381

***** CHANNEL 2 *****
Sample #1 = 3044.00
Sample #2 = 3075.00
Sample #3 = 3056.00
Sample #4 = 3050.00
Average Result = 3063.6667
STD DEV = 12.6623
REL STD DEV = 0.413

***** CHANNEL 2 *****
Sample % RDS (% Abs Ref)
Sample #1 = 9.5430 (-0.0110)
Sample #2 = 9.9360 (-0.0050)
Sample #3 = 9.9170 (0.0110)
Sample #4 = 9.9020 (0.0100)
Avg % RDS = 9.8153 (0.0032)
STD DEV = 0.0140 (0.0039)
REL STD DEV = 0.141 (155.775)

***** CHANNEL 2 *****
Sample % RDS (% Abs Ref)
Sample #1 = 3.6100 (0.0120)
Sample #2 = 3.6080 (0.0340)
Sample #3 = 3.6100 (0.0320)
Sample #4 = 3.6220 (0.0220)
Avg % RDS = 3.6133 (0.0293)
STD DEV = 0.0076 (0.0264)
REL STD DEV = 0.210 (21.917)

***** CHANNEL 2 *****
Sample % RDS (% Abs Ref)
Sample #1 = 0.1580 (0.0010)
Sample #2 = 0.1770 (0.0190)
Sample #3 = 0.1890 (0.0170)
Sample #4 = 0.1900 (0.0190)
Avg % RDS = 0.1683 (0.0170)
STD DEV = 0.0072 (0.0260)
REL STD DEV = 0.420 (11.765)

Sol Value = 0.200 g/210L ***
Fit Value = 0.1955 mg/l ****
Samples Taken = 4, Discarded = 1
***** CHANNEL 1 *****
Sample #1 = 0.1980 (0.0170)
Sample #2 = 0.1970 (0.0290)
Sample #3 = 0.1950 (0.0300)
Sample #4 = 0.1940 (0.0240)
Avg % RDS = 0.1963 (0.0243)
STD DEV = 0.0072 (0.0260)
REL STD DEV = 0.360 (11.765)

Sol Value = 0.200 g/210L ***
Fit Value = 0.1955 mg/l ****
Samples Taken = 4, Discarded = 1
***** CHANNEL 1 *****
Sample #1 = 0.1980 (0.0170)
Sample #2 = 0.1970 (0.0290)
Sample #3 = 0.1950 (0.0300)
Sample #4 = 0.1940 (0.0240)
Avg % RDS = 0.1963 (0.0243)
STD DEV = 0.0072 (0.0260)
REL STD DEV = 0.360 (11.765)

Sol Value = 0.200 g/210L ***
Fit Value = 0.1955 mg/l ****
Samples Taken = 4, Discarded = 1
***** CHANNEL 1 *****
Sample #1 = 0.1980 (0.0170)
Sample #2 = 0.1970 (0.0290)
Sample #3 = 0.1950 (0.0300)
Sample #4 = 0.1940 (0.0240)
Avg % RDS = 0.1963 (0.0243)
STD DEV = 0.0072 (0.0260)
REL STD DEV = 0.360 (11.765)

Sol Value = 0.200 g/210L ***
Fit Value = 0.1955 mg/l ****
Samples Taken = 4, Discarded = 1
***** CHANNEL 1 *****
Sample #1 = 0.1980 (0.0170)
Sample #2 = 0.1970 (0.0290)
Sample #3 = 0.1950 (0.0300)
Sample #4 = 0.1940 (0.0240)
Avg % RDS = 0.1963 (0.0243)
STD DEV = 0.0072 (0.0260)
REL STD DEV = 0.360 (11.765)

Sol Value = 0.200 g/210L ***
Fit Value = 0.1955 mg/l ****
Samples Taken = 4, Discarded = 1
***** CHANNEL 1 *****
Sample #1 = 0.1980 (0.0170)
Sample #2 = 0.1970 (0.0290)
Sample #3 = 0.1950 (0.0300)
Sample #4 = 0.1940 (0.0240)
Avg % RDS = 0.1963 (0.0243)
STD DEV = 0.0072 (0.0260)
REL STD DEV = 0.360 (11.765)

Dry Gas H2O Adjust Results *****
Barometric Pressure = 1014
3 um H2O Adjust (mg/l x 10,000) = 1059
9 um H2O Adjust (mg/l x 10,000) = 745
***** AUTO CAL PASS *****

Solution Stats Quadratic Fit Chan 1
Act Fit Residual
g/210L g/210L g/210L
0.000 0.000 -0.0003
0.040 0.039 0.0006
0.100 0.100 -0.0005
0.200 0.200 0.0002
0.300 0.300 -0.0001

Solution Stats Quadratic Fit Chan 1
Act Fit Residual
g/210L g/210L g/210L
0.000 0.000 -0.0003
0.040 0.039 0.0006
0.100 0.100 -0.0005
0.200 0.200 0.0002
0.300 0.300 -0.0001

Solution Stats Quadratic Fit Chan 1
Act Fit Residual
g/210L g/210L g/210L
0.000 0.000 -0.0003
0.040 0.039 0.0006
0.100 0.100 -0.0005
0.200 0.200 0.0002
0.300 0.300 -0.0001

Solution Stats Quadratic Fit Chan 1
Act Fit Residual
g/210L g/210L g/210L
0.000 0.000 -0.0003
0.040 0.039 0.0006
0.100 0.100 -0.0005
0.200 0.200 0.0002
0.300 0.300 -0.0001

Solution Stats Quadratic Fit Chan 1
Act Fit Residual
g/210L g/210L g/210L
0.000 0.000 -0.0003
0.040 0.039 0.0006
0.100 0.100 -0.0005
0.200 0.200 0.0002
0.300 0.300 -0.0001

Solution Stats Quadratic Fit Chan 1
Act Fit Residual
g/210L g/210L g/210L
0.000 0.000 -0.0003
0.040 0.039 0.0006
0.100 0.100 -0.0005
0.200 0.200 0.0002
0.300 0.300 -0.0001

Solution Stats Quadratic Fit Chan 1
Act Fit Residual
g/210L g/210L g/210L
0.000 0.000 -0.0003
0.040 0.039 0.0006
0.100 0.100 -0.0005
0.200 0.200 0.0002
0.300 0.300 -0.0001

Solution Stats Quadratic Fit Chan 1
Act Fit Residual
g/210L g/210L g/210L
0.000 0.000 -0.0003
0.040 0.039 0.0006
0.100 0.100 -0.0005
0.200 0.200 0.0002
0.300 0.300 -0.0001

Solution Stats Quadratic Fit Chan 1
Act Fit Residual
g/210L g/210L g/210L
0.000 0.000 -0.0003
0.040 0.039 0.0006
0.100 0.100 -0.0005
0.200 0.200 0.0002
0.300 0.300 -0.0001

Optical Calibration	
SN:	80-001647
Agency:	Cape Coral PD
Date:	3/23/2022
Quadratic Fit:	+/- 0.002g/210L
By:	DERR <i>Full</i>

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-001647	Cape Coral Police Department	3/23/2022	DERR <i>DELL</i>

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083
CAPE CORAL PD Intoxilyzer - Alcohol Analyzer Model 8000 03/23/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 07:15 Control Test 0.126 07:16 Air Blank 0.000 07:16 Control Test 0.127 07:17 Air Blank 0.000 07:17 Control Test 0.127 07:18 Air Blank 0.000 07:19 Control Test Status Average 0.1463 Std Dev 0.0006 Rel Std Dev(%) 0.461	CAPE CORAL PD Intoxilyzer - Alcohol Analyzer Model 8000 03/23/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 07:10 Control Test 0.077 07:11 Air Blank 0.000 07:11 Control Test 0.076 07:12 Air Blank 0.000 07:13 Control Test 0.077 07:13 Air Blank 0.000 07:14 Control Test Status Average 0.0767 Std Dev 0.0006 Rel Std Dev(%) 0.7531	CAPE CORAL PD Intoxilyzer - Alcohol Analyzer Model 8000 03/23/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 07:15 Control Test 0.139 07:16 Air Blank 0.000 07:16 Control Test 0.138 07:17 Air Blank 0.000 07:17 Control Test 0.137 07:18 Air Blank 0.000 07:19 Control Test Status Average 0.1500 Std Dev 0.0010 Rel Std Dev(%) 0.5301	CAPE CORAL PD Intoxilyzer - Alcohol Analyzer Model 8000 03/23/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 07:20 Control Test 0.077 07:21 Air Blank 0.000 07:21 Control Test 0.077 07:21 Air Blank 0.000 07:22 Control Test 0.077 07:22 Air Blank 0.000 07:23 Control Test Status Average 0.0770 Std Dev 0.0009 Rel Std Dev(%) 0.6009
<i>DELL</i> Operator's Signature	<i>DELL</i> Operator's Signature	<i>DELL</i> Operator's Signature	<i>DELL</i> Operator's Signature