



INSTRUMENT PROCESSING SHEET

Agency Osceola County SO

S/N 80-001715

Florida Department of Law Enforcement

Date In 02/11/2022 DI Completion Date 02/14/2022

Ship P/U H/D CMI EE

Intake, Quality Checks, Flow Calibration, Maintenance, and Simulator tables. Includes visual inspection checklist and flow verification data.

Calibration Adjustment and Department Inspection sections. Includes barometric pressure gauge tables and instrument compliance checkboxes.

Notes/Suggested Service section with handwritten text and digital signature of Israel Soto. Includes agency inspection checkboxes.

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: OSCEOLA COUNTY S.O.
Time of Inspection: 15:31

Date of Inspection: 02/14/2022

Serial Number: 80-001715
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202010A Exp: 10/05/2022	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202010D Exp: 10/06/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG113403 Exp: 05/14/2023
0.000	0.050	0.083	0.200	0.080
0.000	0.051	0.084	0.203	0.080
0.000	0.051	0.081	0.202	0.079
0.000	0.052	0.082	0.202	0.080
0.000	0.051	0.083	0.202	0.080
0.000	0.051	0.083	0.202	0.080
0.000	0.051	0.083	0.205	0.079
0.000	0.052	0.084	0.205	0.080
0.000	0.051	0.084	0.205	0.080
0.000	0.052	0.084	0.205	0.080


Standard Deviations	0.0006	0.0009	0.0017	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0009 Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



 RICHARD A WILLIAMS
 Signature and Printed Name

02/14/2022
 Date

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: OSCEOLA COUNTY S.O.
Time of Inspection: 13:27

Date of Inspection: 02/14/2022

Serial Number: 80-001715
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK		No	Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK		No	Barometric Pressure Sensor Check: OK		No
Alcohol Free Subject Test: 0.000		No	Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No	Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____

Standard Deviations				
---------------------	--	--	--	--


Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: _____ Number of Simulators Used: _____

Remarks:

MULTIPLE AMBIENT FAILS, DUE MULTIPLE INSTRUMENT INSPECTIONS Non-compliance: - RAW, 02/14/22

The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


RICHARD A WILLIAMS

 Signature and Printed Name

02/14/2022
 Date

OSCEOLA COUNTY S.O.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001715
02/14/2022
Software: 8100.27

OSCEOLA COUNTY S.O.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001715
02/14/2022
Software: 8100.27

OSCEOLA COUNTY S.O.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001715
02/14/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:57
Control Test	0.050	10:58
Air Blank	0.000	10:58
Control Test	0.050	10:59
Air Blank	0.000	11:00
Control Test	0.050	11:00
Air Blank	0.000	11:01
Control Test Stats		
Average	0.0500	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

0.05

PAW

Operator's Signature

Test	g/210L	Time
Air Blank	0.000	11:28
Control Test	0.083	11:29
Air Blank	0.000	11:30
Control Test	0.082	11:30
Air Blank	0.000	11:31
Control Test	0.082	11:31
Air Blank	0.000	11:32
Control Test Stats		
Average	0.0823	
Std Dev	0.0006	
Rel Std Dev(%)	0.7012	

0.08
wet

PAW

Operator's Signature

Test	g/210L	Time
Air Blank	0.000	11:34
Control Test	0.199	11:34
Air Blank	0.000	11:35
Control Test	0.200	11:35
Air Blank	0.000	11:36
Control Test	0.199	11:37
Air Blank	0.000	11:37
Control Test Stats		
Average	0.1993	
Std Dev	0.0006	
Rel Std Dev(%)	0.2895	

0.20

PAW

Operator's Signature

OSCEOLA COUNTY S.O.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001715
02/14/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:39
Control Test	0.081	11:39
Air Blank	0.000	11:40
Control Test	0.081	11:40
Air Blank	0.000	11:40
Control Test	0.080	11:41
Air Blank	0.000	11:41
Control Test Stats		
Average	0.0807	
Std Dev	0.0006	
Rel Std Dev(%)	0.7157	

0.08
DGS

PAW

Operator's Signature

Stability Checks
80-001715

02/14/22

PAW



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2729 Fort Knox Blvd.
Bldg. 2, Suite 1300
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001715, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001715</u>	UNCERTAINTY* \pm	
Owning Agency:	<u>OSCEOLA COUNTY S.O.</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>02/14/2022</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>15:31</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Richard A Williams
Digitally signed by Richard A Williams
Date: 2022.02.14 16:21:27 -05'00'

02/14/2022

Date

RICHARD A WILLIAMS,
Department Inspector

FDLE/ATP Form 69 January 2022

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality