



INSTRUMENT PROCESSING SHEET

Agency Lee CSOS/N 80-001722Florida Department of
Law EnforcementDate In 10/25/2022 DI Completion Date 10/25/2022☐ Ship ☒ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By TDG	Quality Checks	By TDG	Date	10/25/2022	Flow Calibration	By TDG	Date	10/25/2022																																													
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes: _____ _____ _____ _____ _____ _____ _____ _____ _____ _____		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>213</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP104</u> 32 mm <u>0.132*</u> (.139 - .169) 36 mm <u>0.144*</u> (.156 - .190) 53 mm <u>0.222*</u> (.228 - .278) 103 mm <u>0.507</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>68639</u> ☒ Stability Checks				Flow Column # <u>ATP106</u> ☒ 5L/min - 17mm ☒ 15L/min - 53mm ☒ 30L/min - 103mm ☒ R-Value <u>201</u> ☒ Post Calibration Verification (L/s) Flow Column # <u>ATP104</u> 32 mm <u>0.152</u> (.139 - .169) 36 mm <u>0.167</u> (.156 - .190) 53 mm <u>0.238</u> (.228 - .278) 103 mm <u>0.515</u> (.447 - .547)																																																
		<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP6286</td><td>202201C 01/11/2024</td></tr><tr><td>0.080</td><td>MP6287</td><td>202201D 01/18/2024</td></tr><tr><td>0.200</td><td>MP6288</td><td>202201E 01/18/2024</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>00521080A2 02/05/2023</td></tr></tbody></table>		Simulator	Serial #	Lot #/Exp	0.050	MP6286	202201C 01/11/2024	0.080	MP6287	202201D 01/18/2024	0.200	MP6288	202201E 01/18/2024	0.080 DGS	N/A	00521080A2 02/05/2023			<table border="1"><thead><tr><th colspan="2">Maintenance</th><th>By</th></tr></thead><tbody><tr><td colspan="2">☐ Battery Replacement</td><td></td></tr><tr><td colspan="2">☐ Dry Gas Regulator Replacement</td><td></td></tr><tr><td colspan="2">☐ Breath Tube Replacement</td><td></td></tr><tr><td colspan="2">☐ Other _____</td><td></td></tr><tr><td colspan="2"> </td><td></td></tr><tr><td colspan="2"> </td><td></td></tr><tr><td colspan="2"> </td><td></td></tr><tr><td colspan="2"> </td><td></td></tr><tr><td colspan="2"> </td><td></td></tr></tbody></table>		Maintenance		By	☐ Battery Replacement			☐ Dry Gas Regulator Replacement			☐ Breath Tube Replacement			☐ Other _____																			
Simulator	Serial #	Lot #/Exp																																																				
0.050	MP6286	202201C 01/11/2024																																																				
0.080	MP6287	202201D 01/18/2024																																																				
0.200	MP6288	202201E 01/18/2024																																																				
0.080 DGS	N/A	00521080A2 02/05/2023																																																				
Maintenance		By																																																				
☐ Battery Replacement																																																						
☐ Dry Gas Regulator Replacement																																																						
☐ Breath Tube Replacement																																																						
☐ Other _____																																																						
Calibration Adjustment		By _____		Department Inspection		By <u>TDG</u>																																																
Barometric Pressure Gauge _____ ID # _____				Barometric Pressure ID# <u>28199</u>																																																		
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A					Gauge <u>1016</u> Instrument <u>1013</u>																						
Simulator	Serial #	Lot #	Expiration																																																			
0.000		N/A	N/A																																																			
0.040																																																						
0.100																																																						
0.200																																																						
0.300																																																						
0.080 DGS	N/A																																																					
				Mouth Alcohol Solution Lot # <u>2021-D</u>																																																		
				Acetone Stock Solution Lot # <u>2021-C</u>																																																		
				<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP6284</td></tr><tr><td>Interferent</td><td>MP6285</td></tr><tr><td>0.050</td><td>MP6286</td></tr><tr><td>0.080</td><td>MP6287</td></tr><tr><td>0.200</td><td>MP6288</td></tr></tbody></table>		Simulator	Serial Number	0.000	MP6284	Interferent	MP6285	0.050	MP6286	0.080	MP6287	0.200	MP6288																																					
Simulator	Serial Number																																																					
0.000	MP6284																																																					
Interferent	MP6285																																																					
0.050	MP6286																																																					
0.080	MP6287																																																					
0.200	MP6288																																																					
☐ Post Calibration Adjustment Stability Checks				Attachments																																																		
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A					☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Calibration Adjustment		☐ Post-Stability Checks ☒ Flow Calibration ☐ Form 40 ☐ Other _____																												
Simulator	Serial #	Lot #	Expiration																																																			
0.050																																																						
0.080																																																						
0.200																																																						
0.080 DGS	N/A																																																					
Notes/Suggested Service: *Flow values outside nominal range. Will perform flow calibration. (TDG) _____ _____ _____ _____ _____ _____ _____ _____ _____				☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																																																		
				David Eliezer Reyes-Rivera Digitally signed by David Eliezer Reyes-Rivera Date: 2022.10.25 12:46:17 -0400		25 13:04:3																																																
				Tech Review / Date		Admin Review / Date																																																

Flow Calibration

80-001722

10/25/2022

ML

LEE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001722
10/25/2022
Software: 8100.27

Flow Rate Calibration*****

1: Rate (Liters/min) = 5

SQRT(Diff) = 6.781

2: Rate (Liters/min) = 15

SQRT(Diff) = 11.660

3: Rate (Liters/min) = 30

SQRT(Diff) = 20.566

Dependent Data Scale Factor = 100000 L/min

Independent Data Scale Factor = 256

Rounded Slope = 702

Rounded Intercept = -670837

Correlation = 0.99864

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-001722	Lee CSO	10/25/2022	TDG MC

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083
✓	✓	✓	✓
LEE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 10/25/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 09:57 Control Test 0.050 09:58 Air Blank 0.000 09:58 Control Test 0.049 09:59 Air Blank 0.000 10:00 Control Test 0.049 10:00 Air Blank 0.000 10:01 Control Test Stats Average 0.0493 Std Dev 0.0006 Rel Std Dev(%) 1.1703 Operator's Signature 	LEE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 10/25/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 10:04 Control Test 0.080 10:04 Air Blank 0.000 10:05 Control Test 0.079 10:06 Air Blank 0.000 10:06 Control Test 0.078 10:07 Air Blank 0.000 10:07 Control Test Stats Average 0.0790 Std Dev 0.0010 Rel Std Dev(%) 1.2658 Operator's Signature 	LEE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 10/25/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 10:11 Control Test 0.199 10:12 Air Blank 0.000 10:12 Control Test 0.198 10:13 Air Blank 0.000 10:14 Control Test 0.198 10:14 Air Blank 0.000 10:15 Control Test Stats Average 0.1983 Std Dev 0.0006 Rel Std Dev(%) 0.2911 Operator's Signature 	LEE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 10/25/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 10:16 Control Test 0.080 10:16 Air Blank 0.000 10:16 Control Test 0.080 10:17 Air Blank 0.000 10:17 Control Test 0.080 10:18 Air Blank 0.000 10:18 Control Test Stats Average 0.0800 Std Dev 0.0000 Rel Std Dev(%) 0.0000 Operator's Signature 

Comments:

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: LEE COUNTY SO
Time of Inspection: 12:21

Date of Inspection: 10/25/2022

Serial Number: 80-001722
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:00521080A2 Exp: 02/05/2023
0.000	0.050	0.080	0.200	0.080
0.000	0.049	0.080	0.200	0.080
0.000	0.049	0.080	0.200	0.080
0.000	0.050	0.080	0.201	0.080
0.000	0.049	0.080	0.201	0.080
0.000	0.049	0.080	0.200	0.080
0.000	0.049	0.080	0.200	0.080
0.000	0.049	0.081	0.200	0.079
0.000	0.049	0.081	0.200	0.080
0.000	0.050	0.080	0.199	0.080

Standard Deviations	0.0004	0.0004	0.0005	0.0003
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.


I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



Signature and Printed Name

TAYLOR D GUTSCHOW

10/25/2022
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001722, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001722</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>LEE COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>10/25/2022</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>12:21</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

10/25/2022

Date


TAYLOR D GUTSCHOW,

Department Inspector

FDLE/ATP Form 69 December 2021
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality