

## INSTRUMENT PROCESSING SHEET

Agency Pinellas Park PD

S/N 80-001732

Florida Department of  
Law Enforcement

Date In 2/21/2022      DI Completion Date 3/1/2022

☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

<b>Intake</b> <input type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE  Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight  Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable  Notes: _____ _____ _____ _____ _____ _____	<b>Quality Checks</b> By TDG Date 2/21/2022 <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value 204 <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # ATP104 32 mm 0.148 (.139 - .169) 36 mm 0.167 (.156 - .190) 53 mm 0.238 (.228 - .278) 103 mm 0.500 (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # 68639 <input checked="" type="checkbox"/> Stability Checks	<b>Flow Calibration</b> By _____ Date _____ Flow Column # _____ <input type="checkbox"/> 5L/min – 17mm <input type="checkbox"/> 15L/min – 53mm <input type="checkbox"/> 30L/min – 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) _____ Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)	<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td rowspan="2">0.050</td> <td rowspan="2">MP5092</td> <td>202010A</td> </tr> <tr> <td>10/05/2022</td> </tr> <tr> <td rowspan="2">0.080</td> <td rowspan="2">MP4864</td> <td>202010B</td> </tr> <tr> <td>10/05/2022</td> </tr> <tr> <td rowspan="2">0.200</td> <td rowspan="2">MP5094</td> <td>202010D</td> </tr> <tr> <td>10/06/2022</td> </tr> <tr> <td rowspan="2">0.080 DGS</td> <td rowspan="2">N/A</td> <td>AG115904</td> </tr> <tr> <td>06/08/2023</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	MP5092	202010A	10/05/2022	0.080	MP4864	202010B	10/05/2022	0.200	MP5094	202010D	10/06/2022	0.080 DGS	N/A	AG115904	06/08/2023
Simulator	Serial #	Lot #/Exp																				
0.050	MP5092	202010A																				
		10/05/2022																				
0.080	MP4864	202010B																				
		10/05/2022																				
0.200	MP5094	202010D																				
		10/06/2022																				
0.080 DGS	N/A	AG115904																				
		06/08/2023																				

<b>Calibration Adjustment</b> By _____ Barometric Pressure Gauge _____ ID# _____ <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> <input type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>	Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			<b>Department Inspection</b> By TDG Barometric Pressure ID# 28199 Gauge 1016                      Instrument 1016 Mouth Alcohol Solution Lot # 2021-B Acetone Stock Solution Lot # 2021-C <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP6284</td> </tr> <tr> <td>Interferent</td> <td>MP6285</td> </tr> <tr> <td>0.050</td> <td>MP6286</td> </tr> <tr> <td>0.080</td> <td>MP6287</td> </tr> <tr> <td>0.200</td> <td>MP6288</td> </tr> </tbody> </table>	Simulator	Serial Number	0.000	MP6284	Interferent	MP6285	0.050	MP6286	0.080	MP6287	0.200	MP6288
Simulator	Serial #	Lot #	Expiration																																																										
0.000		N/A	N/A																																																										
0.040																																																													
0.100																																																													
0.200																																																													
0.300																																																													
0.080 DGS	N/A																																																												
Simulator	Serial #	Lot #	Expiration																																																										
0.050																																																													
0.080																																																													
0.200																																																													
0.080 DGS	N/A																																																												
Simulator	Serial Number																																																												
0.000	MP6284																																																												
Interferent	MP6285																																																												
0.050	MP6286																																																												
0.080	MP6287																																																												
0.200	MP6288																																																												

Notes/Suggested Service: _____ _____ _____ _____ _____ _____ _____	<b>Attachments</b> <div style="display: flex; justify-content: space-between;"> <ul style="list-style-type: none"> <li><input checked="" type="checkbox"/> Form 41</li> <li><input checked="" type="checkbox"/> Stability Checks</li> <li><input checked="" type="checkbox"/> Calibration Certificate</li> <li><input type="checkbox"/> Calibration Adjustment</li> </ul> <ul style="list-style-type: none"> <li><input type="checkbox"/> Post-Stability Checks</li> <li><input type="checkbox"/> Flow Calibration</li> <li><input type="checkbox"/> Form 40</li> <li><input type="checkbox"/> Other _____</li> </ul> </div> <hr/> <input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use <div style="text-align: right;">             Digitally signed by David Eliezer Reyes-Rivera  <small>Date: 2022.03.03 06:17:54 -05'00'</small>                07              15:28:27              05/09           </div>
--	--

Tech Review / Date \_\_\_\_\_ Admin Review / Date \_\_\_\_\_

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-001732	Pineallas Park PD	02/21/2022	TDG <i>MC</i>

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083
✓	✓	✓	✓
<p>PINEALLAS PARK P.D. Intoxilyzer - Alcohol Analyzer Model 8000 02/21/2022 Software: 8100.27</p> <p>SN 80-001732</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 14:59</p> <p>Control Test 0.049 15:00</p> <p>Air Blank 0.000 15:00</p> <p>Control Test 0.049 15:01</p> <p>Air Blank 0.000 15:01</p> <p>Control Test 0.048 15:02</p> <p>Air Blank 0.000 15:02</p> <p>Control Test Stats</p> <p>Average 0.0487</p> <p>Std Dev 0.0006</p> <p>Rel Std Dev(%) 1.1863</p>	<p>PINEALLAS PARK P.D. Intoxilyzer - Alcohol Analyzer Model 8000 02/21/2022 Software: 8100.27</p> <p>SN 80-001732</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 15:19</p> <p>Control Test 0.079 15:20</p> <p>Air Blank 0.000 15:21</p> <p>Control Test 0.078 15:21</p> <p>Air Blank 0.000 15:22</p> <p>Control Test 0.079 15:23</p> <p>Air Blank 0.000 15:23</p> <p>Control Test Stats</p> <p>Average 0.0787</p> <p>Std Dev 0.0006</p> <p>Rel Std Dev(%) 0.7339</p>	<p>PINEALLAS PARK P.D. Intoxilyzer - Alcohol Analyzer Model 8000 02/21/2022 Software: 8100.27</p> <p>SN 80-001732</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 14:52</p> <p>Control Test 0.199 14:52</p> <p>Air Blank 0.000 14:53</p> <p>Control Test 0.198 14:54</p> <p>Air Blank 0.000 14:54</p> <p>Control Test 0.197 14:55</p> <p>Air Blank 0.000 14:55</p> <p>Control Test Stats</p> <p>Average 0.1980</p> <p>Std Dev 0.0010</p> <p>Rel Std Dev(%) 0.5051</p>	<p><i>Q65</i></p> <p>PINEALLAS PARK P.D. Intoxilyzer - Alcohol Analyzer Model 8000 02/21/2022 Software: 8100.27</p> <p>SN 80-001732</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 14:46</p> <p>Control Test 0.082 14:47</p> <p>Air Blank 0.000 14:47</p> <p>Control Test 0.081 14:47</p> <p>Air Blank 0.000 14:48</p> <p>Control Test 0.081 14:48</p> <p>Air Blank 0.000 14:49</p> <p>Control Test Stats</p> <p>Average 0.0813</p> <p>Std Dev 0.0006</p> <p>Rel Std Dev(%) 0.7099</p>
<p><i>MC</i></p> <p>Operator's Signature</p>	<p><i>MC</i></p> <p>Operator's Signature</p>	<p><i>MC</i></p> <p>Operator's Signature</p>	<p><i>MC</i></p> <p>Operator's Signature</p>

Comments:



# Florida Department of Law Enforcement Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: PINELLAS PARK P.D.  
Time of Inspection: 16:21

Date of Inspection: 03/01/2022

Serial Number: 80-001732  
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202010A Exp: 10/05/2022	0.08g/210L Test (g/210L) Lot#:202010B Exp: 10/05/2022	0.20g/210L Test (g/210L) Lot#:202010D Exp: 10/06/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG115904 Exp: 06/08/2023
0.000	0.049	0.079	0.197	0.079
0.000	0.049	0.079	0.197	0.079
0.000	0.048	0.079	0.197	0.079
0.000	0.049	0.080	0.197	0.079
0.000	0.049	0.080	0.197	0.079
0.000	0.049	0.080	0.197	0.079
0.000	0.049	0.080	0.197	0.079
0.000	0.049	0.079	0.198	0.079
0.000	0.050	0.080	0.197	0.079
0.000	0.049	0.079	0.197	0.079


Standard Deviations	0.0004	0.0005	0.0003	0.0000
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies ( ☒ ) does not comply ( ☐ ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

  
\_\_\_\_\_  
Signature and Printed Name

TAYLOR D GUTSCHOW

03/01/2022  
Date



# Calibration Certificate

Florida Department of Law Enforcement  
Alcohol Testing Program  
4700 Terminal Drive, Suite 1  
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001732, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001732</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>PINELLAS PARK P.D.</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>03/01/2022</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>16:21</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within  $\pm 0.005$  or 5%, whichever is greater, of the target alcohol concentration.

\*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ( $k=3$ ).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

## TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

03/01/2022 Date  
Taylor D Gutschow TAYLOR D GUTSCHOW,  
Department Inspector

FDLE/ATP Form 69 December 2021  
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality



## Return Material Authorization

**Ship to:** ☒ CMI, Inc.

☐ Enforcement Electronics

Shipment to repair facility authorized by: Jacob Rolleston on 4/4/2022

**Items Returned:** Instrument ☒ Supplies ☐ Other ☐ Describe: \_\_\_\_\_

Instrument Model: Intoxilyzer 8000 Serial Number: 80-001732

**Bill To Address:**

Pinellas Park PD

Attn: Jacob Rolleston

**Ship to Address:**

Florida Department of Law Enforcement

Fort Myers Regional Operations Center

Attn: Alcohol Testing Program

4700 Terminal Drive, Suite 1

Fort Myers, FL 33907

**Reason for Return:**

Agency reports the instrument would not come out of Standby Mode due to repeated DSP Fails.

Could not duplicate at FDLE. Sending for evaluation.

**Please choose one of the following options:**

☐ 1. I \_\_\_\_\_, authorize all repairs.

☐ 2. I \_\_\_\_\_, authorize repairs up to \$\_\_\_\_\_.

☒ 3. I require an estimate **BEFORE** any repairs will be authorized and/ or conducted.

Please contact: Name: Jacob Rolleston

Phone #: 727-946-4084 Email: Rolleston@pinellas-park.com

ATP Contact Name: Taylor Gutschow ATP Email: TaylorGutschow@fdle.state.fl.us







## INSTRUMENT PROCESSING SHEET

Agency Pinellas Park PDS/N 80-001732Florida Department of  
Law EnforcementDate In 11/08/2022 DI Completion Date 11/15/2022☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By TDG	Quality Checks	By TDG	Date <u>11/08/2022</u>	Flow Calibration	By	Date																												
<input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input checked="" type="checkbox"/> Return from CMI / EE  Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight  Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable  Notes: _____ _____ _____ _____ _____ _____ _____ _____ _____ _____		<input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>217</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP104</u> 32 mm <u>0.156</u> (.139 - .169) 36 mm <u>0.164</u> (.156 - .190) 53 mm <u>0.234</u> (.228 - .278) 103 mm <u>0.503</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>68639</u> <input checked="" type="checkbox"/> Stability Checks			Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																														
					<b>Maintenance</b> By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ _____ _____ _____ _____ _____																														
<b>Calibration Adjustment</b> By _____				<b>Department Inspection</b> By <u>TDG</u>																															
Barometric Pressure Gauge ID # _____				Barometric Pressure ID# <u>28199</u>																															
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>				Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Gauge <u>1018</u> Instrument <u>1019</u> Mouth Alcohol Solution Lot # <u>2021-D</u> Acetone Stock Solution Lot # <u>2021-C</u>			
Simulator	Serial #	Lot #	Expiration																																
0.000		N/A	N/A																																
0.040																																			
0.100																																			
0.200																																			
0.300																																			
0.080 DGS	N/A																																		
<input type="checkbox"/> Post Calibration Adjustment Stability Checks				<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP6284</td></tr><tr><td>Interferent</td><td>MP6285</td></tr><tr><td>0.050</td><td>MP6286</td></tr><tr><td>0.080</td><td>MP6287</td></tr><tr><td>0.200</td><td>MP6288</td></tr></tbody></table>				Simulator	Serial Number	0.000	MP6284	Interferent	MP6285	0.050	MP6286	0.080	MP6287	0.200	MP6288																
Simulator	Serial Number																																		
0.000	MP6284																																		
Interferent	MP6285																																		
0.050	MP6286																																		
0.080	MP6287																																		
0.200	MP6288																																		
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>				Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			<b>Attachments</b> <input checked="" type="checkbox"/> Form 41 <input type="checkbox"/> Post-Stability Checks <input checked="" type="checkbox"/> Stability Checks <input type="checkbox"/> Flow Calibration <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Form 40 <input type="checkbox"/> Calibration Adjustment <input type="checkbox"/> Other _____											
Simulator	Serial #	Lot #	Expiration																																
0.050																																			
0.080																																			
0.200																																			
0.080 DGS	N/A																																		
Notes/Suggested Service: _____ _____ _____ _____ _____ _____ _____ _____ _____				<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use																															
Phil Nicodemo <small>Digitally signed by Phil Nicodemo Date: 2022.11.17 11:05:48 -05'00'</small>				Israel Soto <small>Digitally signed by Israel Soto Date: 2022.11.17 11:26:50 -05'00'</small>																															
Tech Review / Date _____				Admin Review / Date _____																															

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-00 1732	Pinellas Park PD	11/08/2022	TDG TML

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																																
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083 ≤0.003 of Wet																																																																																																																																																
✓	✓	✓	✓																																																																																																																																																
<p>PINELLAS PARK P.D. Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001732 11/08/2022 Software: 8100.27</p> <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>15:09</td></tr><tr><td>Control Test</td><td>0.049</td><td>15:10</td></tr><tr><td>Air Blank</td><td>0.000</td><td>15:11</td></tr><tr><td>Control Test</td><td>0.049</td><td>15:11</td></tr><tr><td>Air Blank</td><td>0.000</td><td>15:12</td></tr><tr><td>Control Test</td><td>0.049</td><td>15:12</td></tr><tr><td>Air Blank</td><td>0.000</td><td>15:13</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0490</td><td></td></tr><tr><td>Std Dev</td><td>0.0000</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr></tbody></table> <p>Operator's Signature <i>TMG</i></p>	Test	g/210L	Time	Air Blank	0.000	15:09	Control Test	0.049	15:10	Air Blank	0.000	15:11	Control Test	0.049	15:11	Air Blank	0.000	15:12	Control Test	0.049	15:12	Air Blank	0.000	15:13	Control Test Stats			Average	0.0490		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<p>PINELLAS PARK P.D. Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001732 11/08/2022 Software: 8100.27</p> <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>15:00</td></tr><tr><td>Control Test</td><td>0.078</td><td>15:01</td></tr><tr><td>Air Blank</td><td>0.000</td><td>15:02</td></tr><tr><td>Control Test</td><td>0.078</td><td>15:02</td></tr><tr><td>Air Blank</td><td>0.000</td><td>15:03</td></tr><tr><td>Control Test</td><td>0.078</td><td>15:03</td></tr><tr><td>Air Blank</td><td>0.000</td><td>15:04</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0780</td><td></td></tr><tr><td>Std Dev</td><td>0.0000</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr></tbody></table> <p>Operator's Signature <i>TMG</i></p>	Test	g/210L	Time	Air Blank	0.000	15:00	Control Test	0.078	15:01	Air Blank	0.000	15:02	Control Test	0.078	15:02	Air Blank	0.000	15:03	Control Test	0.078	15:03	Air Blank	0.000	15:04	Control Test Stats			Average	0.0780		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<p>PINELLAS PARK P.D. Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001732 11/08/2022 Software: 8100.27</p> <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>15:25</td></tr><tr><td>Control Test</td><td>0.198</td><td>15:26</td></tr><tr><td>Air Blank</td><td>0.000</td><td>15:26</td></tr><tr><td>Control Test</td><td>0.198</td><td>15:27</td></tr><tr><td>Air Blank</td><td>0.000</td><td>15:27</td></tr><tr><td>Control Test</td><td>0.198</td><td>15:28</td></tr><tr><td>Air Blank</td><td>0.000</td><td>15:29</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.1980</td><td></td></tr><tr><td>Std Dev</td><td>0.0000</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr></tbody></table> <p>Operator's Signature <i>TMG</i></p>	Test	g/210L	Time	Air Blank	0.000	15:25	Control Test	0.198	15:26	Air Blank	0.000	15:26	Control Test	0.198	15:27	Air Blank	0.000	15:27	Control Test	0.198	15:28	Air Blank	0.000	15:29	Control Test Stats			Average	0.1980		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<p><i>DGS</i></p> <p>PINELLAS PARK P.D. Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001732 11/08/2022 Software: 8100.27</p> <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>14:52</td></tr><tr><td>Control Test</td><td>0.079</td><td>14:52</td></tr><tr><td>Air Blank</td><td>0.000</td><td>14:53</td></tr><tr><td>Control Test</td><td>0.078</td><td>14:53</td></tr><tr><td>Air Blank</td><td>0.000</td><td>14:53</td></tr><tr><td>Control Test</td><td>0.078</td><td>14:54</td></tr><tr><td>Air Blank</td><td>0.000</td><td>14:54</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0783</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.7370</td><td></td></tr></tbody></table> <p>Operator's Signature <i>TMG</i></p>	Test	g/210L	Time	Air Blank	0.000	14:52	Control Test	0.079	14:52	Air Blank	0.000	14:53	Control Test	0.078	14:53	Air Blank	0.000	14:53	Control Test	0.078	14:54	Air Blank	0.000	14:54	Control Test Stats			Average	0.0783		Std Dev	0.0006		Rel Std Dev(%)	0.7370	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	15:09																																																																																																																																																	
Control Test	0.049	15:10																																																																																																																																																	
Air Blank	0.000	15:11																																																																																																																																																	
Control Test	0.049	15:11																																																																																																																																																	
Air Blank	0.000	15:12																																																																																																																																																	
Control Test	0.049	15:12																																																																																																																																																	
Air Blank	0.000	15:13																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0490																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	15:00																																																																																																																																																	
Control Test	0.078	15:01																																																																																																																																																	
Air Blank	0.000	15:02																																																																																																																																																	
Control Test	0.078	15:02																																																																																																																																																	
Air Blank	0.000	15:03																																																																																																																																																	
Control Test	0.078	15:03																																																																																																																																																	
Air Blank	0.000	15:04																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0780																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	15:25																																																																																																																																																	
Control Test	0.198	15:26																																																																																																																																																	
Air Blank	0.000	15:26																																																																																																																																																	
Control Test	0.198	15:27																																																																																																																																																	
Air Blank	0.000	15:27																																																																																																																																																	
Control Test	0.198	15:28																																																																																																																																																	
Air Blank	0.000	15:29																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1980																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	14:52																																																																																																																																																	
Control Test	0.079	14:52																																																																																																																																																	
Air Blank	0.000	14:53																																																																																																																																																	
Control Test	0.078	14:53																																																																																																																																																	
Air Blank	0.000	14:53																																																																																																																																																	
Control Test	0.078	14:54																																																																																																																																																	
Air Blank	0.000	14:54																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0783																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7370																																																																																																																																																		

Comments:



# Florida Department of Law Enforcement Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: PINELLAS PARK P.D.  
Time of Inspection: 15:25

Date of Inspection: 11/15/2022

Serial Number: 80-001732  
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:00521080A2 Exp: 02/05/2023
0.000	0.049	0.078	0.198	0.079
0.000	0.048	0.078	0.198	0.079
0.000	0.048	0.078	0.197	0.078
0.000	0.048	0.078	0.197	0.078
0.000	0.048	0.078	0.197	0.079
0.000	0.048	0.078	0.198	0.079
0.000	0.048	0.078	0.197	0.079
0.000	0.048	0.078	0.198	0.079
0.000	0.048	0.078	0.197	0.078
0.000	0.049	0.078	0.197	0.079
Standard Deviations	0.0004	0.0000	0.0005	0.0004

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies ( ☒ ) does not comply ( ☐ ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

11/15/2022  
Date



## Calibration Certificate

Florida Department of Law Enforcement  
Alcohol Testing Program  
4700 Terminal Drive, Suite 1  
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001732, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001732</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>PINELLAS PARK P.D.</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>11/15/2022</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>15:25</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within  $\pm 0.005$  or 5%, whichever is greater, of the target alcohol concentration.

\*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ( $k=3$ ).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

### TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

11/15/2022

Date

  
TAYLOR D GUTSCHOW,  
Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Page 1 of 1