




INSTRUMENT PROCESSING SHEET

Agency Port St. Lucie Police DepartmentS/N 80-001962Florida Department of
Law EnforcementDate In 3/31/2022 DI Completion Date 4/4/2022☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By DERR	Quality Checks	By DERR	Date 4/4/2022	Flow Calibration	By	Date																												
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: <u>Instrument has a note</u> <u>"It keeps getting ambient fails</u> <u>during inspections." Will evaluate</u> <u>to determine if there is an issue.</u> <u>No ambient fails during the</u> <u>inspection process. DERR</u>		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>221</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP106</u> 32 mm <u>0.156</u> (.139 - .169) 36 mm <u>0.179</u> (.156 - .190) 53 mm <u>0.246</u> (.228 - .278) 103 mm <u>0.523</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28663</u> ☒ Stability Checks			Flow Column # _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																														
		<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP6286</td><td>202201C 01/11/2024</td></tr><tr><td>0.080</td><td>MP6287</td><td>202201D 01/18/2024</td></tr><tr><td>0.200</td><td>MP6288</td><td>202201E 01/18/2024</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG115904 06/08/2023</td></tr></tbody></table>	Simulator	Serial #	Lot #/Exp	0.050	MP6286	202201C 01/11/2024	0.080	MP6287	202201D 01/18/2024	0.200	MP6288	202201E 01/18/2024	0.080 DGS	N/A	AG115904 06/08/2023			Maintenance By _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____															
Simulator	Serial #	Lot #/Exp																																	
0.050	MP6286	202201C 01/11/2024																																	
0.080	MP6287	202201D 01/18/2024																																	
0.200	MP6288	202201E 01/18/2024																																	
0.080 DGS	N/A	AG115904 06/08/2023																																	
Calibration Adjustment By DERR		Department Inspection By DERR																																	
Barometric Pressure Gauge <u>1014</u> ID # <u>68639</u>		Barometric Pressure ID# <u>28199</u>																																	
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td>MP5095</td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td>MP5096</td><td>21070</td><td>03/01/2023</td></tr><tr><td>0.100</td><td>MP5098</td><td>21080</td><td>03/08/2023</td></tr><tr><td>0.200</td><td>MP5100</td><td>20510</td><td>12/03/2022</td></tr><tr><td>0.300</td><td>MP5101</td><td>21030</td><td>02/02/2023</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>00521080A2</td><td>02/05/2023</td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.000	MP5095	N/A	N/A	0.040	MP5096	21070	03/01/2023	0.100	MP5098	21080	03/08/2023	0.200	MP5100	20510	12/03/2022	0.300	MP5101	21030	02/02/2023	0.080 DGS	N/A	00521080A2	02/05/2023	Gauge <u>1014</u> Instrument <u>1014</u> Mouth Alcohol Solution Lot # <u>2021-D</u> Acetone Stock Solution Lot # <u>2021-C</u>					
Simulator	Serial #	Lot #	Expiration																																
0.000	MP5095	N/A	N/A																																
0.040	MP5096	21070	03/01/2023																																
0.100	MP5098	21080	03/08/2023																																
0.200	MP5100	20510	12/03/2022																																
0.300	MP5101	21030	02/02/2023																																
0.080 DGS	N/A	00521080A2	02/05/2023																																
☒ Post Calibration Adjustment Stability Checks		<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP6284</td></tr><tr><td>Interferent</td><td>MP6285</td></tr><tr><td>0.050</td><td>MP6286</td></tr><tr><td>0.080</td><td>MP6287</td></tr><tr><td>0.200</td><td>MP6288</td></tr></tbody></table>		Simulator	Serial Number	0.000	MP6284	Interferent	MP6285	0.050	MP6286	0.080	MP6287	0.200	MP6288	Attachments ☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☒ Calibration Adjustment		☒ Post-Stability Checks ☐ Flow Calibration ☒ Form 40 ☐ Other _____																	
Simulator	Serial Number																																		
0.000	MP6284																																		
Interferent	MP6285																																		
0.050	MP6286																																		
0.080	MP6287																																		
0.200	MP6288																																		
Notes/Suggested Service: <u>Bypassed AI to bring to Ready.</u> <u>Instrument was calibrated to bring values closer to</u> <u>nominal. Upload failed, I had to go into the level 2 menu</u> <u>to upload. Instrument locked on me during upload, I had</u> <u>to restart the instrument to upload the data. DERR</u>		☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																																	
		Taylor Gutschow Digitally signed by Taylor Gutschow Date: 2022.04.04 14:48:59		05 14:48:59 0400																															
		Tech Review / Date		Admin Review / Date																															



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001962 , manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001962</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>PORT ST LUCIE PD</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>04/04/2022</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>12:36</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.
This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

04/04/2022

Date

DAVID E REXES-RIVERA,

Department Inspector

Service • Integrity • Respect • Quality

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: PORT ST LUCIE PD

Time of Inspection: 12:36

Date of Inspection: 04/04/2022

Serial Number: 80-001962

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG115904 Exp: 06/08/2023
0.000	0.049	0.078	0.198	0.080
0.000	0.050	0.079	0.198	0.080
0.000	0.049	0.079	0.200	0.080
0.000	0.050	0.079	0.199	0.079
0.000	0.049	0.079	0.199	0.079
0.000	0.050	0.079	0.199	0.079
0.000	0.050	0.079	0.199	0.079
0.000	0.050	0.079	0.199	0.079
0.000	0.049	0.079	0.199	0.079
0.000	0.050	0.078	0.198	0.079


Standard Deviations	0.0005	0.0004	0.0006	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


DAVID E REYES-RIVERA

 Signature and Printed Name

04/04/2022
 Date

Type of Test	Serial Number	Agency	Date	Performed By
Post Stabilities	80-001962	Port St. Lucie Police Department	4/4/2022	DERR <i>pull</i>

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																																
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083																																																																																																																																																
PORT ST LUCIE PD Intoxilyzer - Alcohol Analyzer Model 8000 04/04/2022 Software: 8100.27 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>10:50</td></tr><tr><td>Control Test</td><td>0.049</td><td>10:51</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:52</td></tr><tr><td>Control Test</td><td>0.049</td><td>10:52</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:53</td></tr><tr><td>Control Test</td><td>0.049</td><td>10:53</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:54</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0490</td><td></td></tr><tr><td>Std Dev</td><td>0.0000</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr></tbody></table>	Test	g/210L	Time	Air Blank	0.000	10:50	Control Test	0.049	10:51	Air Blank	0.000	10:52	Control Test	0.049	10:52	Air Blank	0.000	10:53	Control Test	0.049	10:53	Air Blank	0.000	10:54	Control Test Stats			Average	0.0490		Std Dev	0.0000		Rel Std Dev(%)	0.0000		PORT ST LUCIE PD Intoxilyzer - Alcohol Analyzer Model 8000 04/04/2022 Software: 8100.27 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>10:55</td></tr><tr><td>Control Test</td><td>0.079</td><td>10:56</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:56</td></tr><tr><td>Control Test</td><td>0.079</td><td>10:57</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:57</td></tr><tr><td>Control Test</td><td>0.079</td><td>10:58</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:58</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0787</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.7339</td><td></td></tr></tbody></table>	Test	g/210L	Time	Air Blank	0.000	10:55	Control Test	0.079	10:56	Air Blank	0.000	10:56	Control Test	0.079	10:57	Air Blank	0.000	10:57	Control Test	0.079	10:58	Air Blank	0.000	10:58	Control Test Stats			Average	0.0787		Std Dev	0.0006		Rel Std Dev(%)	0.7339		PORT ST LUCIE PD Intoxilyzer - Alcohol Analyzer Model 8000 04/04/2022 Software: 8100.27 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>11:00</td></tr><tr><td>Control Test</td><td>0.198</td><td>11:01</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:01</td></tr><tr><td>Control Test</td><td>0.198</td><td>11:02</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:02</td></tr><tr><td>Control Test</td><td>0.198</td><td>11:03</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:04</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.1980</td><td></td></tr><tr><td>Std Dev</td><td>0.0000</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr></tbody></table>	Test	g/210L	Time	Air Blank	0.000	11:00	Control Test	0.198	11:01	Air Blank	0.000	11:01	Control Test	0.198	11:02	Air Blank	0.000	11:02	Control Test	0.198	11:03	Air Blank	0.000	11:04	Control Test Stats			Average	0.1980		Std Dev	0.0000		Rel Std Dev(%)	0.0000		PORT ST LUCIE PD Intoxilyzer - Alcohol Analyzer Model 8000 04/04/2022 Software: 8100.27 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>11:06</td></tr><tr><td>Control Test</td><td>0.079</td><td>11:06</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:07</td></tr><tr><td>Control Test</td><td>0.080</td><td>11:07</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:07</td></tr><tr><td>Control Test</td><td>0.080</td><td>11:08</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:08</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0797</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.7247</td><td></td></tr></tbody></table>	Test	g/210L	Time	Air Blank	0.000	11:06	Control Test	0.079	11:06	Air Blank	0.000	11:07	Control Test	0.080	11:07	Air Blank	0.000	11:07	Control Test	0.080	11:08	Air Blank	0.000	11:08	Control Test Stats			Average	0.0797		Std Dev	0.0006		Rel Std Dev(%)	0.7247	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	10:50																																																																																																																																																	
Control Test	0.049	10:51																																																																																																																																																	
Air Blank	0.000	10:52																																																																																																																																																	
Control Test	0.049	10:52																																																																																																																																																	
Air Blank	0.000	10:53																																																																																																																																																	
Control Test	0.049	10:53																																																																																																																																																	
Air Blank	0.000	10:54																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0490																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	10:55																																																																																																																																																	
Control Test	0.079	10:56																																																																																																																																																	
Air Blank	0.000	10:56																																																																																																																																																	
Control Test	0.079	10:57																																																																																																																																																	
Air Blank	0.000	10:57																																																																																																																																																	
Control Test	0.079	10:58																																																																																																																																																	
Air Blank	0.000	10:58																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0787																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7339																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	11:00																																																																																																																																																	
Control Test	0.198	11:01																																																																																																																																																	
Air Blank	0.000	11:01																																																																																																																																																	
Control Test	0.198	11:02																																																																																																																																																	
Air Blank	0.000	11:02																																																																																																																																																	
Control Test	0.198	11:03																																																																																																																																																	
Air Blank	0.000	11:04																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1980																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	11:06																																																																																																																																																	
Control Test	0.079	11:06																																																																																																																																																	
Air Blank	0.000	11:07																																																																																																																																																	
Control Test	0.080	11:07																																																																																																																																																	
Air Blank	0.000	11:07																																																																																																																																																	
Control Test	0.080	11:08																																																																																																																																																	
Air Blank	0.000	11:08																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0797																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7247																																																																																																																																																		
pull Operator's Signature	pull Operator's Signature	pull Operator's Signature	pull Operator's Signature																																																																																																																																																

PORT ST LUCIE PD
Intoxilyzer - Alcohol Analyzer
Model 8000
04/04/2022

SN 80-001962
10:04:44

Auto Calibration

Max Power Res Value = 98
Auto Range Res Value = 72

Sol Value = 0.000 g/210L ***
Fit value = 0.0020 mg/l 222%
Samples Taken = 4, Discarded = 1
Sum Io = 12762, Sum Io = 12955

Sample % Abs (% Abs Ref)
Sample #1 = 0.2390 (-0.0260)
Sample #2 = 0.2230 (0.0120)
Sample #3 = 0.2250 (0.0120)
Sample #4 = 0.2190 (0.0180)
Avg % Abs = 0.2223 (0.0140)
STD DEV = 0.0031 (0.0035)
REL STD DEV = 1.394 (24.744)

Sample % Abs (% Abs Ref)
Sample #1 = 0.4770 (-0.0200)
Sample #2 = 0.4760 (0.0100)
Sample #3 = 0.4710 (0.0150)
Sample #4 = 0.4750 (0.0180)
Avg % Abs = 0.4740 (0.0143)
STD DEV = 0.0026 (0.0040)
REL STD DEV = 0.558 (28.196)

Sol Value = 0.040 g/210L ***
Fit value = 0.1905 mg/l 222%
Samples Taken = 4, Discarded = 1
Sum Io = 12758, Sum Io = 12956

Sample % Abs (% Abs Ref)
Sample #1 = 0.9380 (-0.0210)
Sample #2 = 0.9170 (0.0080)
Sample #3 = 0.9010 (0.0080)
Sample #4 = 0.9020 (0.0300)
Avg % Abs = 0.9067 (0.0127)
STD DEV = 0.0090 (0.0155)
REL STD DEV = 0.989 (122.644)

Sample % Abs (% Abs Ref)
Sample #1 = 1.8030 (0.0200)
Sample #2 = 1.8170 (0.0120)
Sample #3 = 1.7940 (0.0190)
Sample #4 = 1.7880 (0.0330)
Avg % Abs = 1.7997 (0.0213)
STD DEV = 0.0153 (0.0107)
REL STD DEV = 0.851 (50.122)

Sol Value = 0.100 g/210L ***
Fit value = 0.4762 mg/l 222%
Samples Taken = 4, Discarded = 1
Sum Io = 12756, Sum Io = 12952

Sample % Abs (% Abs Ref)
Sample #1 = 1.9370 (-0.0090)
Sample #2 = 1.9220 (0.0030)
Sample #3 = 1.9110 (0.0190)
Sample #4 = 1.9210 (0.0200)
Avg % Abs = 1.9180 (0.0140)
STD DEV = 0.0061 (0.0095)
REL STD DEV = 0.317 (58.139)

Sample % Abs (% Abs Ref)
Sample #1 = 3.7370 (0.0040)
Sample #2 = 3.7330 (-0.0120)
Sample #3 = 3.7220 (0.0090)
Sample #4 = 3.7070 (0.0030)
Avg % Abs = 3.7207 (-0.0050)
STD DEV = 0.0131 (0.0079)
REL STD DEV = 0.351 (132.288)

Sol Value = 0.200 g/210L ***
Fit value = 0.9524 mg/l 222%
Samples Taken = 4, Discarded = 1
Sum Io = 12756, Sum Io = 12954

Sample % Abs (% Abs Ref)
Sample #1 = 3.5560 (-0.0050)
Sample #2 = 3.5640 (0.0230)
Sample #3 = 3.5570 (0.0090)
Sample #4 = 3.5500 (0.0160)
Avg % Abs = 3.5570 (0.0160)
STD DEV = 0.0070 (0.0070)
REL STD DEV = 0.197 (43.750)

Sample % Abs (% Abs Ref)
Sample #1 = 6.7860 (0.0090)
Sample #2 = 6.7910 (0.0250)
Sample #3 = 6.7850 (0.0120)
Sample #4 = 6.7710 (0.0370)
Avg % Abs = 6.7823 (0.0247)
STD DEV = 0.0103 (0.0125)
REL STD DEV = 0.151 (50.689)

Sol Value = 0.300 g/210L ***
Fit value = 1.4286 mg/l 222%
Samples Taken = 4, Discarded = 1
Sum Io = 12752, Sum Io = 12950

Sample % Abs (% Abs Ref)
Sample #1 = 5.1610 (-0.0090)
Sample #2 = 5.1620 (-0.0100)
Sample #3 = 5.1620 (0.0050)
Sample #4 = 5.1700 (-0.0030)
Avg % Abs = 5.1647 (-0.0027)
STD DEV = 0.0046 (0.0075)
REL STD DEV = 0.089 (281.458)

Sample % Abs (% Abs Ref)
Sample #1 = 9.7400 (-0.0140)
Sample #2 = 9.7160 (-0.0020)
Sample #3 = 9.7190 (0.0000)
Sample #4 = 9.7350 (-0.0090)
Avg % Abs = 9.7233 (-0.0037)
STD DEV = 0.0102 (0.0047)
REL STD DEV = 0.105 (128.886)

Sol Value = 0.400 g/210L ***
Fit value = 0.4762 mg/l 222%
Samples Taken = 4, Discarded = 1
Sum Io = 12756, Sum Io = 12954

Sample % Abs (% Abs Ref)
Sample #1 = 3.5560 (-0.0050)
Sample #2 = 3.5640 (0.0230)
Sample #3 = 3.5570 (0.0090)
Sample #4 = 3.5500 (0.0160)
Avg % Abs = 3.5570 (0.0160)
STD DEV = 0.0070 (0.0070)
REL STD DEV = 0.197 (43.750)

Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.222
Std Dev = 0.00 Rel Std Dev = 1.37
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.907
Std Dev = 0.01 Rel Std Dev = 0.99
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 1.918
Std Dev = 0.01 Rel Std Dev = 0.32
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 3.557
Std Dev = 0.01 Rel Std Dev = 0.20
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 5.165
Std Dev = 0.01 Rel Std Dev = 0.09
Zero Order Coef = -620.08
First Order Coef = 2765.52
Second Order Coef = 23.47
Standard Deviation = 7.147159

Sample #1 = 2600.00
Sample #2 = 2668.00
Sample #3 = 2698.00
Sample #4 = 2692.00
Average Result = 2686.0000
STD DEV = 15.8745
REL STD DEV = 0.591

Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.474
Std Dev = 0.00 Rel Std Dev = 0.56
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 1.880
Std Dev = 0.02 Rel Std Dev = 0.85
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 3.721
Std Dev = 0.01 Rel Std Dev = 0.35
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 6.782
Std Dev = 0.01 Rel Std Dev = 0.15
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 9.723
Std Dev = 0.01 Rel Std Dev = 0.11
Zero Order Coef = -681.92
First Order Coef = 1418.61
Second Order Coef = 12.46
Standard Deviation = 8.295056

Dry Gas H2O Adjust Results *****
Barometric Pressure = 1014
3 um H2O Adjust (mg/l*10,000) = 11.23
9 um H2O Adjust (mg/l*10,000) = 1467
**** AUTO CAL PASS

Solution Stats Quadratic Fit Chan 2
Act Fit Residual
g/210L g/210L g/210L
0.000 -0.000 0.000
-0.040 0.040 -0.000
0.100 0.100 -0.000
0.200 0.200 0.000
0.300 0.300 -0.000

Sol Value = 0.080 g/210L ***
Fit value = 0.3810 mg/l 222%
Samples Taken = 4, Discarded = 1
**** CHANNEL 1
Sample #1 = 2600.00
Sample #2 = 2668.00
Sample #3 = 2698.00
Sample #4 = 2692.00
Average Result = 2686.0000
STD DEV = 15.8745
REL STD DEV = 0.591

Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.474
Std Dev = 0.00 Rel Std Dev = 0.56
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 1.880
Std Dev = 0.02 Rel Std Dev = 0.85
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 3.721
Std Dev = 0.01 Rel Std Dev = 0.35
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 6.782
Std Dev = 0.01 Rel Std Dev = 0.15
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 9.723
Std Dev = 0.01 Rel Std Dev = 0.11
Zero Order Coef = -681.92
First Order Coef = 1418.61
Second Order Coef = 12.46
Standard Deviation = 8.295056

Dry Gas H2O Adjust Results *****
Barometric Pressure = 1014
3 um H2O Adjust (mg/l*10,000) = 11.23
9 um H2O Adjust (mg/l*10,000) = 1467
**** AUTO CAL PASS

Optical Calibration	
SN:	80-001962
Agency: Port St. Lucie PD	
Date:	4/4/2022
Quadratic Fit: +/- 0.002g/210L	
By:	DERR

Solution Stats Quadratic Fit Chan 1
Act Fit Residual
g/210L g/210L g/210L
0.000 -0.000 0.000
0.040 0.040 -0.000
0.100 0.100 -0.000
0.200 0.200 0.000
0.300 0.300 -0.000

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-001962	Port St. Lucie Police Department	4/4/2022	DERR <i>[Signature]</i>

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																																
0.047 to 0.053 ☒	0.077 to 0.083 ☒	0.194 to 0.206 ☒	0.077 to 0.083 ☒																																																																																																																																																
PORT ST LUCIE PD Intoxilyzer - Alcohol Analyzer Model 8000 04/04/2022 Software: 8100.27 SN 80-001962 <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>07:32</td></tr><tr><td>Control Test</td><td>0.047</td><td>07:33</td></tr><tr><td>Air Blank</td><td>0.000</td><td>07:34</td></tr><tr><td>Control Test</td><td>0.046</td><td>07:34</td></tr><tr><td>Air Blank</td><td>0.000</td><td>07:35</td></tr><tr><td>Control Test</td><td>0.048</td><td>07:35</td></tr><tr><td>Air Blank</td><td>0.000</td><td>07:36</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0470</td><td></td></tr><tr><td>Std Dev</td><td>0.0010</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>2.1277</td><td></td></tr></table> <i>[Signature]</i> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	07:32	Control Test	0.047	07:33	Air Blank	0.000	07:34	Control Test	0.046	07:34	Air Blank	0.000	07:35	Control Test	0.048	07:35	Air Blank	0.000	07:36	Control Test Stats			Average	0.0470		Std Dev	0.0010		Rel Std Dev(%)	2.1277		PORT ST LUCIE PD Intoxilyzer - Alcohol Analyzer Model 8000 04/04/2022 Software: 8100.27 SN 80-001962 <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>07:38</td></tr><tr><td>Control Test</td><td>0.077</td><td>07:38</td></tr><tr><td>Air Blank</td><td>0.000</td><td>07:39</td></tr><tr><td>Control Test</td><td>0.077</td><td>07:40</td></tr><tr><td>Air Blank</td><td>0.000</td><td>07:40</td></tr><tr><td>Control Test</td><td>0.077</td><td>07:41</td></tr><tr><td>Air Blank</td><td>0.000</td><td>07:41</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0770</td><td></td></tr><tr><td>Std Dev</td><td>0.0000</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr></table> <i>[Signature]</i> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	07:38	Control Test	0.077	07:38	Air Blank	0.000	07:39	Control Test	0.077	07:40	Air Blank	0.000	07:40	Control Test	0.077	07:41	Air Blank	0.000	07:41	Control Test Stats			Average	0.0770		Std Dev	0.0000		Rel Std Dev(%)	0.0000		PORT ST LUCIE PD Intoxilyzer - Alcohol Analyzer Model 8000 04/04/2022 Software: 8100.27 SN 80-001962 <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>07:43</td></tr><tr><td>Control Test</td><td>0.198</td><td>07:43</td></tr><tr><td>Air Blank</td><td>0.000</td><td>07:44</td></tr><tr><td>Control Test</td><td>0.198</td><td>07:45</td></tr><tr><td>Air Blank</td><td>0.000</td><td>07:45</td></tr><tr><td>Control Test</td><td>0.198</td><td>07:46</td></tr><tr><td>Air Blank</td><td>0.000</td><td>07:46</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.1980</td><td></td></tr><tr><td>Std Dev</td><td>0.0000</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr></table> <i>[Signature]</i> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	07:43	Control Test	0.198	07:43	Air Blank	0.000	07:44	Control Test	0.198	07:45	Air Blank	0.000	07:45	Control Test	0.198	07:46	Air Blank	0.000	07:46	Control Test Stats			Average	0.1980		Std Dev	0.0000		Rel Std Dev(%)	0.0000		PORT ST LUCIE PD Intoxilyzer - Alcohol Analyzer Model 8000 04/04/2022 Software: 8100.27 SN 80-001962 <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>07:47</td></tr><tr><td>Control Test</td><td>0.077</td><td>07:48</td></tr><tr><td>Air Blank</td><td>0.000</td><td>07:48</td></tr><tr><td>Control Test</td><td>0.077</td><td>07:49</td></tr><tr><td>Air Blank</td><td>0.000</td><td>07:49</td></tr><tr><td>Control Test</td><td>0.077</td><td>07:49</td></tr><tr><td>Air Blank</td><td>0.000</td><td>07:50</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0770</td><td></td></tr><tr><td>Std Dev</td><td>0.0000</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr></table> <i>[Signature]</i> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	07:47	Control Test	0.077	07:48	Air Blank	0.000	07:48	Control Test	0.077	07:49	Air Blank	0.000	07:49	Control Test	0.077	07:49	Air Blank	0.000	07:50	Control Test Stats			Average	0.0770		Std Dev	0.0000		Rel Std Dev(%)	0.0000	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	07:32																																																																																																																																																	
Control Test	0.047	07:33																																																																																																																																																	
Air Blank	0.000	07:34																																																																																																																																																	
Control Test	0.046	07:34																																																																																																																																																	
Air Blank	0.000	07:35																																																																																																																																																	
Control Test	0.048	07:35																																																																																																																																																	
Air Blank	0.000	07:36																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0470																																																																																																																																																		
Std Dev	0.0010																																																																																																																																																		
Rel Std Dev(%)	2.1277																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	07:38																																																																																																																																																	
Control Test	0.077	07:38																																																																																																																																																	
Air Blank	0.000	07:39																																																																																																																																																	
Control Test	0.077	07:40																																																																																																																																																	
Air Blank	0.000	07:40																																																																																																																																																	
Control Test	0.077	07:41																																																																																																																																																	
Air Blank	0.000	07:41																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0770																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	07:43																																																																																																																																																	
Control Test	0.198	07:43																																																																																																																																																	
Air Blank	0.000	07:44																																																																																																																																																	
Control Test	0.198	07:45																																																																																																																																																	
Air Blank	0.000	07:45																																																																																																																																																	
Control Test	0.198	07:46																																																																																																																																																	
Air Blank	0.000	07:46																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1980																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	07:47																																																																																																																																																	
Control Test	0.077	07:48																																																																																																																																																	
Air Blank	0.000	07:48																																																																																																																																																	
Control Test	0.077	07:49																																																																																																																																																	
Air Blank	0.000	07:49																																																																																																																																																	
Control Test	0.077	07:49																																																																																																																																																	
Air Blank	0.000	07:50																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0770																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PORT ST LUCIE PD
Time of Inspection: 06:50

Date of Inspection: 04/04/2022

Serial Number: 80-001962
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:_____ Exp:_____	0.08g/210L Test (g/210L) Lot#:_____ Exp:_____	0.20g/210L Test (g/210L) Lot#:_____ Exp:_____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:_____ Exp:_____


Number of Simulators Used: _____

Remarks:

COMPLIANCE NOT DETERMINED, AI NOT CONDUCTED.

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



DAVID E REYES-RIVERA

Signature and Printed Name

04/04/2022
Date