



# INSTRUMENT PROCESSING SHEET

Agency NAS Jacksonville PD RAW, 1/6/22S/N 80-002097

Florida Department of Law Enforcement

Date In 1/3/2022DI Completion Date 1/3/20221/4/22, RAW Ship  P/U  H/D  CMI  EE

<b>Intake</b> By RAW _____ <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE  Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight  Other Equipment/ Accessories: <input checked="" type="checkbox"/> Power cord <input checked="" type="checkbox"/> Printer Cable <input type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable  Notes: _____ _____ _____ _____ _____ _____ _____ _____	<b>Quality Checks</b> By RAW _____ Date <u>01/03/22</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>209</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP105</u> 32 mm <u>.160</u> (.139 - .169) 36 mm <u>.171</u> (.156 - .190) 53 mm <u>.238</u> (.228 - .278) 103 mm <u>.515</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>30793</u> <input checked="" type="checkbox"/> Stability Checks	<b>Flow Calibration</b> By _____ Date _____ Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																								
	<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td rowspan="2">MP6291</td> <td>202010A</td> </tr> <tr> <td></td> <td>10/05/22</td> </tr> <tr> <td>0.080</td> <td rowspan="2">MP6292</td> <td>202010B</td> </tr> <tr> <td></td> <td>10/05/22</td> </tr> <tr> <td>0.200</td> <td rowspan="2">MP6293</td> <td>202010D</td> </tr> <tr> <td></td> <td>10/06/22</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG113403</td> </tr> <tr> <td></td> <td></td> <td>05/14/23</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	MP6291	202010A		10/05/22	0.080	MP6292	202010B		10/05/22	0.200	MP6293	202010D		10/06/22	0.080 DGS	N/A	AG113403			05/14/23	<b>Maintenance</b> By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____
Simulator	Serial #	Lot #/Exp																								
0.050	MP6291	202010A																								
		10/05/22																								
0.080	MP6292	202010B																								
		10/05/22																								
0.200	MP6293	202010D																								
		10/06/22																								
0.080 DGS	N/A	AG113403																								
		05/14/23																								
		<b>DI Temp. Checks</b> By RAW _____ <input checked="" type="checkbox"/> Lab Temp °C <u>22.27</u> External Digital Therm. ID#: <u>381189</u> <input checked="" type="checkbox"/> 34°C +/-2 Serial #: <u>MP6291</u> <input checked="" type="checkbox"/> 34°C +/-2 Serial #: <u>MP6292</u> <input checked="" type="checkbox"/> 34°C +/-2 Serial #: <u>MP6293</u>																								

<b>Calibration Adjustment</b> By _____ Barometric Pressure Gauge _____ ID # _____ <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> <input type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>	Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			<b>Department Inspection</b> By RAW _____ Barometric Pressure ID# <u>30793</u> Gauge <u>1025</u> Instrument <u>1016</u> Mouth Alcohol Solution Lot # <u>2021-D</u> Acetone Stock Solution Lot # <u>2021-C</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP6289</td> </tr> <tr> <td>Interferent</td> <td>MP6290</td> </tr> <tr> <td>0.050</td> <td>MP6291</td> </tr> <tr> <td>0.080</td> <td>MP6292</td> </tr> <tr> <td>0.200</td> <td>MP6293</td> </tr> </tbody> </table>	Simulator	Serial Number	0.000	MP6289	Interferent	MP6290	0.050	MP6291	0.080	MP6292	0.200	MP6293
Simulator	Serial #	Lot #	Expiration																																																										
0.000		N/A	N/A																																																										
0.040																																																													
0.100																																																													
0.200																																																													
0.300																																																													
0.080 DGS	N/A																																																												
Simulator	Serial #	Lot #	Expiration																																																										
0.050																																																													
0.080																																																													
0.200																																																													
0.080 DGS	N/A																																																												
Simulator	Serial Number																																																												
0.000	MP6289																																																												
Interferent	MP6290																																																												
0.050	MP6291																																																												
0.080	MP6292																																																												
0.200	MP6293																																																												
	<b>Attachments</b> <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Stability Checks <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Calibration Adjustment <input type="checkbox"/> Post-Stability Checks <input type="checkbox"/> Flow Calibration <input checked="" type="checkbox"/> Form 40 <input type="checkbox"/> Other _____																																																												

Notes/Suggested Service: _____ Tech Review: Updated DI completion date RAW, 01/04/22 Admin Review Updated agency name to reflect that of instrument set-up. RAW, 1/6/22 _____ _____ _____ _____ _____	<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use  <div style="text-align: right;"> <small>2022.01.06 14:34:04</small> </div>
	Israel Soto <small>Digitally signed by Israel Soto Date: 2022.01.05 10:42:55 -0500</small> Tech Review / Date _____ Admin Review / Date _____

# Florida Department of Law Enforcement Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: NAS JACKSONVILLE PD  
Time of Inspection: 11:29

Date of Inspection: 01/04/2022

Serial Number: 80-002097  
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202010A Exp: 10/05/2022	0.08g/210L Test (g/210L) Lot#:202010B Exp: 10/05/2022	0.20g/210L Test (g/210L) Lot#:202010D Exp: 10/06/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG113403 Exp: 05/14/2023
0.000	0.050	0.079	0.201	0.081
0.000	0.049	0.079	0.200	0.081
0.000	0.050	0.079	0.201	0.081
0.000	0.050	0.079	0.201	0.080
0.000	0.050	0.080	0.201	0.080
0.000	0.050	0.079	0.201	0.080
0.000	0.049	0.080	0.201	0.081
0.000	0.050	0.080	0.201	0.081
0.000	0.050	0.079	0.201	0.081
0.000	0.050	0.080	0.201	0.081


Standard Deviations	0.0004	0.0005	0.0003	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



RICHARD A WILLIAMS

Signature and Printed Name

01/04/2022  
Date

NAS JACKSONVILLE PD  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-002097  
 01/03/2022  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	16:37
Control Test	0.049	16:38
Air Blank	0.000	16:38
Control Test	0.048	16:39
Air Blank	0.000	16:39
Control Test	0.048	16:40
Air Blank	0.000	16:41
Control Test Stats		
Average	0.0483	
Std Dev	0.0006	
Rel Std Dev(%)	1.1945	

Stability Checks  
 80-002097  
 01/03/22  
 (PAW)

.05

(PAW)

Operator's Signature

NAS JACKSONVILLE PD  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-002097  
 01/03/2022  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	16:43
Control Test	0.079	16:43
Air Blank	0.000	16:44
Control Test	0.078	16:45
Air Blank	0.000	16:45
Control Test	0.078	16:46
Air Blank	0.000	16:47
Control Test Stats		
Average	0.0783	
Std Dev	0.0006	
Rel Std Dev(%)	0.7370	

.08  
wet

(PAW)

Operator's Signature

NAS JACKSONVILLE PD  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-002097  
 01/03/2022  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	16:49
Control Test	0.199	16:50
Air Blank	0.000	16:50
Control Test	0.197	16:51
Air Blank	0.000	16:52
Control Test	0.196	16:52
Air Blank	0.000	16:53
Control Test Stats		
Average	0.1973	
Std Dev	0.0015	
Rel Std Dev(%)	0.7741	

.20

(PAW)

Operator's Signature

NAS JACKSONVILLE PD  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-002097  
 01/03/2022  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	16:55
Control Test	0.080	16:56
Air Blank	0.000	16:56
Control Test	0.091	16:56
Air Blank	0.000	16:57
Control Test	0.081	16:57
Air Blank	0.000	16:58
Control Test Stats		
Average	0.0807	
Std Dev	0.0006	
Rel Std Dev(%)	0.7157	

.08  
DGS

(PAW)

Operator's Signature

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: NAS JACKSONVILLE PD  
Time of Inspection: 16:35

Date of Inspection: 01/03/2022

Serial Number: 80-002097  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____

Number of Simulators Used: \_\_\_\_\_


**Remarks:**

BYPASSED AI TO OPERATE INSTRUMENT, COMPLIANCE NOT DETERMINED. **RAW, 01/03/22**

X Not Determined

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

  
 \_\_\_\_\_  
 Signature and Printed Name

RICHARD A WILLIAMS

01/03/2022  
Date



# Calibration Certificate

Florida Department of Law Enforcement  
Alcohol Testing Program  
2729 Fort Knox Blvd.  
Bldg. 2, Suite 1300  
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-002097, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-002097</u>	UNCERTAINTY* ±	
Owning Agency:	<u>NAS JACKSONVILLE PD</u>	0.050 g/ 210 L	0.005
Calibration Date:	<u>01/04/2022</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>11:29</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

\*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

## TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Richard A Williams  
Digitally signed by Richard A Williams  
Date: 2022.01.04 13:58:19 -05'00'

01/04/2022

Date

RICHARD A WILLIAMS,  
Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality