



INSTRUMENT PROCESSING SHEET

Agency Florida Fish and Wildlife Conservation Commission Conservation IS 04-04-2022 S/N 80-003197Florida Department of
Law EnforcementDate In 04-01-2022 DI Completion Date 04-04-2022 ☐ Ship ☒ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By IS	Quality Checks	By IS	Date 04-04-2022	Flow Calibration	By	Date																																																												
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes: Agency Inspector stated that instrument lost forms. Tech Review: Corrected agency name IS 04-04-2022		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>204</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP-102</u> 32 mm <u>0.152</u> (.139 - .169) 36 mm <u>0.167</u> (.156 - .190) 53 mm <u>0.234</u> (.228 - .278) 103 mm <u>0.503</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>30793</u> ☒ Stability Checks			☐ Flow Column # _____ ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																																														
		<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP6291</td><td>202201C 01-11-2024</td></tr><tr><td>0.080</td><td>MP6292</td><td>202201D 01-18-2024</td></tr><tr><td>0.200</td><td>MP6293</td><td>202201E 01-18-2024</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG113403 05-14-2023</td></tr></tbody></table>	Simulator	Serial #	Lot #/Exp	0.050	MP6291	202201C 01-11-2024	0.080	MP6292	202201D 01-18-2024	0.200	MP6293	202201E 01-18-2024	0.080 DGS	N/A	AG113403 05-14-2023			Maintenance By IS ☒ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☒ Other CMI uploaded forms remotely through phone line.																																															
Simulator	Serial #	Lot #/Exp																																																																	
0.050	MP6291	202201C 01-11-2024																																																																	
0.080	MP6292	202201D 01-18-2024																																																																	
0.200	MP6293	202201E 01-18-2024																																																																	
0.080 DGS	N/A	AG113403 05-14-2023																																																																	
Calibration Adjustment By _____				Department Inspection By IS _____																																																															
Barometric Pressure Gauge _____ ID # _____ <table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table> ☐ Post Calibration Adjustment Stability Checks <table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>				Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Barometric Pressure ID# <u>28421</u> Gauge <u>1014</u> Instrument <u>1012</u> Mouth Alcohol Solution Lot # <u>2021-D</u> Acetone Stock Solution Lot # <u>2021-C</u> <table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP6289</td></tr><tr><td>Interferent</td><td>MP6290</td></tr><tr><td>0.050</td><td>MP6291</td></tr><tr><td>0.080</td><td>MP6292</td></tr><tr><td>0.200</td><td>MP6293</td></tr></tbody></table> Attachments ☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Calibration Adjustment ☐ Post-Stability Checks ☐ Flow Calibration ☒ Form 40 x2 ☒ Other Extra Stabilities				Simulator	Serial Number	0.000	MP6289	Interferent	MP6290	0.050	MP6291	0.080	MP6292	0.200	MP6293
Simulator	Serial #	Lot #	Expiration																																																																
0.000		N/A	N/A																																																																
0.040																																																																			
0.100																																																																			
0.200																																																																			
0.300																																																																			
0.080 DGS	N/A																																																																		
Simulator	Serial #	Lot #	Expiration																																																																
0.050																																																																			
0.080																																																																			
0.200																																																																			
0.080 DGS	N/A																																																																		
Simulator	Serial Number																																																																		
0.000	MP6289																																																																		
Interferent	MP6290																																																																		
0.050	MP6291																																																																		
0.080	MP6292																																																																		
0.200	MP6293																																																																		
Notes/Suggested Service: Forms uploaded prior to Quality Checks and Department Inspection. Battery Change done prior to Quality Checks, forms upload, and Department Inspection. Instrument had Interferent Detect during DI H2O test, double checked that correct simulator was connected and repeated with nominal results. Performed extra DI H2O stability tests to ensure accuracy/reliability. IS				☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use 2022.04. David Eliezer Reyes-Rivera Digitally signed by David Eliezer Reyes-Rivera Date: 2022.04.04 13:43:16 -04'00' Tech Review / Date _____ Admin Review / Date _____																																																															

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FFWCC

Time of Inspection: 11:18

Date of Inspection: 04/04/2022

Serial Number: 80-003197

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG113403 Exp: 05/14/2023
INT / 0.000	0.050	0.080	0.201	0.079
/ 0.000	0.051	0.080	0.201	0.079
/ 0.000	0.050	0.080	0.201	0.079
/ 0.000	0.050	0.080	0.201	0.078
/ 0.000	0.050	0.080	0.202	0.079
/ 0.000	0.050	0.080	0.202	0.079
/ 0.000	0.050	0.080	0.201	0.079
/ 0.000	0.051	0.080	0.202	0.079
/ 0.000	0.050	0.080	0.201	0.079
/ 0.000	0.051	0.080	0.201	0.079

Standard Deviations	0.0004	0.0000	0.0004	0.0003
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0002 Number of Simulators Used: 5

Remarks:

00: Interferent Detect DOUBLE CHECKED CORRECT SIM.

will run additional DI H₂O stability tests

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Israel Soto

ISRAEL SOTO

Signature and Printed Name

04/04/2022
Date

Stability checks

FFWCC
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-003197
04/04/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:19
Control Test	0.050	07:20
Air Blank	0.000	07:20
Control Test	0.050	07:21
Air Blank	0.000	07:21
Control Test	0.049	07:22
Air Blank	0.000	07:23
Control Test Stats		
Average	0.0497	
Std Dev	0.0006	
Rel Std Dev(%)	1.1625	

Operator's Signature

FFWCC
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-003197
04/04/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:24
Control Test	0.080	07:24
Air Blank	0.000	07:25
Control Test	0.081	07:25
Air Blank	0.000	07:26
Control Test	0.080	07:27
Air Blank	0.000	07:27
Control Test Stats		
Average	0.0803	
Std Dev	0.0006	
Rel Std Dev(%)	0.7187	

Operator's Signature

FFWCC
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-003197
04/04/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:28
Control Test	0.201	07:29
Air Blank	0.000	07:29
Control Test	0.201	07:30
Air Blank	0.000	07:30
Control Test	0.201	07:31
Air Blank	0.000	07:32
Control Test Stats		
Average	0.2010	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature

FFWCC
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-003197
04/04/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:33
Control Test	0.080	07:33
Air Blank	0.000	07:34
Control Test	0.079	07:34
Air Blank	0.000	07:34
Control Test	0.080	07:35
Air Blank	0.000	07:35
Control Test Stats		
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	

Operator's Signature

Dry



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road.
Suite B1032
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-003197, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-003197</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>FFWCC</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>04/04/2022</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>11:18</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.
This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Israel Soto
Soto
Date: 2022.04.04 13:25:06
-04'00'

Date

ISRAEL SOTO,

FDLE/ATP Form 69 March 2022

Issuing Authority: Alcohol Testing Program

Department Inspector

Service • Integrity • Respect • Quality

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FFWCC

Serial Number: 80-003197

Time of Inspection: 09:23

Date of Inspection: 04/01/2022

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot #: Exp:

Number of Simulators Used: _____

Remarks:

BYPASSED AI TO OPERATE INSTRUMENT

Instrument lost forms and could not print, reprinted from database

The above instrument complies (☒) does not comply () with Chapter 11D-8, FAC.

N/A compliance not determined

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Israel Soto

ISRAEL SOTO

Signature and Printed Name

04/01/2022

Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FFWCC

Time of Inspection: 07:09

Date of Inspection: 04/04/2022

Serial Number: 80-003197

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:_____ Exp:_____	0.08g/210L Test (g/210L) Lot#:_____ Exp:_____	0.20g/210L Test (g/210L) Lot#:_____ Exp:_____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:_____ Exp:_____

Number of Simulators Used: _____

Remarks:

CHECKING FORMS UPLOAD

The above instrument complies (N/A) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Israel Soto

ISRAEL SOTO

Signature and Printed Name

04/04/2022

Date

FFWCC
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-003197
04/04/2022
Software: 8.00.27

Test	g/210L	Time
Air Blank	0.000	11:20
Control Test	0.000	11:21
Air Blank	0.000	11:21
Control Test	0.000	11:22
Air Blank	0.000	11:22
Control Test	0.000	11:23
Air Blank	0.000	11:23
Control Test	0.000	11:24
Air Blank	0.000	11:25
Control Test	0.000	11:25
Air Blank	0.000	11:26
Control Test	0.000	11:26
Air Blank	0.000	11:27
Control Test	0.000	11:28
Air Blank	0.000	11:28
Control Test	0.000	11:29
Air Blank	0.000	11:29
Control Test	0.000	11:30
Air Blank	0.000	11:30
Control Test	0.000	11:31
Air Blank	0.000	11:31
Control Test	0.000	11:32
Air Blank	0.000	11:33
Control Test	0.000	11:33
Air Blank	0.000	11:34
Control Test	0.000	11:34
Air Blank	0.000	11:35
Control Test	0.000	11:36
Air Blank	0.000	11:36
Control Test	0.000	11:37
Air Blank	0.000	11:37
Control Test	0.000	11:38
Air Blank	0.000	11:38
Control Test	0.000	11:39
Air Blank	0.000	11:40
Control Test	0.000	11:40
Air Blank	0.000	11:41
Control Test	0.000	11:41
Air Blank	0.000	11:42
Control Test	0.000	11:43
Air Blank	0.000	11:43
Control Test Stats		
Average	0.0000	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Extra DI H₂O
Stability Tests


Operator's Signature