



INSTRUMENT PROCESSING SHEET

Agency Pasco CSOS/N 80-003347Florida Department of
Law EnforcementDate In 3/7/2022DI Completion Date 3/7/2022☐ Ship ☒ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By TDG	Quality Checks	By TDG	Date <u>3/7/2022</u>	Flow Calibration	By	Date																																						
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes: _____ _____ _____ _____ _____ _____ _____ _____		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>126</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP104</u> 32 mm <u>0.144</u> (.139 - .169) 36 mm <u>0.164</u> (.156 - .190) 53 mm <u>0.234</u> (.228 - .278) 103 mm <u>0.511</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>68639</u> ☒ Stability Checks			Flow Column # _____ ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																								
		<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP6286</td><td>202010A 10/05/2022</td></tr><tr><td>0.080</td><td>MP6287</td><td>202010B 10/05/2022</td></tr><tr><td>0.200</td><td>MP6288</td><td>202010D 10/06/2022</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG115904 06/08/2023</td></tr></tbody></table>	Simulator	Serial #	Lot #/Exp	0.050	MP6286	202010A 10/05/2022	0.080	MP6287	202010B 10/05/2022	0.200	MP6288	202010D 10/06/2022	0.080 DGS	N/A	AG115904 06/08/2023			Maintenance By _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____ _____ _____ _____ _____																									
Simulator	Serial #	Lot #/Exp																																											
0.050	MP6286	202010A 10/05/2022																																											
0.080	MP6287	202010B 10/05/2022																																											
0.200	MP6288	202010D 10/06/2022																																											
0.080 DGS	N/A	AG115904 06/08/2023																																											
Calibration Adjustment By _____		Department Inspection By <u>TDG</u>																																											
Barometric Pressure Gauge _____ ID # _____		Barometric Pressure ID# <u>28199</u> Gauge <u>1020</u> Instrument <u>1020</u> Mouth Alcohol Solution Lot # <u>2021-B</u> Acetone Stock Solution Lot # <u>2021-C</u>																																											
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP6284</td></tr><tr><td>Interferent</td><td>MP6285</td></tr><tr><td>0.050</td><td>MP6286</td></tr><tr><td>0.080</td><td>MP6287</td></tr><tr><td>0.200</td><td>MP6288</td></tr></tbody></table>				Simulator	Serial Number	0.000	MP6284	Interferent	MP6285	0.050	MP6286	0.080	MP6287	0.200	MP6288
Simulator	Serial #	Lot #	Expiration																																										
0.000		N/A	N/A																																										
0.040																																													
0.100																																													
0.200																																													
0.300																																													
0.080 DGS	N/A																																												
Simulator	Serial Number																																												
0.000	MP6284																																												
Interferent	MP6285																																												
0.050	MP6286																																												
0.080	MP6287																																												
0.200	MP6288																																												
☐ Post Calibration Adjustment Stability Checks		Attachments ☒ Form 41 ☐ Post-Stability Checks ☒ Stability Checks ☐ Flow Calibration ☒ Calibration Certificate ☐ Form 40 ☐ Calibration Adjustment ☐ Other _____																																											
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use 2022.03.																							
Simulator	Serial #	Lot #	Expiration																																										
0.050																																													
0.080																																													
0.200																																													
0.080 DGS	N/A																																												
Notes/Suggested Service: _____ _____ _____ _____ _____ _____ _____ _____		David Eliezer Digitally signed by David Eliezer Reyes-Rivera Date: 2022.03.07 14:45:23 +05'00' Reyes-Rivera Tech Review / Date <u>07 15:01:22</u> Admin Review / Date <u>05:00</u>																																											

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-003347	Pasco CSO	03/07/2022	TDG <i>ML</i>

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083
✓	✓	✓	✓
PASCO COUNTY SO Intoxilyzer - Alcolon Analyzer Model 8000 03/07/2022 Software: 8100.27 SN 80-003347 Test g/210L Time Air Blank 0.000 10:50 Control Test 0.049 10:51 Air Blank 0.000 10:52 Control Test 0.048 10:53 Air Blank 0.000 10:53 Control Test 0.049 10:54 Control Test Stats Average 0.0487 Std Dev 0.0006 Rel Std Dev(%) 1.1863	PASCO COUNTY SO Intoxilyzer - Alcolon Analyzer Model 8000 03/07/2022 Software: 8100.27 SN 80-003347 Test g/210L Time Air Blank 0.000 10:58 Control Test 0.078 10:59 Air Blank 0.000 10:59 Control Test 0.078 10:59 Air Blank 0.000 11:00 Control Test 0.078 11:01 Air Blank 0.000 11:01 Control Test Stats Average 0.0780 Std Dev 0.0000 Rel Std Dev(%) 0.0000	PASCO COUNTY SO Intoxilyzer - Alcolon Analyzer Model 8000 03/07/2022 Software: 8100.27 SN 80-003347 Test g/210L Time Air Blank 0.000 11:05 Control Test 0.196 11:06 Air Blank 0.000 11:06 Control Test 0.196 11:07 Air Blank 0.000 11:08 Control Test 0.197 11:08 Air Blank 0.000 11:09 Control Test Stats Average 0.1963 Std Dev 0.0006 Rel Std Dev(%) 0.2941	PASCO COUNTY SO Intoxilyzer - Alcolon Analyzer Model 8000 03/07/2022 Software: 8100.27 SN 80-003347 Test g/210L Time Air Blank 0.000 11:12 Control Test 0.080 11:12 Air Blank 0.000 11:13 Control Test 0.081 11:13 Air Blank 0.000 11:14 Control Test 0.081 11:14 Air Blank 0.000 11:15 Control Test Stats Average 0.0807 Std Dev 0.0006 Rel Std Dev(%) 0.7157
<i>ML</i> Operator's Signature	<i>ML</i> Operator's Signature	<i>ML</i> Operator's Signature	<i>ML</i> Operator's Signature

Comments:

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: PASCO COUNTY SO
Time of Inspection: 14:05

Date of Inspection: 03/07/2022

Serial Number: 80-003347
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202010A Exp: 10/05/2022	0.08g/210L Test (g/210L) Lot#:202010B Exp: 10/05/2022	0.20g/210L Test (g/210L) Lot#:202010D Exp: 10/06/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG115904 Exp: 06/08/2023
0.000	0.049	0.078	0.196	0.081
0.000	0.049	0.078	0.197	0.081
0.000	0.049	0.078	0.197	0.080
0.000	0.049	0.078	0.197	0.081
0.000	0.048	0.078	0.196	0.080
0.000	0.048	0.078	0.196	0.081
0.000	0.049	0.078	0.197	0.080
0.000	0.048	0.079	0.196	0.080
0.000	0.048	0.079	0.197	0.080
0.000	0.048	0.078	0.196	0.081


Standard Deviations	0.0005	0.0004	0.0005	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

03/07/2022
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-003347, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-003347</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>PASCO COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>03/07/2022</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>14:05</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

03/07/2022

Date


TAYLOR D GUTSCHOW,

Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality