



INSTRUMENT PROCESSING SHEET

Agency Florida Highway Patrol Troop KS/N 80-003405Florida Department of
Law EnforcementDate In 12/14/2022DI Completion Date 12/14/2022☐ Ship ☒ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By DERR	Quality Checks	By DERR	Date	12/14/2022	Flow Calibration	By	Date																																						
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes: _____ _____ _____ _____ _____ _____ _____ _____ _____ _____		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>154</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP101</u> 32 mm <u>0.148</u> (.139 - .169) 36 mm <u>0.167</u> (.156 - .190) 53 mm <u>0.230</u> (.228 - .278) 103 mm <u>0.500</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28199</u> ☒ Stability Checks				Flow Column # _____ ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																								
		<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP6286</td><td>202201C 01/11/2024</td></tr><tr><td>0.080</td><td>MP6287</td><td>202201D 01/18/2024</td></tr><tr><td>0.200</td><td>MP6288</td><td>202201E 01/18/2024</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>00521080A2 02/05/2023</td></tr></tbody></table>			Simulator	Serial #	Lot #/Exp	0.050	MP6286	202201C 01/11/2024	0.080	MP6287	202201D 01/18/2024	0.200	MP6288	202201E 01/18/2024	0.080 DGS	N/A	00521080A2 02/05/2023	<table border="1"><thead><tr><th colspan="2">Maintenance</th><th>By</th></tr></thead><tbody><tr><td colspan="2">☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____</td><td>_____</td></tr></tbody></table>			Maintenance		By	☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____		_____																		
Simulator	Serial #	Lot #/Exp																																												
0.050	MP6286	202201C 01/11/2024																																												
0.080	MP6287	202201D 01/18/2024																																												
0.200	MP6288	202201E 01/18/2024																																												
0.080 DGS	N/A	00521080A2 02/05/2023																																												
Maintenance		By																																												
☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____		_____																																												
Calibration Adjustment		By _____		Department Inspection				By DERR _____																																						
Barometric Pressure Gauge _____ ID # _____				Barometric Pressure ID# <u>28663</u> Gauge <u>1017</u> Instrument <u>1014</u> Mouth Alcohol Solution Lot # <u>2021-D</u> Acetone Stock Solution Lot # <u>2021-C</u>																																										
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A					<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP6284</td></tr><tr><td>Interferent</td><td>MP6285</td></tr><tr><td>0.050</td><td>MP6286</td></tr><tr><td>0.080</td><td>MP6287</td></tr><tr><td>0.200</td><td>MP6288</td></tr></tbody></table>		Simulator	Serial Number	0.000	MP6284	Interferent	MP6285	0.050	MP6286	0.080	MP6287	0.200	MP6288	
Simulator	Serial #	Lot #	Expiration																																											
0.000		N/A	N/A																																											
0.040																																														
0.100																																														
0.200																																														
0.300																																														
0.080 DGS	N/A																																													
Simulator	Serial Number																																													
0.000	MP6284																																													
Interferent	MP6285																																													
0.050	MP6286																																													
0.080	MP6287																																													
0.200	MP6288																																													
☐ Post Calibration Adjustment Stability Checks				Attachments																																										
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A					☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Calibration Adjustment		☐ Post-Stability Checks ☐ Flow Calibration ☐ Form 40 ☐ Other _____																				
Simulator	Serial #	Lot #	Expiration																																											
0.050																																														
0.080																																														
0.200																																														
0.080 DGS	N/A																																													
Notes/Suggested Service: _____ _____ _____ _____ _____ _____ _____ _____ _____ _____				☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																																										
		Taylor Gutschow		Digitally signed by Taylor Gutschow Date: 2022.12.14 10:17:02 -0500		Israel Soto		Digitally signed by Israel Soto Date: 2022.12.14 10:42:43 -0500																																						
		Tech Review / Date				Admin Review / Date																																								



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-003405, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-003405</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>FHP TROOP K</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>12/14/2022</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>10:01</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

12/14/2022

Date


DAVID E REYES-RIVERA,
Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Page 1 of 1

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FHP TROOP K
Time of Inspection: 10:01

Date of Inspection: 12/14/2022

Serial Number: 80-003405
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:00521080A2 Exp: 02/05/2023
0.000	0.050	0.079	0.200	0.081
0.000	0.050	0.079	0.201	0.081
0.000	0.050	0.079	0.201	0.081
0.000	0.050	0.079	0.200	0.080
0.000	0.050	0.079	0.201	0.081
0.000	0.050	0.080	0.200	0.080
0.000	0.051	0.080	0.200	0.080
0.000	0.051	0.079	0.201	0.080
0.000	0.051	0.080	0.201	0.080
0.000	0.051	0.080	0.201	0.080


Standard Deviations	0.0005	0.0005	0.0005	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0005 Number of Simulators Used: 5


Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 DAVID E REYES-RIVERA
Signature and Printed Name

12/14/2022
Date

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-003405	Florida Highway Patrol Troop K	12/14/2022	DERR 

0.05g/210L 0.047 to 0.053 ☒	0.08g/210L 0.077 to 0.083 ☒	0.20g/210L 0.194 to 0.206 ☒	DGS 0.08g/210L 0.077 to 0.083 ☒																																																																																																																																																
FHP TROOP K Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-003405 12/14/2022 Software: 8100.27 <table> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> <tr><td>Air Blank</td><td>0.000</td><td>08:22</td></tr> <tr><td>Control Test</td><td>0.050</td><td>08:23</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:24</td></tr> <tr><td>Control Test</td><td>0.049</td><td>08:24</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:25</td></tr> <tr><td>Control Test</td><td>0.050</td><td>08:26</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:26</td></tr> <tr><td colspan="3">Control Test Stats</td></tr> <tr><td>Average</td><td>0.0497</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.1625</td><td></td></tr> </table> Operator's Signature 	Test	g/210L	Time	Air Blank	0.000	08:22	Control Test	0.050	08:23	Air Blank	0.000	08:24	Control Test	0.049	08:24	Air Blank	0.000	08:25	Control Test	0.050	08:26	Air Blank	0.000	08:26	Control Test Stats			Average	0.0497		Std Dev	0.0006		Rel Std Dev(%)	1.1625		FHP TROOP K Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-003405 12/14/2022 Software: 8100.27 <table> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> <tr><td>Air Blank</td><td>0.000</td><td>08:27</td></tr> <tr><td>Control Test</td><td>0.079</td><td>08:28</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:28</td></tr> <tr><td>Control Test</td><td>0.079</td><td>08:29</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:29</td></tr> <tr><td>Control Test</td><td>0.079</td><td>08:30</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:31</td></tr> <tr><td colspan="3">Control Test Stats</td></tr> <tr><td>Average</td><td>0.0790</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </table> Operator's Signature 	Test	g/210L	Time	Air Blank	0.000	08:27	Control Test	0.079	08:28	Air Blank	0.000	08:28	Control Test	0.079	08:29	Air Blank	0.000	08:29	Control Test	0.079	08:30	Air Blank	0.000	08:31	Control Test Stats			Average	0.0790		Std Dev	0.0000		Rel Std Dev(%)	0.0000		FHP TROOP K Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-003405 12/14/2022 Software: 8100.27 <table> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> <tr><td>Air Blank</td><td>0.000</td><td>08:32</td></tr> <tr><td>Control Test</td><td>0.200</td><td>08:32</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:33</td></tr> <tr><td>Control Test</td><td>0.200</td><td>08:33</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:34</td></tr> <tr><td>Control Test</td><td>0.200</td><td>08:35</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:35</td></tr> <tr><td colspan="3">Control Test Stats</td></tr> <tr><td>Average</td><td>0.2000</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </table> Operator's Signature 	Test	g/210L	Time	Air Blank	0.000	08:32	Control Test	0.200	08:32	Air Blank	0.000	08:33	Control Test	0.200	08:33	Air Blank	0.000	08:34	Control Test	0.200	08:35	Air Blank	0.000	08:35	Control Test Stats			Average	0.2000		Std Dev	0.0000		Rel Std Dev(%)	0.0000		FHP TROOP K Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-003405 12/14/2022 Software: 8100.27 <table> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> <tr><td>Air Blank</td><td>0.000</td><td>08:36</td></tr> <tr><td>Control Test</td><td>0.081</td><td>08:37</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:37</td></tr> <tr><td>Control Test</td><td>0.081</td><td>08:37</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:38</td></tr> <tr><td>Control Test</td><td>0.081</td><td>08:38</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:39</td></tr> <tr><td colspan="3">Control Test Stats</td></tr> <tr><td>Average</td><td>0.0810</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </table> Operator's Signature 	Test	g/210L	Time	Air Blank	0.000	08:36	Control Test	0.081	08:37	Air Blank	0.000	08:37	Control Test	0.081	08:37	Air Blank	0.000	08:38	Control Test	0.081	08:38	Air Blank	0.000	08:39	Control Test Stats			Average	0.0810		Std Dev	0.0000		Rel Std Dev(%)	0.0000	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	08:22																																																																																																																																																	
Control Test	0.050	08:23																																																																																																																																																	
Air Blank	0.000	08:24																																																																																																																																																	
Control Test	0.049	08:24																																																																																																																																																	
Air Blank	0.000	08:25																																																																																																																																																	
Control Test	0.050	08:26																																																																																																																																																	
Air Blank	0.000	08:26																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0497																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	1.1625																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	08:27																																																																																																																																																	
Control Test	0.079	08:28																																																																																																																																																	
Air Blank	0.000	08:28																																																																																																																																																	
Control Test	0.079	08:29																																																																																																																																																	
Air Blank	0.000	08:29																																																																																																																																																	
Control Test	0.079	08:30																																																																																																																																																	
Air Blank	0.000	08:31																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0790																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	08:32																																																																																																																																																	
Control Test	0.200	08:32																																																																																																																																																	
Air Blank	0.000	08:33																																																																																																																																																	
Control Test	0.200	08:33																																																																																																																																																	
Air Blank	0.000	08:34																																																																																																																																																	
Control Test	0.200	08:35																																																																																																																																																	
Air Blank	0.000	08:35																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.2000																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	08:36																																																																																																																																																	
Control Test	0.081	08:37																																																																																																																																																	
Air Blank	0.000	08:37																																																																																																																																																	
Control Test	0.081	08:37																																																																																																																																																	
Air Blank	0.000	08:38																																																																																																																																																	
Control Test	0.081	08:38																																																																																																																																																	
Air Blank	0.000	08:39																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0810																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		