



## Agency Osceola County SO

S/N 80-003937

Date In 02/10/2022      DI Completion Date 02/14/2022

☒ Ship    ☐ P/U    ☐ H/D    ☐ CMI    ☐ EE

FDLE/ATP Form 48 January 2022  
Issuing Authority: Alcohol Testing Program

Page 1 of 1

# Florida Department of Law Enforcement Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: OSCEOLA COUNTY S.O.  
Time of Inspection: 13:25

Date of Inspection: 02/14/2022

Serial Number: 80-003937  
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202010A Exp: 10/05/2022	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202010D Exp: 10/06/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG113403 Exp: 05/14/2023
0.000	0.050	0.080	0.200	0.078
0.000	0.049	0.080	0.200	0.079
0.000	0.049	0.080	0.200	0.079
0.000	0.049	0.080	0.200	0.079
0.000	0.050	0.080	0.200	0.079
0.000	0.049	0.080	0.199	0.079
0.000	0.050	0.080	0.200	0.079
0.000	0.050	0.080	0.200	0.078
0.000	0.050	0.080	0.200	0.079
0.000	0.050	0.080	0.199	0.079


Standard Deviations	0.0005	0.0000	0.0004	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies ( ☒ ) does not comply ( ☐ ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 RICHARD A WILLIAMS  
Signature and Printed Name

02/14/2022  
Date

OSCEOLA COUNTY S.O.  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-003937  
02/14/2022  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:57
Control Test	0.050	09:58
Air Blank	0.000	09:58
Control Test	0.050	09:59
Air Blank	0.000	09:59
Control Test	0.050	10:00
Air Blank	0.000	10:01
Control Test Stats		
Average	0.0500	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

0.05

PAW

Operator's Signature

OSCEOLA COUNTY S.O.  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-003937  
02/14/2022  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:04
Control Test	0.079	10:05
Air Blank	0.000	10:05
Control Test	0.080	10:06
Air Blank	0.000	10:06
Control Test	0.080	10:07
Air Blank	0.000	10:08
Control Test Stats		
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	

0.08  
wet

PAW

Operator's Signature

OSCEOLA COUNTY S.O.  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-003937  
02/14/2022  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:10
Control Test	0.199	10:10
Air Blank	0.000	10:11
Control Test	0.200	10:12
Air Blank	0.000	10:12
Control Test	0.200	10:13
Air Blank	0.000	10:13
Control Test Stats		
Average	0.1997	
Std Dev	0.0006	
Rel Std Dev(%)	0.2892	

0.20

PAW

Operator's Signature

OSCEOLA COUNTY S.O.  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-003937  
02/14/2022  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:18
Control Test	0.078	10:18
Air Blank	0.000	10:19
Control Test	0.079	10:19
Air Blank	0.000	10:20
Control Test	0.078	10:20
Air Blank	0.000	10:20
Control Test Stats		
Average	0.0783	
Std Dev	0.0006	
Rel Std Dev(%)	0.7370	

0.08  
DGS

PAW

Operator's Signature

Stability Checks

80-003937

02/14/22

PAW



# Calibration Certificate

Florida Department of Law Enforcement  
Alcohol Testing Program  
2729 Fort Knox Blvd.  
Bldg. 2, Suite 1300  
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-003937, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-003937</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>OSCEOLA COUNTY S.O.</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>02/14/2022</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>13:25</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within  $\pm 0.005$  or 5%, whichever is greater, of the target alcohol concentration.

\*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ( $k=3$ ).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

## TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

**Richard A Williams**  
Digitally signed by Richard A Williams  
Date: 2022.02.14 16:26:30 -05'00'

**02/14/2022**

Date

**RICHARD A WILLIAMS,**  
**Department Inspector**

FDLE/ATP Form 69 January 2022

Issuing Authority: Alcohol Testing Program

*Service • Integrity • Respect • Quality*





## Agency Osceola County S.O.

S/N 80-003937

Date In 1/6/2022      DI Completion Date 1/6/2022

☒ Ship   ☐ P/U   ☐ H/D   ☐ CMI   ☐ EE

FDLE/ATP Form 48 January 2021  
Issuing Authority: Alcohol Testing Program

Page 1 of 1

# Florida Department of Law Enforcement Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: OSCEOLA COUNTY S.O.  
Time of Inspection: 16:10

Date of Inspection: 01/06/2022

Serial Number: 80-003937  
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202010A Exp: 10/05/2022	0.08g/210L Test (g/210L) Lot#:202010B Exp: 10/05/2022	0.20g/210L Test (g/210L) Lot#:202010D Exp: 10/06/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG113403 Exp: 05/14/2023
0.000	0.049	0.078	0.197	0.077
0.000	0.049	0.079	0.198	0.076
0.000	0.049	0.079	0.198	0.077
0.000	0.049	0.078	0.200	0.077
0.000	0.049	0.079	0.199	0.077
0.000	0.049	0.079	0.200	0.077
0.000	0.049	0.079	0.200	0.077
0.000	0.049	0.078	0.200	0.077
0.000	0.049	0.079	0.200	0.077
0.000	0.049	0.079	0.200	0.077


Standard Deviations	0.0000	0.0004	0.0011	0.0003
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies ( ☒ ) does not comply ( ☐ ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



RICHARD A WILLIAMS  
Signature and Printed Name

01/06/2022  
Date

OSCEOLA COUNTY S.O.  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-003937  
01/06/2022  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:29
Control Test	0.049	13:29
Air Blank	0.000	13:30
Control Test	0.048	13:31
Air Blank	0.000	13:31
Control Test	0.049	13:32
Air Blank	0.000	13:32
Control Test Stats		
Average	0.0487	
Std Dev	0.0006	
Rel Std Dev(%)	1.1863	

.05

PAW

Operator's Signature

OSCEOLA COUNTY S.O.  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-003937  
01/06/2022  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:41
Control Test	0.078	13:41
Air Blank	0.000	13:42
Control Test	0.078	13:43
Air Blank	0.000	13:43
Control Test	0.078	13:44
Air Blank	0.000	13:44
Control Test Stats		
Average	0.0780	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

wet .08

PAW

Operator's Signature

OSCEOLA COUNTY S.O.  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-003937  
01/06/2022  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:53
Control Test	0.197	13:54
Air Blank	0.000	13:54
Control Test	0.197	13:55
Air Blank	0.000	13:56
Control Test	0.197	13:56
Air Blank	0.000	13:57
Control Test Stats		
Average	0.1970	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

.20

PAW

Operator's Signature

OSCEOLA COUNTY S.O.  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-003937  
01/06/2022  
Software: 8100.27

ITP CONTROL TEST

Test	g/210L	Time
Air Blank	0.000	13:39
ITP Check	0.079	13:40
Air Blank	0.000	13:40

PAW

Operator's Signature

OSCEOLA COUNTY S.O.  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-003937  
01/06/2022  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:07
Control Test	0.078	14:08
Air Blank	0.000	14:08
Control Test	0.078	14:09
Air Blank	0.000	14:09
Control Test	0.078	14:09
Air Blank	0.000	14:10
Control Test Stats		
Average	0.0780	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

.08

DGS

PAW

Operator's Signature

Stability checks

80-003937

1/6/22

PAW

# Return Material Authorization

Ship to: ☒ CMI, Inc.

☐ Enforcement Electronics

Shipment to repair facility authorized by: Owen Gayle on 03-31-2021

Items Returned: Instrument ☒ Supplies ☐ Other ☐ Describe: \_\_\_\_\_

Instrument Model: Intoxilyzer 8000 Serial Number: 80-003937

Bill To Address:

Osceola County Sheriff's Office

Ship to Address:

Alcohol Testing Program

FDLE Tallahassee, Florida

Reason for Return:

Modem port on back of instrument is loose and pushed in, phone line cannot be connected for uploads.

Records have already been uploaded.

**Please choose one of the following options:**

☐ 1. I \_\_\_\_\_, authorize all repairs.

☐ 2. I \_\_\_\_\_, authorize repairs up to \$ \_\_\_\_\_.

☒ 3. I require an estimate **BEFORE** any repairs will be authorized and/ or conducted.

Please contact: Name: Owen Gayle

Phone #: 407-690-9811

Email: ogay@osceola.org

ATP Contact Name: Israel Soto

ATP Email: israelsoto@fdle.state.fl.us







# Calibration Certificate

Florida Department of Law Enforcement  
Alcohol Testing Program  
2729 Fort Knox Blvd.  
Bldg. 2, Suite 1300  
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-003937, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-003937</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>OSCEOLA COUNTY S.O.</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>01/06/2022</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>16:10</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within  $\pm 0.005$  or 5%, whichever is greater, of the target alcohol concentration.

\*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ( $k=3$ ).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

## TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Richard A  
Williams  
Digitally signed by Richard  
A Williams  
Date: 2022.01.06 17:10:38  
-05'00'

01/06/2022

Date

RICHARD A WILLIAMS,  
Department Inspector

FDLE/ATP Form 69 January 2022

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality