



Agency Crestview Police Department

S/N 80-005058

Date In 7/5/2022

DI Completion Date 07-06-2022

☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake		By <u>PN</u>		Quality Checks		By <u>PN</u>		Date <u>7/6/2022</u>		Flow Calibration		By <u> </u>		Date <u> </u>																																									
☒ Annual ☐ Registration ☒ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: _____ _____ _____ _____ _____ _____ _____ _____ _____ _____				☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>240</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP 103</u> 32 mm <u>.156</u> (.139 - .169) 36 mm <u>.164</u> (.156 - .190) 53 mm <u>.230</u> (.228 - .278) 103 mm <u>.492</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28662</u> ☒ Stability Checks				Flow Column # _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)				<table border="1"> <thead> <tr> <th colspan="2">Maintenance</th> <th colspan="2">By <u> </u></th> </tr> </thead> <tbody> <tr> <td colspan="4">☐ Battery Replacement</td> </tr> <tr> <td colspan="4">☐ Dry Gas Regulator Replacement</td> </tr> <tr> <td colspan="4">☐ Breath Tube Replacement</td> </tr> <tr> <td colspan="4">☐ Other _____</td> </tr> <tr><td colspan="4">_____</td></tr> <tr><td colspan="4">_____</td></tr> <tr><td colspan="4">_____</td></tr> <tr><td colspan="4">_____</td></tr> <tr><td colspan="4">_____</td></tr> </tbody> </table>				Maintenance		By <u> </u>		☐ Battery Replacement				☐ Dry Gas Regulator Replacement				☐ Breath Tube Replacement				☐ Other _____				_____				_____				_____				_____				_____			
Maintenance		By <u> </u>																																																					
☐ Battery Replacement																																																							
☐ Dry Gas Regulator Replacement																																																							
☐ Breath Tube Replacement																																																							
☐ Other _____																																																							

				<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td rowspan="2">0.050</td> <td rowspan="2">MP6291</td> <td>202201C</td> </tr> <tr> <td>01/11/2024</td> </tr> <tr> <td rowspan="2">0.080</td> <td rowspan="2">MP6292</td> <td>202201D</td> </tr> <tr> <td>01/18/2024</td> </tr> <tr> <td rowspan="2">0.200</td> <td rowspan="2">MP6293</td> <td>202201E</td> </tr> <tr> <td>01/18/2024</td> </tr> <tr> <td rowspan="2">0.080 DGS</td> <td rowspan="2">N/A</td> <td>AG121002</td> </tr> <tr> <td>07/29/2023</td> </tr> </tbody> </table>		Simulator	Serial #	Lot #/Exp	0.050	MP6291	202201C	01/11/2024	0.080	MP6292	202201D	01/18/2024	0.200	MP6293	202201E	01/18/2024	0.080 DGS	N/A	AG121002	07/29/2023																															
Simulator	Serial #	Lot #/Exp																																																					
0.050	MP6291	202201C																																																					
		01/11/2024																																																					
0.080	MP6292	202201D																																																					
		01/18/2024																																																					
0.200	MP6293	202201E																																																					
		01/18/2024																																																					
0.080 DGS	N/A	AG121002																																																					
		07/29/2023																																																					

Calibration Adjustment				By _____	
Barometric Pressure Gauge _____				ID # _____	
Simulator	Serial #	Lot #	Expiration		
0.000		N/A	N/A		
0.040					
0.100					
0.200					
0.300					
0.080 DGS	N/A				
☐ Post Calibration Adjustment Stability Checks					
Simulator	Serial #	Lot #	Expiration		
0.050					
0.080					
0.200					
0.080 DGS	N/A				
Notes/Suggested Service: _____ _____ _____ _____ _____ _____ _____					

Department Inspection		By IS _____	
Barometric Pressure ID# <u>30793</u>			
Gauge <u>1015</u>	Instrument <u>1015</u>		
Mouth Alcohol Solution Lot # <u>2022-A</u>			
Acetone Stock Solution Lot # <u>2021-C</u>			
Simulator	Serial Number		
0.000	MP6289		
Interferent	MP6290		
0.050	MP6291		
0.080	MP6292		
0.200	MP6293		
Attachments			
☒ Form 41		☐ Post-Stability Checks	
☒ Stability Checks		☐ Flow Calibration	
☒ Calibration Certificate		☐ Form 40	
☐ Calibration Adjustment		☐ Other _____	
☒ Instrument Complies with Chapter 11D-8, FAC			
☐ Instrument Does Not Comply with Chapter 11D-8, FAC			
☒ Return to/Place into Evidentiary Use			
☐ Remain Out of Evidentiary Use			
☒ Conduct an Agency Inspection Before Evidentiary Use			
Taylor Gutschow		2022.07.06	
Digitally signed by Taylor Gutschow Date: 2022.07.06 11:36:11 -0400		12:35:24	
Tech Review / Date		Admin Review / Date	

Florida Department of Law Enforcement

Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: CRESTVIEW PD
Time of Inspection: 09:48

Date of Inspection: 07/06/2022

Serial Number: 80-005058
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted	Yes	
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG113403 Exp: 05/14/2023
0.000	0.050	0.079	0.201	0.081
0.000	0.050	0.080	0.201	0.080
0.000	0.050	0.080	0.202	0.080
0.000	0.050	0.080	0.202	0.080
0.000	0.050	0.080	0.201	0.080
0.000	0.050	0.080	0.201	0.080
0.000	0.050	0.079	0.201	0.081
0.000	0.050	0.079	0.201	0.080
0.000	0.050	0.080	0.201	0.080
0.000	0.050	0.080	0.201	0.080

Standard Deviations	0.0000	0.0004	0.0004	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

Time-Date changed. CENTRAL TIME ZONE

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Israel Soto ISRAEL SOTO
Signature and Printed Name

07/06/2022
Date

Type of Test	Serial Number	Agency	Date	Performed By
Stability Checks	80-005058	Crestview Police Department	7/6/2022	PN

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083
CRESTVIEW PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-005058 07/06/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 08:22 Control Test 0.050 08:23 Air Blank 0.000 08:23 Control Test 0.050 08:24 Air Blank 0.000 08:25 Control Test 0.050 08:25 Air Blank 0.000 08:26 Control Test Stats Average 0.0500 Std Dev 0.0000 Rel Std Dev(%) 0.0000 Operator's Signature 	CRESTVIEW PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-005058 07/06/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 08:27 Control Test 0.080 08:28 Air Blank 0.000 08:28 Control Test 0.080 08:29 Air Blank 0.000 08:30 Control Test 0.080 08:30 Air Blank 0.000 08:31 Control Test Stats Average 0.0800 Std Dev 0.0000 Rel Std Dev(%) 0.0000 Operator's Signature 	CRESTVIEW PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-005058 07/06/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 08:33 Control Test 0.200 08:33 Air Blank 0.000 08:34 Control Test 0.201 08:34 Air Blank 0.000 08:35 Control Test 0.201 08:36 Air Blank 0.000 08:36 Control Test Stats Average 0.2007 Std Dev 0.0006 Rel Std Dev(%) 0.2877 Operator's Signature 	CRESTVIEW PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-005058 07/06/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 08:37 Control Test 0.080 08:38 Air Blank 0.000 08:38 Control Test 0.080 08:39 Air Blank 0.000 08:39 Control Test 0.081 08:39 Air Blank 0.000 08:40 Control Test Stats Average 0.0803 Std Dev 0.0006 Rel Std Dev(%) 0.7187 Operator's Signature 

Comments: N/A



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road.
Suite B1032
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-005058, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-005058</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>CRESTVIEW PD</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>07/06/2022</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>09:48</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Israel Soto
Soto
Digitally signed by Israel Soto
Date: 2022.07.06 11:19:34 -04'00'

07/06/2022

Date

ISRAEL SOTO,

Department Inspector

FDLE/ATP Form 69 March 2022

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality