



INSTRUMENT PROCESSING SHEET

Agency Florida Fish and Wildlife Conservation Commission


S/N 80-005248

Florida Department of
Law Enforcement

Date In 7/26/2022 DI Completion Date 7/27/2022

☐ Ship ☒ P/U ☐ H/D ☐ CMI ☐ EE

Intake By <u>DERR</u> ☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☒ Power cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes: <u>Pelican Size 50 hard case, DGS. DERR</u>	Quality Checks By <u>DERR</u> Date <u>7/27/2022</u> ☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>106</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP101</u> 32 mm <u>0.125</u> (.139 - .169) 36 mm <u>0.132</u> (.156 - .190) 53 mm <u>0.210</u> (.228 - .278) 103 mm <u>0.507</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28199</u> ☒ Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> <tr> <td>0.050</td> <td>MP5092</td> <td>202201C 01/11/2024</td> </tr> <tr> <td>0.080</td> <td>MP5093</td> <td>202201D 01/18/2024</td> </tr> <tr> <td>0.200</td> <td>MP5094</td> <td>202201E 01/18/2024</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG115904 06/08/2023</td> </tr> </table>	Simulator	Serial #	Lot #/Exp	0.050	MP5092	202201C 01/11/2024	0.080	MP5093	202201D 01/18/2024	0.200	MP5094	202201E 01/18/2024	0.080 DGS	N/A	AG115904 06/08/2023	Flow Calibration By <u>DERR</u> Date <u>7/27/2022</u> Flow Column # <u>ATP104</u> ☒ 5L/min - 17mm ☒ 15L/min - 53mm ☒ 30L/min - 103mm ☒ R-Value <u>106/106</u> ☒ Post Calibration Verification (L/s) Flow Column # <u>ATP106</u> 32 mm <u>0.140/140</u> (.139 - .169) 36 mm <u>0.160/160</u> (.156 - .190) 53 mm <u>0.222/234</u> (.228 - .278) 103 mm <u>0.500/503</u> (.447 - .547)																																	
Simulator	Serial #	Lot #/Exp																																																
0.050	MP5092	202201C 01/11/2024																																																
0.080	MP5093	202201D 01/18/2024																																																
0.200	MP5094	202201E 01/18/2024																																																
0.080 DGS	N/A	AG115904 06/08/2023																																																
Maintenance By _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____																																																		
Calibration Adjustment By _____ Barometric Pressure Gauge _____ ID # _____ <table border="1" style="width:100%; border-collapse: collapse;"> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </table> ☐ Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </table>			Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A		
Simulator	Serial #	Lot #	Expiration																																															
0.000		N/A	N/A																																															
0.040																																																		
0.100																																																		
0.200																																																		
0.300																																																		
0.080 DGS	N/A																																																	
Simulator	Serial #	Lot #	Expiration																																															
0.050																																																		
0.080																																																		
0.200																																																		
0.080 DGS	N/A																																																	
Department Inspection By <u>DERR</u> Barometric Pressure ID# <u>26932</u> Gauge <u>1017</u> Instrument <u>1021</u> Mouth Alcohol Solution Lot # <u>2021-D</u> Acetone Stock Solution Lot # <u>2021-C</u> <table border="1" style="width:100%; border-collapse: collapse;"> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> <tr> <td>0.000</td> <td>MP5095</td> </tr> <tr> <td>Interferent</td> <td>MP5097</td> </tr> <tr> <td>0.050</td> <td>MP5092</td> </tr> <tr> <td>0.080</td> <td>MP5093</td> </tr> <tr> <td>0.200</td> <td>MP5094</td> </tr> </table>			Simulator	Serial Number	0.000	MP5095	Interferent	MP5097	0.050	MP5092	0.080	MP5093	0.200	MP5094																																				
Simulator	Serial Number																																																	
0.000	MP5095																																																	
Interferent	MP5097																																																	
0.050	MP5092																																																	
0.080	MP5093																																																	
0.200	MP5094																																																	
Attachments ☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Calibration Adjustment ☐ Post-Stability Checks ☒ Flow Calibration ☐ Form 40 ☐ Other _____																																																		
Notes/Suggested Service: <u>Performed flow calibration twice using the same flow meters as indicated to bring values closer to nominal. DERR</u>																																																		
☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																																																		
Israel Soto Digitally signed by Israel Soto Date: 2022.07.27 10:14:12 -0400		Date: <u>2022.07.27</u> <u>11:11:24 -04'00'</u>																																																
Tech Review / Date		Admin Review / Date																																																



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-005248, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-005248</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>FISH & WILDLIFE CC</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>07/27/2022</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>09:48</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.
This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

FDLE/ATP Form 69 December 2021
Issuing Authority: Alcohol Testing Program

07/27/2022

Date

DAVID E REYES-RIVERA,

Department Inspector

Service • Integrity • Respect • Quality

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FISH & WILDLIFE CC
Time of Inspection: 09:48

Date of Inspection: 07/27/2022

Serial Number: 80-005248
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG115904 Exp: 06/08/2023
0.000	0.048	0.078	0.203	0.080
0.000	0.048	0.078	0.203	0.080
0.000	0.048	0.077	0.203	0.080
0.000	0.049	0.078	0.202	0.081
0.000	0.049	0.078	0.203	0.080
0.000	0.049	0.078	0.202	0.080
0.000	0.049	0.078	0.203	0.080
0.000	0.048	0.078	0.203	0.081
0.000	0.048	0.078	0.203	0.080
0.000	0.049	0.078	0.202	0.080


Standard Deviations	0.0005	0.0003	0.0004	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 DAVID E REYES-RIVERA
Signature and Printed Name

07/27/2022
Date

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-005248	Florida Fish and Wildlife Conservation Commission	7/27/2022	DERR <i>full</i>

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																																
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083																																																																																																																																																
FISH & WILDLIFE CC Intoxilyzer - Alcohol Analyzer Model 8000 07/27/2022 Software: 8100.27 SN 80-005248 <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>07:16</td></tr><tr><td>Control Test</td><td>0.049</td><td>07:17</td></tr><tr><td>Air Blank</td><td>0.000</td><td>07:17</td></tr><tr><td>Control Test</td><td>0.049</td><td>07:18</td></tr><tr><td>Air Blank</td><td>0.000</td><td>07:18</td></tr><tr><td>Control Test</td><td>0.048</td><td>07:19</td></tr><tr><td>Air Blank</td><td>0.000</td><td>07:20</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0487</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>1.1863</td><td></td></tr></table> Operator's Signature <i>full</i>	Test	g/210L	Time	Air Blank	0.000	07:16	Control Test	0.049	07:17	Air Blank	0.000	07:17	Control Test	0.049	07:18	Air Blank	0.000	07:18	Control Test	0.048	07:19	Air Blank	0.000	07:20	Control Test Stats			Average	0.0487		Std Dev	0.0006		Rel Std Dev(%)	1.1863		FISH & WILDLIFE CC Intoxilyzer - Alcohol Analyzer Model 8000 07/27/2022 Software: 8100.27 SN 80-005248 <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>07:21</td></tr><tr><td>Control Test</td><td>0.079</td><td>07:22</td></tr><tr><td>Air Blank</td><td>0.000</td><td>07:22</td></tr><tr><td>Control Test</td><td>0.078</td><td>07:23</td></tr><tr><td>Air Blank</td><td>0.000</td><td>07:24</td></tr><tr><td>Control Test</td><td>0.078</td><td>07:24</td></tr><tr><td>Air Blank</td><td>0.000</td><td>07:25</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0783</td><td></td></tr><tr><td>Std Dev</td><td>0.0005</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.7379</td><td></td></tr></table> Operator's Signature <i>full</i>	Test	g/210L	Time	Air Blank	0.000	07:21	Control Test	0.079	07:22	Air Blank	0.000	07:22	Control Test	0.078	07:23	Air Blank	0.000	07:24	Control Test	0.078	07:24	Air Blank	0.000	07:25	Control Test Stats			Average	0.0783		Std Dev	0.0005		Rel Std Dev(%)	0.7379		FISH & WILDLIFE CC Intoxilyzer - Alcohol Analyzer Model 8000 07/27/2022 Software: 8100.27 SN 80-005248 <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>07:26</td></tr><tr><td>Control Test</td><td>0.203</td><td>07:27</td></tr><tr><td>Air Blank</td><td>0.000</td><td>07:27</td></tr><tr><td>Control Test</td><td>0.201</td><td>07:28</td></tr><tr><td>Air Blank</td><td>0.000</td><td>07:29</td></tr><tr><td>Control Test</td><td>0.201</td><td>07:29</td></tr><tr><td>Air Blank</td><td>0.000</td><td>07:30</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.2017</td><td></td></tr><tr><td>Std Dev</td><td>0.0012</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.5726</td><td></td></tr></table> Operator's Signature <i>full</i>	Test	g/210L	Time	Air Blank	0.000	07:26	Control Test	0.203	07:27	Air Blank	0.000	07:27	Control Test	0.201	07:28	Air Blank	0.000	07:29	Control Test	0.201	07:29	Air Blank	0.000	07:30	Control Test Stats			Average	0.2017		Std Dev	0.0012		Rel Std Dev(%)	0.5726		FISH & WILDLIFE CC Intoxilyzer - Alcohol Analyzer Model 8000 07/27/2022 Software: 8100.27 SN 80-005248 <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>07:31</td></tr><tr><td>Control Test</td><td>0.080</td><td>07:31</td></tr><tr><td>Air Blank</td><td>0.000</td><td>07:32</td></tr><tr><td>Control Test</td><td>0.081</td><td>07:32</td></tr><tr><td>Air Blank</td><td>0.000</td><td>07:33</td></tr><tr><td>Control Test</td><td>0.081</td><td>07:33</td></tr><tr><td>Air Blank</td><td>0.000</td><td>07:33</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0807</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.7157</td><td></td></tr></table> Operator's Signature <i>full</i>	Test	g/210L	Time	Air Blank	0.000	07:31	Control Test	0.080	07:31	Air Blank	0.000	07:32	Control Test	0.081	07:32	Air Blank	0.000	07:33	Control Test	0.081	07:33	Air Blank	0.000	07:33	Control Test Stats			Average	0.0807		Std Dev	0.0006		Rel Std Dev(%)	0.7157	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	07:16																																																																																																																																																	
Control Test	0.049	07:17																																																																																																																																																	
Air Blank	0.000	07:17																																																																																																																																																	
Control Test	0.049	07:18																																																																																																																																																	
Air Blank	0.000	07:18																																																																																																																																																	
Control Test	0.048	07:19																																																																																																																																																	
Air Blank	0.000	07:20																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0487																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	1.1863																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	07:21																																																																																																																																																	
Control Test	0.079	07:22																																																																																																																																																	
Air Blank	0.000	07:22																																																																																																																																																	
Control Test	0.078	07:23																																																																																																																																																	
Air Blank	0.000	07:24																																																																																																																																																	
Control Test	0.078	07:24																																																																																																																																																	
Air Blank	0.000	07:25																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0783																																																																																																																																																		
Std Dev	0.0005																																																																																																																																																		
Rel Std Dev(%)	0.7379																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	07:26																																																																																																																																																	
Control Test	0.203	07:27																																																																																																																																																	
Air Blank	0.000	07:27																																																																																																																																																	
Control Test	0.201	07:28																																																																																																																																																	
Air Blank	0.000	07:29																																																																																																																																																	
Control Test	0.201	07:29																																																																																																																																																	
Air Blank	0.000	07:30																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.2017																																																																																																																																																		
Std Dev	0.0012																																																																																																																																																		
Rel Std Dev(%)	0.5726																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	07:31																																																																																																																																																	
Control Test	0.080	07:31																																																																																																																																																	
Air Blank	0.000	07:32																																																																																																																																																	
Control Test	0.081	07:32																																																																																																																																																	
Air Blank	0.000	07:33																																																																																																																																																	
Control Test	0.081	07:33																																																																																																																																																	
Air Blank	0.000	07:33																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0807																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7157																																																																																																																																																		

#1

FISH & WILDLIFE CC
Intoxilizer - Alconol Analyzer
Model 6000 SN 80-005248
07/27/2022
Software: 8100.27

Flow Rate Calibration*****

1: Rate (Liters/min) = 5
SORT(DIFF) = 5.914
2: Rate (Liters/min) = 15
SORT(DIFF) = 10.816
3: Rate (Liters/min) = 30
SORT(DIFF) = 20.516
Dependent Data Scale Factor = 10000 L/min
Independent Data Scale Factor = 256
Rounded Slope = 660
Rounded Intercept = -430146
Correlation = 0.99737



#2

FISH & WILDLIFE CC
Intoxilizer - Alconol Analyzer
Model 6000 SN 80-005248
07/27/2022
Software: 8100.27

Flow Rate Calibration*****

1: Rate (Liters/min) = 5
SORT(DIFF) = 5.566
2: Rate (Liters/min) = 15
SORT(DIFF) = 10.556
3: Rate (Liters/min) = 30
SORT(DIFF) = 20.637
Dependent Data Scale Factor = 10000 L/min
Independent Data Scale Factor = 256
Rounded Slope = 638
Rounded Intercept = -333607
Correlation = 0.99587

Flow Calibration	
SN:	80-005248
Agency:	Florida Fish and Wildlife CC
Date:	7/27/2022
By:	DERR <i>[Signature]</i>