

**FLORIDA DEPARTMENT OF LAW ENFORCEMENT
ALCOHOL TESTING PROGRAM
BREATH ALCOHOL TEST AFFIDAVIT**

Instrument Type: Intoxilyzer 8000
Instrument Registered To: MIAMI DADE PD
Instrument Serial Number: 80-005550 Software: 8100.27
Date of Test: 05/17/2022

Date of Last Agency Inspection: 05/17/2022
Observation Period Began: 14:00
Subject's Name: SAMPLE D LICENSE

DOB: 02/14/1965 Sex: F

The subject was observed for at least twenty-minutes prior to the administration of the breath test to ensure that the subject did not take anything orally and did not regurgitate.

Results:	Test	g/210L	Time
	Diagnostics Check	OK	14:36
	Air Blank	0.000	14:36
	Control Test	0.082	14:36
	Air Blank	0.000	14:37
	Subject Sample #1	REF*	14:37
	Air Blank	0.000	14:38
	Control Test	0.082	14:38
	Air Blank	0.000	14:38
	Diagnostics Check	OK	14:39

*Subject Test Refused

Testing form 38, used a training license for data entry.
DEER

Cylinder Lot: AG115904
Exp: 06/08/2023

State of Florida, County of _____,

Personally appeared before me the undersigned authority, who () is personally known to me or () produced _____ as identification, and who after being placed under oath, states:

I DAVID E REYES-RIVERA, hold a valid Breath Test Operator permit issued by the Florida Department of Law Enforcement, I administered the above breath test to the subject named above in accordance with Chapter 11D-8, Florida Administrative Code, and this form is a true and accurate report of that breath test.

Breath Test Operator: _____ Date: _____
Signature

Sworn to (or affirmed) before me this _____ day of _____, _____

Signature of Notary Public-State of Florida Printed Name of Notary Public-State of Florida

Note: Pursuant to section 117.10, Florida Statutes, law enforcement officers, correctional officers, traffic accident investigation officers and traffic infraction enforcement officers are notaries public when engaged in the performance of official duties. In accordance with section 316.1934(5), F.S., this completed form is admissible without further authentication and is presumptive proof of the results herein. To be used in accordance with Section 316.1934(5), F.S., and in administrative proceedings pursuant to 322.2615, F.S.

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: MIAMI DADE PD

Time of Inspection: 14:31

Date of Inspection: 05/17/2022

Serial Number: 80-005550

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK		No	Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK		No	Barometric Pressure Sensor Check: OK		No
Alcohol Free Subject Test: 0.000		No	Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No	Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Standard Deviations				
---------------------	--	--	--	--


Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: _____ Number of Simulators Used: _____

Remarks:

TESTING FORM 41

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


DAVID E REYES-RIVERA

 Signature and Printed Name

05/17/2022
 Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: MIAMI DADE PD

Time of Inspection: 14:29

Date of Inspection: 05/17/2022

Serial Number: 80-005550

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____

Number of Simulators Used: _____

Remarks:

TESTING THE FORM 40

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



DAVID E REYES-RIVERA

Signature and Printed Name

05/17/2022
Date

**FLORIDA DEPARTMENT OF LAW ENFORCEMENT
ALCOHOL TESTING PROGRAM
BREATH ALCOHOL TEST AFFIDAVIT**

Instrument Type: Intoxilyzer 8000
Instrument Registered To: MIAMI DADE PD
Instrument Serial Number: 80-005550 Software: 8100.27
Date of Test: 05/17/2022

Date of Last Agency Inspection: 05/17/2022

Observation Period Began: 11:34

Subject's Name: DRIVER L SAMPLE

DOB: 02/14/1965 Sex: F

The subject was observed for at least twenty-minutes prior to the administration of the breath test to ensure that the subject did not take anything orally and did not regurgitate.

Results:	Test	g/210L	Time
	Diagnostics Check	OK	11:54
	Air Blank	0.000	11:55
	Control Test	0.082	11:55
	Air Blank	0.000	11:55
	Subject Sample #1	0.000	11:56
	Air Blank	0.000	11:56
	Air Blank	0.000	11:58
	Subject Sample #2	REF*	11:59
	Air Blank	0.000	11:59
	Control Test	0.081	11:59
	Air Blank	0.000	12:00
	Diagnostics Check	OK	12:00

*Subject Test Refused

*Conducted test to insure that Form 38 was printing.
DERR*

Cylinder Lot: AG115904
Exp: 06/08/2023

State of Florida, County of _____,

Personally appeared before me the undersigned authority, who () is personally known to me or () produced _____ as identification, and who after being placed under oath, states:

I DAVID E REYES-RIVERA, hold a valid Breath Test Operator permit issued by the Florida Department of Law Enforcement, I administered the above breath test to the subject named above in accordance with Chapter 11D-8, Florida Administrative Code, and this form is a true and accurate report of that breath test.

Breath Test Operator: _____ Date: _____
Signature

Sworn to (or affirmed) before me this _____ day of _____, _____

Signature of Notary Public-State of Florida _____ Printed Name of Notary Public-State of Florida _____

Note: Pursuant to section 117.10, Florida Statutes, law enforcement officers, correctional officers, traffic accident investigation officers and traffic infraction enforcement officers are notaries public when engaged in the performance of official duties. In accordance with section 316.1934(5), F.S., this completed form is admissible without further authentication and is presumptive proof of the results herein. To be used in accordance with Section 316.1934(5), F.S., and in administrative proceedings pursuant to 322.2615, F.S.

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: MIAMI DADE PD

Time of Inspection: 10:31

Date of Inspection: 05/17/2022

Serial Number: 80-005550

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:_____ Exp:_____	0.08g/210L Test (g/210L) Lot#:_____ Exp:_____	0.20g/210L Test (g/210L) Lot#:_____ Exp:_____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:_____ Exp:_____

Number of Simulators Used: _____

Remarks:

TEST TO MAKE SURE IT RETAINED FORM

Compliance not determined, AI Not conducted DERR

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

David E Reyes-Rivera DAVID E REYES-RIVERA
Signature and Printed Name

05/17/2022
Date

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: MIAMI DADE PD

Time of Inspection: 10:33

Date of Inspection: 05/17/2022

Serial Number: 80-005550

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK		No	Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK		No	Barometric Pressure Sensor Check: OK		No
Alcohol Free Subject Test: 0.000		No	Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No	Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Standard Deviations				
---------------------	--	--	--	--

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: _____ Number of Simulators Used: _____

Remarks:

TEST TO MAKE SURE IT RETAINED FORM

Compliance not determined AI not conducted DERR

The above instrument complies (☒) does not comply () with Chapter 11D-8, FAC.


I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

David Reyes Rivera

DAVID E REYES-RIVERA

Signature and Printed Name

05/17/2022
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-005550, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-005550</u>	UNCERTAINTY* \pm	
Owning Agency:	<u>MIAMI DADE PD</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>05/17/2022</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>09:12</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.
This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

FDLE/ATP Form 69 December 2021
Issuing Authority: Alcohol Testing Program

05/17/2022

Date


DAVID E REYES-RIVERA,
Department Inspector

Service • Integrity • Respect • Quality

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: MIAMI DADE PD

Serial Number: 80-005550

Time of Inspection: 09:12

Date of Inspection: 05/17/2022

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#: 202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#: 202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: AG115904 Exp: 06/08/2023
0.000	0.049	0.080	0.203	0.079
0.000	0.050	0.080	0.204	0.079
0.000	0.050	0.080	0.203	0.080
0.000	0.050	0.080	0.203	0.080
0.000	0.051	0.080	0.203	0.079
0.000	0.050	0.081	0.203	0.079
0.000	0.051	0.080	0.203	0.080
0.000	0.051	0.080	0.202	0.080
0.000	0.051	0.081	0.203	0.079
0.000	0.051	0.080	0.203	0.080

Standard Deviations	0.0006	0.0004	0.0004	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004

Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

David E Reyes-Rivera

DAVID E REYES-RIVERA

Signature and Printed Name

05/17/2022

Date

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-005550	Miami-Dade Police Department	5/17/2022	DERR <i>[Signature]</i>

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053 <input checked="" type="checkbox"/>	0.077 to 0.083 <input checked="" type="checkbox"/>	0.194 to 0.206 <input checked="" type="checkbox"/>	0.077 to 0.083 <input checked="" type="checkbox"/>
<p>MIAMI DADE PD Intoxilyzer - Alcohol Analyzer Model 8000 05/17/2022 Software: 8100.27</p> <p>SN 80-005550</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 07:04</p> <p>Control Test 0.050 07:05</p> <p>Air Blank 0.000 07:05</p> <p>Control Test 0.049 07:06</p> <p>Air Blank 0.000 07:06</p> <p>Control Test 0.049 07:07</p> <p>Air Blank 0.000 07:08</p> <p>Control Test Stats</p> <p>Average 0.0433</p> <p>Std Dev 0.0006</p> <p>Rel Std Dev(%) 1.1703</p>	<p>MIAMI DADE PD Intoxilyzer - Alcohol Analyzer Model 8000 05/17/2022 Software: 8100.27</p> <p>SN 80-005550</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 07:09</p> <p>Control Test 0.080 07:10</p> <p>Air Blank 0.000 07:10</p> <p>Control Test 0.080 07:11</p> <p>Air Blank 0.000 07:11</p> <p>Control Test 0.079 07:12</p> <p>Air Blank 0.000 07:13</p> <p>Control Test Stats</p> <p>Average 0.0797</p> <p>Std Dev 0.0006</p> <p>Rel Std Dev(%) 0.7247</p>	<p>MIAMI DADE PD Intoxilyzer - Alcohol Analyzer Model 8000 05/17/2022 Software: 8100.27</p> <p>SN 80-005550</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 07:14</p> <p>Control Test 0.200 07:15</p> <p>Air Blank 0.000 07:15</p> <p>Control Test 0.200 07:16</p> <p>Air Blank 0.000 07:16</p> <p>Control Test 0.200 07:17</p> <p>Air Blank 0.000 07:18</p> <p>Control Test Stats</p> <p>Average 0.2019</p> <p>Std Dev 0.0000</p> <p>Rel Std Dev(%) 0.0000</p>	<p>MIAMI DADE PD Intoxilyzer - Alcohol Analyzer Model 8000 05/17/2022 Software: 8100.27</p> <p>SN 80-005550</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 07:19</p> <p>Control Test 0.081 07:19</p> <p>Air Blank 0.000 07:20</p> <p>Control Test 0.080 07:20</p> <p>Air Blank 0.000 07:21</p> <p>Control Test 0.081 07:21</p> <p>Air Blank 0.000 07:21</p> <p>Control Test Stats</p> <p>Average 0.0807</p> <p>Std Dev 0.0006</p> <p>Rel Std Dev(%) 0.7157</p>
<i>[Signature]</i> Operator's Signature	<i>[Signature]</i> Operator's Signature	<i>[Signature]</i> Operator's Signature	<i>[Signature]</i> Operator's Signature