



INSTRUMENT PROCESSING SHEET

Agency Miami-Dade Police Department

S/N 80-005551

Florida Department of
Law Enforcement

Date In 7/24/2022


DI Completion Date 7/22/2022

☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

DERR 7/21/22

DERR 7/21/22

Intake By <u>DERR</u> ☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: 	Quality Checks By <u>DERR</u> Date <u>7/22/2022</u> ☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>125</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP101</u> 32 mm <u>0.140</u> (.139 - .169) 36 mm <u>0.164</u> (.156 - .190) 53 mm <u>0.242</u> (.228 - .278) 103 mm <u>0.515</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28199</u> ☒ Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>SD3963</td> <td>202201C 01/11/2024</td> </tr> <tr> <td>0.080</td> <td>SD3967</td> <td>202201D 01/18/2024</td> </tr> <tr> <td>0.200</td> <td>SD3968</td> <td>202201E 01/18/2024</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG115904 06/08/2023</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	SD3963	202201C 01/11/2024	0.080	SD3967	202201D 01/18/2024	0.200	SD3968	202201E 01/18/2024	0.080 DGS	N/A	AG115904 06/08/2023	Flow Calibration By <u>DERR</u> Date <u>7/22/2022</u> Flow Column # <u>ATP101</u> ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value <u>125</u> ☐ Post Calibration Verification (L/s) Flow Column # <u>ATP101</u> 32 mm <u>0.140</u> (.139 - .169) 36 mm <u>0.164</u> (.156 - .190) 53 mm <u>0.242</u> (.228 - .278) 103 mm <u>0.515</u> (.447 - .547) Maintenance By <u>DERR</u> ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☒ Other <u>Replaced printer paper</u> 																																														
Simulator	Serial #	Lot #/Exp																																																													
0.050	SD3963	202201C 01/11/2024																																																													
0.080	SD3967	202201D 01/18/2024																																																													
0.200	SD3968	202201E 01/18/2024																																																													
0.080 DGS	N/A	AG115904 06/08/2023																																																													
Calibration Adjustment By <u>DERR</u> Barometric Pressure Gauge <u>1016</u> ID # <u>68639</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP5099</td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td>MP5096</td> <td>21070</td> <td>03/01/2023</td> </tr> <tr> <td>0.100</td> <td>MP5098</td> <td>21080</td> <td>03/08/2023</td> </tr> <tr> <td>0.200</td> <td>MP5100</td> <td>20510</td> <td>12/03/2022</td> </tr> <tr> <td>0.300</td> <td>MP5101</td> <td>21030</td> <td>02/02/2023</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>00521080A2</td> <td>02/05/2023</td> </tr> </tbody> </table> ☒ Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>SD3963</td> <td>202201C</td> <td>01/11/2024</td> </tr> <tr> <td>0.080</td> <td>SD3967</td> <td>202201D</td> <td>01/18/2024</td> </tr> <tr> <td>0.200</td> <td>SD3968</td> <td>202201E</td> <td>01/18/2024</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG115904</td> <td>06/08/2023</td> </tr> </tbody> </table>		Simulator	Serial #	Lot #	Expiration	0.000	MP5099	N/A	N/A	0.040	MP5096	21070	03/01/2023	0.100	MP5098	21080	03/08/2023	0.200	MP5100	20510	12/03/2022	0.300	MP5101	21030	02/02/2023	0.080 DGS	N/A	00521080A2	02/05/2023	Simulator	Serial #	Lot #	Expiration	0.050	SD3963	202201C	01/11/2024	0.080	SD3967	202201D	01/18/2024	0.200	SD3968	202201E	01/18/2024	0.080 DGS	N/A	AG115904	06/08/2023	Department Inspection By <u>DERR</u> Barometric Pressure ID# <u>26932</u> Gauge <u>1017</u> Instrument <u>1017</u> Mouth Alcohol Solution Lot # <u>2021-D</u> Acetone Stock Solution Lot # <u>2021-C</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>SD3933</td> </tr> <tr> <td>Interferent</td> <td>SD3962</td> </tr> <tr> <td>0.050</td> <td>SD3963</td> </tr> <tr> <td>0.080</td> <td>SD3967</td> </tr> <tr> <td>0.200</td> <td>SD3968</td> </tr> </tbody> </table> Attachments ☒ Form 41 ☒ Post-Stability Checks ☒ Stability Checks ☐ Flow Calibration ☒ Calibration Certificate ☐ Form 40 ☒ Calibration Adjustment ☐ Other <u> </u> ☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use Israel Soto Digitally signed by Israel Soto Date: 2022.07.21 13:04:06 +04'00' Tech Review / Date <u>7/22/2022</u> Admin. Review / Date <u>7/22/2022</u>		Simulator	Serial Number	0.000	SD3933	Interferent	SD3962	0.050	SD3963	0.080	SD3967	0.200	SD3968
Simulator	Serial #	Lot #	Expiration																																																												
0.000	MP5099	N/A	N/A																																																												
0.040	MP5096	21070	03/01/2023																																																												
0.100	MP5098	21080	03/08/2023																																																												
0.200	MP5100	20510	12/03/2022																																																												
0.300	MP5101	21030	02/02/2023																																																												
0.080 DGS	N/A	00521080A2	02/05/2023																																																												
Simulator	Serial #	Lot #	Expiration																																																												
0.050	SD3963	202201C	01/11/2024																																																												
0.080	SD3967	202201D	01/18/2024																																																												
0.200	SD3968	202201E	01/18/2024																																																												
0.080 DGS	N/A	AG115904	06/08/2023																																																												
Simulator	Serial Number																																																														
0.000	SD3933																																																														
Interferent	SD3962																																																														
0.050	SD3963																																																														
0.080	SD3967																																																														
0.200	SD3968																																																														
Notes/Suggested Service: <u>Barometric offset.</u> <u>Instrument was calibrated to bring barometric values</u> <u>closer to nominal. DERR Tech review:</u> <u>Adjusted dates in and DI completion</u> <u>of dates DERR 7/21/22</u> 		22 12:35:02 04:00																																																													



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-005551, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-005551</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>MIAMI DADE PD</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>07/21/2022</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>11:47</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration. *Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

07/21/2022

Date

DAVID E REYES-RIVERA,

Department Inspector

FDLE/ATP Form 69 December 2021
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: MIAMI DADE PD

Time of Inspection: 11:47

Date of Inspection: 07/21/2022

Serial Number: 80-005551

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG115904 Exp: 06/08/2023
0.000	0.048	0.078	0.197	0.000 / 0.078
0.000	0.048	0.079	0.199	0.000 / 0.078
0.000	0.048	0.079	0.199	0.000 / 0.078
0.000	0.049	0.079	0.199	0.000 / 0.079
0.000	0.048	0.079	0.199	0.000 / 0.078
0.000	0.048	0.079	0.200	0.000 / 0.078
0.000	0.049	0.079	0.200	0.000 / 0.078
0.000	0.049	0.079	0.200	0.000 / 0.078
0.000	0.049	0.079	0.200	0.000 / 0.078
0.000	0.049	0.079	0.200	0.000 / 0.078
0.000	0.049	0.079	0.199	0.000 / 0.078
Standard Deviations	0.0005	0.0003	0.0009	0.0000 / 0.0003


Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0005 Number of Simulators Used: 5

Remarks:

08: Control Outside Tolerance GAS LINE NOT CONNECTED.

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 DAVID E REYES-RIVERA
Signature and Printed Name

07/21/2022
Date

Type of Test	Serial Number	Agency	Date	Performed By
Post Stabilities	80-005551	Miami-Dade Police Department	7/21/2022	DERR <i>[Signature]</i>

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083
MIAMI DADE PD Intoxilyzer - Alcohol Analyzer Model 8000 07/21/2022 Software: 8100.27 SN 80-005551 Test g/210L Time Air Blank 0.000 09:10 Control Test 0.048 09:11 Air Blank 0.000 09:12 Control Test 0.048 09:12 Air Blank 0.000 09:13 Control Test 0.048 09:14 Air Blank 0.000 09:14 Control Test Stats Average 0.0480 Std Dev 0.0000 Rel Std Dev(%) 0.0000 <i>[Signature]</i> Operator's Signature	MIAMI DADE PD Intoxilyzer - Alcohol Analyzer Model 8000 07/21/2022 Software: 8100.27 SN 80-005551 Test g/210L Time Air Blank 0.000 09:16 Control Test 0.078 09:16 Air Blank 0.000 09:17 Control Test 0.078 09:18 Air Blank 0.000 09:18 Control Test 0.078 09:19 Air Blank 0.000 09:19 Control Test Stats Average 0.0780 Std Dev 0.0000 Rel Std Dev(%) 0.0000 <i>[Signature]</i> Operator's Signature	MIAMI DADE PD Intoxilyzer - Alcohol Analyzer Model 8000 07/21/2022 Software: 8100.27 SN 80-005551 Test g/210L Time Air Blank 0.000 09:21 Control Test 0.197 09:21 Air Blank 0.000 09:22 Control Test 0.198 09:23 Air Blank 0.000 09:23 Control Test 0.199 09:24 Air Blank 0.000 09:25 Control Test Stats Average 1.1980 Std Dev 1.0010 Rel Std Dev(%) 1.5051 <i>[Signature]</i> Operator's Signature	MIAMI DADE PD Intoxilyzer - Alcohol Analyzer Model 8000 07/21/2022 Software: 8100.27 SN 80-005551 Test g/210L Time Air Blank 0.000 09:26 Control Test 0.079 09:26 Air Blank 0.000 09:27 Control Test 0.080 09:27 Air Blank 0.000 09:27 Control Test 0.079 09:28 Air Blank 0.000 09:28 Control Test Stats Average 0.0793 Std Dev 0.0006 Rel Std Dev(%) 0.7277 <i>[Signature]</i> Operator's Signature

<<<< CHANNEL 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 1.5320 (-0.0030)
Sample #2 = 1.5350 (0.0130)
Sample #3 = 1.5250 (0.0160)
Sample #4 = 1.5550 (0.0290)
Avg % Abs = 1.5383 (0.0193)
STD DEV = 0.0153 (0.0085)
REL STD DEV = 0.993 (43.991)

MIAMI Dade PD
Intoxilyzer - Alcohol Analyzer
Model 8000
07/21/2022
SN 80-005551
08:16:05

Auto Calibration
Max Power Res Value = 87
Auto Range Res Value = 46
Sol Value = 0.000 g/210L ***
Fit value = 0.0000 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12537, Sum Io = 14087
<<<< CHANNEL 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 1.9020 (-0.0230)
Sample #2 = 1.8200 (0.0430)
Sample #3 = 1.8330 (0.0380)
Sample #4 = 1.8480 (0.0570)
Avg % Abs = 1.8337 (0.0460)
STD DEV = 0.0140 (0.0098)
REL STD DEV = 0.764 (21.411)

<<<< CHANNEL 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 3.6080 (-0.0020)
Sample #2 = 3.5320 (0.0350)
Sample #3 = 3.5580 (0.0370)
Sample #4 = 3.5560 (0.0530)
Avg % Abs = 3.5487 (0.0417)
STD DEV = 0.0145 (0.0099)
REL STD DEV = 0.408 (23.678)

Sol Value = 0.200 g/210L ***
Fit value = 0.9524 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12492, Sum Io = 14068
<<<< CHANNEL 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 3.5960 (-0.0100)
Sample #2 = 3.5690 (0.0190)
Sample #3 = 3.5620 (0.0640)
Sample #4 = 3.5280 (0.0820)
Avg % Abs = 3.5530 (0.0550)
STD DEV = 0.0219 (0.0324)
REL STD DEV = 0.617 (59.000)

Optical Calibration	
SN:	80-005551
Agency:	Miami-Dade PD
Date:	7/21/2022
Quadratic Fit:	+/- 0.002g/210L
By:	DERR

<<<< CHANNEL 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 6.8510 (-0.0060)
Sample #2 = 6.7920 (0.0670)
Sample #3 = 6.7580 (0.0980)
Sample #4 = 6.7260 (0.1280)
Avg % Abs = 6.7587 (0.0977)
STD DEV = 0.0330 (0.0305)
REL STD DEV = 0.488 (31.230)
Sol Value = 0.300 g/210L ***
Fit value = 1.4286 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12488, Sum Io = 14063
<<<< CHANNEL 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 5.2940 (-0.0070)
Sample #2 = 5.2070 (0.0940)
Sample #3 = 5.1920 (0.1170)
Sample #4 = 5.1830 (0.1340)
Avg % Abs = 5.1940 (0.1150)
STD DEV = 0.0121 (0.0201)
REL STD DEV = 0.233 (17.456)

<<<< CHANNEL 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 9.9240 (-0.0100)
Sample #2 = 9.8070 (0.1350)
Sample #3 = 9.7790 (0.1560)
Sample #4 = 9.7770 (0.1630)
Avg % Abs = 9.7877 (0.1513)
STD DEV = 0.0168 (0.0146)
REL STD DEV = 0.171 (9.629)

<<<< CHANNEL 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.1470 (-0.0010)
Sample #2 = 0.1420 (0.0140)
Sample #3 = 0.1360 (0.0100)
Sample #4 = 0.1550 (0.0190)
Avg % Abs = 0.1443 (0.0143)
STD DEV = 0.0097 (0.0045)
REL STD DEV = 6.729 (31.460)
Sol Value = 0.040 g/210L ***
Fit value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12513, Sum Io = 14078
<<<< CHANNEL 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.7730 (-0.0060)
Sample #2 = 0.7870 (0.0280)
Sample #3 = 0.7630 (0.0440)
Sample #4 = 0.7920 (0.0630)
Avg % Abs = 0.7807 (0.0450)
STD DEV = 0.0155 (0.0175)
REL STD DEV = 1.986 (36.936)

***** AUTO CAL DATA *****
<<<< CHANNEL 1 >>>>
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.078
Std Dev = 0.02 Rel Std Dev = 30.25
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.781
Std Dev = 0.02 Rel Std Dev = 1.99
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 1.834
Std Dev = 0.01 Rel Std Dev = 0.76
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 3.553
Std Dev = 0.02 Rel Std Dev = 0.62
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 5.194
Std Dev = 0.01 Rel Std Dev = 0.23
Zero Order Coef = -191.86
First Order Coef = 2644.03
Second Order Coef = 27.27
Standard Deviation = 17.884008

<<<< CHANNEL 2 >>>>
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.144
Std Dev = 0.01 Rel Std Dev = 6.73
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 1.538
Std Dev = 0.02 Rel Std Dev = 0.99
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 3.549
Std Dev = 0.01 Rel Std Dev = 0.41
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 6.759
Std Dev = 0.03 Rel Std Dev = 0.49
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 9.788
Std Dev = 0.02 Rel Std Dev = 0.17
Zero Order Coef = -195.86
First Order Coef = 1347.62
Second Order Coef = 13.46
Standard Deviation = 4.180796

<<<< CHANNEL 1 >>>>
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.000
Std Dev = 0.000
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.040
Std Dev = 0.000
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 0.100
Std Dev = 0.000
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 0.200
Std Dev = 0.000
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 0.300
Std Dev = 0.000

Solution Stats Quadratic Fit Chan 2
Act Fit Residual
g/210L g/210L g/210L
0.000 -0.000 0.0000
0.040 0.040 -0.0001
0.100 0.100 0.0001
0.200 0.200 -0.0001
0.300 0.300 0.0000

Sol Value = 0.080 g/210L ***
Fit value = 0.3810 mg/l %%%
Samples Taken = 4, Discarded = 1
***** CHANNEL 1 *****
Sample #1 = 3322.00
Sample #2 = 3301.00
Sample #3 = 3363.00
Sample #4 = 3296.00
Average Result = 3320.0000
STD DEV = 37.3229
REL STD DEV = 1.124
***** CHANNEL 2 *****
Sample #1 = 3259.00
Sample #2 = 3298.00
Sample #3 = 3352.00
Sample #4 = 3331.00
Average Result = 3327.0000
STD DEV = 27.2213
REL STD DEV = 0.818
***** CHANNEL 1 *****
Dry Gas H2O Adjust Results *****
Barometric Pressure = 1017
3 um H2O Adjust (mg/l*10,000) = 489
9 um H2O Adjust (mg/l*10,000) = 482
***** AUTO CAL PASS *****

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-005551	Miami-Dade Police Department	7/21/2022	DERR

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083
MIAMI DADE PD Intoxilyzer - Alcohol Analyzer Model 8000 07/21/2022 Software: 8100.27 SN 80-005551 Test g/210L Time Air Blank 0.000 07:25 Control Test 0.049 07:26 Air Blank 0.000 07:27 Control Test 0.048 07:27 Air Blank 0.000 07:28 Control Test 0.049 07:29 Air Blank 0.000 07:29 Control Test Stats Average 0.0487 Std Dev 0.0006 Rel Std Dev(%) 1.1863	MIAMI DADE PD Intoxilyzer - Alcohol Analyzer Model 8000 07/21/2022 Software: 8100.27 SN 80-005551 Test g/210L Time Air Blank 0.000 07:31 Control Test 0.079 07:31 Air Blank 0.000 07:32 Control Test 0.079 07:33 Air Blank 0.000 07:33 Control Test 0.078 07:34 Air Blank 0.000 07:34 Control Test Stats Average 0.0787 Std Dev 0.0006 Rel Std Dev(%) 0.7339	MIAMI DADE PD Intoxilyzer - Alcohol Analyzer Model 8000 07/21/2022 Software: 8100.27 SN 80-005551 Test g/210L Time Air Blank 0.000 07:35 Control Test 0.201 07:36 Air Blank 0.000 07:37 Control Test 0.200 07:37 Air Blank 0.000 07:38 Control Test 0.200 07:39 Air Blank 0.000 07:39 Control Test Stats Average 0.2003 Std Dev 0.0006 Rel Std Dev(%) 0.2882	MIAMI DADE PD Intoxilyzer - Alcohol Analyzer Model 8000 07/21/2022 Software: 8100.27 SN 80-005551 Test g/210L Time Air Blank 0.000 07:41 Control Test 0.079 07:42 Air Blank 0.000 07:42 Control Test 0.079 07:42 Air Blank 0.000 07:43 Control Test 0.079 07:43 Air Blank 0.000 07:44 Control Test Stats Average 0.0790 Std Dev 0.0000 Rel Std Dev(%) 0.0000
Operator's Signature	Operator's Signature	Operator's Signature	Operator's Signature