



INSTRUMENT PROCESSING SHEET

Agency Osceola County SOS/N 80-005935Florida Department of
Law EnforcementDate In 02/11/2022 DI Completion Date 02/14/2022☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By RAW	Quality Checks	By RAW	Date 02/14/22	Flow Calibration	By	Date																																								
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes:		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>229</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP-105</u> 32 mm <u>.164</u> (.139 - .169) 36 mm <u>.179</u> (.156 - .190) 53 mm <u>.246</u> (.228 - .278) 103 mm <u>.523</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28421</u> ☒ Stability Checks			Flow Column # ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value ☐ Post Calibration Verification (L/s) Flow Column # 32 mm (.139 - .169) 36 mm (.156 - .190) 53 mm (.228 - .278) 103 mm (.447 - .547)																																										
		<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP5088</td><td>202010A 10/05/2022</td></tr><tr><td>0.080</td><td>MP5089</td><td>202201D 01/18/2024</td></tr><tr><td>0.200</td><td>MP5090</td><td>202010D 10/06/2022</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG113403 05/14/2023</td></tr></tbody></table>	Simulator	Serial #	Lot #/Exp	0.050	MP5088	202010A 10/05/2022	0.080	MP5089	202201D 01/18/2024	0.200	MP5090	202010D 10/06/2022	0.080 DGS	N/A	AG113403 05/14/2023																														
Simulator	Serial #	Lot #/Exp																																													
0.050	MP5088	202010A 10/05/2022																																													
0.080	MP5089	202201D 01/18/2024																																													
0.200	MP5090	202010D 10/06/2022																																													
0.080 DGS	N/A	AG113403 05/14/2023																																													
Calibration Adjustment		By		Department Inspection By RAW																																											
Barometric Pressure Gauge ID #				Barometric Pressure ID# <u>26932</u> Gauge <u>1027</u> Instrument <u>1025</u> Mouth Alcohol Solution Lot # <u>2021-D</u> Acetone Stock Solution Lot # <u>2021-C</u>																																											
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A					<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP5086</td></tr><tr><td>Interferent</td><td>MP5087</td></tr><tr><td>0.050</td><td>MP5088</td></tr><tr><td>0.080</td><td>MP5089</td></tr><tr><td>0.200</td><td>MP5090</td></tr></tbody></table>		Simulator	Serial Number	0.000	MP5086	Interferent	MP5087	0.050	MP5088	0.080	MP5089	0.200	MP5090		
Simulator	Serial #	Lot #	Expiration																																												
0.000		N/A	N/A																																												
0.040																																															
0.100																																															
0.200																																															
0.300																																															
0.080 DGS	N/A																																														
Simulator	Serial Number																																														
0.000	MP5086																																														
Interferent	MP5087																																														
0.050	MP5088																																														
0.080	MP5089																																														
0.200	MP5090																																														
☐ Post Calibration Adjustment Stability Checks				Attachments																																											
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A					☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Calibration Adjustment		☐ Post-Stability Checks ☐ Flow Calibration ☐ Form 40 ☐ Other																					
Simulator	Serial #	Lot #	Expiration																																												
0.050																																															
0.080																																															
0.200																																															
0.080 DGS	N/A																																														
Notes/Suggested Service: <u>Mistyped .08 lot number and date, corrected lot number and date to match solution used. -RAW, 02/14/22</u> <u>Admin Review! Added check marks to intake section. -RAW, 02/16/22</u>				☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use 2022.02. Israel Soto Digitally signed by Israel Soto Date: 2022.02.15 07:18:16 +05'00' Tech Review / Date Admin Review Date <u>05/09/2022</u>																																											

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: OSCEOLA COUNTY SO
Time of Inspection: 13:11

Date of Inspection: 02/14/2022

Serial Number: 80-005935
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 202010A Exp: 10/05/2022	0.08g/210L Test (g/210L) Lot#: 202010D 202201D Exp: 10/06/2022 01/18/2024 ^{RAW, 02/14/22}	0.20g/210L Test (g/210L) Lot#: 202010D Exp: 10/06/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: AG113403 Exp: 05/14/2023
0.000	0.049	0.080	0.200	0.080
0.000	0.049	0.079	0.200	0.081
0.000	0.050	0.080	0.200	0.081
0.000	0.050	0.079	0.200	0.081
0.000	0.050	0.079	0.200	0.081
0.000	0.049	0.079	0.200	0.081
0.000	0.050	0.080	0.200	0.081
0.000	0.050	0.080	0.199	0.080
0.000	0.050	0.080	0.199	0.080
0.000	0.050	0.080	0.200	0.080

Standard Deviations	0.0004	0.0005	0.0004	0.0005
---------------------	--------	--------	--------	--------


Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

For .08 wet solution (TAW, 02/14/22)
MISTYPED LOT NUMBER AND DATE, CORRECTED NUMBERS. - RAW, 02/14/22
RAW, 02/14/22

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



RICHARD A WILLIAMS
Signature and Printed Name

02/14/2022
Date

OSCEOLA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005935
02/14/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:31
Control Test	0.050	10:32
Air Blank	0.000	10:33
Control Test	0.049	10:33
Air Blank	0.000	10:34
Control Test	0.049	10:35
Air Blank	0.000	10:35
Control Test Stats		
Average	0.0493	
Std Dev	0.0006	
Rel Std Dev(%)	1.1703	

005

RAW

Operator's Signature

OSCEOLA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005935
02/14/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:36
Control Test	0.081	10:37
Air Blank	0.000	10:37
Control Test	0.079	10:38
Air Blank	0.000	10:39
Control Test	0.080	10:39
Air Blank	0.000	10:40
Control Test Stats		
Average	0.0800	
Std Dev	0.0010	
Rel Std Dev(%)	1.2500	

108
wet

RAW

Operator's Signature

OSCEOLA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005935
02/14/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:41
Control Test	0.199	10:42
Air Blank	0.000	10:43
Control Test	0.198	10:43
Air Blank	0.000	10:44
Control Test	0.199	10:44
Air Blank	0.000	10:45
Control Test Stats		
Average	0.1987	
Std Dev	0.0006	
Rel Std Dev(%)	0.2906	

120

RAW

Operator's Signature

OSCEOLA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005935
02/14/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:47
Control Test	0.081	10:48
Air Blank	0.000	10:48
Control Test	0.081	10:48
Air Blank	0.000	10:49
Control Test	0.080	10:49
Air Blank	0.000	10:50
Control Test Stats		
Average	0.0807	
Std Dev	0.0006	
Rel Std Dev(%)	0.7157	

108
DGS

RAW

Operator's Signature

Stability Checks

80-005935

02/14/22

RAW



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2729 Fort Knox Blvd.
Bldg. 2, Suite 1300
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-005935, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-005935</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>OSCEOLA COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>02/14/2022</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>13:11</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Richard A
Williams
Digitally signed by Richard A
Williams
Date: 2022.02.14 16:17:42
-05'00'

02/14/2022

Date

RICHARD A WILLIAMS,
Department Inspector

FDLE/ATP Form 69 January 2022

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality