

80-006027

1/6/2022

MG

PALM BEACH CO SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006027
01/06/2022
Software: 8100.27

Flow Rate Calibration*****

1: Rate (Liters/min) = 5

SQRT(Diff) = 7.211

2: Rate (Liters/min) = 15

SQRT(Diff) = 11.789

3: Rate (Liters/min) = 30

SQRT(Diff) = 20.734

Dependent Data Scale Factor = 100000 L/min

Independent Data Scale Factor = 256

Rounded Slope = 713

Rounded Intercept = -750508

Correlation = 0.99760

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-006027	Palm Beach CSO	01/06/2022	TDG ML

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083
✓	✓	✓	✓
PALM BEACH CO SO Intoxilyzer - Alcohol Analyzer Model: 8000 SN 80-006027 01/06/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 11:11 Control Test 0.049 11:12 Air Blank 0.000 11:13 Control Test 0.049 11:13 Air Blank 0.000 11:14 Control Test 0.049 11:14 Air Blank 0.000 11:15 Control Test Stats Average 0.0490 Std Dev 0.0000 Rel Std Dev(%) 0.0000 Operator's Signature 	PALM BEACH CO SO Intoxilyzer - Alcohol Analyzer Model: 8000 SN 80-006027 01/06/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 11:18 Control Test 0.079 11:19 Air Blank 0.000 11:19 Control Test 0.080 11:20 Air Blank 0.000 11:21 Control Test 0.080 11:21 Air Blank 0.000 11:22 Control Test Stats Average 0.0797 Std Dev 0.0006 Rel Std Dev(%) 0.7247 Operator's Signature 	PALM BEACH CO SO Intoxilyzer - Alcohol Analyzer Model: 8000 SN 80-006027 01/06/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 11:26 Control Test 0.200 11:27 Air Blank 0.000 11:27 Control Test 0.199 11:28 Air Blank 0.000 11:28 Control Test 0.200 11:29 Air Blank 0.000 11:30 Control Test Stats Average 0.1997 Std Dev 0.0006 Rel Std Dev(%) 0.2892 Operator's Signature 	065 PALM BEACH CO SO Intoxilyzer - Alcohol Analyzer Model: 8000 SN 80-006027 01/06/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 11:04 Control Test 0.079 11:04 Air Blank 0.000 11:05 Control Test 0.080 11:05 Air Blank 0.000 11:06 Control Test 0.080 11:06 Air Blank 0.000 11:07 Control Test Stats Average 0.0797 Std Dev 0.0006 Rel Std Dev(%) 0.7247 Operator's Signature 

Comments:

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: PALM BEACH CO SO
Time of Inspection: 14:56

Date of Inspection: 01/06/2022

Serial Number: 80-006027
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202010A Exp: 10/05/2022	0.08g/210L Test (g/210L) Lot#:202010B Exp: 10/05/2022	0.20g/210L Test (g/210L) Lot#:202010D Exp: 10/06/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG115904 Exp: 06/08/2023
0.000	0.048	0.079	0.198	0.079
0.000	0.049	0.079	0.199	0.078
0.000	0.048	0.078	0.199	0.079
0.000	0.048	0.078	0.199	0.079
0.000	0.048	0.079	0.199	0.079
0.000	0.049	0.078	0.199	0.079
0.000	0.049	0.078	0.199	0.079
0.000	0.048	0.079	0.198	0.078
0.000	0.049	0.079	0.199	0.079
0.000	0.048	0.079	0.199	0.079


Standard Deviations	0.0005	0.0005	0.0004	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

01/06/2022
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-006238, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-006238</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>PALM BEACH CO SO</u>	0.050 g/ 210 L 0.005
Calibration Date:	<u>01/06/2022</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>13:54</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

01/06/2022

Date

TAYLOR D GUTSCHOW,

Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality