

INSTRUMENT PROCESSING SHEET

Agency Palm Beach CSO

S/N 80-006029

Florida Department of
Law Enforcement

Date In 1/13/2022 DI Completion Date 1/13/2022

☐ Ship ☒ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By TDG	Quality Checks	Date 1/13/2022	Flow Calibration	By	Date																																							
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes: _____ _____ _____ _____ _____ _____ _____ _____		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value 217 ☒ Flow Verification (L/s) Flow Column # ATP104 32 mm 0.152 (.139 - .169) 36 mm 0.167 (.156 - .190) 53 mm 0.234 (.228 - .278) 103 mm 0.496 (.447 - .547) ☒ Barometric Pressure Check Gauge ID # 68639 ☒ Stability Checks		Flow Column # _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																									
		<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th style="width: 15%;">Simulator</th><th style="width: 25%;">Serial #</th><th style="width: 60%;">Lot #/Exp</th></tr> </thead> <tbody> <tr> <td rowspan="2">0.050</td><td rowspan="2">MP5092</td><td>202010A</td></tr> <tr> <td>10/05/2022</td></tr> <tr> <td rowspan="2">0.080</td><td rowspan="2">MP4864</td><td>202010B</td></tr> <tr> <td>10/05/2022</td></tr> <tr> <td rowspan="2">0.200</td><td rowspan="2">MP5094</td><td>202010D</td></tr> <tr> <td>10/06/2022</td></tr> <tr> <td rowspan="2">0.080 DGS</td><td rowspan="2">N/A</td><td>AG115904</td></tr> <tr> <td>06/08/2023</td></tr> </tbody> </table>		Simulator	Serial #	Lot #/Exp	0.050	MP5092	202010A	10/05/2022	0.080	MP4864	202010B	10/05/2022	0.200	MP5094	202010D	10/06/2022	0.080 DGS	N/A	AG115904	06/08/2023	Maintenance By DERR ☒ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____ _____ _____ _____ _____																						
Simulator	Serial #	Lot #/Exp																																											
0.050	MP5092	202010A																																											
		10/05/2022																																											
0.080	MP4864	202010B																																											
		10/05/2022																																											
0.200	MP5094	202010D																																											
		10/06/2022																																											
0.080 DGS	N/A	AG115904																																											
		06/08/2023																																											
Calibration Adjustment		By		Department Inspection By TDG																																									
Barometric Pressure Gauge _____ ID # _____		Barometric Pressure ID# 28663																																											
<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th style="width: 20%;">Simulator</th><th style="width: 20%;">Serial #</th><th style="width: 20%;">Lot #</th><th style="width: 40%;">Expiration</th></tr> </thead> <tbody> <tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr> <tr><td>0.040</td><td></td><td></td><td></td></tr> <tr><td>0.100</td><td></td><td></td><td></td></tr> <tr><td>0.200</td><td></td><td></td><td></td></tr> <tr><td>0.300</td><td></td><td></td><td></td></tr> <tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr> </tbody> </table>		Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Gauge 1017 Instrument 1014 Mouth Alcohol Solution Lot # 2021-B Acetone Stock Solution Lot # 2020-A <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th style="width: 60%;">Simulator</th><th style="width: 40%;">Serial Number</th></tr> </thead> <tbody> <tr><td>0.000</td><td>SD3963</td></tr> <tr><td>Interferent</td><td>SD1017</td></tr> <tr><td>0.050</td><td>MP5092</td></tr> <tr><td>0.080</td><td>MP4864</td></tr> <tr><td>0.200</td><td>MP5094</td></tr> </tbody> </table>				Simulator	Serial Number	0.000	SD3963	Interferent	SD1017	0.050	MP5092	0.080	MP4864	0.200	MP5094
Simulator	Serial #	Lot #	Expiration																																										
0.000		N/A	N/A																																										
0.040																																													
0.100																																													
0.200																																													
0.300																																													
0.080 DGS	N/A																																												
Simulator	Serial Number																																												
0.000	SD3963																																												
Interferent	SD1017																																												
0.050	MP5092																																												
0.080	MP4864																																												
0.200	MP5094																																												
☐ Post Calibration Adjustment Stability Checks <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th style="width: 20%;">Simulator</th><th style="width: 20%;">Serial #</th><th style="width: 20%;">Lot #</th><th style="width: 40%;">Expiration</th></tr> </thead> <tbody> <tr><td>0.050</td><td></td><td></td><td></td></tr> <tr><td>0.080</td><td></td><td></td><td></td></tr> <tr><td>0.200</td><td></td><td></td><td></td></tr> <tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr> </tbody> </table>		Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Attachments ☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Calibration Adjustment ☐ Post-Stability Checks ☐ Flow Calibration ☐ Form 40 ☐ Other _____																							
Simulator	Serial #	Lot #	Expiration																																										
0.050																																													
0.080																																													
0.200																																													
0.080 DGS	N/A																																												
Notes/Suggested Service: _____ _____ _____ _____ _____		☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																																											
		Israel Soto Digitally signed by Israel Soto Date: 2022.01.13 14:04:56 -05'00' Tech Review / Date _____ Admin Review / Date _____																																											

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-006029	Palm Beach CSD	01/13/2022	TDG <i>ML</i>

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083
PALM BEACH CO SO Intoxilyzer - Alcohol Analyzer Model 8000 01/13/2022 Software: 8100.27 SN 80-006029 Test g/210L Time Air Blank 0.000 10:22 Control Test 0.051 10:22 Air Blank 0.000 10:23 Control Test 0.050 10:24 Air Blank 0.000 10:24 Control Test 0.050 10:25 Air Blank 0.000 10:25 Control Test Stats Average 0.0503 Std Dev 0.0006 Rel Std Dev(%) 1.1471	PALM BEACH CO SO Intoxilyzer - Alcohol Analyzer Model 8000 01/13/2022 Software: 8100.27 SN 80-006029 Test g/210L Time Air Blank 0.000 10:30 Control Test 0.090 10:31 Air Blank 0.000 10:32 Control Test 0.080 10:32 Air Blank 0.000 10:33 Control Test 0.081 10:34 Air Blank 0.000 10:34 Control Test Stats Average 0.0803 Std Dev 0.0006 Rel Std Dev(%) 0.7187	PALM BEACH CO SO Intoxilyzer - Alcohol Analyzer Model 8000 01/13/2022 Software: 8100.27 SN 80-006029 Test g/210L Time Air Blank 0.000 10:38 Control Test 0.205 10:39 Air Blank 0.000 10:39 Control Test 0.203 10:40 Air Blank 0.000 10:40 Control Test 0.203 10:41 Air Blank 0.000 10:42 Control Test Stats Average 0.2037 Std Dev 0.0012 Rel Std Dev(%) 0.5670	PALM BEACH CO SO Intoxilyzer - Alcohol Analyzer Model 8000 01/13/2022 Software: 8100.27 SN 80-006029 Test g/210L Time Air Blank 0.000 10:43 Control Test 0.080 10:43 Air Blank 0.000 10:43 Control Test 0.079 10:44 Air Blank 0.000 10:44 Control Test 0.080 10:45 Air Blank 0.000 10:45 Control Test Stats Average 0.0797 Std Dev 0.0006 Rel Std Dev(%) 0.7247
<i>ML</i> Operator's Signature	<i>ML</i> Operator's Signature	<i>ML</i> Operator's Signature	<i>ML</i> Operator's Signature

Comments:

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: PALM BEACH CO SO
Time of Inspection: 13:42

Date of Inspection: 01/13/2022

Serial Number: 80-006029
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202010A Exp: 10/05/2022	0.08g/210L Test (g/210L) Lot#:202010B Exp: 10/05/2022	0.20g/210L Test (g/210L) Lot#:202010D Exp: 10/06/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG115904 Exp: 06/08/2023
0.000	0.051	0.080	0.202	0.081
0.000	0.050	0.081	0.203	0.080
0.000	0.051	0.080	0.202	0.081
0.000	0.050	0.080	0.202	0.081
0.000	0.051	0.080	0.202	0.081
0.000	0.051	0.080	0.202	0.080
0.000	0.051	0.081	0.202	0.081
0.000	0.051	0.080	0.202	0.081
0.000	0.051	0.080	0.202	0.081
0.000	0.051	0.080	0.201	0.081

Standard Deviations	0.0004	0.0004	0.0004	0.0004
---------------------	--------	--------	--------	--------


Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

Int Det:DID NOT ATTACH SIM TO SIM RETURN VALVE.

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

01/13/2022
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-006029, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-006029</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>PALM BEACH CO SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>01/13/2022</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>13:42</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.



01/13/2022

Date
TAYLOR D GUTSCHOW,
Department Inspector

FDLE/ATP Form 69 December 2021
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality