



INSTRUMENT PROCESSING SHEET

Agency Martin CSO

S/N 80-006168

Florida Department of
Law EnforcementDate In 1/18/2022 DI Completion Date 2/2/2022☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake	Quality Checks	Flow Calibration															
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes: _____ _____ _____ _____ _____ _____ _____ _____ 	☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>197</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP104</u> 32 mm <u>0.160</u> (.139 - .169) 36 mm <u>0.175</u> (.156 - .190) 53 mm <u>0.246</u> (.228 - .278) 103 mm <u>0.511</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>68639</u> ☒ Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP5092</td> <td>202010A 10/05/2022</td> </tr> <tr> <td>0.080</td> <td>MP4864</td> <td>202010B 10/05/2022</td> </tr> <tr> <td>0.200</td> <td>MP5094</td> <td>202010D 10/06/2022</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG115904 06/08/2023</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	MP5092	202010A 10/05/2022	0.080	MP4864	202010B 10/05/2022	0.200	MP5094	202010D 10/06/2022	0.080 DGS	N/A	AG115904 06/08/2023	Flow Column # _____ ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547) Maintenance By <u>DERR</u> ☒ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____ _____ _____ _____ _____
Simulator	Serial #	Lot #/Exp															
0.050	MP5092	202010A 10/05/2022															
0.080	MP4864	202010B 10/05/2022															
0.200	MP5094	202010D 10/06/2022															
0.080 DGS	N/A	AG115904 06/08/2023															

Calibration Adjustment	Department Inspection																																																												
Barometric Pressure Gauge _____ ID # _____ <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> ☐ Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> Notes/Suggested Service: <u>Added a static bag in the box for shipping back to agency (TDG)</u> _____ <u>Checked breath tube screen and replaced o-rings on 1/21/2022 (TDG)</u> _____ <u>DERR replaced the battery on 2/2/2022 (TDG)</u> _____ _____	Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Barometric Pressure ID# <u>28663</u> Gauge <u>1019</u> Instrument <u>1017</u> Mouth Alcohol Solution Lot # <u>2021-B</u> Acetone Stock Solution Lot # <u>2020-A</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>SD3963</td> </tr> <tr> <td>Interferent</td> <td>SD1017</td> </tr> <tr> <td>0.050</td> <td>MP5092</td> </tr> <tr> <td>0.080</td> <td>MP4864</td> </tr> <tr> <td>0.200</td> <td>MP5094</td> </tr> </tbody> </table> Attachments ☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Calibration Adjustment ☐ Post-Stability Checks ☐ Flow Calibration ☐ Form 40 ☐ Other _____ ☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before <u>2022-02-09</u> Evidentiary Use Israel Soto Digitally signed by Israel Soto Date: 2022.02.03 12:17:01 -0500 Tech Review / Date _____ <u>13:58:44</u> Admin Review / Date _____	Simulator	Serial Number	0.000	SD3963	Interferent	SD1017	0.050	MP5092	0.080	MP4864	0.200	MP5094
Simulator	Serial #	Lot #	Expiration																																																										
0.000		N/A	N/A																																																										
0.040																																																													
0.100																																																													
0.200																																																													
0.300																																																													
0.080 DGS	N/A																																																												
Simulator	Serial #	Lot #	Expiration																																																										
0.050																																																													
0.080																																																													
0.200																																																													
0.080 DGS	N/A																																																												
Simulator	Serial Number																																																												
0.000	SD3963																																																												
Interferent	SD1017																																																												
0.050	MP5092																																																												
0.080	MP4864																																																												
0.200	MP5094																																																												

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-006168	Martin CSO	02/02/2022	TDG <i>MC</i>

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083
MARTIN COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006168 02/02/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 12:35 Control Test 0.050 12:36 Air Blank 0.000 12:37 Control Test 0.050 12:37 Air Blank 0.000 12:38 Control Test 0.050 12:39 Air Blank 0.000 12:39 Control Test Stats Average 0.0500 Std Dev 0.0000 Rel Std Dev(%) 0.0000	MARTIN COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006168 02/02/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 12:47 Control Test 0.080 12:48 Air Blank 0.000 12:49 Control Test 0.080 12:49 Air Blank 0.000 12:50 Control Test 0.080 12:51 Air Blank 0.000 12:51 Control Test Stats Average 0.0800 Std Dev 0.0000 Rel Std Dev(%) 0.0000	MARTIN COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006168 02/02/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 13:00 Control Test 0.199 13:01 Air Blank 0.000 13:01 Control Test 0.198 13:02 Air Blank 0.000 13:02 Control Test 0.199 13:03 Air Blank 0.000 13:04 Control Test Stats Average 0.1987 Std Dev 0.0006 Rel Std Dev(%) 0.2906	MARTIN COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006168 02/02/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 12:30 Control Test 0.080 12:30 Air Blank 0.000 12:31 Control Test 0.080 12:31 Air Blank 0.000 12:31 Control Test 0.080 12:32 Air Blank 0.000 12:32 Control Test Stats Average 0.0800 Std Dev 0.0000 Rel Std Dev(%) 0.0000
<i>MC</i> Operator's Signature	<i>MC</i> Operator's Signature	<i>MC</i> Operator's Signature	<i>MC</i> Operator's Signature

Comments:

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: MARTIN COUNTY SO
Time of Inspection: 15:53

Date of Inspection: 02/02/2022

Serial Number: 80-006168
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202010A Exp: 10/05/2022	0.08g/210L Test (g/210L) Lot#:202010B Exp: 10/05/2022	0.20g/210L Test (g/210L) Lot#:202010D Exp: 10/06/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG115904 Exp: 06/08/2023
0.000	0.050	0.079	0.197	0.079
0.000	0.050	0.079	0.198	0.079
0.000	0.050	0.079	0.198	0.079
0.000	0.050	0.079	0.198	0.079
0.000	0.051	0.079	0.198	0.079
0.000	0.050	0.080	0.198	0.079
0.000	0.050	0.079	0.198	0.079
0.000	0.051	0.080	0.199	0.079
0.000	0.051	0.079	0.199	0.078
0.000	0.051	0.079	0.198	0.079

Standard Deviations	0.0005	0.0004	0.0005	0.0003
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

02/02/2022
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-006168, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-006168</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>MARTIN COUNTY SO</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>02/02/2022</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>15:53</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

02/02/2022

Date

Taylor D Gutschow

TAYLOR D GUTSCHOW,

Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-006168, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-006168</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>MARTIN COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>02/23/2022</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>11:24</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

02/23/2022

Date

David E Reyes-Rivera
DAVID E REYES-RIVERA,

Department Inspector

Service • Integrity • Respect • Quality

Florida Department of Law Enforcement

Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: MARTIN COUNTY SO
Time of Inspection: 11:24

Date of Inspection: 02/23/2022

Serial Number: 80-006168
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202010A Exp: 10/05/2022	0.08g/210L Test (g/210L) Lot#:202010B Exp: 10/05/2022	0.20g/210L Test (g/210L) Lot#:202010D Exp: 10/06/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG115904 Exp: 06/08/2023
0.000	0.050	0.080	0.198	0.079
0.000	0.050	0.079	0.199	0.080
0.000	0.050	0.079	0.199	0.079
0.000	0.050	0.079	0.199	0.079
0.000	0.050	0.080	0.199	0.078
0.000	0.050	0.080	0.199	0.080
0.000	0.049	0.080	0.199	0.079
0.000	0.049	0.079	0.199	0.079
0.000	0.050	0.080	0.199	0.080
0.000	0.050	0.079	0.199	0.079

Standard Deviations	0.0004	0.0005	0.0003	0.0006
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

David E Reyes-Rivera DAVID E REYES-RIVERA
Signature and Printed Name

02/23/2022
Date

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-006168	Martin County Sheriff's Office	2/23/2022	DERR <i>DER</i>

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																																			
0.047 to 0.053 ☒	0.077 to 0.083 ☒	0.194 to 0.206 ☒	0.077 to 0.083 ☒																																																																																																																																																			
MARTIN COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 02/23/2022 Software: 8100.27 <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>09:10</td></tr><tr><td>Control Test</td><td>0.050</td><td>09:10</td></tr><tr><td>Air Blank</td><td>0.000</td><td>09:11</td></tr><tr><td>Control Test</td><td>0.050</td><td>09:12</td></tr><tr><td>Air Blank</td><td>0.000</td><td>09:12</td></tr><tr><td>Control Test</td><td>0.050</td><td>09:13</td></tr><tr><td>Air Blank</td><td>0.000</td><td>09:13</td></tr><tr><td>Control Test</td><td>0.050</td><td>09:13</td></tr><tr><td>Control Test Status</td><td></td><td></td></tr><tr><td>Average</td><td>0.0500</td><td></td></tr><tr><td>Std Dev</td><td>0.0000</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr></table>	Test	g/210L	Time	Air Blank	0.000	09:10	Control Test	0.050	09:10	Air Blank	0.000	09:11	Control Test	0.050	09:12	Air Blank	0.000	09:12	Control Test	0.050	09:13	Air Blank	0.000	09:13	Control Test	0.050	09:13	Control Test Status			Average	0.0500		Std Dev	0.0000		Rel Std Dev(%)	0.0000		MARTIN COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 02/23/2022 Software: 8100.27 <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>09:14</td></tr><tr><td>Control Test</td><td>0.080</td><td>09:15</td></tr><tr><td>Air Blank</td><td>0.000</td><td>09:16</td></tr><tr><td>Control Test</td><td>0.080</td><td>09:16</td></tr><tr><td>Air Blank</td><td>0.000</td><td>09:17</td></tr><tr><td>Control Test</td><td>0.079</td><td>09:18</td></tr><tr><td>Air Blank</td><td>0.000</td><td>09:18</td></tr><tr><td>Control Test Status</td><td></td><td></td></tr><tr><td>Average</td><td>0.0797</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.7247</td><td></td></tr></table>	Test	g/210L	Time	Air Blank	0.000	09:14	Control Test	0.080	09:15	Air Blank	0.000	09:16	Control Test	0.080	09:16	Air Blank	0.000	09:17	Control Test	0.079	09:18	Air Blank	0.000	09:18	Control Test Status			Average	0.0797		Std Dev	0.0006		Rel Std Dev(%)	0.7247		MARTIN COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 02/23/2022 Software: 8100.27 <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>09:19</td></tr><tr><td>Control Test</td><td>0.198</td><td>09:20</td></tr><tr><td>Air Blank</td><td>0.000</td><td>09:21</td></tr><tr><td>Control Test</td><td>0.197</td><td>09:21</td></tr><tr><td>Air Blank</td><td>0.000</td><td>09:22</td></tr><tr><td>Control Test</td><td>0.197</td><td>09:23</td></tr><tr><td>Air Blank</td><td>0.000</td><td>09:23</td></tr><tr><td>Control Test Status</td><td></td><td></td></tr><tr><td>Average</td><td>0.1973</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.2926</td><td></td></tr></table>	Test	g/210L	Time	Air Blank	0.000	09:19	Control Test	0.198	09:20	Air Blank	0.000	09:21	Control Test	0.197	09:21	Air Blank	0.000	09:22	Control Test	0.197	09:23	Air Blank	0.000	09:23	Control Test Status			Average	0.1973		Std Dev	0.0006		Rel Std Dev(%)	0.2926		MARTIN COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 02/23/2022 Software: 8100.27 <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>09:24</td></tr><tr><td>Control Test</td><td>0.079</td><td>09:25</td></tr><tr><td>Air Blank</td><td>0.000</td><td>09:25</td></tr><tr><td>Control Test</td><td>0.079</td><td>09:25</td></tr><tr><td>Air Blank</td><td>0.000</td><td>09:26</td></tr><tr><td>Control Test</td><td>0.079</td><td>09:26</td></tr><tr><td>Air Blank</td><td>0.000</td><td>09:27</td></tr><tr><td>Control Test Status</td><td></td><td></td></tr><tr><td>Average</td><td>0.0790</td><td></td></tr><tr><td>Std Dev</td><td>0.0000</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr></table>	Test	g/210L	Time	Air Blank	0.000	09:24	Control Test	0.079	09:25	Air Blank	0.000	09:25	Control Test	0.079	09:25	Air Blank	0.000	09:26	Control Test	0.079	09:26	Air Blank	0.000	09:27	Control Test Status			Average	0.0790		Std Dev	0.0000		Rel Std Dev(%)	0.0000	
Test	g/210L	Time																																																																																																																																																				
Air Blank	0.000	09:10																																																																																																																																																				
Control Test	0.050	09:10																																																																																																																																																				
Air Blank	0.000	09:11																																																																																																																																																				
Control Test	0.050	09:12																																																																																																																																																				
Air Blank	0.000	09:12																																																																																																																																																				
Control Test	0.050	09:13																																																																																																																																																				
Air Blank	0.000	09:13																																																																																																																																																				
Control Test	0.050	09:13																																																																																																																																																				
Control Test Status																																																																																																																																																						
Average	0.0500																																																																																																																																																					
Std Dev	0.0000																																																																																																																																																					
Rel Std Dev(%)	0.0000																																																																																																																																																					
Test	g/210L	Time																																																																																																																																																				
Air Blank	0.000	09:14																																																																																																																																																				
Control Test	0.080	09:15																																																																																																																																																				
Air Blank	0.000	09:16																																																																																																																																																				
Control Test	0.080	09:16																																																																																																																																																				
Air Blank	0.000	09:17																																																																																																																																																				
Control Test	0.079	09:18																																																																																																																																																				
Air Blank	0.000	09:18																																																																																																																																																				
Control Test Status																																																																																																																																																						
Average	0.0797																																																																																																																																																					
Std Dev	0.0006																																																																																																																																																					
Rel Std Dev(%)	0.7247																																																																																																																																																					
Test	g/210L	Time																																																																																																																																																				
Air Blank	0.000	09:19																																																																																																																																																				
Control Test	0.198	09:20																																																																																																																																																				
Air Blank	0.000	09:21																																																																																																																																																				
Control Test	0.197	09:21																																																																																																																																																				
Air Blank	0.000	09:22																																																																																																																																																				
Control Test	0.197	09:23																																																																																																																																																				
Air Blank	0.000	09:23																																																																																																																																																				
Control Test Status																																																																																																																																																						
Average	0.1973																																																																																																																																																					
Std Dev	0.0006																																																																																																																																																					
Rel Std Dev(%)	0.2926																																																																																																																																																					
Test	g/210L	Time																																																																																																																																																				
Air Blank	0.000	09:24																																																																																																																																																				
Control Test	0.079	09:25																																																																																																																																																				
Air Blank	0.000	09:25																																																																																																																																																				
Control Test	0.079	09:25																																																																																																																																																				
Air Blank	0.000	09:26																																																																																																																																																				
Control Test	0.079	09:26																																																																																																																																																				
Air Blank	0.000	09:27																																																																																																																																																				
Control Test Status																																																																																																																																																						
Average	0.0790																																																																																																																																																					
Std Dev	0.0000																																																																																																																																																					
Rel Std Dev(%)	0.0000																																																																																																																																																					
DER Operator's Signature	DER Operator's Signature	DER Operator's Signature	DER Operator's Signature																																																																																																																																																			