



INSTRUMENT PROCESSING SHEET

Agency Flagler County Sheriff's OfficeS/N 80-006616Florida Department of
Law EnforcementDate In 04-29-2022 DI Completion Date 04-29-2022☐ Ship ☒ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By IS	Quality Checks	By PN	Date <u>4/29/2022</u>	Flow Calibration	By	Date																																								
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes: <u>Breath tube has slight fraying</u> <u>approximately 2/3 the length</u> <u>from the mouth port. PN</u> <u>Instrument was displaying Error</u> <u>10 Contact CMI. IS</u>		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>225</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP 102</u> 32 mm <u>.156</u> (.139 - .169) 36 mm <u>.171</u> (.156 - .190) 53 mm <u>.238</u> (.228 - .278) 103 mm <u>.503</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28662</u> ☒ Stability Checks			☐ Flow Column # _____ ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																										
		<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP5088</td><td>202201C 01/11/2024</td></tr><tr><td>0.080</td><td>MP5089</td><td>202201D 01/18/2024</td></tr><tr><td>0.200</td><td>MP5090</td><td>202201E 01/18/2024</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG113403 05/14/2023</td></tr></tbody></table>			Simulator	Serial #	Lot #/Exp	0.050	MP5088	202201C 01/11/2024	0.080	MP5089	202201D 01/18/2024	0.200	MP5090	202201E 01/18/2024	0.080 DGS	N/A	AG113403 05/14/2023	Maintenance By <u>IS</u> ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☒ Other <u>CMI cleared error code</u> <u>remotely through phone line.</u>																											
Simulator	Serial #	Lot #/Exp																																													
0.050	MP5088	202201C 01/11/2024																																													
0.080	MP5089	202201D 01/18/2024																																													
0.200	MP5090	202201E 01/18/2024																																													
0.080 DGS	N/A	AG113403 05/14/2023																																													
Calibration Adjustment By _____		Department Inspection By <u>IS</u>																																													
Barometric Pressure Gauge _____ ID # _____		Barometric Pressure ID# <u>26932</u> Gauge <u>1016</u> Instrument <u>1012</u> Mouth Alcohol Solution Lot # <u>2021-D</u> Acetone Stock Solution Lot # <u>2021-C</u>																																													
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP5086</td></tr><tr><td>Interferent</td><td>MP5087</td></tr><tr><td>0.050</td><td>MP5088</td></tr><tr><td>0.080</td><td>MP5089</td></tr><tr><td>0.200</td><td>MP5090</td></tr></tbody></table>						Simulator	Serial Number	0.000	MP5086	Interferent	MP5087	0.050	MP5088	0.080	MP5089	0.200	MP5090
Simulator	Serial #	Lot #	Expiration																																												
0.000		N/A	N/A																																												
0.040																																															
0.100																																															
0.200																																															
0.300																																															
0.080 DGS	N/A																																														
Simulator	Serial Number																																														
0.000	MP5086																																														
Interferent	MP5087																																														
0.050	MP5088																																														
0.080	MP5089																																														
0.200	MP5090																																														
☐ Post Calibration Adjustment Stability Checks		Attachments ☒ Form 41 ☐ Post-Stability Checks ☒ Stability Checks ☐ Flow Calibration ☒ Calibration Certificate ☐ Form 40 ☐ Calibration Adjustment ☐ Other _____																																													
Notes/Suggested Service: <u>Error code was cleared</u> <u>prior to Quality Checks and Department Inspection. IS</u>		☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use <u>2022.04.</u>																																													
		Taylor Gutschow Digitally signed by Taylor Gutschow Date: 2022.04.29 14:34:26 -0400 <u>29</u> 15:05:45 Tech Review / Date <u>04/29/22</u> Admin Review / Date _____																																													

Florida Department of Law Enforcement

Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FLAGLER COUNTY SO
Time of Inspection: 13:59

Date of Inspection: 04/29/2022

Serial Number: 80-006616
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG113403 Exp: 05/14/2023
0.000	0.050	0.080	0.200	0.081
0.000	0.050	0.081	0.200	0.081
0.000	0.050	0.081	0.200	0.082
0.000	0.050	0.080	0.200	0.081
0.000	0.050	0.080	0.199	0.082
0.000	0.050	0.080	0.200	0.082
0.000	0.050	0.080	0.199	0.082
0.000	0.051	0.080	0.200	0.082
0.000	0.051	0.081	0.200	0.082
0.000	0.050	0.081	0.199	0.082

Standard Deviations	0.0004	0.0005	0.0004	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Israel Soto ISRAEL SOTO
Signature and Printed Name

04/29/2022
Date

Stability Checks 80-006616

FLAGLER COUNTY SO
Intoxilyzer - Alconci Analyzer
Model 8000 SN 80-006616
04/29/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:19
Control Test	0.050	12:20
Air Blank	0.000	12:20
Control Test	0.050	12:21
Air Blank	0.000	12:22
Control Test	0.051	12:22
Air Blank	0.000	12:23
Control Test Stats		
Average	0.0503	
Std Dev	0.0006	
Rel Std Dev(%)	1.1471	

FLAGLER COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006616
04/29/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:24
Control Test	0.081	12:24
Air Blank	0.000	12:25
Control Test	0.081	12:26
Air Blank	0.000	12:26
Control Test	0.081	12:27
Air Blank	0.000	12:27
Control Test Stats		
Average	0.0810	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

FLAGLER COUNTY SO
Intoxilyzer - Alconci Analyzer
Model 8000 SN 80-006616
04/29/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:29
Control Test	0.201	12:30
Air Blank	0.000	12:30
Control Test	0.200	12:31
Air Blank	0.000	12:31
Control Test	0.201	12:32
Air Blank	0.000	12:33
Control Test Stats		
Average	0.2007	
Std Dev	0.0006	
Rel Std Dev(%)	0.2877	

FLAGLER COUNTY SO
Intoxilyzer - Alconci Analyzer
Model 8000 SN 80-006616
04/29/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:33
Control Test	0.081	12:34
Air Blank	0.000	12:34
Control Test	0.080	12:35
Air Blank	0.000	12:35
Control Test	0.080	12:36
Air Blank	0.000	12:36
Control Test Stats		
Average	0.0803	
Std Dev	0.0006	
Rel Std Dev(%)	0.7187	

D6S

Pat Madsen

Operator's Signature

Pat Madsen

Operator's Signature

Pat Madsen

Operator's Signature

Pat Madsen

Operator's Signature



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road.
Suite B1032
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-006616, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-006616</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>FLAGLER COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>04/29/2022</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>13:59</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Israel Soto
Soto
Date: 2022.04.29 14:24:58
-04'00'

Date

ISRAEL SOTO,

FDLE/ATP Form 69 March 2022

Issuing Authority: Alcohol Testing Program

Department Inspector

Service • Integrity • Respect • Quality