

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKE COUNTY SO
Time of Inspection: 11:04

Date of Inspection: 11/08/2022

Serial Number: 80-006222
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot #: Exp:

Number of Simulators Used: _____

Remarks:

BYPASSED AI TO OPERATE. COMPLIANCE NOT DETERMINED.

Not determined

MG
11/08/2022

The above instrument complies (☒) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Taylor D Gutschow

TAYLOR D GUTSCHOW

Signature and Printed Name

11/08/2022

Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKE COUNTY SO
Time of Inspection: 09:28

Date of Inspection: 11/15/2022

Serial Number: 80-006222
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Simulators Used: _____

Remarks:

FORMS TEST AFTER FORMS RELOAD. AI NOT CONDUCTED.

Not determined *MG*
11/15/2022

The above instrument complies (☒) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Taylor D Gutschow

TAYLOR D GUTSCHOW
Signature and Printed Name

11/15/2022
Date

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-004222	Lake CSO	11/08/2022	TDG MK

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083 ≤0.003 of Wet
LAKE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-116222 11/08/2022 Software: 8100.27	LAKE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006222 11/08/2022 Software: 8100.27	LAKE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006222 11/08/2022 Software: 8100.27	DGS LAKE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006222 11/08/2022 Software: 8100.27
Test g/210L Time	Test g/210L Time	Test g/210L Time	Test g/210L Time
Air Blank 0.000 13:45	Air Blank 0.000 14:00	Air Blank 0.000 14:07	Air Blank 0.000 13:49
Control Test 0.048 13:46	Control Test 0.078 14:01	Control Test 0.200 14:07	Control Test 0.078 13:50
Air Blank 0.000 13:46	Air Blank 0.000 14:02	Air Blank 0.000 14:08	Air Blank 0.000 13:51
Control Test 0.048 13:47	Control Test 0.079 14:02	Control Test 0.199 14:09	Control Test 0.079 13:51
Air Blank 0.000 13:47	Air Blank 0.000 14:03	Air Blank 0.000 14:09	Air Blank 0.000 13:51
Control Test 0.049 13:48	Control Test 0.078 14:03	Control Test 0.199 14:10	Control Test 0.077 13:51
Air Blank 0.000 13:49	Air Blank 0.000 14:03	Air Blank 0.000 14:10	Air Blank 0.000 13:52
Control Test Stats	Control Test Stats	Control Test Stats	Control Test Stats
Average 0.0483	Average 0.0783	Average 0.1993	Average 0.0780
Std Dev 0.0006	Std Dev 0.0006	Std Dev 0.0006	Std Dev 0.0010
Rel Std Dev(%) 1.1945	Rel Std Dev(%) 0.7370	Rel Std Dev(%) 0.2896	Rel Std Dev(%) 1.2821
Operator's Signature	Operator's Signature	Operator's Signature	Operator's Signature

Comments:

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKE COUNTY SO
Time of Inspection: 14:42

Date of Inspection: 11/15/2022

Serial Number: 80-006222
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:00521080A2 Exp: 02/05/2023
0.000	0.049	0.078	0.198	0.078
0.000	0.048	0.078	0.198	0.078
0.000	0.049	0.078	0.198	0.078
0.000	0.048	0.078	0.198	0.079
0.000	0.048	0.078	0.198	0.078
0.000	0.049	0.078	0.198	0.078
0.000	0.048	0.078	0.199	0.078
0.000	0.048	0.078	0.199	0.078
0.000	0.049	0.078	0.199	0.078
0.000	0.049	0.079	0.198	0.078


Standard Deviations	0.0005	0.0003	0.0004	0.0003
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 TAYLOR D GUTSCHOW
Signature and Printed Name

11/15/2022
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-006222, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-006222</u>	UNCERTAINTY* \pm	
Owning Agency:	<u>LAKE COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>11/15/2022</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>14:42</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

11/15/2022

Date

TAYLOR D GUTSCHOW,
Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Page 1 of 1