



INSTRUMENT PROCESSING SHEET

Agency Hendry CSOS/N 80-006234Florida Department of
Law EnforcementDate In 4/7/2022DI Completion Date 4/18/2022☐ Ship ☒ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By TDG	Quality Checks	By TDG	Date 4/18/2022	Flow Calibration	By	Date																										
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes: AI reports a suspected DGS leak.		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>193</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP104</u> 32 mm <u>0.148</u> (.139 - .169) 36 mm <u>0.160</u> (.156 - .190) 53 mm <u>0.234</u> (.228 - .278) 103 mm <u>0.500</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>68639</u> ☒ Stability Checks			Flow Column # _____ ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																												
		<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP6286</td><td>202201C 01/11/2024</td></tr><tr><td>0.080</td><td>MP6287</td><td>202201D 01/18/2024</td></tr><tr><td>0.200</td><td>MP6288</td><td>202201E 01/18/2024</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG115904 06/08/2023</td></tr></tbody></table>	Simulator	Serial #	Lot #/Exp	0.050	MP6286	202201C 01/11/2024	0.080	MP6287	202201D 01/18/2024	0.200	MP6288	202201E 01/18/2024	0.080 DGS	N/A	AG115904 06/08/2023			Maintenance By _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____													
Simulator	Serial #	Lot #/Exp																															
0.050	MP6286	202201C 01/11/2024																															
0.080	MP6287	202201D 01/18/2024																															
0.200	MP6288	202201E 01/18/2024																															
0.080 DGS	N/A	AG115904 06/08/2023																															
Calibration Adjustment By _____		Department Inspection By TDG _____																															
Barometric Pressure Gauge _____ ID # _____		Barometric Pressure ID# <u>28199</u>																															
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Gauge <u>1014</u> Instrument <u>1011</u> Mouth Alcohol Solution Lot # <u>2021-B</u> Acetone Stock Solution Lot # <u>2021-C</u>			
Simulator	Serial #	Lot #	Expiration																														
0.000		N/A	N/A																														
0.040																																	
0.100																																	
0.200																																	
0.300																																	
0.080 DGS	N/A																																
☐ Post Calibration Adjustment Stability Checks		<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP6284</td></tr><tr><td>Interferent</td><td>MP6285</td></tr><tr><td>0.050</td><td>MP6286</td></tr><tr><td>0.080</td><td>MP6287</td></tr><tr><td>0.200</td><td>MP6288</td></tr></tbody></table>		Simulator	Serial Number	0.000	MP6284	Interferent	MP6285	0.050	MP6286	0.080	MP6287	0.200	MP6288																		
Simulator	Serial Number																																
0.000	MP6284																																
Interferent	MP6285																																
0.050	MP6286																																
0.080	MP6287																																
0.200	MP6288																																
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Attachments ☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Calibration Adjustment ☐ Post-Stability Checks ☐ Flow Calibration ☐ Form 40 ☐ Other _____											
Simulator	Serial #	Lot #	Expiration																														
0.050																																	
0.080																																	
0.200																																	
0.080 DGS	N/A																																
Notes/Suggested Service: Tested for DGS leak. No loss of psi between initial test date (4/8) and end test date (4/18). (TDG)		☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																															
		Israel Soto Tech Review / Date		13:48:45 Admin Review / Date																													

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-006234	Handy CSO	04/18/2022	TDG <i>ML</i>

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083																																																																																																
0.05 HENRY COUNTY SO Intoxilyzer - Alcotest Analyzer Model 8000 04/18/2022 Software: 8100.27 Test Time <table><tr><td>Air Blank</td><td>0.000</td><td>11:59</td></tr><tr><td>Control Test</td><td>0.050</td><td>12:03</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:04</td></tr><tr><td>Control Test</td><td>0.049</td><td>12:08</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:09</td></tr><tr><td>Control Test</td><td>0.050</td><td>12:13</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:14</td></tr><tr><td>Control Test</td><td>0.049</td><td>12:17</td></tr></table> Average 0.0497 Std Dev 0.0006 Rel Std Dev(%) 1.1625 ML Operator's Signature	Air Blank	0.000	11:59	Control Test	0.050	12:03	Air Blank	0.000	12:04	Control Test	0.049	12:08	Air Blank	0.000	12:09	Control Test	0.050	12:13	Air Blank	0.000	12:14	Control Test	0.049	12:17	HENRY COUNTY SO Intoxilyzer - Alcotest Analyzer Model 8000 04/18/2022 Software: 8100.27 Test Time <table><tr><td>Air Blank</td><td>0.000</td><td>11:53</td></tr><tr><td>Control Test</td><td>0.080</td><td>11:57</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:58</td></tr><tr><td>Control Test</td><td>0.080</td><td>12:02</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:03</td></tr><tr><td>Control Test</td><td>0.079</td><td>12:06</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:07</td></tr><tr><td>Control Test</td><td>0.080</td><td>12:10</td></tr></table> Average 0.0797 Std Dev 0.0006 Rel Std Dev(%) 0.7247 ML Operator's Signature	Air Blank	0.000	11:53	Control Test	0.080	11:57	Air Blank	0.000	11:58	Control Test	0.080	12:02	Air Blank	0.000	12:03	Control Test	0.079	12:06	Air Blank	0.000	12:07	Control Test	0.080	12:10	HENRY COUNTY SO Intoxilyzer - Alcotest Analyzer Model 8000 04/18/2022 Software: 8100.27 Test Time <table><tr><td>Air Blank</td><td>0.000</td><td>11:53</td></tr><tr><td>Control Test</td><td>0.201</td><td>11:57</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:58</td></tr><tr><td>Control Test</td><td>0.201</td><td>12:02</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:03</td></tr><tr><td>Control Test</td><td>0.200</td><td>12:06</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:07</td></tr><tr><td>Control Test</td><td>0.200</td><td>12:10</td></tr></table> Average 0.2007 Std Dev 0.0006 Rel Std Dev(%) 0.2877 ML Operator's Signature	Air Blank	0.000	11:53	Control Test	0.201	11:57	Air Blank	0.000	11:58	Control Test	0.201	12:02	Air Blank	0.000	12:03	Control Test	0.200	12:06	Air Blank	0.000	12:07	Control Test	0.200	12:10	HENRY COUNTY SO Intoxilyzer - Alcotest Analyzer Model 8000 04/18/2022 Software: 8100.27 Test Time <table><tr><td>Air Blank</td><td>0.000</td><td>11:21</td></tr><tr><td>Control Test</td><td>0.082</td><td>11:25</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:26</td></tr><tr><td>Control Test</td><td>0.083</td><td>11:30</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:31</td></tr><tr><td>Control Test</td><td>0.081</td><td>11:35</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:36</td></tr><tr><td>Control Test</td><td>0.080</td><td>11:39</td></tr></table> Average 0.0820 Std Dev 0.0010 Rel Std Dev(%) 1.2195 ML Operator's Signature	Air Blank	0.000	11:21	Control Test	0.082	11:25	Air Blank	0.000	11:26	Control Test	0.083	11:30	Air Blank	0.000	11:31	Control Test	0.081	11:35	Air Blank	0.000	11:36	Control Test	0.080	11:39
Air Blank	0.000	11:59																																																																																																	
Control Test	0.050	12:03																																																																																																	
Air Blank	0.000	12:04																																																																																																	
Control Test	0.049	12:08																																																																																																	
Air Blank	0.000	12:09																																																																																																	
Control Test	0.050	12:13																																																																																																	
Air Blank	0.000	12:14																																																																																																	
Control Test	0.049	12:17																																																																																																	
Air Blank	0.000	11:53																																																																																																	
Control Test	0.080	11:57																																																																																																	
Air Blank	0.000	11:58																																																																																																	
Control Test	0.080	12:02																																																																																																	
Air Blank	0.000	12:03																																																																																																	
Control Test	0.079	12:06																																																																																																	
Air Blank	0.000	12:07																																																																																																	
Control Test	0.080	12:10																																																																																																	
Air Blank	0.000	11:53																																																																																																	
Control Test	0.201	11:57																																																																																																	
Air Blank	0.000	11:58																																																																																																	
Control Test	0.201	12:02																																																																																																	
Air Blank	0.000	12:03																																																																																																	
Control Test	0.200	12:06																																																																																																	
Air Blank	0.000	12:07																																																																																																	
Control Test	0.200	12:10																																																																																																	
Air Blank	0.000	11:21																																																																																																	
Control Test	0.082	11:25																																																																																																	
Air Blank	0.000	11:26																																																																																																	
Control Test	0.083	11:30																																																																																																	
Air Blank	0.000	11:31																																																																																																	
Control Test	0.081	11:35																																																																																																	
Air Blank	0.000	11:36																																																																																																	
Control Test	0.080	11:39																																																																																																	

Comments:

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: HENDRY COUNTY SO
Time of Inspection: 13:39

Date of Inspection: 04/18/2022

Serial Number: 80-006234
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG115904 Exp: 06/08/2023
0.000	0.049	0.080	0.201	0.082
0.000	0.048	0.080	0.201	0.082
0.000	0.048	0.079	0.201	0.081
0.000	0.049	0.079	0.201	0.082
0.000	0.049	0.079	0.201	0.082
0.000	0.049	0.079	0.201	0.081
0.000	0.049	0.079	0.201	0.081
0.000	0.049	0.080	0.201	0.081
0.000	0.049	0.079	0.201	0.081
0.000	0.049	0.080	0.201	0.082


Standard Deviations	0.0004	0.0005	0.0000	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.


I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

04/18/2022
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-006234, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-006234</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>HENDRY COUNTY SO</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>04/18/2022</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>13:39</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

04/18/2022

Date


TAYLOR D GUTSCHOW,

Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality