



INSTRUMENT PROCESSING SHEET

Agency Hillsborough CSO

S/N 80-006567

Florida Department of
Law Enforcement

Date In 4/20/2022 DI Completion Date 4/25/2022

☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By TDG	Quality Checks	By TDG	Date <u>4/22/2022</u>	Flow Calibration	By	Date																																								
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: _____ _____ _____ _____ _____ _____ _____ _____		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>204</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP104</u> 32 mm <u>0.152</u> (.139 - .169) 36 mm <u>0.164</u> (.156 - .190) 53 mm <u>0.238</u> (.228 - .278) 103 mm <u>0.496</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>68639</u> ☒ Stability Checks			Flow Column # _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																										
					Maintenance By _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____ _____ _____ _____ _____																																										
Calibration Adjustment By TDG _____ Barometric Pressure Gauge <u>1023</u> ID # <u>28199</u>				Department Inspection By TDG _____ Barometric Pressure ID# <u>68639</u> Gauge <u>1019</u> Instrument <u>1019</u> Mouth Alcohol Solution Lot # <u>2021-B</u> Acetone Stock Solution Lot # <u>2021-C</u>																																											
<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP5095</td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td>MP5096</td> <td>21070</td> <td>03/01/2023</td> </tr> <tr> <td>0.100</td> <td>MP5098</td> <td>21080</td> <td>03/08/2023</td> </tr> <tr> <td>0.200</td> <td>MP5100</td> <td>20510</td> <td>12/03/2022</td> </tr> <tr> <td>0.300</td> <td>MP5101</td> <td>21030</td> <td>02/02/2023</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>00521080A2</td> <td>02/05/2023</td> </tr> </tbody> </table>				Simulator	Serial #	Lot #	Expiration	0.000	MP5095	N/A	N/A	0.040	MP5096	21070	03/01/2023	0.100	MP5098	21080	03/08/2023	0.200	MP5100	20510	12/03/2022	0.300	MP5101	21030	02/02/2023	0.080 DGS	N/A	00521080A2	02/05/2023	<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP6284</td> </tr> <tr> <td>Interferent</td> <td>MP6285</td> </tr> <tr> <td>0.050</td> <td>MP6286</td> </tr> <tr> <td>0.080</td> <td>MP6287</td> </tr> <tr> <td>0.200</td> <td>MP6288</td> </tr> </tbody> </table>				Simulator	Serial Number	0.000	MP6284	Interferent	MP6285	0.050	MP6286	0.080	MP6287	0.200	MP6288
Simulator	Serial #	Lot #	Expiration																																												
0.000	MP5095	N/A	N/A																																												
0.040	MP5096	21070	03/01/2023																																												
0.100	MP5098	21080	03/08/2023																																												
0.200	MP5100	20510	12/03/2022																																												
0.300	MP5101	21030	02/02/2023																																												
0.080 DGS	N/A	00521080A2	02/05/2023																																												
Simulator	Serial Number																																														
0.000	MP6284																																														
Interferent	MP6285																																														
0.050	MP6286																																														
0.080	MP6287																																														
0.200	MP6288																																														
☒ Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP5092</td> <td>202010A</td> <td>10/05/2022</td> </tr> <tr> <td>0.080</td> <td>MP5093</td> <td>202010B</td> <td>10/05/2022</td> </tr> <tr> <td>0.200</td> <td>MP5094</td> <td>202010D</td> <td>10/06/2022</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG115904</td> <td>06/08/2023</td> </tr> </tbody> </table>				Simulator	Serial #	Lot #	Expiration	0.050	MP5092	202010A	10/05/2022	0.080	MP5093	202010B	10/05/2022	0.200	MP5094	202010D	10/06/2022	0.080 DGS	N/A	AG115904	06/08/2023	Attachments ☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☒ Calibration Adjustment ☒ Post-Stability Checks ☐ Flow Calibration ☐ Form 40 ☐ Other _____																							
Simulator	Serial #	Lot #	Expiration																																												
0.050	MP5092	202010A	10/05/2022																																												
0.080	MP5093	202010B	10/05/2022																																												
0.200	MP5094	202010D	10/06/2022																																												
0.080 DGS	N/A	AG115904	06/08/2023																																												
Notes/Suggested Service: <u>Instrument failed Baro Pressure Check during the Quality Checks (the instrument read 1001 and the gauge read 1022). After the Optical Cal, the gauge and the instrument both read 1022. (TDG)</u> _____ _____ _____ _____				☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use David Eliezer Reyes-Rivera Digitally signed by David Eliezer Reyes-Rivera Date: 2022.04.25 14:24:50 -0400 2022.04 <u>26</u> 13:30:3 Tech Review / Date Admin Review / Date																																											

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-006567	Hillsborough CSO	04/22/2022	TDG MC

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083
✓	✓	✓	✓
HILLSBOROUGH CO SO Intoxilyzer - Alcohol Analyzer Model 8000 04/22/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 10:51 Control Test 0.051 10:52 Air Blank 0.000 10:52 Control Test 0.050 10:53 Air Blank 0.000 10:54 Control Test 0.050 10:54 Air Blank 0.000 10:55 Control Test Stats Average 0.0503 Std Dev 0.0006 Rel Std Dev(%) 1.1471	HILLSBOROUGH CO SO Intoxilyzer - Alcohol Analyzer Model 8000 04/22/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 10:58 Control Test 0.080 10:59 Air Blank 0.000 10:59 Control Test 0.081 11:00 Air Blank 0.000 11:00 Control Test 0.081 11:01 Air Blank 0.000 11:02 Control Test Stats Average 0.0807 Std Dev 0.0006 Rel Std Dev(%) 0.7157	HILLSBOROUGH CO SO Intoxilyzer - Alcohol Analyzer Model 8000 04/22/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 11:05 Control Test 0.198 11:06 Air Blank 0.000 11:06 Control Test 0.198 11:07 Air Blank 0.000 11:08 Control Test 0.198 11:08 Air Blank 0.000 11:09 Control Test Stats Average 0.1980 Std Dev 0.0000 Rel Std Dev(%) 0.0000	HILLSBOROUGH CO SO Intoxilyzer - Alcohol Analyzer Model 8000 04/22/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 11:11 Control Test 0.081 11:11 Air Blank 0.000 11:12 Control Test 0.081 11:12 Air Blank 0.000 11:13 Control Test 0.082 11:13 Air Blank 0.000 11:14 Control Test Stats Average 0.0813 Std Dev 0.0006 Rel Std Dev(%) 0.7099
MC Operator's Signature	MC Operator's Signature	MC Operator's Signature	MC Operator's Signature

Comments:


```

*****
Dry Gas H2O Adjust Results *****
Barometric Pressure = 1022
3 um H2O Adjust (mg/l*10,000) = 240
9 um H2O Adjust (mg/l*10,000) = 40
**** AUTO CAL PASS

```

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities (Post-Cal)	80-006567	Hillsborough CSO	04/22/2022	TDG MB

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083
✓	✓	✓	✓
HILLSBOROUGH CO SO Intoxilyzer - Alcohol Analyzer Model 8000 04/22/2022 Software: 8100.27 SN 80-006567 Test g/210L Time Air Blank 0.000 12:51 Control Test 0.048 12:52 Air Blank 0.000 12:52 Control Test 0.048 12:53 Air Blank 0.000 12:54 Control Test 0.048 12:54 Air Blank 0.000 12:55 Control Test Stats Average 0.0480 Std Dev 0.0000 Rel Std Dev(%) 0.0000	HILLSBOROUGH CO SO Intoxilyzer - Alcohol Analyzer Model 8000 04/22/2022 Software: 8100.27 SN 80-006567 Test g/210L Time Air Blank 0.000 13:10 Control Test 0.078 13:11 Air Blank 0.000 13:12 Control Test 0.077 13:12 Air Blank 0.000 13:13 Control Test 0.077 13:14 Air Blank 0.000 13:14 Control Test Stats Average 0.0773 Std Dev 0.0006 Rel Std Dev(%) 0.7466	HILLSBOROUGH CO SO Intoxilyzer - Alcohol Analyzer Model 8000 04/22/2022 Software: 8100.27 SN 80-006567 Test g/210L Time Air Blank 0.000 13:29 Control Test 0.196 13:30 Air Blank 0.000 13:31 Control Test 0.196 13:31 Air Blank 0.000 13:32 Control Test 0.196 13:33 Air Blank 0.000 13:33 Control Test Stats Average 0.1960 Std Dev 0.0000 Rel Std Dev(%) 0.0000	HILLSBOROUGH CO SO Intoxilyzer - Alcohol Analyzer Model 8000 04/22/2022 Software: 8100.27 SN 80-006567 Test g/210L Time Air Blank 0.000 12:57 Control Test 0.079 12:57 Air Blank 0.000 12:58 Control Test 0.079 12:58 Air Blank 0.000 12:58 Control Test 0.079 12:59 Air Blank 0.000 12:59 Control Test Stats Average 0.0790 Std Dev 0.0000 Rel Std Dev(%) 0.0000
Operator's Signature MB	Operator's Signature MB	Operator's Signature MB	Operator's Signature MB

Comments:

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: HILLSBOROUGH CO SO
Time of Inspection: 13:02

Date of Inspection: 04/25/2022

Serial Number: 80-006567
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG115904 Exp: 06/08/2023
0.000	0.048	0.077	0.197	0.078
0.000	0.048	0.077	0.197	0.078
0.000	0.048	0.077	0.197	0.078
0.000	0.048	0.078	0.197	0.078
0.000	0.048	0.078	0.197	0.078
0.000	0.048	0.078	0.196	0.078
0.000	0.049	0.078	0.198	0.078
0.000	0.048	0.077	0.197	0.077
0.000	0.049	0.077	0.197	0.077
0.000	0.049	0.077	0.196	0.078


Standard Deviations	0.0004	0.0005	0.0005	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

04/25/2022
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-006567, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-006567</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>HILLSBOROUGH CO SO</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>04/25/2022</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>13:02</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

04/25/2022

Date


TAYLOR D GUTSCHOW,
Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality



INSTRUMENT PROCESSING SHEET

Agency Hillsborough CSO

S/N 80-006567

Florida Department of
Law Enforcement

Date In 9/9/2022

DI Completion Date 9/21/2022

☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By TDG	Quality Checks	By TDG	Date 9/21/2022	Flow Calibration	By	Date																												
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: AI Dendy reports low DGS readings and a Control Outside Tolerance.		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value 210 ☒ Flow Verification (L/s) Flow Column # ATP104 32 mm 0.144 (.139 - .169) 36 mm 0.167 (.156 - .190) 53 mm 0.238 (.228 - .278) 103 mm 0.500 (.447 - .547) ☒ Barometric Pressure Check Gauge ID # 68639 ☒ Stability Checks			Flow Column # ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value ☐ Post Calibration Verification (L/s) Flow Column # 32 mm (.139 - .169) 36 mm (.156 - .190) 53 mm (.228 - .278) 103 mm (.447 - .547)																														
		<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP6286</td><td>202201C 01/11/2024</td></tr><tr><td>0.080</td><td>MP6287</td><td>202201D 01/18/2024</td></tr><tr><td>0.200</td><td>MP6288</td><td>202201E 01/18/2024</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>00521080A2 02/05/2023</td></tr></tbody></table>	Simulator	Serial #	Lot #/Exp	0.050	MP6286	202201C 01/11/2024	0.080	MP6287	202201D 01/18/2024	0.200	MP6288	202201E 01/18/2024	0.080 DGS	N/A	00521080A2 02/05/2023																		
Simulator	Serial #	Lot #/Exp																																	
0.050	MP6286	202201C 01/11/2024																																	
0.080	MP6287	202201D 01/18/2024																																	
0.200	MP6288	202201E 01/18/2024																																	
0.080 DGS	N/A	00521080A2 02/05/2023																																	
Calibration Adjustment		By		Department Inspection			By TDG																												
Barometric Pressure Gauge ID #		Barometric Pressure ID# 28199																																	
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Gauge 1012 Instrument 1008 Mouth Alcohol Solution Lot # 2021-D Acetone Stock Solution Lot # 2021-C					
Simulator	Serial #	Lot #	Expiration																																
0.000		N/A	N/A																																
0.040																																			
0.100																																			
0.200																																			
0.300																																			
0.080 DGS	N/A																																		
☐ Post Calibration Adjustment Stability Checks		<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP6284</td></tr><tr><td>Interferent</td><td>MP6285</td></tr><tr><td>0.050</td><td>MP6286</td></tr><tr><td>0.080</td><td>MP6287</td></tr><tr><td>0.200</td><td>MP6288</td></tr></tbody></table>						Simulator	Serial Number	0.000	MP6284	Interferent	MP6285	0.050	MP6286	0.080	MP6287	0.200	MP6288																
Simulator	Serial Number																																		
0.000	MP6284																																		
Interferent	MP6285																																		
0.050	MP6286																																		
0.080	MP6287																																		
0.200	MP6288																																		
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Attachments ☒ Form 41 ☐ Post-Stability Checks ☒ Stability Checks ☐ Flow Calibration ☒ Calibration Certificate ☐ Form 40 ☐ Calibration Adjustment ☐ Other													
Simulator	Serial #	Lot #	Expiration																																
0.050																																			
0.080																																			
0.200																																			
0.080 DGS	N/A																																		
Notes/Suggested Service: All ARS and DGS values were in acceptable range. Reviewed breath tests and Agency Inspection data for the last few months and did not find a trend of COT exception messages for either. (TDG)		☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																																	
		Phil Nicodemo Digitally signed by Phil Nicodemo Date: 2022.09.22 15:45:11 -0400																																	
		Tech Review / Date Admin Review / Date																																	

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-00 6567	Hillsborough CSO	09/21/2022	TDG <i>ML</i>

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083
HILLSBOROUGH CO SO Intoxilyzer - Alcohol Analyzer Model 8000 09/21/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 10:54 Control Test 0.049 10:54 Air Blank 0.000 10:55 Control Test 0.048 10:55 Air Blank 0.000 10:56 Control Test 0.048 10:57 Air Blank 0.000 10:57 Control Test Stats Average 0.0483 Std Dev 0.0006 Rel Std Dev(%) 1.1945 Operator's Signature <i>ML</i>	HILLSBOROUGH CO SO Intoxilyzer - Alcohol Analyzer Model 8000 09/21/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 11:00 Control Test 0.079 11:01 Air Blank 0.000 11:01 Control Test 0.078 11:02 Air Blank 0.000 11:03 Control Test 0.078 11:03 Air Blank 0.000 11:04 Control Test Stats Average 0.0783 Std Dev 0.0006 Rel Std Dev(%) 0.7370 Operator's Signature <i>ML</i>	HILLSBOROUGH CO SO Intoxilyzer - Alcohol Analyzer Model 8000 09/21/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 11:09 Control Test 0.198 11:09 Air Blank 0.000 11:10 Control Test 0.198 11:11 Air Blank 0.000 11:11 Control Test 0.199 11:12 Air Blank 0.000 11:12 Control Test Stats Average 0.1983 Std Dev 0.0006 Rel Std Dev(%) 0.2911 Operator's Signature <i>ML</i>	HILLSBOROUGH CO SO Intoxilyzer - Alcohol Analyzer Model 8000 09/21/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 11:13 Control Test 0.078 11:13 Air Blank 0.000 11:14 Control Test 0.078 11:14 Air Blank 0.000 11:15 Control Test 0.079 11:15 Air Blank 0.000 11:15 Control Test Stats Average 0.0783 Std Dev 0.0006 Rel Std Dev(%) 0.7370 Operator's Signature <i>ML</i>

Comments:

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: HILLSBOROUGH CO SO
Time of Inspection: 14:42

Date of Inspection: 09/21/2022

Serial Number: 80-006567
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:00521080A2 Exp: 02/05/2023
0.000	0.048	0.077	0.197	0.078
0.000	0.048	0.077	0.197	0.078
0.000	0.049	0.077	0.197	0.077
0.000	0.049	0.077	0.197	0.078
0.000	0.049	0.077	0.197	0.078
0.000	0.048	0.077	0.197	0.077
0.000	0.049	0.077	0.197	0.078
0.000	0.049	0.077	0.196	0.077
0.000	0.048	0.077	0.197	0.077
0.000	0.048	0.077	0.197	0.078


Standard Deviations	0.0005	0.0000	0.0003	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5


Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 TAYLOR D GUTSCHOW
Signature and Printed Name

09/21/2022
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-006567, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-006567</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>HILLSBOROUGH CO SO</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>09/21/2022</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>14:42</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

09/21/2022

Date

TAYLOR D GUTSCHOW,

Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality