




INSTRUMENT PROCESSING SHEET

Agency Boca Raton Police DepartmentS/N 80-006622Florida Department of
Law EnforcementDate In 9/15/2022DI Completion Date 9/21/2022☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake By <u>DERR</u> ☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: <u>Affixed missing CMI</u> <u>authorized service agency label.</u>	Quality Checks By <u>DERR</u> Date <u>9/19/2022</u> ☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>ATP101 220 Derr 9/26/22</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP101</u> 32 mm <u>0.160</u> (.139 - .169) 36 mm <u>0.179</u> (.156 - .190) 53 mm <u>0.250</u> (.228 - .278) 103 mm <u>0.519</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28199</u> ☒ Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> <tr> <td>0.050</td> <td>MP6286</td> <td>202201C 01/11/2024</td> </tr> <tr> <td>0.080</td> <td>MP6287</td> <td>202201D 01/18/2024</td> </tr> <tr> <td>0.200</td> <td>MP6288</td> <td>202201E 01/18/2024</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>00521080A2 02/05/2023</td> </tr> </table>	Simulator	Serial #	Lot #/Exp	0.050	MP6286	202201C 01/11/2024	0.080	MP6287	202201D 01/18/2024	0.200	MP6288	202201E 01/18/2024	0.080 DGS	N/A	00521080A2 02/05/2023	Flow Calibration By _____ Date _____ Flow Column # _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																														
Simulator	Serial #	Lot #/Exp																																																													
0.050	MP6286	202201C 01/11/2024																																																													
0.080	MP6287	202201D 01/18/2024																																																													
0.200	MP6288	202201E 01/18/2024																																																													
0.080 DGS	N/A	00521080A2 02/05/2023																																																													
		Maintenance By <u>DERR</u> ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☒ Other <u>Had CMI clear error 012</u>																																																													
Calibration Adjustment By <u>DERR</u> Barometric Pressure Gauge <u>1010</u> ID # <u>68639</u> <table border="1" style="width:100%; border-collapse: collapse;"> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> <tr> <td>0.000</td> <td>MP5099</td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td>MP5096</td> <td>21070</td> <td>03/01/2023</td> </tr> <tr> <td>0.100</td> <td>MP5098</td> <td>21080</td> <td>03/08/2023</td> </tr> <tr> <td>0.200</td> <td>MP5100</td> <td>20510</td> <td>12/03/2022</td> </tr> <tr> <td>0.300</td> <td>MP5101</td> <td>21030</td> <td>02/02/2023</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG115904</td> <td>06/08/2023</td> </tr> </table> ☒ Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> <tr> <td>0.050</td> <td>MP6286</td> <td>202201C</td> <td>01/11/2024</td> </tr> <tr> <td>0.080</td> <td>MP6287</td> <td>202201D</td> <td>01/18/2024</td> </tr> <tr> <td>0.200</td> <td>MP6288</td> <td>202201E</td> <td>01/18/2024</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>00521080A2</td> <td>02/05/2023</td> </tr> </table>		Simulator	Serial #	Lot #	Expiration	0.000	MP5099	N/A	N/A	0.040	MP5096	21070	03/01/2023	0.100	MP5098	21080	03/08/2023	0.200	MP5100	20510	12/03/2022	0.300	MP5101	21030	02/02/2023	0.080 DGS	N/A	AG115904	06/08/2023	Simulator	Serial #	Lot #	Expiration	0.050	MP6286	202201C	01/11/2024	0.080	MP6287	202201D	01/18/2024	0.200	MP6288	202201E	01/18/2024	0.080 DGS	N/A	00521080A2	02/05/2023	Department Inspection By <u>DERR</u> Barometric Pressure ID# <u>26932</u> Gauge <u>1011</u> Instrument <u>1011</u> Mouth Alcohol Solution Lot # <u>2021-D</u> Acetone Stock Solution Lot # <u>2021-C</u> <table border="1" style="width:100%; border-collapse: collapse;"> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> <tr> <td>0.000</td> <td>MP6284</td> </tr> <tr> <td>Interferent</td> <td>MP6285</td> </tr> <tr> <td>0.050</td> <td>MP6286</td> </tr> <tr> <td>0.080</td> <td>MP6287</td> </tr> <tr> <td>0.200</td> <td>MP6288</td> </tr> </table>		Simulator	Serial Number	0.000	MP6284	Interferent	MP6285	0.050	MP6286	0.080	MP6287	0.200	MP6288
Simulator	Serial #	Lot #	Expiration																																																												
0.000	MP5099	N/A	N/A																																																												
0.040	MP5096	21070	03/01/2023																																																												
0.100	MP5098	21080	03/08/2023																																																												
0.200	MP5100	20510	12/03/2022																																																												
0.300	MP5101	21030	02/02/2023																																																												
0.080 DGS	N/A	AG115904	06/08/2023																																																												
Simulator	Serial #	Lot #	Expiration																																																												
0.050	MP6286	202201C	01/11/2024																																																												
0.080	MP6287	202201D	01/18/2024																																																												
0.200	MP6288	202201E	01/18/2024																																																												
0.080 DGS	N/A	00521080A2	02/05/2023																																																												
Simulator	Serial Number																																																														
0.000	MP6284																																																														
Interferent	MP6285																																																														
0.050	MP6286																																																														
0.080	MP6287																																																														
0.200	MP6288																																																														
Notes/Suggested Service: <u>Instrument came with a note indicating that it needed to be uploaded, that it was showing error 012, that was in dissabled mode, and needed a DI. After CMI cleared the error 012, I conducted quality checks. Conducted optical calibration to bring values closer to nominal. DERR</u> <u>Corrected R value and resubmitted Derr 9/26/22 Corrected per admin review</u>		Attachments ☒ Form 41 ☒ Post-Stability Checks ☒ Stability Checks ☐ Flow Calibration ☒ Calibration Certificate ☐ Form 40 ☒ Calibration Adjustment ☐ Other _____																																																													
☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use		Israel Soto Digitally signed by Israel Soto Date: 2022.09.27 08:27:47 +04'00' Tech Review / Date _____ Admin Review / Date <u>2022.10.12 12:37:50</u>																																																													

DERR Digitally signed by DERR
Date: 2022.10.10
09:39:21 -04'00'



Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

Calibration Certificate

This is to certify the calibration of Intoxilyzer 8000 serial number 80-006622, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-006622</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>BOCA RATON PD</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>09/21/2022</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>08:46</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

09/21/2022

Date

DAVID E REYES-RIVERA,

Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Page 1 of 1

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: BOCA RATON PD
Time of Inspection: 08:46

Date of Inspection: 09/21/2022

Serial Number: 80-006622
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:00521080A2 Exp: 02/05/2023
0.000	0.048	0.078	0.198	0.080
0.000	0.049	0.077	0.198	0.080
0.000	0.049	0.078	0.198	0.080
0.000	0.049	0.078	0.198	0.080
0.000	0.049	0.078	0.198	0.080
0.000	0.049	0.078	0.198	0.080
0.000	0.049	0.078	0.198	0.081
0.000	0.049	0.078	0.198	0.081
0.000	0.049	0.078	0.198	0.081
0.000	0.049	0.078	0.198	0.081


Standard Deviations	0.0003	0.0003	0.0000	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0002 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that, I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


Signature and Printed Name

DAVID E REYES-RIVERA

09/21/2022
Date

Type of Test	Serial Number	Agency	Date	Performed By
Post Stabilities	80-006622	Boca Raton Police Department	9/21/2022	DERR <i>[Signature]</i>

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053 ☒	0.077 to 0.083 ☒	0.194 to 0.206 ☒	0.077 to 0.083 ☒
BOCA RATON PD Intoxilyzer - Alcohol Analyzer Model 8000 09/21/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 07:05 Control Test 0.049 07:06 Air Blank 0.000 07:07 Control Test 0.049 07:07 Air Blank 0.000 07:08 Control Test 0.048 07:09 Air Blank 0.000 07:09 Control Test Stats Average 0.0487 Std Dev 0.0006 Rel Std Dev(%) 1.1863	BOCA RATON PD Intoxilyzer - Alcohol Analyzer Model 8000 09/21/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 07:10 Control Test 0.080 07:11 Air Blank 0.000 07:11 Control Test 0.079 07:12 Air Blank 0.000 07:12 Control Test 0.078 07:13 Air Blank 0.000 07:14 Control Test Stats Average 0.0793 Std Dev 0.0011 Rel Std Dev(%) 1.2658	BOCA RATON PD Intoxilyzer - Alcohol Analyzer Model 8000 09/21/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 07:15 Control Test 0.199 07:15 Air Blank 0.000 07:16 Control Test 0.199 07:16 Air Blank 0.000 07:17 Control Test 0.199 07:18 Air Blank 0.000 07:18 Control Test Stats Average 0.1990 Std Dev 0.0000 Rel Std Dev(%) 0.0000	BOCA RATON PD Intoxilyzer - Alcohol Analyzer Model 8000 09/21/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 07:19 Control Test 0.080 07:19 Air Blank 0.000 07:20 Control Test 0.080 07:20 Air Blank 0.000 07:21 Control Test 0.079 07:21 Air Blank 0.000 07:22 Control Test Stats Average 0.0797 Std Dev 0.0006 Rel Std Dev(%) 0.7247
<i>[Signature]</i> Operator's Signature	<i>[Signature]</i> Operator's Signature	<i>[Signature]</i> Operator's Signature	<i>[Signature]</i> Operator's Signature

***** AUTO CAL DATA *****

<<<< CHANNEL 1 >>>>
 Sol Val = 0.0000 mg/l or 0.000 g/210L
 % Abs = 0.076
 Std Dev = 0.03 Rel Std Dev = 33.65
 Sol Val = 0.1905 mg/l or 0.040 g/210L
 % Abs = 0.736
 Std Dev = 0.01 Rel Std Dev = 1.70
 Sol Val = 0.4762 mg/l or 0.100 g/210L
 % Abs = 1.780
 Std Dev = 0.03 Rel Std Dev = 1.48
 Sol Val = 0.9524 mg/l or 0.200 g/210L
 % Abs = 3.412
 Std Dev = 0.02 Rel Std Dev = 0.47
 Sol Val = 1.4286 mg/l or 0.300 g/210L
 % Abs = 4.972
 Std Dev = 0.03 Rel Std Dev = 3.51
 Zero Order Coef = -171.35
 First Order Coef = 2711.53
 Second Order Coef = 39.20
 Standard Deviation = 36.945148

 ***** CHANNEL 2 *****
 Sol Val = 0.0000 mg/l or 0.000 g/210L
 % Abs = 0.255
 Std Dev = 0.01 Rel Std Dev = 5.30
 Sol Val = 0.1905 mg/l or 0.040 g/210L
 % Abs = 1.383
 Std Dev = 0.00 Rel Std Dev = 0.04
 Sol Val = 0.4762 mg/l or 0.100 g/210L
 % Abs = 3.555
 Std Dev = 0.02 Rel Std Dev = 0.60
 Sol Val = 0.9524 mg/l or 0.200 g/210L
 % Abs = 6.593
 Std Dev = 0.01 Rel Std Dev = 0.20
 Sol Val = 1.4286 mg/l or 0.300 g/210L
 % Abs = 9.642
 Std Dev = 0.02 Rel Std Dev = 0.19
 Zero Order Coef = -339.31
 First Order Coef = 1304.78
 Second Order Coef = 13.62
 Standard Deviation = 14.785054

<<<< CHANNEL 2 >>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 6.6970 (-0.0140)
 Sample #2 = 6.7020 (-0.0160)
 Sample #3 = 6.7000 (-0.0100)
 Sample #4 = 6.6760 (-0.0150)
 Avg % Abs = 6.6833 (-0.0037)
 STD DEV = 0.0133 (-0.0164)
 REL STD DEV = 0.199 (448.413)

Sol Value = 0.000 g/210L ***
 Fit value = 0.3810 mg/l %
 Samples Taken = 4, Discarded = 1
 ***** CHANNEL 1 *****
 Sample #1 = 3346.00
 Sample #2 = 3395.00
 Sample #3 = 3353.00
 Sample #4 = 3326.00
 Average Result = 3356.0000
 STD DEV = 31.6070
 REL STD DEV = 0.942

 ***** CHANNEL 2 *****
 Sample #1 = 3156.00
 Sample #2 = 3144.00
 Sample #3 = 3130.00
 Sample #4 = 3113.00
 Average Result = 3129.0000
 STD DEV = 15.5242
 REL STD DEV = 0.496

 Dry Gas H2O Adjust Results *****
 Barometric Pressure = 1011
 3 um H2O Adjust (mg/l*10,000) = 453
 9 um H2O Adjust (mg/l*10,000) = 680
 ***** AUTO CAL PASS *****

Sol Value = 0.310 g/210L ***
 Fit value = 1.4286 mg/l %
 Samples Taken = 4, Discarded = 1
 Sum Io = 12563, Sum Io = 12679
 <<<< CHANNEL 1 >>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 4.9710 (0.0030)
 Sample #2 = 4.9630 (0.0270)
 Sample #3 = 5.1000 (0.0220)
 Sample #4 = 4.5520 (0.0540)
 Avg % Abs = 4.9717 (0.0343)
 STD DEV = 0.0251 (0.0172)
 REL STD DEV = 0.506 (50.139)

<<<< CHANNEL 2 >>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 9.6490 (-0.0070)
 Sample #2 = 9.6450 (0.0060)
 Sample #3 = 9.6580 (0.0030)
 Sample #4 = 9.6220 (0.0230)
 Avg % Abs = 9.6417 (0.0087)
 STD DEV = 0.0182 (0.0125)
 REL STD DEV = 0.89 (144.265)

<<<< CHANNEL 2 >>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 3.5730 (-0.0180)
 Sample #2 = 3.5790 (-0.0170)
 Sample #3 = 3.5380 (-0.0070)
 Sample #4 = 3.5490 (-0.0160)
 Avg % Abs = 3.5553 (-0.0133)
 STD DEV = 0.0212 (0.0055)
 REL STD DEV = 0.597 (41.307)

Sol Value = 0.280 g/210L ***
 Fit value = 0.9524 mg/l %
 Samples Taken = 4, Discarded = 1
 Sum Io = 12571, Sum Io = 12683
 <<<< CHANNEL 1 >>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 3.3960 (0.0100)
 Sample #2 = 3.4180 (0.0100)
 Sample #3 = 3.4240 (0.0200)
 Sample #4 = 3.3940 (0.0500)
 Avg % Abs = 3.4120 (0.0267)
 STD DEV = 0.0159 (0.0208)
 REL STD DEV = 0.465 (78.062)

<<<< CHANNEL 2 >>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 0.7420 (-0.0040)
 Sample #2 = 0.7500 (0.0040)
 Sample #3 = 0.7300 (0.0400)
 Sample #4 = 0.7270 (0.0500)
 Avg % Abs = 0.7357 (0.0343)
 STD DEV = 0.0125 (0.0279)
 REL STD DEV = 1.700 (81.362)

Optical Calibration	
SN:	80-006622
Agency:	Boca Raton PD
Date:	9/21/2022
Quadratic Fit:	+/- 0.002g/210L
By:	DERR <i>[Signature]</i>

Sol Value = 0.040 g/210L ***
 Fit value = 0.1905 mg/l %
 Samples Taken = 4, Discarded = 1
 Sum Io = 12590, Sum Io = 12630
 <<<< CHANNEL 1 >>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 0.0890 (-0.0100)
 Sample #2 = 0.0490 (0.0380)
 Sample #3 = 0.1000 (0.0360)
 Sample #4 = 0.0900 (0.0860)
 Avg % Abs = 0.0763 (0.0533)
 STD DEV = 0.0257 (0.0283)
 REL STD DEV = 33.664 (53.077)

Sol Value = 0.040 g/210L ***
 Fit value = 0.1905 mg/l %
 Samples Taken = 4, Discarded = 1
 Sum Io = 12590, Sum Io = 12630
 <<<< CHANNEL 1 >>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 0.7420 (-0.0040)
 Sample #2 = 0.7500 (0.0040)
 Sample #3 = 0.7300 (0.0400)
 Sample #4 = 0.7270 (0.0500)
 Avg % Abs = 0.7357 (0.0343)
 STD DEV = 0.0125 (0.0279)
 REL STD DEV = 1.700 (81.362)

BOCA RATON PD
 Intoxilyzer - Alcohol Analyzer
 Model 8000
 09/21/2022
 SN 80-006622
 06:22:41
 Auto Calibration
 Max Power Res Value = 35
 Auto Range Res Value = 21

Sol Value = 0.000 g/210L ***
 Fit value = 0.0000 mg/l %
 Samples Taken = 4, Discarded = 1
 Sum Io = 12605, Sum Io = 12701
 <<<< CHANNEL 1 >>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 0.0890 (-0.0100)
 Sample #2 = 0.0490 (0.0380)
 Sample #3 = 0.1000 (0.0360)
 Sample #4 = 0.0900 (0.0860)
 Avg % Abs = 0.0763 (0.0533)
 STD DEV = 0.0257 (0.0283)
 REL STD DEV = 33.664 (53.077)

Sol Value = 0.040 g/210L ***
 Fit value = 0.1905 mg/l %
 Samples Taken = 4, Discarded = 1
 Sum Io = 12590, Sum Io = 12630
 <<<< CHANNEL 1 >>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 0.0890 (-0.0100)
 Sample #2 = 0.0490 (0.0380)
 Sample #3 = 0.1000 (0.0360)
 Sample #4 = 0.0900 (0.0860)
 Avg % Abs = 0.0763 (0.0533)
 STD DEV = 0.0257 (0.0283)
 REL STD DEV = 33.664 (53.077)

Sol Value = 0.040 g/210L ***
 Fit value = 0.1905 mg/l %
 Samples Taken = 4, Discarded = 1
 Sum Io = 12590, Sum Io = 12630
 <<<< CHANNEL 1 >>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 0.0890 (-0.0100)
 Sample #2 = 0.0490 (0.0380)
 Sample #3 = 0.1000 (0.0360)
 Sample #4 = 0.0900 (0.0860)
 Avg % Abs = 0.0763 (0.0533)
 STD DEV = 0.0257 (0.0283)
 REL STD DEV = 33.664 (53.077)

Sol Value = 0.040 g/210L ***
 Fit value = 0.1905 mg/l %
 Samples Taken = 4, Discarded = 1
 Sum Io = 12590, Sum Io = 12630
 <<<< CHANNEL 1 >>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 0.0890 (-0.0100)
 Sample #2 = 0.0490 (0.0380)
 Sample #3 = 0.1000 (0.0360)
 Sample #4 = 0.0900 (0.0860)
 Avg % Abs = 0.0763 (0.0533)
 STD DEV = 0.0257 (0.0283)
 REL STD DEV = 33.664 (53.077)

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-006622	Boca Raton Police Department	9/19/2022	DERR <i>hdu</i>

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
☒	☒	☒	☒
BOCA RATON PD Intoxilyzer - Alcohol Analyzer Model 8000 09/19/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 10:56 Control Test 0.049 10:56 Air Blank 0.000 10:57 Control Test 0.048 10:58 Air Blank 0.000 10:58 Control Test 0.048 10:59 Air Blank 0.000 10:59 Control Test Stats Average 0.0483 Std Dev 0.0006 Rel Std Dev(%) 1.1945 Operator's Signature <i>hdu</i>	BOCA RATON PD Intoxilyzer - Alcohol Analyzer Model 8000 09/19/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 11:00 Control Test 0.077 11:01 Air Blank 0.000 11:02 Control Test 0.077 11:02 Air Blank 0.000 11:03 Control Test 0.077 11:03 Air Blank 0.000 11:04 Control Test Stats Average 0.0770 Std Dev 0.0000 Rel Std Dev(%) 0.0000 Operator's Signature <i>hdu</i>	BOCA RATON PD Intoxilyzer - Alcohol Analyzer Model 8000 09/19/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 11:05 Control Test 0.200 11:06 Air Blank 0.000 11:06 Control Test 0.199 11:07 Air Blank 0.000 11:07 Control Test 0.199 11:08 Air Blank 0.000 11:09 Control Test Stats Average 0.1993 Std Dev 0.0106 Rel Std Dev(%) 0.2896 Operator's Signature <i>hdu</i>	BOCA RATON PD Intoxilyzer - Alcohol Analyzer Model 8000 09/19/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 11:10 Control Test 0.081 11:10 Air Blank 0.000 11:10 Control Test 0.080 11:11 Air Blank 0.000 11:11 Control Test 0.081 11:12 Air Blank 0.000 11:12 Control Test Stats Average 0.0807 Std Dev 0.0006 Rel Std Dev(%) 0.7157 Operator's Signature <i>hdu</i>