



INSTRUMENT PROCESSING SHEET

Agency Manatee CSOS/N 80-006631Florida Department of
Law EnforcementDate In 3/23/2022 DI Completion Date 4/5/2022☐ Ship ☒ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By TDG	Quality Checks	By TDG	Date <u>3/29/2022</u>	Flow Calibration	By	Date																																	
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes: _____ _____ _____ _____ _____ _____ _____ _____ _____ _____		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>213</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP104</u> 32 mm <u>0.152</u> (.139 - .169) 36 mm <u>0.171</u> (.156 - .190) 53 mm <u>0.238</u> (.228 - .278) 103 mm <u>0.523</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>68639</u> ☒ Stability Checks			Flow Column # _____ ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																			
				<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP6286</td><td>202010A 10/05/2022</td></tr><tr><td>0.080</td><td>MP6287</td><td>202010B 10/05/2022</td></tr><tr><td>0.200</td><td>MP6288</td><td>202010D 10/06/2022</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG115904 06/08/2023</td></tr></tbody></table>	Simulator	Serial #	Lot #/Exp	0.050	MP6286	202010A 10/05/2022	0.080	MP6287	202010B 10/05/2022	0.200	MP6288	202010D 10/06/2022	0.080 DGS	N/A	AG115904 06/08/2023	<table border="1"><thead><tr><th>Maintenance</th><th>By</th></tr></thead><tbody><tr><td>☐ Battery Replacement</td><td></td></tr><tr><td>☐ Dry Gas Regulator Replacement</td><td></td></tr><tr><td>☐ Breath Tube Replacement</td><td></td></tr><tr><td>☐ Other _____</td><td></td></tr><tr><td>_____</td><td></td></tr><tr><td>_____</td><td></td></tr><tr><td>_____</td><td></td></tr><tr><td>_____</td><td></td></tr></tbody></table>			Maintenance	By	☐ Battery Replacement		☐ Dry Gas Regulator Replacement		☐ Breath Tube Replacement		☐ Other _____		_____		_____		_____		_____	
Simulator	Serial #	Lot #/Exp																																						
0.050	MP6286	202010A 10/05/2022																																						
0.080	MP6287	202010B 10/05/2022																																						
0.200	MP6288	202010D 10/06/2022																																						
0.080 DGS	N/A	AG115904 06/08/2023																																						
Maintenance	By																																							
☐ Battery Replacement																																								
☐ Dry Gas Regulator Replacement																																								
☐ Breath Tube Replacement																																								
☐ Other _____																																								

Calibration Adjustment By _____				Department Inspection By <u>TDG</u>																																				
Barometric Pressure Gauge _____ ID # _____				Barometric Pressure ID# <u>28663</u>																																				
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>				Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Gauge <u>1014</u> Instrument <u>1013</u> Mouth Alcohol Solution Lot # <u>2021-B</u> Acetone Stock Solution Lot # <u>2021-C</u>								
Simulator	Serial #	Lot #	Expiration																																					
0.000		N/A	N/A																																					
0.040																																								
0.100																																								
0.200																																								
0.300																																								
0.080 DGS	N/A																																							
☐ Post Calibration Adjustment Stability Checks				<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP6284</td></tr><tr><td>Interferent</td><td>MP6285</td></tr><tr><td>0.050</td><td>MP6286</td></tr><tr><td>0.080</td><td>MP6287</td></tr><tr><td>0.200</td><td>MP6288</td></tr></tbody></table>				Simulator	Serial Number	0.000	MP6284	Interferent	MP6285	0.050	MP6286	0.080	MP6287	0.200	MP6288																					
Simulator	Serial Number																																							
0.000	MP6284																																							
Interferent	MP6285																																							
0.050	MP6286																																							
0.080	MP6287																																							
0.200	MP6288																																							
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>				Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Attachments ☒ Form 41 ☐ Post-Stability Checks ☒ Stability Checks ☐ Flow Calibration ☒ Calibration Certificate ☐ Form 40 ☐ Calibration Adjustment ☒ Other <u>Extra Stabilities</u>																
Simulator	Serial #	Lot #	Expiration																																					
0.050																																								
0.080																																								
0.200																																								
0.080 DGS	N/A																																							
Notes/Suggested Service: _____ _____ _____ _____ _____ _____ _____ _____ _____ _____				☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use 2022.04.05 15:52:47 Israel Soto Digitally signed by Israel Soto Date: 2022.04.05 13:18:49 +0400 Tech Review / Date _____ Admin Review / Date _____																																				

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-00 6631	Manatee CSO	03/29/2022	TDG <i>ML</i>

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083
✓	✓	✓	✓
MANATEE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006631 03/29/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 14:17 Control Test 0.049 14:18 Air Blank 0.000 14:19 Control Test 0.049 14:19 Air Blank 0.000 14:20 Control Test 0.049 14:21 Air Blank 0.000 14:21 Control Test Stats Average 0.0490 Std Dev 0.0000 Rel Std Dev(%) 0.0000	MANATEE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006631 03/29/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 14:34 Control Test 0.078 14:35 Air Blank 0.000 14:35 Control Test 0.078 14:36 Air Blank 0.000 14:37 Control Test 0.078 14:37 Air Blank 0.000 14:38 Control Test Stats Average 0.0780 Std Dev 0.0000 Rel Std Dev(%) 0.0000	MANATEE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006631 03/29/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 14:58 Control Test 0.196 14:59 Air Blank 0.000 15:00 Control Test 0.196 15:00 Air Blank 0.000 15:01 Control Test 0.195 15:02 Air Blank 0.000 15:02 Control Test Stats Average 0.1957 Std Dev 0.0006 Rel Std Dev(%) 0.2951	MANATEE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006631 03/29/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 14:10 Control Test 0.078 14:11 Air Blank 0.000 14:11 Control Test 0.079 14:12 Air Blank 0.000 14:12 Control Test 0.078 14:12 Air Blank 0.000 14:13 Control Test Stats Average 0.0783 Std Dev 0.0006 Rel Std Dev(%) 0.7370
<i>ML</i> Operator's Signature	<i>ML</i> Operator's Signature	<i>ML</i> Operator's Signature	<i>ML</i> Operator's Signature

Comments:

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-006631	Manatee CSO	03/29/2022	TDG <i>ML</i>

ML

0.05g/210L	0.08g/210L #2	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083
	✓		≤0.003 of Wet

MANATEE COUNTY SO
Intoxilizer - Alcohohol Analyzer
Model 8000 SN 80-006631
03/29/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:52
Control Test	0.078	14:53
Air Blank	0.000	14:53
Control Test	0.079	14:54
Air Blank	0.000	14:55
Control Test	0.078	14:55
Air Blank	0.000	14:56
Control Test Stats		
Average	0.0783	
Std Dev	0.0006	
Rel Std Dev(%)	0.7370	

ML

Operator's Signature

Comments: Inadvertently ran an extra Stability Test on the 0.08 ARS. All results within acceptable range. *ML* 03/29/2022

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: MANATEE COUNTY SO
Time of Inspection: 12:10

Date of Inspection: 04/05/2022

Serial Number: 80-006631
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG115904 Exp: 06/08/2023
0.000	0.048	0.078	0.197	0.078
0.000	0.048	0.078	0.197	0.078
0.000	0.048	0.078	0.197	0.078
0.000	0.049	0.078	0.197	0.078
0.000	0.049	0.078	0.197	0.078
0.000	0.049	0.078	0.197	0.078
0.000	0.049	0.078	0.197	0.077
0.000	0.049	0.078	0.197	0.078
0.000	0.048	0.078	0.197	0.078
0.000	0.049	0.078	0.197	0.078


Standard Deviations	0.0005	0.0000	0.0000	0.0003
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0002 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

04/05/2022
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-006631, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	80-006631	UNCERTAINTY* $\pm$
Owning Agency:	MANATEE COUNTY SO	0.050 g/ 210 L 0.004
Calibration Date:	04/05/2022	0.080 g/ 210 L 0.004
Calibration Time:	12:10	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.



04/05/2022

Date

TAYLOR D GUTSCHOW,

Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality