

# INSTRUMENT PROCESSING SHEET

Agency Holmes Beach Police Department

S/N 80-006692

Florida Department of  
Law Enforcement

Date In 10/31/2022 DI Completion Date 11/01/2022

☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By <u>DERR</u>	Quality Checks	By <u>DERR</u> Date <u>11/01/2022</u>																																																												
<input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE  Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight  Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable  Notes: _____ _____ _____ _____ _____ _____ _____ _____ _____ _____	<input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>157</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP101</u> 32 mm <u>0.156</u> (.139 - .169) 36 mm <u>0.167</u> (.156 - .190) 53 mm <u>0.234</u> (.228 - .278) 103 mm <u>0.500</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28199</u> <input checked="" type="checkbox"/> Stability Checks	Flow Column # _____ <input type="checkbox"/> 5L/min – 17mm <input type="checkbox"/> 15L/min – 53mm <input type="checkbox"/> 30L/min – 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)	Maintenance By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ _____ _____ _____ _____ _____																																																												
<b>Calibration Adjustment</b> By _____		<b>Department Inspection</b> By <u>DERR</u>																																																													
Barometric Pressure Gauge _____ ID # _____ <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> <input type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>		Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Barometric Pressure ID# <u>26932</u> Gauge <u>1016</u> Instrument <u>1016</u> Mouth Alcohol Solution Lot # <u>2021-D</u> Acetone Stock Solution Lot # <u>2021-C</u> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP6284</td> </tr> <tr> <td>Interferent</td> <td>MP6285</td> </tr> <tr> <td>0.050</td> <td>MP6286</td> </tr> <tr> <td>0.080</td> <td>MP6287</td> </tr> <tr> <td>0.200</td> <td>MP6288</td> </tr> </tbody> </table>		Simulator	Serial Number	0.000	MP6284	Interferent	MP6285	0.050	MP6286	0.080	MP6287	0.200	MP6288
Simulator	Serial #	Lot #	Expiration																																																												
0.000		N/A	N/A																																																												
0.040																																																															
0.100																																																															
0.200																																																															
0.300																																																															
0.080 DGS	N/A																																																														
Simulator	Serial #	Lot #	Expiration																																																												
0.050																																																															
0.080																																																															
0.200																																																															
0.080 DGS	N/A																																																														
Simulator	Serial Number																																																														
0.000	MP6284																																																														
Interferent	MP6285																																																														
0.050	MP6286																																																														
0.080	MP6287																																																														
0.200	MP6288																																																														
Notes/Suggested Service: _____ _____ _____ _____ _____ _____ _____ _____ _____		<b>Attachments</b> <div style="display: flex; justify-content: space-between;"> <div> <input checked="" type="checkbox"/> Form 41  <input checked="" type="checkbox"/> Stability Checks  <input checked="" type="checkbox"/> Calibration Certificate  <input type="checkbox"/> Calibration Adjustment             </div> <div> <input type="checkbox"/> Post-Stability Checks  <input type="checkbox"/> Flow Calibration  <input type="checkbox"/> Form 40  <input type="checkbox"/> Other _____             </div> </div>																																																													
		<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use																																																													
Phil Nicodemo <small>Digitally signed by Phil Nicodemo Date: 2022.11.01 15:08:12 -04'00'</small>		Israel Soto <small>Digitally signed by Israel Soto Date: 2022.11.02 09:44:26 -04'00'</small>																																																													
Tech Review / Date		Admin Review / Date																																																													





## Calibration Certificate

Florida Department of Law Enforcement  
Alcohol Testing Program  
4700 Terminal Drive, Suite 1  
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-006692, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number: 80-006692

Owning Agency: HOLMES BEACH PD

Calibration Date: 11/01/2022

Calibration Time: 08:52

### UNCERTAINTY\* $\pm$

0.050 g/ 210 L	0.004
0.080 g/ 210 L	0.004
0.200 g/ 210 L	0.007
0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within  $\pm 0.005$  or 5%, whichever is greater, of the target alcohol concentration.

\*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ( $k=3$ ).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

### TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

11/01/2022

Date

  
DAVID E REYES-RIVERA,  
Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Page 1 of 1

# Florida Department of Law Enforcement Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: HOLMES BEACH PD  
Time of Inspection: 08:52

Date of Inspection: 11/01/2022

Serial Number: 80-006692  
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:00521080A2 Exp: 02/05/2023
0.000	0.047	0.077	0.197	0.077
0.000	0.047	0.077	0.198	0.077
0.000	0.048	0.077	0.198	0.078
0.000	0.047	0.077	0.197	0.077
0.000	0.048	0.077	0.198	0.077
0.000	0.048	0.077	0.198	0.078
0.000	0.048	0.077	0.198	0.077
0.000	0.048	0.077	0.198	0.077
0.000	0.048	0.077	0.198	0.077
0.000	0.048	0.077	0.198	0.077
0.000	0.048	0.077	0.198	0.077


Standard Deviations	0.0004	0.0000	0.0004	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies ( ☒ ) does not comply ( ☐ ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 \_\_\_\_\_  
Signature and Printed Name

DAVID E REYES-RIVERA

11/01/2022  
Date



Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-006692	Holmes Beach Police Department	11/01/2022	DERR <i>[Signature]</i>

0.05g/210L 0.047 to 0.053 <input checked="" type="checkbox"/>	0.08g/210L 0.077 to 0.083 <input checked="" type="checkbox"/>	0.20g/210L 0.194 to 0.206 <input checked="" type="checkbox"/>	DGS 0.08g/210L 0.077 to 0.083 <input checked="" type="checkbox"/>																																																																																																																																																
<p>HOLMES BEACH PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006692 11/01/2022 Software: 8100.27</p> <table> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>06:50</td></tr> <tr><td>Control Test</td><td>0.048</td><td>06:51</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:51</td></tr> <tr><td>Control Test</td><td>0.048</td><td>06:52</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:52</td></tr> <tr><td>Control Test</td><td>0.048</td><td>06:53</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:54</td></tr> <tr><td colspan="3">Control Test Stats</td></tr> <tr><td>Average</td><td>0.0488</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> <p><i>[Signature]</i> Operator's Signature</p>	Test	g/210L	Time	Air Blank	0.000	06:50	Control Test	0.048	06:51	Air Blank	0.000	06:51	Control Test	0.048	06:52	Air Blank	0.000	06:52	Control Test	0.048	06:53	Air Blank	0.000	06:54	Control Test Stats			Average	0.0488		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<p>HOLMES BEACH PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006692 11/01/2022 Software: 8100.27</p> <table> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>06:55</td></tr> <tr><td>Control Test</td><td>0.078</td><td>06:56</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:56</td></tr> <tr><td>Control Test</td><td>0.078</td><td>06:57</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:57</td></tr> <tr><td>Control Test</td><td>0.077</td><td>06:58</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:59</td></tr> <tr><td colspan="3">Control Test Stats</td></tr> <tr><td>Average</td><td>0.0777</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7434</td><td></td></tr> </tbody> </table> <p><i>[Signature]</i> Operator's Signature</p>	Test	g/210L	Time	Air Blank	0.000	06:55	Control Test	0.078	06:56	Air Blank	0.000	06:56	Control Test	0.078	06:57	Air Blank	0.000	06:57	Control Test	0.077	06:58	Air Blank	0.000	06:59	Control Test Stats			Average	0.0777		Std Dev	0.0006		Rel Std Dev(%)	0.7434		<p>HOLMES BEACH PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006692 11/01/2022 Software: 8100.27</p> <table> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>07:00</td></tr> <tr><td>Control Test</td><td>0.197</td><td>07:00</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:01</td></tr> <tr><td>Control Test</td><td>0.198</td><td>07:02</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:02</td></tr> <tr><td>Control Test</td><td>0.198</td><td>07:03</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:03</td></tr> <tr><td colspan="3">Control Test Stats</td></tr> <tr><td>Average</td><td>0.1977</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.2921</td><td></td></tr> </tbody> </table> <p><i>[Signature]</i> Operator's Signature</p>	Test	g/210L	Time	Air Blank	0.000	07:00	Control Test	0.197	07:00	Air Blank	0.000	07:01	Control Test	0.198	07:02	Air Blank	0.000	07:02	Control Test	0.198	07:03	Air Blank	0.000	07:03	Control Test Stats			Average	0.1977		Std Dev	0.0006		Rel Std Dev(%)	0.2921		<p>HOLMES BEACH PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006692 11/01/2022 Software: 8100.27</p> <table> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>07:08</td></tr> <tr><td>Control Test</td><td>0.077</td><td>07:08</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:09</td></tr> <tr><td>Control Test</td><td>0.077</td><td>07:09</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:09</td></tr> <tr><td>Control Test</td><td>0.077</td><td>07:10</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:10</td></tr> <tr><td colspan="3">Control Test Stats</td></tr> <tr><td>Average</td><td>0.0770</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> <p><i>[Signature]</i> Operator's Signature</p>	Test	g/210L	Time	Air Blank	0.000	07:08	Control Test	0.077	07:08	Air Blank	0.000	07:09	Control Test	0.077	07:09	Air Blank	0.000	07:09	Control Test	0.077	07:10	Air Blank	0.000	07:10	Control Test Stats			Average	0.0770		Std Dev	0.0000		Rel Std Dev(%)	0.0000	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	06:50																																																																																																																																																	
Control Test	0.048	06:51																																																																																																																																																	
Air Blank	0.000	06:51																																																																																																																																																	
Control Test	0.048	06:52																																																																																																																																																	
Air Blank	0.000	06:52																																																																																																																																																	
Control Test	0.048	06:53																																																																																																																																																	
Air Blank	0.000	06:54																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0488																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	06:55																																																																																																																																																	
Control Test	0.078	06:56																																																																																																																																																	
Air Blank	0.000	06:56																																																																																																																																																	
Control Test	0.078	06:57																																																																																																																																																	
Air Blank	0.000	06:57																																																																																																																																																	
Control Test	0.077	06:58																																																																																																																																																	
Air Blank	0.000	06:59																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0777																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7434																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	07:00																																																																																																																																																	
Control Test	0.197	07:00																																																																																																																																																	
Air Blank	0.000	07:01																																																																																																																																																	
Control Test	0.198	07:02																																																																																																																																																	
Air Blank	0.000	07:02																																																																																																																																																	
Control Test	0.198	07:03																																																																																																																																																	
Air Blank	0.000	07:03																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1977																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.2921																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	07:08																																																																																																																																																	
Control Test	0.077	07:08																																																																																																																																																	
Air Blank	0.000	07:09																																																																																																																																																	
Control Test	0.077	07:09																																																																																																																																																	
Air Blank	0.000	07:09																																																																																																																																																	
Control Test	0.077	07:10																																																																																																																																																	
Air Blank	0.000	07:10																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0770																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		