



INSTRUMENT PROCESSING SHEET

Agency FHP Troop FS/N 80-006761Florida Department of
Law EnforcementDate In 8/18/2022 DI Completion Date 8/30/2022☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By TDG	Quality Checks	By TDG	Date 8/29/2022	Flow Calibration	By TDG	Date 8/29/2022																												
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: _____ _____ _____ _____ _____ _____ _____ _____ _____ _____		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>187</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP104</u> 32 mm <u>0.148</u> (.139 - .169) 36 mm <u>0.160</u> (.156 - .190) 53 mm <u>0.226*</u> (.228 - .278) 103 mm <u>0.472</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>68639</u> ☒ Stability Checks			Flow Column # <u>ATP106</u> ☒ 5L/min - 17mm ☒ 15L/min - 53mm ☒ 30L/min - 103mm ☒ R-Value <u>187</u> ☒ Post Calibration Verification (L/s) Flow Column # <u>ATP104</u> 32 mm <u>0.148</u> (.139 - .169) 36 mm <u>0.164</u> (.156 - .190) 53 mm <u>0.234</u> (.228 - .278) 103 mm <u>0.500</u> (.447 - .547)																														
					Maintenance By _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____ _____ _____ _____ _____ _____																														
Calibration Adjustment By _____				Department Inspection By <u>TDG</u>																															
Barometric Pressure Gauge ID # _____				Barometric Pressure ID# <u>28199</u>																															
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>				Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Gauge <u>1015</u> Instrument <u>1012</u> Mouth Alcohol Solution Lot # <u>2021-D</u> Acetone Stock Solution Lot # <u>2021-C</u>			
Simulator	Serial #	Lot #	Expiration																																
0.000		N/A	N/A																																
0.040																																			
0.100																																			
0.200																																			
0.300																																			
0.080 DGS	N/A																																		
☐ Post Calibration Adjustment Stability Checks				<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP5095</td></tr><tr><td>Interferent</td><td>SD3966</td></tr><tr><td>0.050</td><td>MP5092</td></tr><tr><td>0.080</td><td>MP5093</td></tr><tr><td>0.200</td><td>MP5094</td></tr></tbody></table>				Simulator	Serial Number	0.000	MP5095	Interferent	SD3966	0.050	MP5092	0.080	MP5093	0.200	MP5094																
Simulator	Serial Number																																		
0.000	MP5095																																		
Interferent	SD3966																																		
0.050	MP5092																																		
0.080	MP5093																																		
0.200	MP5094																																		
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>				Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Attachments ☒ Form 41 ☐ Post-Stability Checks ☒ Stability Checks ☒ Flow Calibration ☒ Calibration Certificate ☐ Form 40 ☐ Calibration Adjustment ☐ Other _____											
Simulator	Serial #	Lot #	Expiration																																
0.050																																			
0.080																																			
0.200																																			
0.080 DGS	N/A																																		
Notes/Suggested Service: *Flow value outside acceptable range. Will perform a flow cal. (TDG) <u>On 8/29, instrument froze up during a reference reading during the Alcohol Free Test. Instrument simply stopped in the middle of drawing a sample. It would not continue the DI or advance to the next screen, and neither would it abort. Had to power off. Gauge 28199 read 1014 psi, and instrument read 1011 psi. Used Mouth Alcohol Lot 2021-D and Acetone Lot 2021-C. (TDG)</u>				☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																															
Israel Soto Digitally signed by Israel Soto Date: 2022.08.31 16:53:23 +0400				Date: 2022.08.31 12:58:00 -04'00'																															
Tech Review / Date				Admin Review / Date																															

80-00 6761

Flow Cal

8/29/2022

MG

FL HIGHWAY PATROL

Intoxilyzer - Alcohol Analyzer

Model 8000

SN 80-006761

08/29/2022

Software: 8100.27

Flow Rate Calibration*****

1: Rate (Liters/min) = 5

SQRT(Diff) = 7.000

2: Rate (Liters/min) = 15

SQRT(Diff) = 11.832

3: Rate (Liters/min) = 30

SQRT(Diff) = 21.047

Dependent Data Scale Factor = 100000 L/min

Independent Data Scale Factor = 256

Rounded Slope = 687

Rounded Intercept = -672460

Correlation = 0.99800

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-00 6761	FTP	08/29/2022	TDG ML

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083
✓	✓	✓	✓
FL HIGHWAY PATROL Intoxilyzer - Alcohol Analyzer Model 8000 08/29/2022 Software: 8100.27 SN 80-006761 Test g/210L Time Air Blank 0.000 13:30 Control Test 0.048 13:31 Air Blank 0.000 13:31 Control Test 0.048 13:32 Air Blank 0.000 13:32 Control Test 0.048 13:33 Air Blank 0.000 13:34 Control Test Stats Average 0.0480 Std Dev 0.0000 Rel Std Dev(%) 0.0000	FL HIGHWAY PATROL Intoxilyzer - Alcohol Analyzer Model 8000 08/29/2022 Software: 8100.27 SN 80-006761 Test g/210L Time Air Blank 0.000 13:37 Control Test 0.079 13:38 Air Blank 0.000 13:38 Control Test 0.078 13:39 Air Blank 0.000 13:39 Control Test 0.078 13:40 Air Blank 0.000 13:40 Control Test Stats Average 0.0783 Std Dev 0.0006 Rel Std Dev(%) 0.7370	FL HIGHWAY PATROL Intoxilyzer - Alcohol Analyzer Model 8000 08/29/2022 Software: 8100.27 SN 80-006761 Test g/210L Time Air Blank 0.000 13:45 Control Test 0.201 13:46 Air Blank 0.000 13:46 Control Test 0.200 13:47 Air Blank 0.000 13:47 Control Test 0.200 13:48 Air Blank 0.000 13:49 Control Test Stats Average 0.2003 Std Dev 0.0006 Rel Std Dev(%) 0.2882	FL HIGHWAY PATROL Intoxilyzer - Alcohol Analyzer Model 8000 08/29/2022 Software: 8100.27 SN 80-006761 Test g/210L Time Air Blank 0.000 13:24 Control Test 0.078 13:25 Air Blank 0.000 13:25 Control Test 0.078 13:25 Air Blank 0.000 13:26 Control Test 0.079 13:26 Air Blank 0.000 13:27 Control Test Stats Average 0.0783 Std Dev 0.0006 Rel Std Dev(%) 0.7370
ML Operator's Signature	ML Operator's Signature	ML Operator's Signature	ML Operator's Signature

Comments:

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL
Time of Inspection: 11:18

Date of Inspection: 08/30/2022

Serial Number: 80-006761
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:00521080A2 Exp: 02/05/2023
0.000	0.049	0.079	0.200	0.078
0.000	0.049	0.078	0.200	0.078
0.000	0.049	0.079	0.200	0.078
0.000	0.049	0.079	0.200	0.079
0.000	0.049	0.078	0.200	0.078
0.000	0.050	0.078	0.200	0.078
0.000	0.050	0.078	0.200	0.078
0.000	0.050	0.079	0.201	0.079
0.000	0.050	0.079	0.200	0.078
0.000	0.050	0.079	0.201	0.078

Standard Deviations	0.0005	0.0005	0.0004	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

08/30/2022
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-006761, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-006761</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>FL HIGHWAY PATROL</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>08/30/2022</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>11:18</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

08/30/2022

Date

Taylor D Gutschow,
Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality