

Return Material Authorization

Ship to:



CMI, Inc.



Enforcement Electronics

Shipment to repair facility authorized by: Captain Strickland on 1/10/2022

Items Returned: Instrument ☒ Supplies ☐ Other ☐ Describe: _____

Instrument Model: Intoxilyzerv 8000 Serial Number: 80-006763

Bill to Address: Florida Highway Patrol Troop L

ATTN: Captain Strickland

Lake Worth Svc. Plz. MM94

Law Enforcement Bldg. #9330

Lake Worth, FL 33467

Ship to address: FDLE FMROC

4700 Terminal Drive Suite 1

Fort Myers, Florida 33991

Reason for Return: I received 80-006763 (Florida Highway Patrol Troop L). The instrument is not working properly. During the stabilities testing the instrument had test exemptions (interferent detect, ambient fail) on all of the values. DVM also shows a great spike on the 3 filter, as high as S: 54.

Please choose one of the following options:

- ☐ 1. I _____, authorize all repairs.
- ☐ 2. I _____, authorize repairs up to \$_____.
- ☒ 3. I require an estimate **BEFORE** any repairs will be authorized and/ or conducted.

Please contact: Name: Captain Strickland

Phone #: 561-357-4040 Email: kevinstrickland@flhsmv.gov

ATP Contact Name: David Reyes ATP Email: davidreyes@fdle.state.fl.us

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: Florida Highway Patrol
Time of Inspection: n/a

Date of Inspection: 2022

Serial Number: 80-006763
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK			Date and/or Time Adjusted		
Minimum Sample Volume Check: OK			Barometric Pressure Sensor Check: OK		
Alcohol Free Subject Test: 0.000			Mouth Alcohol Test: Slope Not Met		
Interferent Detect Test: Interferent Detect			Diagnostic Check (Post-Inspection): OK		

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Standard Deviations				
---------------------	--	--	--	--

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: _____ Number of Simulators Used: _____

Remarks:

Instrument did not receive a Department Inspection for 2022. It was sent to repair in January 2022 and did not return within the calendar year.

The above instrument complies (☐) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Taylor Gutschow

Digitally signed by Taylor Gutschow
Date: 2022.12.29 15:04:39 -05'00'

Taylor D. Gutschow

Signature and Printed Name

12/29/2022
Date