



Agency FHP Troop G

S/N 80-006765

Date In 02/11/2022 DI Completion Date 02/14/2022

☐ Ship ☒ P/U ☐ H/D ☐ CMI ☐ EE

Calibration Adjustment

By

☐ Post Calibration Adjustment Stability Checks

Notes/Suggested Service:

Tech Review: Updated check boxes and .08 ARS lot numbers. RAW.02/15/22.

Department Inspection

By RAW

Attachments

■ Instrument Complies with Chapter 11D-8, FAC

☐ Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use

☐ **Remain Out of Evidentiary Use**

■ Conduct an Agency Inspection Before Evidentiary Use

2022.02.15

Taylor Gutschow Digitally signed by Taylor Gutschow
Date: 2022.02.15 14:54:32 -0500

Tech Review / Date

Admin Review / Date

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FHP TROOP G

Time of Inspection: 13:38

Date of Inspection: 02/14/2022

Serial Number: 80-006765

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test:			Diagnostic Check		
Interferent Detect	Yes		(Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202010A Exp: 10/05/2022	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202010D Exp: 10/06/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG113403 Exp: 05/14/2023
0.000	0.048	0.079	0.201	0.080
0.000	0.049	0.079	0.201	0.081
0.000	0.049	0.079	0.201	0.080
0.000	0.048	0.079	0.201	0.080
0.000	0.049	0.079	0.201	0.080
0.000	0.049	0.079	0.201	0.080
0.000	0.049	0.080	0.201	0.080
0.000	0.048	0.080	0.201	0.080
0.000	0.049	0.079	0.201	0.080
0.000	0.049	0.079	0.201	0.080


Standard Deviations	0.0004	0.0004	0.0000	0.0003
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0002 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 RICHARD A WILLIAMS
Signature and Printed Name

02/14/2022
Date

FHP TROOP G
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006765
02/14/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:03
Control Test	0.049	10:04
Air Blank	0.000	10:04
Control Test	0.048	10:05
Air Blank	0.000	10:05
Control Test	0.049	10:06
Air Blank	0.000	10:07
Control Test Stats		
Average	0.0487	
Std Dev	0.0006	
Rel Std Dev(%)	1.1863	

.05

PAW

Operator's Signature

FHP TROOP G
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006765
02/14/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:09
Control Test	0.079	10:10
Air Blank	0.000	10:10
Control Test	0.079	10:11
Air Blank	0.000	10:11
Control Test	0.079	10:12
Air Blank	0.000	10:13
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

.08
wct

PAW

Operator's Signature

FHP TROOP G
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006765
02/14/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:14
Control Test	0.200	10:15
Air Blank	0.000	10:15
Control Test	0.200	10:16
Air Blank	0.000	10:16
Control Test	0.200	10:17
Air Blank	0.000	10:18
Control Test Stats		
Average	0.2000	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

.20

PAW

Operator's Signature

FHP TROOP G
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006765
02/14/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:21
Control Test	0.081	10:21
Air Blank	0.000	10:22
Control Test	0.081	10:22
Air Blank	0.000	10:23
Control Test	0.081	10:23
Air Blank	0.000	10:23
Control Test Stats		
Average	0.0810	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

.08
DGS

PAW

Operator's Signature

Stability Check
80-006765
02/14/22
PAW



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2729 Fort Knox Blvd.
Bldg. 2, Suite 1300
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-006765, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-006765</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>FHP TROOP G</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>02/14/2022</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>13:38</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Richard A
Williams
Digitally signed by Richard
A Williams
Date: 2022.02.14 16:02:34
-05'00'

02/14/2022

Date

RICHARD A WILLIAMS,
Department Inspector

FDLE/ATP Form 69 January 2022

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality



INSTRUMENT PROCESSING SHEET


Agency FHP Troop GS/N 80-006765Florida Department of
Law EnforcementDate In 2/19/2022DI Completion Date 02-23-2022☐ Ship ☒ P/U ☐ H/D ☐ CMI ☐ EE

Intake By <u>TJG</u> ☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes: Trooper reported he received a control outside tolerance during a breath test.	Quality Checks By <u>IS</u> Date <u>02-23-2022</u> ☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>148</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP-102</u> 32 mm <u>0.156</u> (.139 - .169) 36 mm <u>0.171</u> (.156 - .190) 53 mm <u>0.242</u> (.228 - .278) 103 mm <u>0.496</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>26932</u> ☒ Stability Checks	Flow Calibration By _____ Date _____ Flow Column # _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																																											
	<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP5088</td> <td>202201C 01-11-2024</td> </tr> <tr> <td>0.080</td> <td>MP5089</td> <td>202201D 01-18-2024</td> </tr> <tr> <td>0.200</td> <td>MP5090</td> <td>202010D 10-06-2022</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG113403 05-14-2023</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	MP5088	202201C 01-11-2024	0.080	MP5089	202201D 01-18-2024	0.200	MP5090	202010D 10-06-2022	0.080 DGS	N/A	AG113403 05-14-2023	Maintenance By _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____																																												
Simulator	Serial #	Lot #/Exp																																																											
0.050	MP5088	202201C 01-11-2024																																																											
0.080	MP5089	202201D 01-18-2024																																																											
0.200	MP5090	202010D 10-06-2022																																																											
0.080 DGS	N/A	AG113403 05-14-2023																																																											
Calibration Adjustment By _____ Barometric Pressure Gauge _____ ID # _____ <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> ☐ Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>	Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Department Inspection By <u>IS</u> Barometric Pressure ID# <u>28421</u> Gauge <u>1025</u> Instrument <u>1024</u> Mouth Alcohol Solution Lot # <u>2021-D</u> Acetone Stock Solution Lot # <u>2021-C</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP5086</td> </tr> <tr> <td>Interferent</td> <td>MP5087</td> </tr> <tr> <td>0.050</td> <td>MP5088</td> </tr> <tr> <td>0.080</td> <td>MP5089</td> </tr> <tr> <td>0.200</td> <td>MP5090</td> </tr> </tbody> </table>	Simulator	Serial Number	0.000	MP5086	Interferent	MP5087	0.050	MP5088	0.080	MP5089	0.200	MP5090
Simulator	Serial #	Lot #	Expiration																																																										
0.000		N/A	N/A																																																										
0.040																																																													
0.100																																																													
0.200																																																													
0.300																																																													
0.080 DGS	N/A																																																												
Simulator	Serial #	Lot #	Expiration																																																										
0.050																																																													
0.080																																																													
0.200																																																													
0.080 DGS	N/A																																																												
Simulator	Serial Number																																																												
0.000	MP5086																																																												
Interferent	MP5087																																																												
0.050	MP5088																																																												
0.080	MP5089																																																												
0.200	MP5090																																																												
Notes/Suggested Service: I logged into the instrument to verify the instrument setup. Trooper Farley then conducted a breath test using a sample ID. The instrument produced a control outside tolerance. While performing the control test, the instrument was taking a sample through the breath tube. This produced a control test value of 0.000. See attached paperwork. Changed setup to have control test be dry gas. TJG	Attachments ☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Calibration Adjustment ☐ Post-Stability Checks ☐ Flow Calibration ☐ Form 40 ☒ Other <u>Breath Test</u>																																																												
	☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																																																												
	David Eliezer Reyes-Rivera Tech Review / Date _____ 01 10:38:55 Admin Review / Date _____																																																												

FHP Troop G

80-006765

2/19/22
JE



FHP TROOP G
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006765
02/19/2022
Software: 8100.27

LA1 Date = 02/18/2022
Obs Period Began = 16:20
Sub Name =
SAMPLE
D
LICENSE
Sub DOB = 02/14/1965
Sub Sex = F
Drii Lic = L252784655541/FL
Officer =
JOSEPH
FARLEY
Operator =
JOSEPH, E
FARLEY

Test	g/210L	Time

Diagnostics Check OK		16:40
Air Blank	0.000	16:41
Control Test	0.000*	16:41
Air Blank	0.000	16:42

*Control Outside Tolerance

Operator's Signature

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FHP TROOP G
Time of Inspection: 13:06

Date of Inspection: 02/23/2022

Serial Number: 80-006765
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202010D Exp: 10/06/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG113403 Exp: 05/14/2023
0.000	0.048	0.078	0.199	0.080
0.000	0.048	0.078	0.200	0.079
0.000	0.048	0.078	0.199	0.079
0.000	0.048	0.078	0.199	0.080
0.000	0.048	0.078	0.200	0.079
0.000	0.048	0.078	0.199	0.079
0.000	0.048	0.078	0.200	0.079
0.000	0.048	0.078	0.199	0.079
0.000	0.048	0.078	0.199	0.079
0.000	0.048	0.078	0.199	0.079


Standard Deviations	0.0000	0.0000	0.0004	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0002 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



ISRAEL SOTO

Signature and Printed Name

02/23/2022
Date

Stability Checks

FHP TROOP G
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006765
02/23/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:29
Control Test	0.048	10:30
Air Blank	0.000	10:30
Control Test	0.048	10:31
Air Blank	0.000	10:31
Control Test	0.047	10:32
Air Blank	0.000	10:33
Control Test Stats		
Average	0.0477	
Std Dev	0.0006	
Rel Std Dev(%)	1.2112	


Operator's Signature

FHP TROOP G
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006765
02/23/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:34
Control Test	0.078	10:35
Air Blank	0.000	10:36
Control Test	0.077	10:36
Air Blank	0.000	10:37
Control Test	0.078	10:38
Air Blank	0.000	10:38
Control Test Stats		
Average	0.0777	
Std Dev	0.0006	
Rel Std Dev(%)	0.7434	

wet


Operator's Signature

FHP TROOP G
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006765
02/23/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:42
Control Test	0.200	10:42
Air Blank	0.000	10:43
Control Test	0.200	10:44
Air Blank	0.000	10:44
Control Test	0.199	10:45
Air Blank	0.000	10:45
Control Test Stats		
Average	0.1997	
Std Dev	0.0006	
Rel Std Dev(%)	0.2892	


Operator's Signature

FHP TROOP G
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006765
02/23/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:47
Control Test	0.080	10:47
Air Blank	0.000	10:48
Control Test	0.079	10:48
Air Blank	0.000	10:49
Control Test	0.080	10:49
Air Blank	0.000	10:49
Control Test Stats		
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	

DRY


Operator's Signature



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2729 Fort Knox Blvd.
Bldg. 2, Suite 1300
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-006765, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-006765</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>FHP TROOP G</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>02/23/2022</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>13:06</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Israel Soto
Soto
Digitally signed by Israel
Date: 2022.02.23
14:25:40 -05'00'

02/23/2022

Date

ISRAEL SOTO,
Department Inspector

FDLE/ATP Form 69 January 2022

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality