

INSTRUMENT PROCESSING SHEET

Agency Florida Highway Patrol Troop L

S/N <u>80-006</u>	772
-------------------	-----

Florida Department of Law Enforcement

**								
	Date In 11/7/2022	DI Completion Date	11/8/2022	Ship	□P/U	□H/D	□смі	D EF

Intake		DERR	Ouglian	Ch	les Desper	9	1 4410/0000	-1 - 1			
Annual	Dy_		Quality			,L	Date 11/8/2022		oration By		
Registrat	ion		■ Breath Tube Screen ■ Replace External O-Rings					Flow Column #			
	rom CMI / EE							□ 5L/min – 17mm			
- Neturn	strument Set Up Verified			☐ 15L/min – 53mm							
Visual Inspe	ection:	R-Valu			-1		☐ 30L/min – 103mm				
Case	Handle				ication (L/s)			R-Value		And the Assessment Assessment	
Keyboard		# ATP101			☐ Post Calibration Verification (L/s)						
Feet	Breath Tul	be	32 mm	1 <u>0.</u>	156	(.139169)		Flow Colu	mn #	***************************************	
Ports	Screws Tig	tht	36 mm	1 <u>0.</u> 1	171	(.156190)	32 mm		(.139169)	
Other Fauir	ment/ Accessories		53 mm	0.2	238	(.228278)		36 mm	Part of the second seco	(.156190)	
	ord Printer Cal		103 mn	n <u>0.8</u>	507	(.447547)	53 mm		(.228278)	
	g 2 12V DC Ca				c Pressure C	heck		103 mm	-	(.447547)	
			Gauge ID	# 28	8199			7			
Notes:		-	Stabili	ity Cl	necks						
			Simulate	or	Serial #	Lot #/	′Ехр	Maintena	nce	Bv	
			0.050			20	2201C		Replacement		
F					MP6286		01/11/2024		Regulator Repl	acement	
			0.080				2201D		Tube Replaceme		
-			0.000		MP6287	01/18/2024		Other_			
			0.200				2201E				
			0.200		MP6288						
						01/	18/2024				
			0.080 D	GS	N/A	0052	21080A2				
20720000000000				MINERALAGE	THE RESERVE OF THE PARTY OF THE	02/0	05/2023				
Calibration				В	У	Departi	ment Inspect	tion		By DERR	
	Pressure Gauge	7	ID #			Barome	etric Pressure	ID# 26932	2		
Simulator	Serial #	Lot#		Ex	piration				strument <u>1018</u>		
0.000	527		N/A		N/A	Mouth.	Alcohol Solut	tion Lot # 2	021-D		
0.040				T.		Acetone	e Stock Soluti	ion Lot # 2	021-C		
0.100						Simula	ator		Serial Number		
0.200						0.000				3284	
0.300				1		Interfer	ent			5285	
0.080 DGS	N/A	<u> </u>		-		0.050				5286	
D Doot Calib						0.200			MP6		
Simulator	ration Adjustment	Control of the Contro	Checks			- Marie Commission of the Comm			IVIT	0200	
0.050	Serial #	Lot#		Exp	piration		ments				
0.080						For			Post-Stabili		
0.200			When the same of t				oility Checks		☐ Flow Calibr	ation	
						1	bration Certi		☐ Form 40		
0.080 DGS	N/A					L Cali	bration Adjus	stment	Other	A THE RESIDENCE AND A SECOND S	
Notes/Sugge	ested Service:					Inst	rument Com	nlies with (Chapter 11D-8, I	:AC	
					7				ly with Chapter		
						Return to/Place into Evidentiary Use Remain Out of Evidentiary Use					
						-					
						e con			on Before Evide	entiary Use	
						Phil Nicodem	Digitally signed by Pl Nicodemo Date: 2022.11.08 12:1 -05'00'		el Soto Digitally Date: 20	signed by Israel Soto 22.11.09 07:32:16 -05'00'	
				,		Tech R	eview / Dat	e	Admin Review	/ / Date	



Calibration Certificate

Florida Department of Law Enforcement Alcohol Testing Program 4700 Terminal Drive, Suite 1 Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-006772, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	80-006772	UNCERTAINTY* ±	
Owning Agency:	FHP TROOP L	0.050 g/210 L	0.004
Calibration Date:	11/08/2022	0.080 g/210 L	0.004
Calibration Time:	09:27	0.200 g/ 210 L	0.007
		0.080 g/210 L Dry Gas Control	0.005

All results are reported in g/210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration. *Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full.

without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

11/08/2022

Date

FDLE/ATP Form 69 December 2021 Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Page 1 of 1

Department Inspector

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FHP TROOP L

Time of Inspection: 09:27

Date of Inspection: 11/08/2022

Serial Number: 80-006772

Software: 8100.27

Check or Test	YES	NO	Check or Test	T	
Diagnostic Check				YES	NO
(Pre-Inspection): OK	Yes		Date and/or Time Adjusted		
Minimum Sample Volume			Demonstration 1		No
Check: OK	Yes		Barometric Pressure Sensor		
Alcohol Free Subject	165		Check: OK	Yes	
Test: 0.000	1		Mouth Alcohol Test:		
	Yes		Slope Not Met	Yes	
Interferent Detect Test:			Diagnostic Check	165	
Interferent Detect	Yes		(Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:00521080A2 Exp: 02/05/2023
0.000	0.049	0.078	0.198	0.079
0.000	0.048	0.079	0.198	0.079
0.000	0.048	0.079	0.198	0.079
0.000	0.048	0.079	0.198	0.079
0.000	0.048	0.078~	0.198	0.079
0.000	0.047	0.079	0.198	
0.000	0.048	0.078	0.198	
0.000	0.047	0.078	0.198	0.079
0.000	0.048	0.078		0.078
0.000	0.047	0.078	0.198	0.078
	0.01/	0.078	0.197	0.078
Standard Deviations	0.0006	0.0005	0.0003	0.0004

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5 Remarks:

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Signature and Printed Name

11/08/2022

Type of Test	Serial Number	Agency		
Stabilities	80-006772	Florida Highway Patrol Troop L	Date	Performed By
		Fiorida riighway Fati of 1700p L	11/8/2022	DERR ///

0.05g/210L	0.08g/210L	0.20-/240:			
0.047 to 0.053	0.077 to 0.083	0.20g/210L	DGS 0.08g/210L		
	3.077 to 0.003	0.194 to 0.206 V	0.077 to 0.083		
FMP TROOP L Intoxilyzer - Alconol Amelyzer Model 8080 SM 80-026770 11/08/2022 Software: 8.00.27	FHP TROOP L Intoxilyzer - Alconol Analyzer Model 8000 - SN 80-306772 11/08/2022 Software: 8100.27	FHP TROOP L Intoxilyzer - Alconol Analyzer Model 8000 SN 80-006772 11/08/2022 Software: 8100.27	FHP TROOP L Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006772 11/08/2022 Software: 8100.27		
Test gx2:8L Time	Test g/210L Time	Test g/210L Time	Test g/210L Time		
Air Blank 0.000 C7:12 Control Test 0.049 C7:13 Air Blank 0.000 07:13 Control Test 0.049 07:14 Air Blank 0.400 07:15 Control Test 0.049 07:15 Control Test 0.049 07:15 Control Test 0.049 07:15 Control Test 0.000 07:16 Control Test 5tats Average 0.0490 Std Deu 0.0000	Air Blank 0.000 07:17 Control Test 0.078 07:18 Air Blank 0.000 07:18 Control Test 0.079 07:19 Air Blank 0.000 07:19 Control Test 0.078 07:20 Air Blank 0.000 07:21 Control Test 0.078 07:21 Control Test Stats Average 0.0783 Std Dev 0.0006 Re! Std Dev(%) 0.7370	Air Blank 0,000 07:22 Control Test 0.198 07:22 Air Blank 0.000 07:23 Control Test 0.198 07:24 Air Blank 0.000 07:24 Control Test 0.198 07:25 Air Blank 0.000 07:25 Control Test 0.198 07:25 Control Test \$1.000 07:25 Control Test \$1.000 07:25 Control Test \$1.000 07:25 Control Test \$1.000 07:25 Res \$1.000 07:25 Control Test \$1.000 0	Air Biank 0.000 07:26 Control Test 0.079 07:27 Air Blank 0.000 07:27 Control Test 0.078 07:28 Air Blank 0.000 07:28 Control Test 0.078 07:28 Air Blank 0.000 07:28 Control Test 0.078 07:29 Control Test 5.078 Air Blank 0.000 07:29 Control Test Stats Auerage 0.0783 Std Deu 0.0006 Rel Std Deu(%) 0.7370		
Openator's Signature	Operatoris Signature	Operator's Signature	Operator's Signature		