



Agency Bay County Sheriff's Office

S/N 80-006937

Date In 01-28-2022      DI Completion Date 01-31-2022

☒ Ship ☐ P/U ☐ H/D ☒ CMI ☐ EE

IS 01.31.2022

| Intake                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                            |          | By IS                  |  | Quality Checks                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                 |  | By IS |  | Date 01-28-2022                                                                                                                                                                                                                                                                                                                                                                                                    |  | Flow Calibration |  | By IS                                                                                                                                                                                                                                                                                                                                                                                                                                      |  | Date |  |           |          |           |       |        |                       |       |        |                       |       |        |                       |           |     |                        |                                                                                                                                                                                                                                                                                                     |  |  |  |
|-------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|----------|------------------------|--|----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|--|-------|--|--------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|--|------------------|--|--------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|--|------|--|-----------|----------|-----------|-------|--------|-----------------------|-------|--------|-----------------------|-------|--------|-----------------------|-----------|-----|------------------------|-----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|--|--|--|
| <input checked="" type="checkbox"/> Annual<br><input type="checkbox"/> Registration<br><input type="checkbox"/> Return from CMI / EE<br><br>Visual Inspection:<br><input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle<br><input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf<br><input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube<br><input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight<br><br>Other Equipment/ Accessories:<br><input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable<br><input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable<br><br>Notes: _____<br>_____<br>_____<br>_____<br>_____<br>_____<br>_____<br>_____<br>_____<br>_____ |          |                        |  | <input checked="" type="checkbox"/> Breath Tube Screen<br><input checked="" type="checkbox"/> Replace External O-Rings<br><input checked="" type="checkbox"/> Instrument Set Up Verified<br><input checked="" type="checkbox"/> R-Value <u>144</u><br><input checked="" type="checkbox"/> Flow Verification (L/s)<br>Flow Column # <u>ATP-102</u><br>32 mm <u>0.144</u> ( .139 - .169 )<br>36 mm <u>0.160</u> ( .156 - .190 )<br>53 mm <u>0.230</u> ( .228 - .278 )<br>103 mm <u>0.488</u> ( .447 - .547 )<br><input checked="" type="checkbox"/> Barometric Pressure Check<br>Gauge ID # <u>28421</u><br><input checked="" type="checkbox"/> Stability Checks |  |       |  | Flow Column # _____<br><input type="checkbox"/> 5L/min – 17mm<br><input type="checkbox"/> 15L/min – 53mm<br><input type="checkbox"/> 30L/min – 103mm<br><input type="checkbox"/> R-Value _____<br><input type="checkbox"/> Post Calibration Verification (L/s)<br>Flow Column # _____<br>32 mm _____ ( .139 - .169 )<br>36 mm _____ ( .156 - .190 )<br>53 mm _____ ( .228 - .278 )<br>103 mm _____ ( .447 - .547 ) |  |                  |  | <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP5088</td> <td>202010A<br/>10-05-2022</td> </tr> <tr> <td>0.080</td> <td>MP5089</td> <td>202010B<br/>10-05-2022</td> </tr> <tr> <td>0.200</td> <td>MP5090</td> <td>202010D<br/>10-06-2022</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG113403<br/>05-14-2023</td> </tr> </tbody> </table> |  |      |  | Simulator | Serial # | Lot #/Exp | 0.050 | MP5088 | 202010A<br>10-05-2022 | 0.080 | MP5089 | 202010B<br>10-05-2022 | 0.200 | MP5090 | 202010D<br>10-06-2022 | 0.080 DGS | N/A | AG113403<br>05-14-2023 | <b>Maintenance</b> By IS<br><input type="checkbox"/> Battery Replacement<br><input type="checkbox"/> Dry Gas Regulator Replacement<br><input type="checkbox"/> Breath Tube Replacement<br><input checked="" type="checkbox"/> Other <u>Form Upload</u><br>_____<br>_____<br>_____<br>_____<br>_____ |  |  |  |
| Simulator                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                         | Serial # | Lot #/Exp              |  |                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                |  |       |  |                                                                                                                                                                                                                                                                                                                                                                                                                    |  |                  |  |                                                                                                                                                                                                                                                                                                                                                                                                                                            |  |      |  |           |          |           |       |        |                       |       |        |                       |       |        |                       |           |     |                        |                                                                                                                                                                                                                                                                                                     |  |  |  |
| 0.050                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                             | MP5088   | 202010A<br>10-05-2022  |  |                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                |  |       |  |                                                                                                                                                                                                                                                                                                                                                                                                                    |  |                  |  |                                                                                                                                                                                                                                                                                                                                                                                                                                            |  |      |  |           |          |           |       |        |                       |       |        |                       |       |        |                       |           |     |                        |                                                                                                                                                                                                                                                                                                     |  |  |  |
| 0.080                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                             | MP5089   | 202010B<br>10-05-2022  |  |                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                |  |       |  |                                                                                                                                                                                                                                                                                                                                                                                                                    |  |                  |  |                                                                                                                                                                                                                                                                                                                                                                                                                                            |  |      |  |           |          |           |       |        |                       |       |        |                       |       |        |                       |           |     |                        |                                                                                                                                                                                                                                                                                                     |  |  |  |
| 0.200                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                             | MP5090   | 202010D<br>10-06-2022  |  |                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                |  |       |  |                                                                                                                                                                                                                                                                                                                                                                                                                    |  |                  |  |                                                                                                                                                                                                                                                                                                                                                                                                                                            |  |      |  |           |          |           |       |        |                       |       |        |                       |       |        |                       |           |     |                        |                                                                                                                                                                                                                                                                                                     |  |  |  |
| 0.080 DGS                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                         | N/A      | AG113403<br>05-14-2023 |  |                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                |  |       |  |                                                                                                                                                                                                                                                                                                                                                                                                                    |  |                  |  |                                                                                                                                                                                                                                                                                                                                                                                                                                            |  |      |  |           |          |           |       |        |                       |       |        |                       |       |        |                       |           |     |                        |                                                                                                                                                                                                                                                                                                     |  |  |  |

| Calibration Adjustment                                                                                                                                                                                         |          |       |            | By _____   |  |
|----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|----------|-------|------------|------------|--|
| Barometric Pressure Gauge _____                                                                                                                                                                                |          |       |            | ID # _____ |  |
| Simulator                                                                                                                                                                                                      | Serial # | Lot # | Expiration |            |  |
| 0.000                                                                                                                                                                                                          |          | N/A   | N/A        |            |  |
| 0.040                                                                                                                                                                                                          |          |       |            |            |  |
| 0.100                                                                                                                                                                                                          |          |       |            |            |  |
| 0.200                                                                                                                                                                                                          |          |       |            |            |  |
| 0.300                                                                                                                                                                                                          |          |       |            |            |  |
| 0.080 DGS                                                                                                                                                                                                      | N/A      |       |            |            |  |
| <input type="checkbox"/> Post Calibration Adjustment Stability Checks                                                                                                                                          |          |       |            |            |  |
| Simulator                                                                                                                                                                                                      | Serial # | Lot # | Expiration |            |  |
| 0.050                                                                                                                                                                                                          |          |       |            |            |  |
| 0.080                                                                                                                                                                                                          |          |       |            |            |  |
| 0.200                                                                                                                                                                                                          |          |       |            |            |  |
| 0.080 DGS                                                                                                                                                                                                      | N/A      |       |            |            |  |
| Notes/Suggested Service: Instrument lost forms, CMI uploaded forms via phone line. Bypassed an Agency Inspection to ensure forms were uploaded correctly. Form 41 and Form 40 printed correctly. IS 01-31-2022 |          |       |            |            |  |
| Tech Review: Updated shipping status to "CMI", sending instrument to repair as instrument loses forms frequently. IS 01-31-2022                                                                                |          |       |            |            |  |

| Department Inspection                                                                   |               | By IS _____                                    |  |
|-----------------------------------------------------------------------------------------|---------------|------------------------------------------------|--|
| Barometric Pressure ID# 30793                                                           |               |                                                |  |
| Gauge 1021                                                                              |               | Instrument 1018                                |  |
| Mouth Alcohol Solution Lot # 2021-D                                                     |               |                                                |  |
| Acetone Stock Solution Lot # 2021-C                                                     |               |                                                |  |
| Simulator                                                                               | Serial Number |                                                |  |
| 0.000                                                                                   | MP5086        |                                                |  |
| Interferent                                                                             | MP5087        |                                                |  |
| 0.050                                                                                   | MP5088        |                                                |  |
| 0.080                                                                                   | MP5089        |                                                |  |
| 0.200                                                                                   | MP5090        |                                                |  |
| <b>Attachments</b>                                                                      |               |                                                |  |
| <input checked="" type="checkbox"/> Form 41                                             |               | <input type="checkbox"/> Post-Stability Checks |  |
| <input checked="" type="checkbox"/> Stability Checks                                    |               | <input type="checkbox"/> Flow Calibration      |  |
| <input checked="" type="checkbox"/> Calibration Certificate                             |               | <input checked="" type="checkbox"/> Form 40    |  |
| <input type="checkbox"/> Calibration Adjustment                                         |               | <input type="checkbox"/> Other _____           |  |
| <input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC         |               |                                                |  |
| <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC             |               |                                                |  |
| <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use                |               |                                                |  |
| <input type="checkbox"/> Remain Out of Evidentiary Use                                  |               |                                                |  |
| <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use |               |                                                |  |
| Richard A Williams                                                                      |               | 14:37:37                                       |  |
| Digitally signed by Richard A Williams<br>Date: 2022.01.31 11:20:59 -05'00'             |               | 05'00'                                         |  |
| Tech Review / Date                                                                      |               | Admin Review / Date                            |  |

# Florida Department of Law Enforcement Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: BAY COUNTY SO  
Time of Inspection: 08:20

Date of Inspection: 01/31/2022

Serial Number: 80-006937  
Software: 8100.27

| Check or Test                                  | YES | NO | Check or Test                             | YES | NO |
|------------------------------------------------|-----|----|-------------------------------------------|-----|----|
| Diagnostic Check<br>(Pre-Inspection): OK       | Yes |    | Date and/or Time Adjusted                 |     | No |
| Minimum Sample Volume<br>Check: OK             | Yes |    | Barometric Pressure Sensor<br>Check: OK   | Yes |    |
| Alcohol Free Subject<br>Test: 0.000            | Yes |    | Mouth Alcohol Test:<br>Slope Not Met      | Yes |    |
| Interferent Detect Test:<br>Interferent Detect | Yes |    | Diagnostic Check<br>(Post-Inspection): OK | Yes |    |

| Alcohol Free<br>Test<br>(g/210L) | 0.05g/210L Test<br>(g/210L)<br>Lot#:202010A<br>Exp: 10/05/2022 | 0.08g/210L Test<br>(g/210L)<br>Lot#:202010B<br>Exp: 10/05/2022 | 0.20g/210L Test<br>(g/210L)<br>Lot#:202010D<br>Exp: 10/06/2022 | 0.08 g/210L<br>Dry Gas Std Test<br>(g/210L)<br>Lot#:AG113403<br>Exp: 05/14/2023 |
|----------------------------------|----------------------------------------------------------------|----------------------------------------------------------------|----------------------------------------------------------------|---------------------------------------------------------------------------------|
| 0.000                            | 0.051                                                          | 0.081                                                          | 0.201                                                          | 0.082                                                                           |
| 0.000                            | 0.050                                                          | 0.081                                                          | 0.201                                                          | 0.083                                                                           |
| 0.000                            | 0.051                                                          | 0.081                                                          | 0.201                                                          | 0.083                                                                           |
| 0.000                            | 0.051                                                          | 0.081                                                          | 0.201                                                          | 0.083                                                                           |
| 0.000                            | 0.051                                                          | 0.081                                                          | 0.201                                                          | 0.083                                                                           |
| 0.000                            | 0.050                                                          | 0.082                                                          | 0.201                                                          | 0.082                                                                           |
| 0.000                            | 0.051                                                          | 0.081                                                          | 0.201                                                          | 0.083                                                                           |
| 0.000                            | 0.051                                                          | 0.081                                                          | 0.201                                                          | 0.082                                                                           |
| 0.000                            | 0.051                                                          | 0.081                                                          | 0.201                                                          | 0.082                                                                           |
| 0.000                            | 0.051                                                          | 0.081                                                          | 0.201                                                          | 0.082                                                                           |

|                     |        |        |        |        |
|---------------------|--------|--------|--------|--------|
| Standard Deviations | 0.0004 | 0.0003 | 0.0000 | 0.0005 |
|---------------------|--------|--------|--------|--------|

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies ( ☒ ) does not comply ( ☐ ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



ISRAEL SOTO

Signature and Printed Name

01/31/2022  
Date

# Stability checks

BAY COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-006937  
01/28/2022  
Software: 8100.27

| Test               | g/210L | Time  |
|--------------------|--------|-------|
| Air Blank          | 0.000  | 14:22 |
| Control Test       | 0.049  | 14:23 |
| Air Blank          | 0.000  | 14:23 |
| Control Test       | 0.049  | 14:24 |
| Air Blank          | 0.000  | 14:24 |
| Control Test       | 0.048  | 14:25 |
| Air Blank          | 0.000  | 14:26 |
| Control Test Stats |        |       |
| Average            | 0.0487 |       |
| Std Dev            | 0.0006 |       |
| Rel Std Dev(%)     | 1.1863 |       |

Operator's Signature

BAY COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-006937  
01/28/2022  
Software: 8100.27

| Test               | g/210L | Time  |
|--------------------|--------|-------|
| Air Blank          | 0.000  | 14:26 |
| Control Test       | 0.079  | 14:27 |
| Air Blank          | 0.000  | 14:28 |
| Control Test       | 0.080  | 14:28 |
| Air Blank          | 0.000  | 14:29 |
| Control Test       | 0.079  | 14:30 |
| Air Blank          | 0.000  | 14:30 |
| Control Test Stats |        |       |
| Average            | 0.0793 |       |
| Std Dev            | 0.0006 |       |
| Rel Std Dev(%)     | 0.7277 |       |

wet

Operator's Signature

BAY COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-006937  
01/28/2022  
Software: 8100.27

| Test               | g/210L | Time  |
|--------------------|--------|-------|
| Air Blank          | 0.000  | 14:31 |
| Control Test       | 0.199  | 14:31 |
| Air Blank          | 0.000  | 14:32 |
| Control Test       | 0.198  | 14:33 |
| Air Blank          | 0.000  | 14:33 |
| Control Test       | 0.198  | 14:34 |
| Air Blank          | 0.000  | 14:34 |
| Control Test Stats |        |       |
| Average            | 0.1983 |       |
| Std Dev            | 0.0006 |       |
| Rel Std Dev(%)     | 0.2911 |       |

Operator's Signature

BAY COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-006937  
01/28/2022  
Software: 8100.27

| Test               | g/210L | Time  |
|--------------------|--------|-------|
| Air Blank          | 0.000  | 14:36 |
| Control Test       | 0.082  | 14:37 |
| Air Blank          | 0.000  | 14:37 |
| Control Test       | 0.081  | 14:38 |
| Air Blank          | 0.000  | 14:38 |
| Control Test       | 0.081  | 14:38 |
| Air Blank          | 0.000  | 14:39 |
| Control Test Stats |        |       |
| Average            | 0.0813 |       |
| Std Dev            | 0.0006 |       |
| Rel Std Dev(%)     | 0.7099 |       |

Dry

Operator's Signature



# Calibration Certificate

Florida Department of Law Enforcement  
Alcohol Testing Program  
2729 Fort Knox Blvd.  
Bldg. 2, Suite 1300  
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-006937, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

|                   |                      |                                |                  |
|-------------------|----------------------|--------------------------------|------------------|
| Serial Number:    | <u>80-006937</u>     | Serial Number:                 | <u>80-006937</u> |
| Owning Agency:    | <u>BAY COUNTY SO</u> | 0.050 g/ 210 L                 | 0.004            |
| Calibration Date: | <u>01/31/2022</u>    | 0.080 g/ 210 L                 | 0.004            |
| Calibration Time: | <u>08:20</u>         | 0.200 g/ 210 L                 | 0.007            |
|                   |                      | 0.080 g/ 210 L Dry Gas Control | 0.005            |

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within  $\pm 0.005$  or 5%, whichever is greater, of the target alcohol concentration.

\*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ( $k=3$ ).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

## TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

**Israel Soto**  
Soto  
Digitally signed by Israel Soto  
Date: 2022.01.31 10:11:44 -05'00'

01/31/2022

Date

ISRAEL SOTO,  
Department Inspector

FDLE/ATP Form 69 January 2022

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Page 1 of 1

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BAY COUNTY SO  
Time of Inspection: 08:52

Date of Inspection: 01/31/2022

Serial Number: 80-006937  
Software: 8100.27

| Check or Test                               | YES | NO |
|---------------------------------------------|-----|----|
| Date and/or Time Adjusted                   |     | No |
| Diagnostic Check (Pre-Inspection): OK       |     | No |
| Alcohol Free Subject Test: 0.000            |     | No |
| Mouth Alcohol Test: Slope Not Met           |     | No |
| Interferent Detect Test: Interferent Detect |     | No |
| Diagnostic Check (Post-Inspection): OK      |     | No |

|                            |                                           |                                           |                                           |                                                        |
|----------------------------|-------------------------------------------|-------------------------------------------|-------------------------------------------|--------------------------------------------------------|
| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L)<br>Lot#:<br>Exp: | 0.08g/210L Test (g/210L)<br>Lot#:<br>Exp: | 0.20g/210L Test (g/210L)<br>Lot#:<br>Exp: | 0.08 g/210L Dry Gas Std Test (g/210L)<br>Lot#:<br>Exp: |
|                            |                                           |                                           |                                           |                                                        |
|                            |                                           |                                           |                                           |                                                        |
|                            |                                           |                                           |                                           |                                                        |

Number of Simulators Used: \_\_\_\_\_

Remarks:

BYPASSED AI TO CHECK FORM UPLOAD

*N/A Not a Compliance check*

The above instrument complies ( ☒ ) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Israel Soto*

ISRAEL SOTO

Signature and Printed Name

01/31/2022  
Date

## **Return Material Authorization**

**Ship to:** ☒ CMI, Inc.

☐ Enforcement Electronics

Shipment to repair facility authorized by: Brian Miller on 1-31-2022

Items Returned:      Instrument ☒    Supplies ☐    Other ☐ Describe: \_\_\_\_\_

Instrument Model: Intoxilyzer 8000      Serial Number: 80-006937

Bill To Address:  
Warranty Repair

Ship to Address:  
Alcohol Testing Program - FDLE  
Tallahassee, Florida

**Reason for Return:**

Instrument continuously loses forms and is losing inspections. Instrument was sent to repair in June 2021 for this issue and instrument continues to lose forms.

**Please choose one of the following options:**

- ☐ 1. I \_\_\_\_\_, authorize all repairs.
- ☐ 2. I \_\_\_\_\_, authorize repairs up to \$\_\_\_\_\_.
- ☒ 3. I require an estimate **BEFORE** any repairs will be authorized and/ or conducted.

Please contact: Name: Brian Miller

Phone #: 850-215-6806

Email: brian.miller@bayso.org

ATP Contact Name: Israel Soto

ATP Email: israelsoto@fdle.state.fl.us





Agency Bay County Sheriff's Office

S/N 80-006937

Date In 7/5/2022

DI Completion Date 07-06-2022

☐ Ship ☒ P/U ☐ H/D ☐ CMI ☐ EE

## Calibration Adjustment

By \_\_\_\_\_

#### ☐ Post Calibration Adjustment Stability Checks

Notes/Suggested Service:

By IS

- ☒ Instrument Complies with Chapter 11D-8, FAC
- ☐ Instrument Does Not Comply with Chapter 11D-8, FAC

---

- ☒ Return to/Place into Evidentiary Use
- ☐ Remain Out of Evidentiary Use

---

- ☒ Conduct an Agency Inspection Before Evidentiary Use

Taylor Gutschow  
Digitally signed by Taylor Gutschow  
Date: 2022.07.06 11:39:04  
+04'00'

Tech Review / Date Admin Review / Date

# Florida Department of Law Enforcement Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: BAY COUNTY SO  
Time of Inspection: 10:01

Date of Inspection: 07/06/2022

Serial Number: 80-006937  
Software: 8100.27

| Check or Test                                  | YES | NO | Check or Test                             | YES | NO |
|------------------------------------------------|-----|----|-------------------------------------------|-----|----|
| Diagnostic Check<br>(Pre-Inspection): OK       | Yes |    | Date and/or Time Adjusted                 | Yes |    |
| Minimum Sample Volume<br>Check: OK             | Yes |    | Barometric Pressure Sensor<br>Check: OK   | Yes |    |
| Alcohol Free Subject<br>Test: 0.000            | Yes |    | Mouth Alcohol Test:<br>Slope Not Met      | Yes |    |
| Interferent Detect Test:<br>Interferent Detect | Yes |    | Diagnostic Check<br>(Post-Inspection): OK | Yes |    |

| Alcohol Free<br>Test<br>(g/210L) | 0.05g/210L Test<br>(g/210L)<br>Lot#:202201C<br>Exp: 01/11/2024 | 0.08g/210L Test<br>(g/210L)<br>Lot#:202201D<br>Exp: 01/18/2024 | 0.20g/210L Test<br>(g/210L)<br>Lot#:202201E<br>Exp: 01/18/2024 | 0.08 g/210L<br>Dry Gas Std Test<br>(g/210L)<br>Lot#:AG113403<br>Exp: 05/14/2023 |
|----------------------------------|----------------------------------------------------------------|----------------------------------------------------------------|----------------------------------------------------------------|---------------------------------------------------------------------------------|
| 0.000                            | 0.051                                                          | 0.079                                                          | 0.200                                                          | 0.081                                                                           |
| 0.000                            | 0.050                                                          | 0.079                                                          | 0.200                                                          | 0.081                                                                           |
| 0.000                            | 0.050                                                          | 0.079                                                          | 0.200                                                          | 0.081                                                                           |
| 0.000                            | 0.050                                                          | 0.079                                                          | 0.200                                                          | 0.081                                                                           |
| 0.000                            | 0.051                                                          | 0.079                                                          | 0.200                                                          | 0.080                                                                           |
| 0.000                            | 0.050                                                          | 0.079                                                          | 0.200                                                          | 0.081                                                                           |
| 0.000                            | 0.050                                                          | 0.079                                                          | 0.200                                                          | 0.081                                                                           |
| 0.000                            | 0.050                                                          | 0.079                                                          | 0.200                                                          | 0.081                                                                           |
| 0.000                            | 0.051                                                          | 0.079                                                          | 0.200                                                          | 0.081                                                                           |
| 0.000                            | 0.050                                                          | 0.079                                                          | 0.200                                                          | 0.081                                                                           |

|                     |        |        |        |        |
|---------------------|--------|--------|--------|--------|
| Standard Deviations | 0.0004 | 0.0000 | 0.0000 | 0.0003 |
|---------------------|--------|--------|--------|--------|

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0001 Number of Simulators Used: 5

**Remarks:**

Time-Date changed. CENTRAL TIME ZONE

The above instrument complies ( ☒ ) does not comply ( ☐ ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Israel Soto*

ISRAEL SOTO

Signature and Printed Name

07/06/2022  
Date



| Type of Test     | Serial Number | Agency                      | Date     | Performed By |
|------------------|---------------|-----------------------------|----------|--------------|
| Stability Checks | 80-006937     | Bay County Sheriff's Office | 7/6/2022 | PN           |

| 0.05g/210L                                                                                                                                                                                                                                                                                                                                                                                                                                                 | 0.08g/210L                                                                                                                                                                                                                                                                                                                                                                                                                                                 | 0.20g/210L                                                                                                                                                                                                                                                                                                                                                                                                                                                 | DGS 0.08g/210L                                                                                                                                                                                                                                                                                                                                                                                                                                             |
|------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| 0.047 to 0.053                                                                                                                                                                                                                                                                                                                                                                                                                                             | 0.077 to 0.083                                                                                                                                                                                                                                                                                                                                                                                                                                             | 0.194 to 0.206                                                                                                                                                                                                                                                                                                                                                                                                                                             | 0.077 to 0.083                                                                                                                                                                                                                                                                                                                                                                                                                                             |
| ✓                                                                                                                                                                                                                                                                                                                                                                                                                                                          | ✓                                                                                                                                                                                                                                                                                                                                                                                                                                                          | ✓                                                                                                                                                                                                                                                                                                                                                                                                                                                          | ✓                                                                                                                                                                                                                                                                                                                                                                                                                                                          |
| <p>BAY COUNTY SO<br/>Intoxilyzer - Alcohol Analyzer<br/>Model 8000<br/>07/06/2022<br/>Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 08:17</p> <p>Control Test 0.052 08:18</p> <p>Air Blank 0.000 08:18</p> <p>Control Test 0.052 08:19</p> <p>Air Blank 0.000 08:19</p> <p>Control Test 0.051 08:20</p> <p>Air Blank 0.000 08:21</p> <p>Control Test Stats</p> <p>Average 0.0517</p> <p>Std Dev 0.0006</p> <p>Rel Std Dev(%) 1.1175</p> | <p>BAY COUNTY SO<br/>Intoxilyzer - Alcohol Analyzer<br/>Model 8000<br/>07/06/2022<br/>Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 08:22</p> <p>Control Test 0.081 08:23</p> <p>Air Blank 0.000 08:24</p> <p>Control Test 0.080 08:24</p> <p>Air Blank 0.000 08:25</p> <p>Control Test 0.080 08:25</p> <p>Air Blank 0.000 08:26</p> <p>Control Test Stats</p> <p>Average 0.0803</p> <p>Std Dev 0.0006</p> <p>Rel Std Dev(%) 0.7187</p> | <p>BAY COUNTY SO<br/>Intoxilyzer - Alcohol Analyzer<br/>Model 8000<br/>07/06/2022<br/>Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 08:27</p> <p>Control Test 0.201 08:28</p> <p>Air Blank 0.000 08:28</p> <p>Control Test 0.200 08:29</p> <p>Air Blank 0.000 08:30</p> <p>Control Test 0.200 08:30</p> <p>Air Blank 0.000 08:31</p> <p>Control Test Stats</p> <p>Average 0.2003</p> <p>Std Dev 0.0006</p> <p>Rel Std Dev(%) 0.2882</p> | <p>BAY COUNTY SO<br/>Intoxilyzer - Alcohol Analyzer<br/>Model 8000<br/>07/06/2022<br/>Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 08:33</p> <p>Control Test 0.081 08:33</p> <p>Air Blank 0.000 08:33</p> <p>Control Test 0.081 08:34</p> <p>Air Blank 0.000 08:34</p> <p>Control Test 0.080 08:35</p> <p>Air Blank 0.000 08:35</p> <p>Control Test Stats</p> <p>Average 0.0807</p> <p>Std Dev 0.0006</p> <p>Rel Std Dev(%) 0.7157</p> |
| <p>Operator's Signature</p> <p></p>                                                                                                                                                                                                                                                                                                                                   | <p>Operator's Signature</p> <p></p>                                                                                                                                                                                                                                                                                                                                   | <p>Operator's Signature</p> <p></p>                                                                                                                                                                                                                                                                                                                                     | <p>Operator's Signature</p> <p></p>                                                                                                                                                                                                                                                                                                                                     |

Comments: N/A



# Calibration Certificate

Florida Department of Law Enforcement  
Alcohol Testing Program  
2331 Phillips Road.  
Suite B1032  
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-006937, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

|                   |                      |                                |       |
|-------------------|----------------------|--------------------------------|-------|
| Serial Number:    | <u>80-006937</u>     | UNCERTAINTY* $\pm$             |       |
| Owning Agency:    | <u>BAY COUNTY SO</u> | 0.050 g/ 210 L                 | 0.004 |
| Calibration Date: | <u>07/06/2022</u>    | 0.080 g/ 210 L                 | 0.004 |
| Calibration Time: | <u>10:01</u>         | 0.200 g/ 210 L                 | 0.007 |
|                   |                      | 0.080 g/ 210 L Dry Gas Control | 0.005 |

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within  $\pm 0.005$  or 5%, whichever is greater, of the target alcohol concentration.

\*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ( $k=3$ ).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

## TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

**Israel Soto**  
Soto  
Date: 2022.07.06 11:21:45  
-04'00'

Date

07/06/2022

ISRAEL SOTO,

**Department Inspector**

FDLE/ATP Form 69 March 2022

Issuing Authority: Alcohol Testing Program

*Service • Integrity • Respect • Quality*