

INSTRUMENT PROCESSING SHEET

Agency Hillsborough CSO

S/N 80-007163

Florida Department of
Law Enforcement

Date In 2/11/2022 DI Completion Date 2/15/2022

☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By TDG	Quality Checks	By TDG	Date 2/15/2022	Flow Calibration By	Date																																								
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: _____ _____ _____ _____ _____ _____ _____ _____ _____ _____		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>214</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP104</u> 32 mm <u>0.148</u> (.139 - .169) 36 mm <u>0.164</u> (.156 - .190) 53 mm <u>0.234</u> (.228 - .278) 103 mm <u>0.496</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>68639</u> ☒ Stability Checks			Flow Column # _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																									
		<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td rowspan="2">0.050</td> <td rowspan="2">MP6286</td> <td>202010A</td> </tr> <tr> <td>10/05/2022</td> </tr> <tr> <td rowspan="2">0.080</td> <td rowspan="2">MP6287</td> <td>202010B</td> </tr> <tr> <td>10/05/2022</td> </tr> <tr> <td rowspan="2">0.200</td> <td rowspan="2">MP6288</td> <td>202010D</td> </tr> <tr> <td>10/06/2022</td> </tr> <tr> <td rowspan="2">0.080 DGS</td> <td rowspan="2">N/A</td> <td>AG115904</td> </tr> <tr> <td>06/08/2023</td> </tr> </tbody> </table>			Simulator	Serial #	Lot #/Exp	0.050	MP6286	202010A	10/05/2022	0.080	MP6287	202010B	10/05/2022	0.200	MP6288	202010D	10/06/2022	0.080 DGS	N/A	AG115904	06/08/2023	<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Maintenance</th> <th>By</th> </tr> </thead> <tbody> <tr> <td>☐ Battery Replacement</td> <td></td> </tr> <tr> <td>☐ Dry Gas Regulator Replacement</td> <td></td> </tr> <tr> <td>☐ Breath Tube Replacement</td> <td></td> </tr> <tr> <td>☐ Other _____</td> <td></td> </tr> <tr> <td> </td> <td></td> </tr> <tr> <td> </td> <td></td> </tr> <tr> <td> </td> <td></td> </tr> <tr> <td> </td> <td></td> </tr> <tr> <td> </td> <td></td> </tr> </tbody> </table>		Maintenance	By	☐ Battery Replacement		☐ Dry Gas Regulator Replacement		☐ Breath Tube Replacement		☐ Other _____												
Simulator	Serial #	Lot #/Exp																																												
0.050	MP6286	202010A																																												
		10/05/2022																																												
0.080	MP6287	202010B																																												
		10/05/2022																																												
0.200	MP6288	202010D																																												
		10/06/2022																																												
0.080 DGS	N/A	AG115904																																												
		06/08/2023																																												
Maintenance	By																																													
☐ Battery Replacement																																														
☐ Dry Gas Regulator Replacement																																														
☐ Breath Tube Replacement																																														
☐ Other _____																																														
Calibration Adjustment By _____			Department Inspection By TDG																																											
Barometric Pressure Gauge ID # _____ <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>			Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Barometric Pressure ID# <u>28199</u> Gauge <u>1027</u> Instrument <u>1025</u> Mouth Alcohol Solution Lot # <u>2021-B</u> Acetone Stock Solution Lot # <u>2021-C</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP6284</td> </tr> <tr> <td>Interferent</td> <td>MP6285</td> </tr> <tr> <td>0.050</td> <td>MP6286</td> </tr> <tr> <td>0.080</td> <td>MP6287</td> </tr> <tr> <td>0.200</td> <td>MP62868</td> </tr> </tbody> </table> TDG 2/17/2022				Simulator	Serial Number	0.000	MP6284	Interferent	MP6285	0.050	MP6286	0.080	MP6287	0.200	MP62868
Simulator	Serial #	Lot #	Expiration																																											
0.000		N/A	N/A																																											
0.040																																														
0.100																																														
0.200																																														
0.300																																														
0.080 DGS	N/A																																													
Simulator	Serial Number																																													
0.000	MP6284																																													
Interferent	MP6285																																													
0.050	MP6286																																													
0.080	MP6287																																													
0.200	MP62868																																													
☐ Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>			Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th colspan="2">Attachments</th> </tr> </thead> <tbody> <tr> <td style="width:50%; vertical-align: top;"> ☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Calibration Adjustment </td> <td style="width:50%; vertical-align: top;"> ☐ Post-Stability Checks ☐ Flow Calibration ☐ Form 40 ☐ Other _____ </td> </tr> </tbody> </table>				Attachments		☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Calibration Adjustment	☐ Post-Stability Checks ☐ Flow Calibration ☐ Form 40 ☐ Other _____																
Simulator	Serial #	Lot #	Expiration																																											
0.050																																														
0.080																																														
0.200																																														
0.080 DGS	N/A																																													
Attachments																																														
☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Calibration Adjustment	☐ Post-Stability Checks ☐ Flow Calibration ☐ Form 40 ☐ Other _____																																													
Notes/Suggested Service: _____ _____ Admin Review: Corrected sim serial number. (TDG 2/17/2022) _____ _____ _____ _____ _____ _____			☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use Richard A Williams Digitally signed by Richard A Williams Date: 2022.02.15 16:16:53 -05'00' 28 12:56:56 -05'00' Tech Review / Date Admin Review / Date																																											

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-007163	Hillsborough CSO	02/15/2022	TDG <i>ML</i>

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083
HILLSBOROUGH CO SO Intoxilyzer - Alcohol Analyzer Model 8000 02/15/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 09:33 Control Test 0.050 09:34 Air Blank 0.000 09:34 Control Test 0.050 09:35 Air Blank 0.000 09:35 Control Test 0.050 09:36 Air Blank 0.000 09:37 Control Test Stats Average 0.0500 Std Dev 0.0000 Rel Std Dev(%) 0.0000	HILLSBOROUGH CO SO Intoxilyzer - Alcohol Analyzer Model 8000 02/15/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 09:40 Control Test 0.082 09:41 Air Blank 0.000 09:41 Control Test 0.081 09:42 Air Blank 0.000 09:42 Control Test 0.080 09:43 Air Blank 0.000 09:44 Control Test Stats Average 0.0810 Std Dev 0.0010 Rel Std Dev(%) 1.2346	HILLSBOROUGH CO SO Intoxilyzer - Alcohol Analyzer Model 8000 02/15/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 09:46 Control Test 0.200 09:47 Air Blank 0.000 09:47 Control Test 0.200 09:48 Air Blank 0.000 09:49 Control Test 0.200 09:49 Air Blank 0.000 09:50 Control Test Stats Average 0.2000 Std Dev 0.0000 Rel Std Dev(%) 0.0000	HILLSBOROUGH CO SO Intoxilyzer - Alcohol Analyzer Model 8000 02/15/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 09:51 Control Test 0.081 09:52 Air Blank 0.000 09:52 Control Test 0.080 09:52 Air Blank 0.000 09:53 Control Test 0.080 09:53 Air Blank 0.000 09:54 Control Test Stats Average 0.0803 Std Dev 0.0006 Rel Std Dev(%) 0.7187
<i>ML</i> Operator's Signature	<i>ML</i> Operator's Signature	<i>ML</i> Operator's Signature	<i>ML</i> Operator's Signature

Comments:

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: HILLSBOROUGH CO SO
Time of Inspection: 14:34

Date of Inspection: 02/15/2022

Serial Number: 80-007163
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202010A Exp: 10/05/2022	0.08g/210L Test (g/210L) Lot#:202010B Exp: 10/05/2022	0.20g/210L Test (g/210L) Lot#:202010D Exp: 10/06/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG115904 Exp: 06/08/2023
0.000	0.050	0.079	0.199	0.080
0.000	0.049	0.079	0.199	0.080
0.000	0.049	0.079	0.198	0.079
0.000	0.049	0.079	0.199	0.080
0.000	0.049	0.079	0.198	0.079
0.000	0.049	0.079	0.199	0.079
0.000	0.049	0.078	0.199	0.080
0.000	0.050	0.079	0.199	0.079
0.000	0.050	0.079	0.199	0.079
0.000	0.049	0.079	0.199	0.079

Standard Deviations	0.0004	0.0003	0.0004	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5



Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Taylor D Gutschow TAYLOR D GUTSCHOW
Signature and Printed Name

02/15/2022
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-007163, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-007163</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>HILLSBOROUGH CO SO</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>02/15/2022</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>14:34</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.


TAYLOR D GUTSCHOW,
Department Inspector

Date

02/15/2022

FDLE/ATP Form 69 December 2021
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality