



INSTRUMENT PROCESSING SHEET

Agency Monroe County Sheriff's OfficeS/N 80-007170Florida Department of
Law EnforcementDate In 3/9/2022 DI Completion Date 3/9/2022☐ Ship ☒ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By <u>DERR</u>	Quality Checks	By <u>DERR</u>	Date <u>3/9/2022</u>	Flow Calibration	By <u>DERR</u>	Date <u>3/9/2022</u>																																								
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes: _____ _____ _____ _____ _____ _____ _____ _____ _____ _____		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>183</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP106</u> 32 mm <u>0.140</u> (.139 - .169) 36 mm <u>0.164</u> (.156 - .190) 53 mm <u>0.230</u> (.228 - .278) 103 mm <u>0.515</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28663</u> ☒ Stability Checks			Flow Column # _____ ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																										
		<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP6286</td><td>202010A 10/05/2022</td></tr><tr><td>0.080</td><td>MP6287</td><td>202010B 10/05/2022</td></tr><tr><td>0.200</td><td>MP6288</td><td>202010D 10/06/2022</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG115904 06/08/2023</td></tr></tbody></table>	Simulator	Serial #	Lot #/Exp	0.050	MP6286	202010A 10/05/2022	0.080	MP6287	202010B 10/05/2022	0.200	MP6288	202010D 10/06/2022	0.080 DGS	N/A	AG115904 06/08/2023			Maintenance By _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____ _____ _____ _____ _____ _____																											
Simulator	Serial #	Lot #/Exp																																													
0.050	MP6286	202010A 10/05/2022																																													
0.080	MP6287	202010B 10/05/2022																																													
0.200	MP6288	202010D 10/06/2022																																													
0.080 DGS	N/A	AG115904 06/08/2023																																													
Calibration Adjustment By _____		Department Inspection By <u>DERR</u>																																													
Barometric Pressure Gauge _____ ID # _____		Barometric Pressure ID# <u>68639</u> Gauge <u>1016</u> Instrument <u>1014</u> Mouth Alcohol Solution Lot # <u>2021-D</u> Acetone Stock Solution Lot # <u>2021-C</u>																																													
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP6284</td></tr><tr><td>Interferent</td><td>MP6285</td></tr><tr><td>0.050</td><td>MP6286</td></tr><tr><td>0.080</td><td>MP6287</td></tr><tr><td>0.200</td><td>MP6288</td></tr></tbody></table>						Simulator	Serial Number	0.000	MP6284	Interferent	MP6285	0.050	MP6286	0.080	MP6287	0.200	MP6288
Simulator	Serial #	Lot #	Expiration																																												
0.000		N/A	N/A																																												
0.040																																															
0.100																																															
0.200																																															
0.300																																															
0.080 DGS	N/A																																														
Simulator	Serial Number																																														
0.000	MP6284																																														
Interferent	MP6285																																														
0.050	MP6286																																														
0.080	MP6287																																														
0.200	MP6288																																														
☐ Post Calibration Adjustment Stability Checks		<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>						Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Attachments ☒ Form 41 ☐ Post-Stability Checks ☒ Stability Checks ☐ Flow Calibration ☒ Calibration Certificate ☐ Form 40 ☐ Calibration Adjustment ☐ Other _____																			
Simulator	Serial #	Lot #	Expiration																																												
0.050																																															
0.080																																															
0.200																																															
0.080 DGS	N/A																																														
Notes/Suggested Service: _____ _____ _____ _____ _____ _____ _____ _____ _____ _____		☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																																													
		Taylor Gutschow Digitally signed by Taylor Gutschow Date: 2022.03.09 11:23:08 -0500 Tech Review / Date _____ Admin Review / Date _____																																													



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-007170, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-007170</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>MONROE COUNTY SO</u>	0.050 g/210 L	0.004
Calibration Date:	<u>03/09/2022</u>	0.080 g/210 L	0.004
Calibration Time:	<u>10:40</u>	0.200 g/210 L	0.007
		0.080 g/210 L Dry Gas Control	0.005

All results are reported in g/210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

03/09/2022

Date

David E Reyes-Rivera

DAVID E REYES-RIVERA,

Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Page 1 of 1

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: MONROE COUNTY SO
Time of Inspection: 10:40

Date of Inspection: 03/09/2022

Serial Number: 80-007170
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202010A Exp: 10/05/2022	0.08g/210L Test (g/210L) Lot#:202010B Exp: 10/05/2022	0.20g/210L Test (g/210L) Lot#:202010D Exp: 10/06/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG115904 Exp: 06/08/2023
0.000	0.048	0.078	0.195	0.079
0.000	0.047	0.078	0.195	0.078
0.000	0.047	0.077	0.195	0.079
0.000	0.048	0.078	0.195	0.079
0.000	0.047	0.077	0.195	0.079
0.000	0.047	0.077	0.195	0.079
0.000	0.047	0.077	0.195	0.079
0.000	0.048	0.078	0.195	0.079
0.000	0.048	0.078	0.195	0.079
0.000	0.048	0.077	0.195	0.080

Standard Deviations	0.0005	0.0005	0.0000	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



DAVID E REYES-RIVERA

Signature and Printed Name

03/09/2022
Date

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-007170	Monroe County Sheriff's Office	3/9/2022	DERR

0.05g/210L 0.047 to 0.053 MONROE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 03/09/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 08:08 Control Test 0.047 08:08 Air Blank 0.000 08:09 Control Test 0.048 08:09 Air Blank 0.000 08:10 Control Test 0.048 08:11 Air Blank 0.000 08:11 Control Test Stats Average 0.0477 Std Dev 0.0006 Rel Std Dev(%) 1.2112 Operator's Signature	0.08g/210L 0.077 to 0.083 MONROE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 03/09/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 08:13 Control Test 0.078 08:13 Air Blank 0.000 08:14 Control Test 0.078 08:14 Air Blank 0.000 08:15 Control Test 0.077 08:16 Air Blank 0.000 08:16 Control Test Stats Average 0.0777 Std Dev 0.0006 Rel Std Dev(%) 0.7434 Operator's Signature	0.20g/210L 0.194 to 0.206 MONROE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 03/09/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 08:17 Control Test 0.195 08:18 Air Blank 0.000 08:18 Control Test 0.195 08:19 Air Blank 0.000 08:20 Control Test 0.195 08:20 Air Blank 0.000 08:21 Control Test Stats Average 0.1950 Std Dev 0.0000 Rel Std Dev(%) 0.0000 Operator's Signature	0.08g/210L 0.077 to 0.083 MONROE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 03/09/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 08:22 Control Test 0.079 08:22 Air Blank 0.000 08:23 Control Test 0.079 08:23 Air Blank 0.000 08:24 Control Test 0.079 08:24 Air Blank 0.000 08:24 Control Test Stats Average 0.0790 Std Dev 0.0000 Rel Std Dev(%) 0.0000 Operator's Signature	0.05g/210L 0.047 to 0.053 MONROE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 03/09/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 08:22 Control Test 0.047 08:22 Air Blank 0.000 08:23 Control Test 0.048 08:23 Air Blank 0.000 08:24 Control Test 0.048 08:24 Air Blank 0.000 08:24 Control Test Stats Average 0.0477 Std Dev 0.0006 Rel Std Dev(%) 1.2112 Operator's Signature
---	---	---	---	---