

Type of Test	Serial Number	Agency	Date	Performed By
Stability Checks	80-007363	St Johns County Sheriff's Office	10/11/2022	PN

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083
ST JOHNS SO Intoxilyzer - Alcohol Analyzer Model 8000 10/11/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 13:07 Control Test 0.050 13:08 Air Blank 0.000 13:08 Control Test 0.050 13:09 Air Blank 0.000 13:09 Control Test 0.050 13:10 Air Blank 0.000 13:11 Control Test Stats Average 0.0500 Std Dev 0.0000 Rel Std Dev(%) 0.0000 Operator's Signature	ST JOHNS SO Intoxilyzer - Alcohol Analyzer Model 8000 10/11/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 13:12 Control Test 0.079 13:13 Air Blank 0.000 13:14 Control Test 0.079 13:14 Air Blank 0.000 13:15 Control Test 0.079 13:16 Air Blank 0.000 13:16 Control Test Stats Average 0.0790 Std Dev 0.0000 Rel Std Dev(%) 0.0000 Operator's Signature	ST JOHNS SO Intoxilyzer - Alcohol Analyzer Model 8000 10/11/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 13:23 Control Test 0.199 13:24 Air Blank 0.000 13:25 Control Test 0.199 13:25 Air Blank 0.000 13:26 Control Test 0.199 13:26 Air Blank 0.000 13:27 Control Test Stats Average 0.1990 Std Dev 0.0000 Rel Std Dev(%) 0.0000 Operator's Signature	ST JOHNS SO Intoxilyzer - Alcohol Analyzer Model 8000 10/11/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 13:29 Control Test 0.079 13:29 Air Blank 0.000 13:29 Control Test 0.079 13:30 Air Blank 0.000 13:30 Control Test 0.079 13:31 Air Blank 0.000 13:31 Control Test Stats Average 0.0790 Std Dev 0.0000 Rel Std Dev(%) 0.0000 Operator's Signature

Comments: N/A

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: ST JOHNS SO
Time of Inspection: 15:22

Date of Inspection: 10/11/2022

Serial Number: 80-007363
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:08121080A1 Exp: 05/05/2023
0.000	0.050	0.080	0.201	0.079
0.000	0.050	0.081	0.202	0.079
0.000	0.050	0.080	0.202	0.079
0.000	0.051	0.080	0.202	0.079
0.000	0.051	0.080	0.202	0.079
0.000	0.051	0.081	0.201	0.079
0.000	0.051	0.081	0.202	0.079
0.000	0.051	0.081	0.202	0.079
0.000	0.051	0.081	0.201	0.079
0.000	0.051	0.081	0.202	0.079
Standard Deviations	0.0004	0.0005	0.0004	0.0000

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

*Accidentally entered character For middle initial.
P.N 10/11/2022*

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



Signature and Printed Name PHIL / NICODEMO
10/11/2022
Date *P.N 10/11/2022*



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road.
Suite B1032
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-007363, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-007363</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>ST JOHNS SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>10/11/2022</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>15:22</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.
This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

FDLE/ATP Form 69 March 2022

Issuing Authority: Alcohol Testing Program

10/11/2022

Date


PHIL - NICODEMO,

Department Inspector

Service • Integrity • Respect • Quality



INSTRUMENT PROCESSING SHEET

Agency St Johns Sheriff's OfficeS/N 80-007363Florida Department of
Law EnforcementDate In 11/16/2022 DI Completion Date 11/18/2022☐ Ship ☒ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By PN	Quality Checks	By PN	Date <u>11/18/2022</u>	Flow Calibration	By	Date																																							
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes: <u>Agency Inspector noted</u> <u>instrument indicated 'Control</u> <u>Outside Tolerance' on the</u> <u>dry gas standard measurements</u> <u>PN (11/18)</u>		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>160</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP 102</u> 32 mm <u>.164</u> (.139 - .169) 36 mm <u>.175</u> (.156 - .190) 53 mm <u>.242</u> (.228 - .278) 103 mm <u>.488</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28662</u> ☒ Stability Checks			Flow Column # _____ ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																									
		<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP5088</td><td>202201C 01/11/2024</td></tr><tr><td>0.080</td><td>MP5089</td><td>202201D 01/18/2024</td></tr><tr><td>0.200</td><td>MP5084</td><td>202201E 01/18/2024</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG229803 10/25/2024</td></tr></tbody></table>	Simulator	Serial #	Lot #/Exp	0.050	MP5088	202201C 01/11/2024	0.080	MP5089	202201D 01/18/2024	0.200	MP5084	202201E 01/18/2024	0.080 DGS	N/A	AG229803 10/25/2024																													
Simulator	Serial #	Lot #/Exp																																												
0.050	MP5088	202201C 01/11/2024																																												
0.080	MP5089	202201D 01/18/2024																																												
0.200	MP5084	202201E 01/18/2024																																												
0.080 DGS	N/A	AG229803 10/25/2024																																												
Calibration Adjustment		By _____		Department Inspection			By PN _____																																							
Barometric Pressure Gauge _____ ID # _____				Barometric Pressure ID# <u>28662</u> Gauge <u>1023</u> Instrument <u>1023</u> Mouth Alcohol Solution Lot # <u>2022-A</u> Acetone Stock Solution Lot # <u>2022-B</u>																																										
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A					<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP5086</td></tr><tr><td>Interferent</td><td>MP5087</td></tr><tr><td>0.050</td><td>MP5088</td></tr><tr><td>0.080</td><td>MP5089</td></tr><tr><td>0.200</td><td>MP5084</td></tr></tbody></table>			Simulator	Serial Number	0.000	MP5086	Interferent	MP5087	0.050	MP5088	0.080	MP5089	0.200	MP5084
Simulator	Serial #	Lot #	Expiration																																											
0.000		N/A	N/A																																											
0.040																																														
0.100																																														
0.200																																														
0.300																																														
0.080 DGS	N/A																																													
Simulator	Serial Number																																													
0.000	MP5086																																													
Interferent	MP5087																																													
0.050	MP5088																																													
0.080	MP5089																																													
0.200	MP5084																																													
☐ Post Calibration Adjustment Stability Checks				Attachments																																										
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A					<table border="1"><tbody><tr><td>☒ Form 41</td><td>☐ Post-Stability Checks</td></tr><tr><td>☒ Stability Checks</td><td>☐ Flow Calibration</td></tr><tr><td>☒ Calibration Certificate</td><td>☐ Form 40</td></tr><tr><td>☐ Calibration Adjustment</td><td>☒ Other <u>Additional DGS stability checks.</u></td></tr></tbody></table>			☒ Form 41	☐ Post-Stability Checks	☒ Stability Checks	☐ Flow Calibration	☒ Calibration Certificate	☐ Form 40	☐ Calibration Adjustment	☒ Other <u>Additional DGS stability checks.</u>												
Simulator	Serial #	Lot #	Expiration																																											
0.050																																														
0.080																																														
0.200																																														
0.080 DGS	N/A																																													
☒ Form 41	☐ Post-Stability Checks																																													
☒ Stability Checks	☐ Flow Calibration																																													
☒ Calibration Certificate	☐ Form 40																																													
☐ Calibration Adjustment	☒ Other <u>Additional DGS stability checks.</u>																																													
Notes/Suggested Service: <u>Ten additional stability checks were performed</u> <u>using the dry gas standard (see Other attachment).</u> <u>All values obtained were within acceptable range.</u> <u>-PN (11/18)</u>				☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use <u>2022.11.18</u>																																										
				Israel Soto Digitally signed by Israel Soto Date: 2022.11.18 10:58:15 -05'00' 18 Tech Review / Date <u>15:25:41</u> Admin Review / Date <u>05:00</u>																																										

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: ST JOHNS SO

Time of Inspection: 10:27

Date of Inspection: 11/18/2022

Serial Number: 80-007363

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG229803 Exp: 10/25/2024
0.000	0.051	0.081	0.203	0.080
0.000	0.051	0.082	0.203	0.080
0.000	0.052	0.081	0.203	0.079
0.000	0.051	0.081	0.203	0.079
0.000	0.052	0.081	0.203	0.080
0.000	0.051	0.082	0.203	0.080
0.000	0.052	0.081	0.203	0.080
0.000	0.051	0.082	0.203	0.080
0.000	0.051	0.081	0.203	0.079
0.000	0.052	0.082	0.203	0.079


Standard Deviations	0.0005	0.0005	0.0000	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



PHIL NICODEMUS
Signature and Printed Name

11/18/2022
Date

Type of Test	Serial Number	Agency	Date	Performed By
Stability Checks	80-007363	St Johns Sheriff's Office	11/18/2022	PN

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083
ST JOHNS SO Intoxilyzer - Alcohol Analyzer Model 8000 11/18/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 08:34 Control Test 0.051 08:34 Air Blank 0.000 08:35 Control Test 0.051 08:36 Air Blank 0.000 08:36 Control Test 0.051 08:37 Air Blank 0.000 08:37 Control Test Stats Average 0.0510 Std Dev 0.0000 Rel Std Dev(%) 0.0000	ST JOHNS SO Intoxilyzer - Alcohol Analyzer Model 8000 11/18/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 08:38 Control Test 0.081 08:39 Air Blank 0.000 08:39 Control Test 0.081 08:40 Air Blank 0.000 08:41 Control Test 0.080 08:41 Air Blank 0.000 08:42 Control Test Stats Average 0.0807 Std Dev 0.0006 Rel Std Dev(%) 0.7157	ST JOHNS SO Intoxilyzer - Alcohol Analyzer Model 8000 11/18/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 08:43 Control Test 0.203 08:43 Air Blank 0.000 08:44 Control Test 0.203 08:45 Air Blank 0.000 08:45 Control Test 0.203 08:46 Air Blank 0.000 08:46 Control Test Stats Average 0.2030 Std Dev 0.0000 Rel Std Dev(%) 0.0000	ST JOHNS SO Intoxilyzer - Alcohol Analyzer Model 8000 11/18/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 08:48 Control Test 0.080 08:48 Air Blank 0.000 08:49 Control Test 0.079 08:49 Air Blank 0.000 08:49 Control Test 0.080 08:50 Air Blank 0.000 08:50 Control Test Stats Average 0.0797 Std Dev 0.0006 Rel Std Dev(%) 0.7247
ST JOHNS SO Intoxilyzer - Alcohol Analyzer Model 8000 11/18/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 08:34 Control Test 0.051 08:34 Air Blank 0.000 08:35 Control Test 0.051 08:36 Air Blank 0.000 08:36 Control Test 0.051 08:37 Air Blank 0.000 08:37 Control Test Stats Average 0.0510 Std Dev 0.0000 Rel Std Dev(%) 0.0000	ST JOHNS SO Intoxilyzer - Alcohol Analyzer Model 8000 11/18/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 08:38 Control Test 0.081 08:39 Air Blank 0.000 08:39 Control Test 0.081 08:40 Air Blank 0.000 08:41 Control Test 0.080 08:41 Air Blank 0.000 08:42 Control Test Stats Average 0.0807 Std Dev 0.0006 Rel Std Dev(%) 0.7157	ST JOHNS SO Intoxilyzer - Alcohol Analyzer Model 8000 11/18/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 08:43 Control Test 0.203 08:43 Air Blank 0.000 08:44 Control Test 0.203 08:45 Air Blank 0.000 08:45 Control Test 0.203 08:46 Air Blank 0.000 08:46 Control Test Stats Average 0.2030 Std Dev 0.0000 Rel Std Dev(%) 0.0000	ST JOHNS SO Intoxilyzer - Alcohol Analyzer Model 8000 11/18/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 08:48 Control Test 0.080 08:48 Air Blank 0.000 08:49 Control Test 0.079 08:49 Air Blank 0.000 08:49 Control Test 0.080 08:50 Air Blank 0.000 08:50 Control Test Stats Average 0.0797 Std Dev 0.0006 Rel Std Dev(%) 0.7247
Operator's Signature	Operator's Signature	Operator's Signature	Operator's Signature

Comments: N/A



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road.
Suite B1032
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-007363, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-007363</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>ST JOHNS SO</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>11/18/2022</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>10:27</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Phil Nicodemo
Digitally signed by Phil
Nicodemo
Date: 2022.11.18 10:44:37 -05'00'

11/18/2022

Date

PHIL NICODEMO,
Department Inspector

FDLE/ATP Form 69 March 2022
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Additional Dry Gas Standard
Stability Checks

P.N (11/18/2022)

ST JOHNS SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007363
11/18/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:51
Control Test	0.080	08:52
Air Blank	0.000	08:52
Control Test	0.080	08:52
Air Blank	0.000	08:53
Control Test	0.080	08:53
Air Blank	0.000	08:54
Control Test	0.080	08:54
Air Blank	0.000	08:55
Control Test	0.080	08:55
Air Blank	0.000	08:55
Control Test	0.079	08:56
Air Blank	0.000	08:56
Control Test	0.079	08:56
Air Blank	0.000	08:57
Control Test	0.080	08:57
Air Blank	0.000	08:58
Control Test	0.081	08:58
Air Blank	0.000	08:59
Control Test	0.080	08:59
Air Blank	0.000	08:59
Control Test Stats		
Average	0.0799	
Std Dev	0.0006	
Rel Std Dev(%)	0.7104	


Operator's Signature