



INSTRUMENT PROCESSING SHEET

Agency Suwannee County Sheriff's OfficeS/N 80-007382Florida Department of
Law EnforcementDate In 10/3/2022 DI Completion Date 10/3/2022☐ Ship ☒ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By PN	Quality Checks	By PN	Date <u>10/3/2022</u>	Flow Calibration	By	Date																																							
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes: _____ _____ _____ _____ _____ _____ _____ _____ _____ _____		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>257</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP 105</u> 32 mm <u>.167</u> (.139 - .169) 36 mm <u>.187</u> (.156 - .190) 53 mm <u>.253</u> (.228 - .278) 103 mm <u>.507</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28662</u> ☒ Stability Checks			☐ Flow Column # _____ ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																									
		<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP5088</td><td>202201C 01/11/2024</td></tr><tr><td>0.080</td><td>MP5089</td><td>202201D 01/18/2024</td></tr><tr><td>0.200</td><td>MP5084</td><td>202201E 01/18/2024</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>08121080A1 05/05/2023</td></tr></tbody></table>	Simulator	Serial #	Lot #/Exp	0.050	MP5088	202201C 01/11/2024	0.080	MP5089	202201D 01/18/2024	0.200	MP5084	202201E 01/18/2024	0.080 DGS	N/A	08121080A1 05/05/2023																													
Simulator	Serial #	Lot #/Exp																																												
0.050	MP5088	202201C 01/11/2024																																												
0.080	MP5089	202201D 01/18/2024																																												
0.200	MP5084	202201E 01/18/2024																																												
0.080 DGS	N/A	08121080A1 05/05/2023																																												
Calibration Adjustment		By _____		Department Inspection			By PN _____																																							
Barometric Pressure Gauge _____ ID # _____				Barometric Pressure ID# <u>28662</u> Gauge <u>1014</u> Instrument <u>1015</u> Mouth Alcohol Solution Lot # <u>2022-A</u> Acetone Stock Solution Lot # <u>2021-C</u>																																										
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A					<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP5086</td></tr><tr><td>Interferent</td><td>MP5087</td></tr><tr><td>0.050</td><td>MP5088</td></tr><tr><td>0.080</td><td>MP5089</td></tr><tr><td>0.200</td><td>MP5084</td></tr></tbody></table>			Simulator	Serial Number	0.000	MP5086	Interferent	MP5087	0.050	MP5088	0.080	MP5089	0.200	MP5084
Simulator	Serial #	Lot #	Expiration																																											
0.000		N/A	N/A																																											
0.040																																														
0.100																																														
0.200																																														
0.300																																														
0.080 DGS	N/A																																													
Simulator	Serial Number																																													
0.000	MP5086																																													
Interferent	MP5087																																													
0.050	MP5088																																													
0.080	MP5089																																													
0.200	MP5084																																													
☐ Post Calibration Adjustment Stability Checks				Attachments																																										
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A					☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Calibration Adjustment ☐ Post-Stability Checks ☐ Flow Calibration ☒ Form 40 ☐ Other _____																						
Simulator	Serial #	Lot #	Expiration																																											
0.050																																														
0.080																																														
0.200																																														
0.080 DGS	N/A																																													
Notes/Suggested Service: _____ _____ _____ _____ _____ _____ _____ _____ _____				☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use 2022.10.																																										
				Israel Soto Digitally signed by Israel Soto Date: 2022.10.04 09:37:46 +04'00'																																										
				Tech Review / Date _____ Admin Review / Date <u>10:56:44</u> 04'00'																																										

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: SUWANNEE COUNTY SO
Time of Inspection: 13:12

Date of Inspection: 10/03/2022

Serial Number: 80-007382
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:08121080A1 Exp: 05/05/2023
0.000	0.049	0.078	0.200	0.080
0.000	0.049	0.079	0.200	0.079
0.000	0.049	0.079	0.201	0.079
0.000	0.050	0.079	0.201	0.079
0.000	0.049	0.079	0.201	0.079
0.000	0.049	0.079	0.201	0.079
0.000	0.050	0.079	0.201	0.079
0.000	0.050	0.079	0.201	0.079
0.000	0.050	0.079	0.201	0.079
0.000	0.050	0.079	0.201	0.079


Standard Deviations	0.0005	0.0003	0.0004	0.0003
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



PHIL NICODEMO

Signature and Printed Name

10/03/2022
Date

Type of Test	Serial Number	Agency	Date	Performed By
Stability Checks	80-007382	Suwannee County Sheriff's Office	10/3/2022	PN

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																																
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083																																																																																																																																																
SUWANNEE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 10/03/2022 Software: 8100.27 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>10:50</td></tr><tr><td>Control Test</td><td>0.049 *</td><td>10:51</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:51</td></tr><tr><td>Control Test</td><td>0.049 *</td><td>10:52</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:53</td></tr><tr><td>Control Test</td><td>0.049 *</td><td>10:53</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:54</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0490</td><td></td></tr><tr><td>Std Dev</td><td>0.0000</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr></tbody></table> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	10:50	Control Test	0.049 *	10:51	Air Blank	0.000	10:51	Control Test	0.049 *	10:52	Air Blank	0.000	10:53	Control Test	0.049 *	10:53	Air Blank	0.000	10:54	Control Test Stats			Average	0.0490		Std Dev	0.0000		Rel Std Dev(%)	0.0000		SUWANNEE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 10/03/2022 Software: 8100.27 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>10:55</td></tr><tr><td>Control Test</td><td>0.080 *</td><td>10:56</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:57</td></tr><tr><td>Control Test</td><td>0.079 *</td><td>10:57</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:58</td></tr><tr><td>Control Test</td><td>0.080 *</td><td>10:58</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:59</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0797 *</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.7247</td><td></td></tr></tbody></table> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	10:55	Control Test	0.080 *	10:56	Air Blank	0.000	10:57	Control Test	0.079 *	10:57	Air Blank	0.000	10:58	Control Test	0.080 *	10:58	Air Blank	0.000	10:59	Control Test Stats			Average	0.0797 *		Std Dev	0.0006		Rel Std Dev(%)	0.7247		SUWANNEE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 10/03/2022 Software: 8100.27 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>11:01</td></tr><tr><td>Control Test</td><td>0.201 *</td><td>11:02</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:02</td></tr><tr><td>Control Test</td><td>0.201 *</td><td>11:03</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:03</td></tr><tr><td>Control Test</td><td>0.201 *</td><td>11:04</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:05</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.2010</td><td></td></tr><tr><td>Std Dev</td><td>0.0000</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr></tbody></table> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	11:01	Control Test	0.201 *	11:02	Air Blank	0.000	11:02	Control Test	0.201 *	11:03	Air Blank	0.000	11:03	Control Test	0.201 *	11:04	Air Blank	0.000	11:05	Control Test Stats			Average	0.2010		Std Dev	0.0000		Rel Std Dev(%)	0.0000		SUWANNEE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 10/03/2022 Software: 8100.27 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>11:06</td></tr><tr><td>Control Test</td><td>0.080 *</td><td>11:06</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:07</td></tr><tr><td>Control Test</td><td>0.080</td><td>11:07</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:08</td></tr><tr><td>Control Test</td><td>0.079</td><td>11:08</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:09</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0797 *</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.7247</td><td></td></tr></tbody></table> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	11:06	Control Test	0.080 *	11:06	Air Blank	0.000	11:07	Control Test	0.080	11:07	Air Blank	0.000	11:08	Control Test	0.079	11:08	Air Blank	0.000	11:09	Control Test Stats			Average	0.0797 *		Std Dev	0.0006		Rel Std Dev(%)	0.7247	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	10:50																																																																																																																																																	
Control Test	0.049 *	10:51																																																																																																																																																	
Air Blank	0.000	10:51																																																																																																																																																	
Control Test	0.049 *	10:52																																																																																																																																																	
Air Blank	0.000	10:53																																																																																																																																																	
Control Test	0.049 *	10:53																																																																																																																																																	
Air Blank	0.000	10:54																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0490																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	10:55																																																																																																																																																	
Control Test	0.080 *	10:56																																																																																																																																																	
Air Blank	0.000	10:57																																																																																																																																																	
Control Test	0.079 *	10:57																																																																																																																																																	
Air Blank	0.000	10:58																																																																																																																																																	
Control Test	0.080 *	10:58																																																																																																																																																	
Air Blank	0.000	10:59																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0797 *																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7247																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	11:01																																																																																																																																																	
Control Test	0.201 *	11:02																																																																																																																																																	
Air Blank	0.000	11:02																																																																																																																																																	
Control Test	0.201 *	11:03																																																																																																																																																	
Air Blank	0.000	11:03																																																																																																																																																	
Control Test	0.201 *	11:04																																																																																																																																																	
Air Blank	0.000	11:05																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.2010																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	11:06																																																																																																																																																	
Control Test	0.080 *	11:06																																																																																																																																																	
Air Blank	0.000	11:07																																																																																																																																																	
Control Test	0.080	11:07																																																																																																																																																	
Air Blank	0.000	11:08																																																																																																																																																	
Control Test	0.079	11:08																																																																																																																																																	
Air Blank	0.000	11:09																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0797 *																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7247																																																																																																																																																		

Comments: N/A



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road.
Suite B1032
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-007382, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-007382</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>SUWANNEE COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>10/03/2022</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>13:12</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

FDLE/ATP Form 69 March 2022

Issuing Authority: Alcohol Testing Program

10/03/2022

Date

PHIL NICODEMO,
Department Inspector

Service • Integrity • Respect • Quality

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: SUWANNEE COUNTY SO
Time of Inspection: 09:48

Date of Inspection: 10/03/2022

Serial Number: 80-007382
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Simulators Used: _____

Remarks:

BYPASSED AI TO OPERATE INSTRUMENT/COMPLIANCE NOT DETERMINED

The above instrument ~~complies~~ (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



PHIL NICODEMO

Signature and Printed Name

10/03/2022
Date