



INSTRUMENT PROCESSING SHEET

Agency Florida Highway PatrolS/N 80-007455Florida Department of
Law EnforcementDate In 8/29/2022DI Completion Date 8/30/2022☐ Ship ☒ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By PN	Quality Checks	By PN	Date <u>8/30/2022</u>	Flow Calibration	By	Date																																								
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes: <u>Instrument hand delivered. -PN</u>		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>209</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP 105</u> 32 mm <u>0.148</u> (.139 - .169) 36 mm <u>0.164</u> (.156 - .190) 53 mm <u>0.242</u> (.228 - .278) 103 mm <u>0.511</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28662</u> ☒ Stability Checks			Flow Column # _____ ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																										
		<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP5088</td><td>202201C 01/11/2024</td></tr><tr><td>0.080</td><td>MP5089</td><td>202201D 01/18/2024</td></tr><tr><td>0.200</td><td>MP5084</td><td>202201E 01/18/2024</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>08121080A1 05/05/2023</td></tr></tbody></table>			Simulator	Serial #	Lot #/Exp	0.050	MP5088	202201C 01/11/2024	0.080	MP5089	202201D 01/18/2024	0.200	MP5084	202201E 01/18/2024	0.080 DGS	N/A	08121080A1 05/05/2023	Maintenance By _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____																											
Simulator	Serial #	Lot #/Exp																																													
0.050	MP5088	202201C 01/11/2024																																													
0.080	MP5089	202201D 01/18/2024																																													
0.200	MP5084	202201E 01/18/2024																																													
0.080 DGS	N/A	08121080A1 05/05/2023																																													
Calibration Adjustment By _____		Department Inspection By PN _____																																													
Barometric Pressure Gauge _____ ID # _____		Barometric Pressure ID# <u>28662/28662</u> Gauge <u>1011/1010</u> Instrument <u>1013/1012</u> Mouth Alcohol Solution Lot # <u>2022-A</u> Acetone Stock Solution Lot # <u>2021-C</u>																																													
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP5086</td></tr><tr><td>Interferent</td><td>MP5087</td></tr><tr><td>0.050</td><td>MP5088</td></tr><tr><td>0.080</td><td>MP5089</td></tr><tr><td>0.200</td><td>MP5084</td></tr></tbody></table>						Simulator	Serial Number	0.000	MP5086	Interferent	MP5087	0.050	MP5088	0.080	MP5089	0.200	MP5084
Simulator	Serial #	Lot #	Expiration																																												
0.000		N/A	N/A																																												
0.040																																															
0.100																																															
0.200																																															
0.300																																															
0.080 DGS	N/A																																														
Simulator	Serial Number																																														
0.000	MP5086																																														
Interferent	MP5087																																														
0.050	MP5088																																														
0.080	MP5089																																														
0.200	MP5084																																														
☐ Post Calibration Adjustment Stability Checks		Attachments ☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Calibration Adjustment ☐ Post-Stability Checks ☐ Flow Calibration ☐ Form 40 ☐ Other _____																																													
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																									
Simulator	Serial #	Lot #	Expiration																																												
0.050																																															
0.080																																															
0.200																																															
0.080 DGS	N/A																																														
Notes/Suggested Service: _____ <u>Instrument power flickered during MA/AF portion of inspection. Changed power cord and repeated inspection. -PN</u>		Israel Soto Digitally signed by Israel Soto Date: 2022.08.31 08:42:20 -0400 Date: 2022.08.31 13:17:21 -04'00'																																													
		Tech Review / Date _____ Admin Review / Date _____																																													

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FHP

Time of Inspection: 14:19

Date of Inspection: 08/30/2022

Serial Number: 80-007455

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:08121080A1 Exp: 05/05/2023
0.000	0.048	0.077	0.198	0.080
0.000	0.048	0.077	0.198	0.080
0.000	0.048	0.077	0.198	0.080
0.000	0.048	0.076	0.198	0.080
0.000	0.048	0.077	0.199	0.079
0.000	0.048	0.077	0.198	0.079
0.000	0.048	0.077	0.198	0.079
0.000	0.048	0.077	0.199	0.079
0.000	0.048	0.076	0.199	0.079
0.000	0.048	0.077	0.199	0.079

Standard Deviations	0.0000	0.0004	0.0005	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 PHIL NICODEMO
Signature and Printed Name

08/30/2022
Date

Type of Test	Serial Number	Agency	Date	Performed By
Stability Checks	80-007455	Florida Highway Patrol	8/30/2022	PN

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																																
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083																																																																																																																																																
✓	✓	✓	✓																																																																																																																																																
FHP Intoxilyzer - Alcohol Analyzer Model 8000 08/30/2022 Software: 8100.27 SN 80-007455 <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>08:39</td></tr><tr><td>Control Test</td><td>0.050</td><td>08:39</td></tr><tr><td>Air Blank</td><td>0.000</td><td>08:40</td></tr><tr><td>Control Test</td><td>0.049</td><td>08:41</td></tr><tr><td>Air Blank</td><td>0.000</td><td>08:41</td></tr><tr><td>Control Test</td><td>0.049</td><td>08:42</td></tr><tr><td>Air Blank</td><td>0.000</td><td>08:42</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0493</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>1.1703</td><td></td></tr></table>  Operator's Signature	Test	g/210L	Time	Air Blank	0.000	08:39	Control Test	0.050	08:39	Air Blank	0.000	08:40	Control Test	0.049	08:41	Air Blank	0.000	08:41	Control Test	0.049	08:42	Air Blank	0.000	08:42	Control Test Stats			Average	0.0493		Std Dev	0.0006		Rel Std Dev(%)	1.1703		FHP Intoxilyzer - Alcohol Analyzer Model 8000 08/30/2022 Software: 8100.27 SN 80-007455 <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>08:44</td></tr><tr><td>Control Test</td><td>0.078</td><td>08:45</td></tr><tr><td>Air Blank</td><td>0.000</td><td>08:45</td></tr><tr><td>Control Test</td><td>0.078</td><td>08:46</td></tr><tr><td>Air Blank</td><td>0.000</td><td>08:46</td></tr><tr><td>Control Test</td><td>0.078</td><td>08:47</td></tr><tr><td>Air Blank</td><td>0.000</td><td>08:48</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0780</td><td></td></tr><tr><td>Std Dev</td><td>0.0000</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr></table>  Operator's Signature	Test	g/210L	Time	Air Blank	0.000	08:44	Control Test	0.078	08:45	Air Blank	0.000	08:45	Control Test	0.078	08:46	Air Blank	0.000	08:46	Control Test	0.078	08:47	Air Blank	0.000	08:48	Control Test Stats			Average	0.0780		Std Dev	0.0000		Rel Std Dev(%)	0.0000		FHP Intoxilyzer - Alcohol Analyzer Model 8000 08/30/2022 Software: 8100.27 SN 80-007455 <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>08:50</td></tr><tr><td>Control Test</td><td>0.199</td><td>08:50</td></tr><tr><td>Air Blank</td><td>0.000</td><td>08:51</td></tr><tr><td>Control Test</td><td>0.199</td><td>08:51</td></tr><tr><td>Air Blank</td><td>0.000</td><td>08:52</td></tr><tr><td>Control Test</td><td>0.199</td><td>08:53</td></tr><tr><td>Air Blank</td><td>0.000</td><td>08:53</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.1990</td><td></td></tr><tr><td>Std Dev</td><td>0.0000</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr></table>  Operator's Signature	Test	g/210L	Time	Air Blank	0.000	08:50	Control Test	0.199	08:50	Air Blank	0.000	08:51	Control Test	0.199	08:51	Air Blank	0.000	08:52	Control Test	0.199	08:53	Air Blank	0.000	08:53	Control Test Stats			Average	0.1990		Std Dev	0.0000		Rel Std Dev(%)	0.0000		FHP Intoxilyzer - Alcohol Analyzer Model 8000 08/30/2022 Software: 8100.27 SN 80-007455 <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>08:54</td></tr><tr><td>Control Test</td><td>0.079</td><td>08:55</td></tr><tr><td>Air Blank</td><td>0.000</td><td>08:55</td></tr><tr><td>Control Test</td><td>0.079</td><td>08:56</td></tr><tr><td>Air Blank</td><td>0.000</td><td>08:56</td></tr><tr><td>Control Test</td><td>0.079</td><td>08:56</td></tr><tr><td>Air Blank</td><td>0.000</td><td>08:57</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0790</td><td></td></tr><tr><td>Std Dev</td><td>0.0000</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr></table>  Operator's Signature	Test	g/210L	Time	Air Blank	0.000	08:54	Control Test	0.079	08:55	Air Blank	0.000	08:55	Control Test	0.079	08:56	Air Blank	0.000	08:56	Control Test	0.079	08:56	Air Blank	0.000	08:57	Control Test Stats			Average	0.0790		Std Dev	0.0000		Rel Std Dev(%)	0.0000	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	08:39																																																																																																																																																	
Control Test	0.050	08:39																																																																																																																																																	
Air Blank	0.000	08:40																																																																																																																																																	
Control Test	0.049	08:41																																																																																																																																																	
Air Blank	0.000	08:41																																																																																																																																																	
Control Test	0.049	08:42																																																																																																																																																	
Air Blank	0.000	08:42																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0493																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	1.1703																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	08:44																																																																																																																																																	
Control Test	0.078	08:45																																																																																																																																																	
Air Blank	0.000	08:45																																																																																																																																																	
Control Test	0.078	08:46																																																																																																																																																	
Air Blank	0.000	08:46																																																																																																																																																	
Control Test	0.078	08:47																																																																																																																																																	
Air Blank	0.000	08:48																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0780																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	08:50																																																																																																																																																	
Control Test	0.199	08:50																																																																																																																																																	
Air Blank	0.000	08:51																																																																																																																																																	
Control Test	0.199	08:51																																																																																																																																																	
Air Blank	0.000	08:52																																																																																																																																																	
Control Test	0.199	08:53																																																																																																																																																	
Air Blank	0.000	08:53																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1990																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	08:54																																																																																																																																																	
Control Test	0.079	08:55																																																																																																																																																	
Air Blank	0.000	08:55																																																																																																																																																	
Control Test	0.079	08:56																																																																																																																																																	
Air Blank	0.000	08:56																																																																																																																																																	
Control Test	0.079	08:56																																																																																																																																																	
Air Blank	0.000	08:57																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0790																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		

Comments: N/A



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road.
Suite B1032
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-007455, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-007455</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>FHP</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>08/30/2022</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>14:19</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION


This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.
This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

08/30/2022

Date


PHIL NICODEMO,
Department Inspector

FDLE/ATP Form 69 March 2022
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality