




INSTRUMENT PROCESSING SHEET

Agency Florida Highway PatrolS/N 80-007459Florida Department of
Law EnforcementDate In 5/24/2022DI Completion Date 5/31/2022☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By <u>DERR</u>	Quality Checks	By <u>DERR</u>	Date <u>5/31/2022</u>	Flow Calibration	By	Date																																																											
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes: _____ _____ _____ _____ _____ _____ _____ _____		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>203</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP101</u> 32 mm <u>0.152</u> (.139 - .169) 36 mm <u>0.171</u> (.156 - .190) 53 mm <u>0.238</u> (.228 - .278) 103 mm <u>0.503</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28199</u> ☒ Stability Checks <table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP4863</td><td>202201C 01/11/2024</td></tr><tr><td>0.080</td><td>MP4864</td><td>202201D 01/18/2024</td></tr><tr><td>0.200</td><td>MP5097</td><td>202201E 01/18/2024</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG115904 06/08/2023</td></tr></tbody></table>	Simulator	Serial #	Lot #/Exp	0.050	MP4863	202201C 01/11/2024	0.080	MP4864	202201D 01/18/2024	0.200	MP5097	202201E 01/18/2024	0.080 DGS	N/A	AG115904 06/08/2023			Flow Column # _____ ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																														
Simulator	Serial #	Lot #/Exp																																																																
0.050	MP4863	202201C 01/11/2024																																																																
0.080	MP4864	202201D 01/18/2024																																																																
0.200	MP5097	202201E 01/18/2024																																																																
0.080 DGS	N/A	AG115904 06/08/2023																																																																
					Maintenance By _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____ _____ _____ _____ _____																																																													
Calibration Adjustment By _____		Department Inspection By <u>DERR</u>																																																																
Barometric Pressure Gauge _____ ID # _____ <table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table> ☐ Post Calibration Adjustment Stability Checks <table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Barometric Pressure Gauge ID# <u>26932</u> Gauge <u>1016</u> Instrument <u>1015</u> Mouth Alcohol Solution Lot # <u>2021-D</u> Acetone Stock Solution Lot # <u>2021-C</u> <table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>SD1014</td></tr><tr><td>Interferent</td><td>SD1017</td></tr><tr><td>0.050</td><td>MP4863</td></tr><tr><td>0.080</td><td>MP4864</td></tr><tr><td>0.200</td><td>MP5097</td></tr></tbody></table> Attachments ☒ Form 41 ☐ Post-Stability Checks ☒ Stability Checks ☐ Flow Calibration ☒ Calibration Certificate ☒ Form 40 ☐ Calibration Adjustment ☐ Other _____ ☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use Taylor Gutschow Digitally Signed by Taylor Gutschow Date: 2022.05.31 14:07:56 -0400  03 09:51:11 Tech Review / Date _____ Admin Review / Date _____					Simulator	Serial Number	0.000	SD1014	Interferent	SD1017	0.050	MP4863	0.080	MP4864	0.200	MP5097
Simulator	Serial #	Lot #	Expiration																																																															
0.000		N/A	N/A																																																															
0.040																																																																		
0.100																																																																		
0.200																																																																		
0.300																																																																		
0.080 DGS	N/A																																																																	
Simulator	Serial #	Lot #	Expiration																																																															
0.050																																																																		
0.080																																																																		
0.200																																																																		
0.080 DGS	N/A																																																																	
Simulator	Serial Number																																																																	
0.000	SD1014																																																																	
Interferent	SD1017																																																																	
0.050	MP4863																																																																	
0.080	MP4864																																																																	
0.200	MP5097																																																																	
Notes/Suggested Service: _____ _____ _____ _____ _____ _____ _____ _____																																																																		



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-007459, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-007459</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>FHP</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>05/31/2022</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>13:23</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement: Alcohol Testing Program.

05/31/2022

Date



DAVID E REYES-RIVERA,

Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FHP

Time of Inspection: 13:23

Date of Inspection: 05/31/2022

Serial Number: 80-007459

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG115904 Exp: 06/08/2023
0.000	0.048	0.078	0.197	0.080
0.000	0.048	0.078	0.197	0.079
0.000	0.048	0.078	0.197	0.079
0.000	0.049	0.078	0.197	0.079
0.000	0.049	0.078	0.197	0.079
0.000	0.049	0.078	0.197	0.080
0.000	0.049	0.078	0.197	0.079
0.000	0.048	0.078	0.197	0.079
0.000	0.049	0.078	0.197	0.079
0.000	0.048	0.078	0.198	0.079


Standard Deviations	0.0005	0.0000	0.0003	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


 Signature and Printed Name

DAVID E REYES-RIVERA

05/31/2022
 Date

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-007459	Florida Highway Patrol	5/31/2022	DERR <i>[Signature]</i>

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																																
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083																																																																																																																																																
FHP Intoxilyzer - Alcohol Analyzer Model 8000 05/31/2022 Software: 8100.27 SN 80-007459 <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>11:25</td></tr><tr><td>Control Test</td><td>0.049</td><td>11:26</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:27</td></tr><tr><td>Control Test</td><td>0.048</td><td>11:27</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:28</td></tr><tr><td>Control Test</td><td>0.048</td><td>11:29</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:29</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0483</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>1.1945</td><td></td></tr></table> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	11:25	Control Test	0.049	11:26	Air Blank	0.000	11:27	Control Test	0.048	11:27	Air Blank	0.000	11:28	Control Test	0.048	11:29	Air Blank	0.000	11:29	Control Test Stats			Average	0.0483		Std Dev	0.0006		Rel Std Dev(%)	1.1945		FHP Intoxilyzer - Alcohol Analyzer Model 8000 05/31/2022 Software: 8100.27 SN 80-007459 <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>11:30</td></tr><tr><td>Control Test</td><td>0.078</td><td>11:31</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:31</td></tr><tr><td>Control Test</td><td>0.078</td><td>11:32</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:32</td></tr><tr><td>Control Test</td><td>0.078</td><td>11:33</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:34</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0780</td><td></td></tr><tr><td>Std Dev</td><td>0.0000</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr></table> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	11:30	Control Test	0.078	11:31	Air Blank	0.000	11:31	Control Test	0.078	11:32	Air Blank	0.000	11:32	Control Test	0.078	11:33	Air Blank	0.000	11:34	Control Test Stats			Average	0.0780		Std Dev	0.0000		Rel Std Dev(%)	0.0000		FHP Intoxilyzer - Alcohol Analyzer Model 8000 05/31/2022 Software: 8100.27 SN 80-007459 <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>11:35</td></tr><tr><td>Control Test</td><td>0.197</td><td>11:35</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:36</td></tr><tr><td>Control Test</td><td>0.197</td><td>11:37</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:37</td></tr><tr><td>Control Test</td><td>0.197</td><td>11:38</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:38</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.1971</td><td></td></tr><tr><td>Std Dev</td><td>0.0000</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr></table> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	11:35	Control Test	0.197	11:35	Air Blank	0.000	11:36	Control Test	0.197	11:37	Air Blank	0.000	11:37	Control Test	0.197	11:38	Air Blank	0.000	11:38	Control Test Stats			Average	0.1971		Std Dev	0.0000		Rel Std Dev(%)	0.0000		FHP Intoxilyzer - Alcohol Analyzer Model 3000 05/31/2022 Software: 8100.27 SN 80-007459 <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>11:39</td></tr><tr><td>Control Test</td><td>0.080</td><td>11:39</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:40</td></tr><tr><td>Control Test</td><td>0.079</td><td>11:40</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:41</td></tr><tr><td>Control Test</td><td>0.080</td><td>11:41</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:42</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0797</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.7247</td><td></td></tr></table> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	11:39	Control Test	0.080	11:39	Air Blank	0.000	11:40	Control Test	0.079	11:40	Air Blank	0.000	11:41	Control Test	0.080	11:41	Air Blank	0.000	11:42	Control Test Stats			Average	0.0797		Std Dev	0.0006		Rel Std Dev(%)	0.7247	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	11:25																																																																																																																																																	
Control Test	0.049	11:26																																																																																																																																																	
Air Blank	0.000	11:27																																																																																																																																																	
Control Test	0.048	11:27																																																																																																																																																	
Air Blank	0.000	11:28																																																																																																																																																	
Control Test	0.048	11:29																																																																																																																																																	
Air Blank	0.000	11:29																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0483																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	1.1945																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	11:30																																																																																																																																																	
Control Test	0.078	11:31																																																																																																																																																	
Air Blank	0.000	11:31																																																																																																																																																	
Control Test	0.078	11:32																																																																																																																																																	
Air Blank	0.000	11:32																																																																																																																																																	
Control Test	0.078	11:33																																																																																																																																																	
Air Blank	0.000	11:34																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0780																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	11:35																																																																																																																																																	
Control Test	0.197	11:35																																																																																																																																																	
Air Blank	0.000	11:36																																																																																																																																																	
Control Test	0.197	11:37																																																																																																																																																	
Air Blank	0.000	11:37																																																																																																																																																	
Control Test	0.197	11:38																																																																																																																																																	
Air Blank	0.000	11:38																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1971																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	11:39																																																																																																																																																	
Control Test	0.080	11:39																																																																																																																																																	
Air Blank	0.000	11:40																																																																																																																																																	
Control Test	0.079	11:40																																																																																																																																																	
Air Blank	0.000	11:41																																																																																																																																																	
Control Test	0.080	11:41																																																																																																																																																	
Air Blank	0.000	11:42																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0797																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7247																																																																																																																																																		

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FHP

Time of Inspection: 11:14

Date of Inspection: 05/31/2022

Serial Number: 80-007459

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:


Number of Simulators Used: _____

Remarks:

COMPLIANCE NOT DETERMINED, AI NOT CONDUCTED.

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 _____
Signature and Printed Name

DAVID E REYES-RIVERA

05/31/2022
Date

Florida Department of Law Enforcement

Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FHP

Time of Inspection: 11:34

Date of Inspection: 11/01/2022

Serial Number: 80-007459

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:08121080A1 Exp: 05/05/2023
0.000	0.048	0.077	0.198	0.080
0.000	0.048	0.077	0.198	0.080
0.000	0.047	0.077	0.198	0.080
0.000	0.048	0.078	0.198	0.080
0.000	0.048	0.077	0.197	0.080
0.000	0.048	0.078	0.197	0.080
0.000	0.048	0.078	0.197	0.079
0.000	0.048	0.077	0.197	0.079
0.000	0.048	0.077	0.197	0.079
0.000	0.048	0.078	0.197	0.080


Standard Deviations	0.0003	0.0005	0.0005	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



PHIL NICODEMO

Signature and Printed Name

11/01/2022
Date

Type of Test	Serial Number	Agency	Date	Performed By
Stability Checks	80-007459	Florida Highway Patrol	11/1/2022	PN

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083
FHP Intoxilyzer - Alcohol Analyzer Model 8000 11/01/2022 Software: 8100.27 SN 80-007459 Test g/210L Time Air Blank 0.000 09:37 Control Test 0.048 09:37 Air Blank 0.000 09:38 Control Test 0.049 09:39 Air Blank 0.000 09:39 Control Test 0.049 09:40 Air Blank 0.000 09:40 Control Test Stats Average 0.0487 Std Dev 0.0006 Rel Std Dev(%) 1.1863	FHP Intoxilyzer - Alcohol Analyzer Model 8000 11/01/2022 Software: 8100.27 SN 80-007459 Test g/210L Time Air Blank 0.000 09:42 Control Test 0.078 09:43 Air Blank 0.000 09:44 Control Test 0.077 09:44 Air Blank 0.000 09:45 Control Test 0.078 09:46 Air Blank 0.000 09:46 Control Test Stats Average 0.0777 Std Dev 0.0006 Rel Std Dev(%) 0.7424	FHP Intoxilyzer - Alcohol Analyzer Model 8000 11/01/2022 Software: 8100.27 SN 80-007459 Test g/210L Time Air Blank 0.000 09:48 Control Test 0.198 09:48 Air Blank 0.000 09:49 Control Test 0.198 09:50 Air Blank 0.000 09:50 Control Test 0.198 09:51 Air Blank 0.000 09:51 Control Test Stats Average 0.1980 Std Dev 0.0000 Rel Std Dev(%) 0.0000	FHP Intoxilyzer - Alcohol Analyzer Model 8000 11/01/2022 Software: 8100.27 SN 80-007459 Test g/210L Time Air Blank 0.000 09:52 Control Test 0.079 09:52 Air Blank 0.000 09:53 Control Test 0.079 09:53 Air Blank 0.000 09:54 Control Test 0.080 09:54 Air Blank 0.000 09:55 Control Test Stats Average 0.0793 Std Dev 0.0006 Rel Std Dev(%) 0.7277
Operator's Signature	Operator's Signature	Operator's Signature	Operator's Signature

Comments: N/A



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road.
Suite B1032
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-007459, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-007459</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>FHP</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>11/01/2022</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>11:34</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Phil Nicodemo
Digitally signed by Phil
Nicodemo
Date: 2022.11.01 12:24:12
-04'00'

11/01/2022

Date

PHIL NICODEMO,

Department Inspector

FDLE/ATP Form 69 March 2022
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FHP

Time of Inspection: 08:29

Date of Inspection: 11/01/2022

Serial Number: 80-007459

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Simulators Used: _____

Remarks:

BYPASSED AI TO OPERATE INSTRUMENT COMPLIANCE NOT DETERMINED

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



PHIL NICODEMO
Signature and Printed Name

11/01/2022
Date