



Agency Bay County Sheriff's Office

S/N 80-007478

Date In 01-28-2022 DI Completion Date 01-28-2022

☐ Ship ☒ P/U ☐ H/D ☐ CMI ☐ EE

Calibration Adjustment

By _____

Barometric Pressure Gauge	ID #
---------------------------	------

☐ Post Calibration Adjustment Stability ChecksNotes/Suggested Service:

Department Inspection

By IS

Barometric Pressure ID# 30793

Gauge 1017	<u>Instrument 1015</u>
------------	------------------------

Mouth Alcohol Solution Lot # 2021-D

Acetone Stock Solution Lot # 2021-C

Attachments

- | | |
|---|--|
| ☒ Form 41 | ☐ Post-Stability Checks |
| ☒ Stability Checks | ☐ Flow Calibration |
| ☒ Calibration Certificate | ☐ Form 40 |
| ☐ Calibration Adjustment | ☐ Other |

- ☒ Instrument Complies with Chapter 11D-8, FAC
 - ☐ Instrument Does Not Comply with Chapter 11D-8, FAC
 - ☒ Return to/Place into Evidentiary Use
 - ☐ Remain Out of Evidentiary Use
 - ☒ Conduct an Agency Inspection Before Evidentiary Use

Taylor
Gutschow

Digitally signed by Taylor
Gutschow
Date: 2022.01.28 12:23:41
-0500

Tech Review / Date

Admin Review / Date

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: BAY COUNTY SO
Time of Inspection: 09:46

Date of Inspection: 01/28/2022

Serial Number: 80-007478
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202010A Exp: 10/05/2022	0.08g/210L Test (g/210L) Lot#:202010B Exp: 10/05/2022	0.20g/210L Test (g/210L) Lot#:202010D Exp: 10/06/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG113403 Exp: 05/14/2023
0.000	0.048	0.079	0.199	0.081
0.000	0.049	0.079	0.199	0.081
0.000	0.049	0.078	0.199	0.081
0.000	0.049	0.079	0.199	0.081
0.000	0.049	0.079	0.199	0.081
0.000	0.049	0.079	0.199	0.081
0.000	0.050	0.079	0.199	0.081
0.000	0.049	0.079	0.199	0.081
0.000	0.049	0.080	0.199	0.081
0.000	0.050	0.079	0.199	0.081
Standard Deviations	0.0005	0.0004	0.0000	0.0000

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0002 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Israel Soto ISRAEL SOTO
Signature and Printed Name

01/28/2022
Date

Stability Checks

BAY COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007478
01/28/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:59
Control Test	0.048	08:00
Air Blank	0.000	08:00
Control Test	0.048	08:01
Air Blank	0.000	08:01
Control Test	0.049	08:02
Air Blank	0.000	08:03
Control Test Stats		
Average	0.0483	
Std Dev	0.0006	
Rel Std Dev(%)	1.1945	

BAY COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007478
01/28/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:03
Control Test	0.078	08:04
Air Blank	0.000	08:05
Control Test	0.079	08:05
Air Blank	0.000	08:06
Control Test	0.078	08:07
Air Blank	0.000	08:07
Control Test Stats		
Average	0.0783	
Std Dev	0.0006	
Rel Std Dev(%)	0.7370	

BAY COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007478
01/28/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:08
Control Test	0.198	08:09
Air Blank	0.000	08:09
Control Test	0.198	08:10
Air Blank	0.000	08:11
Control Test	0.198	08:11
Air Blank	0.000	08:12
Control Test Stats		
Average	0.1980	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

wet

Operator's Signature

Operator's Signature

Operator's Signature

BAY COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007478
01/28/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:13
Control Test	0.081	08:14
Air Blank	0.000	08:14
Control Test	0.080	08:14
Air Blank	0.000	08:15
Control Test	0.081	08:15
Air Blank	0.000	08:16
Control Test Stats		
Average	0.0807	
Std Dev	0.0006	
Rel Std Dev(%)	0.7157	

Dry

Operator's Signature



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2729 Fort Knox Blvd.
Bldg. 2, Suite 1300
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-007478, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-007478</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>BAY COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>01/28/2022</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>09:46</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Israel Soto
Soto
Digitally signed by Israel
Date: 2022.01.28
11:40:50 -05'00'

01/28/2022

Date

ISRAEL SOTO,
Department Inspector

FDLE/ATP Form 69 January 2022

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality