

Return Material Authorization

Ship to: ☒ CMI, Inc.

☐ Enforcement Electronics

Shipment to repair facility authorized by: Haley Dendy on 4/20/2022

Items Returned: Instrument ☒ Supplies ☐ Other ☐ Describe: _____

Instrument Model: Intoxilyzer 8000 Serial Number: 80-007479

Bill To Address:

Hillsborough CSO

Attn: Haley Dendy

Ship to Address:

Florida Department of Law Enforcement

Fort Myers Regional Operations Center

Attn: Alcohol Testing Program

4700 Terminal Drive, Suite 1

Fort Myers, FL 33907

Reason for Return:

Instrument intermittently freezes on start-up screen during power-up. Verified no records are
on the instrument. Instrument is very new and has only been in service since late July 2021.

Please choose one of the following options:

☐ 1. I _____, authorize all repairs.

☐ 2. I _____, authorize repairs up to \$_____.

☒ 3. I require an estimate **BEFORE** any repairs will be authorized and/ or conducted.

Please contact: Name: Haley Dendy

Phone #: 813-317-5100

Email: hdendy@teamhcso.com

ATP Contact Name: Taylor Gutschow

ATP Email: TaylorGutschow@fdle.state.fl.us



INSTRUMENT PROCESSING SHEET

Agency Hillsborough CSO

S/N 80-007479

Florida Department of
Law Enforcement

Date In 4/20/2022

DI Completion Date

☒ Ship☐ P/U☐ H/D☒ CMI☐ EE

Intake	By TDG	Quality Checks	By	Date	Flow Calibration	By	Date																												
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: AI reports the instrument freezes on start-up screen while powering on		☐ Breath Tube Screen ☐ Replace External O-Rings ☐ Instrument Set Up Verified ☐ R-Value _____ ☐ Flow Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547) ☐ Barometric Pressure Check Gauge ID # _____ ☐ Stability Checks			Flow Column # _____ ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																														
		<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td></tr></tbody></table>			Simulator	Serial #	Lot #/Exp	0.050			0.080			0.200			0.080 DGS	N/A		<table border="1"><thead><tr><th colspan="2">Maintenance</th><th>By</th></tr></thead><tbody><tr><td colspan="2">☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____</td><td></td></tr></tbody></table>			Maintenance		By	☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____									
Simulator	Serial #	Lot #/Exp																																	
0.050																																			
0.080																																			
0.200																																			
0.080 DGS	N/A																																		
Maintenance		By																																	
☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____																																			
Calibration Adjustment		By		Department Inspection			By																												
Barometric Pressure Gauge _____ ID # _____				Barometric Pressure ID# _____																															
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A					Gauge _____ Instrument _____ Mouth Alcohol Solution Lot # _____ Acetone Stock Solution Lot # _____			
Simulator	Serial #	Lot #	Expiration																																
0.000		N/A	N/A																																
0.040																																			
0.100																																			
0.200																																			
0.300																																			
0.080 DGS	N/A																																		
☐ Post Calibration Adjustment Stability Checks				<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td></td></tr><tr><td>Interferent</td><td></td></tr><tr><td>0.050</td><td></td></tr><tr><td>0.080</td><td></td></tr><tr><td>0.200</td><td></td></tr></tbody></table>			Simulator	Serial Number	0.000		Interferent		0.050		0.080		0.200				Attachments ☐ Form 41 ☐ Stability Checks ☐ Calibration Certificate ☐ Calibration Adjustment ☐ Post-Stability Checks ☐ Flow Calibration ☐ Form 40 ☒ Other Form 51														
Simulator	Serial Number																																		
0.000																																			
Interferent																																			
0.050																																			
0.080																																			
0.200																																			
Notes/Suggested Service: Instrument intermittently freezes on start-up screen like AI described. Will send to repair (TDG)				☐ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☐ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☐ Conduct an Agency Inspection Before Evidentiary Use																															
				Tech Review / Date			Admin Review / Date																												

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-00 7479	Hillsborough CSO	01/27/2022	TDG

0.05g/210L		0.08g/210L		0.20g/210L		DGS 0.08g/210L																																																																																																																																																	
0.047 to 0.053		0.077 to 0.083		0.194 to 0.206		0.077 to 0.083																																																																																																																																																	
✓		✓		✓		✓																																																																																																																																																	
HILLSBOROUGH CO SO Intoxilyzer - Alcohol Analyzer Model 8000 01/27/2022 Software: 8100.27 SN 80-007479 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>09:33</td></tr><tr><td>Control Test</td><td>0.050</td><td>09:34</td></tr><tr><td>Air Blank</td><td>0.000</td><td>09:34</td></tr><tr><td>Control Test</td><td>0.050</td><td>09:35</td></tr><tr><td>Air Blank</td><td>0.000</td><td>09:35</td></tr><tr><td>Control Test</td><td>0.050</td><td>09:36</td></tr><tr><td>Air Blank</td><td>0.000</td><td>09:37</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0500</td><td></td></tr><tr><td>Std Dev</td><td>0.0000</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr></tbody></table>		Test	g/210L	Time	Air Blank	0.000	09:33	Control Test	0.050	09:34	Air Blank	0.000	09:34	Control Test	0.050	09:35	Air Blank	0.000	09:35	Control Test	0.050	09:36	Air Blank	0.000	09:37	Control Test Stats			Average	0.0500		Std Dev	0.0000		Rel Std Dev(%)	0.0000		HILLSBOROUGH CO SO Intoxilyzer - Alcohol Analyzer Model 8000 01/27/2022 Software: 8100.27 SN 80-007479 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>09:40</td></tr><tr><td>Control Test</td><td>0.080</td><td>09:41</td></tr><tr><td>Air Blank</td><td>0.000</td><td>09:42</td></tr><tr><td>Control Test</td><td>0.080</td><td>09:42</td></tr><tr><td>Air Blank</td><td>0.000</td><td>09:43</td></tr><tr><td>Control Test</td><td>0.080</td><td>09:44</td></tr><tr><td>Air Blank</td><td>0.000</td><td>09:44</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0800</td><td></td></tr><tr><td>Std Dev</td><td>0.0000</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr></tbody></table>		Test	g/210L	Time	Air Blank	0.000	09:40	Control Test	0.080	09:41	Air Blank	0.000	09:42	Control Test	0.080	09:42	Air Blank	0.000	09:43	Control Test	0.080	09:44	Air Blank	0.000	09:44	Control Test Stats			Average	0.0800		Std Dev	0.0000		Rel Std Dev(%)	0.0000		HILLSBOROUGH CO SO Intoxilyzer - Alcohol Analyzer Model 8000 01/27/2022 Software: 8100.27 SN 80-007479 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>09:48</td></tr><tr><td>Control Test</td><td>0.201</td><td>09:48</td></tr><tr><td>Air Blank</td><td>0.000</td><td>09:49</td></tr><tr><td>Control Test</td><td>0.202</td><td>09:50</td></tr><tr><td>Air Blank</td><td>0.000</td><td>09:50</td></tr><tr><td>Control Test</td><td>0.201</td><td>09:51</td></tr><tr><td>Air Blank</td><td>0.000</td><td>09:51</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.2013</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.2868</td><td></td></tr></tbody></table>		Test	g/210L	Time	Air Blank	0.000	09:48	Control Test	0.201	09:48	Air Blank	0.000	09:49	Control Test	0.202	09:50	Air Blank	0.000	09:50	Control Test	0.201	09:51	Air Blank	0.000	09:51	Control Test Stats			Average	0.2013		Std Dev	0.0006		Rel Std Dev(%)	0.2868		HILLSBOROUGH CO SO Intoxilyzer - Alcohol Analyzer Model 8000 01/27/2022 Software: 8100.27 SN 80-007479 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>09:25</td></tr><tr><td>Control Test</td><td>0.082</td><td>09:25</td></tr><tr><td>Air Blank</td><td>0.000</td><td>09:26</td></tr><tr><td>Control Test</td><td>0.081</td><td>09:26</td></tr><tr><td>Air Blank</td><td>0.000</td><td>09:27</td></tr><tr><td>Control Test</td><td>0.081</td><td>09:27</td></tr><tr><td>Air Blank</td><td>0.000</td><td>09:27</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0813</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.7099</td><td></td></tr></tbody></table>		Test	g/210L	Time	Air Blank	0.000	09:25	Control Test	0.082	09:25	Air Blank	0.000	09:26	Control Test	0.081	09:26	Air Blank	0.000	09:27	Control Test	0.081	09:27	Air Blank	0.000	09:27	Control Test Stats			Average	0.0813		Std Dev	0.0006		Rel Std Dev(%)	0.7099	
Test	g/210L	Time																																																																																																																																																					
Air Blank	0.000	09:33																																																																																																																																																					
Control Test	0.050	09:34																																																																																																																																																					
Air Blank	0.000	09:34																																																																																																																																																					
Control Test	0.050	09:35																																																																																																																																																					
Air Blank	0.000	09:35																																																																																																																																																					
Control Test	0.050	09:36																																																																																																																																																					
Air Blank	0.000	09:37																																																																																																																																																					
Control Test Stats																																																																																																																																																							
Average	0.0500																																																																																																																																																						
Std Dev	0.0000																																																																																																																																																						
Rel Std Dev(%)	0.0000																																																																																																																																																						
Test	g/210L	Time																																																																																																																																																					
Air Blank	0.000	09:40																																																																																																																																																					
Control Test	0.080	09:41																																																																																																																																																					
Air Blank	0.000	09:42																																																																																																																																																					
Control Test	0.080	09:42																																																																																																																																																					
Air Blank	0.000	09:43																																																																																																																																																					
Control Test	0.080	09:44																																																																																																																																																					
Air Blank	0.000	09:44																																																																																																																																																					
Control Test Stats																																																																																																																																																							
Average	0.0800																																																																																																																																																						
Std Dev	0.0000																																																																																																																																																						
Rel Std Dev(%)	0.0000																																																																																																																																																						
Test	g/210L	Time																																																																																																																																																					
Air Blank	0.000	09:48																																																																																																																																																					
Control Test	0.201	09:48																																																																																																																																																					
Air Blank	0.000	09:49																																																																																																																																																					
Control Test	0.202	09:50																																																																																																																																																					
Air Blank	0.000	09:50																																																																																																																																																					
Control Test	0.201	09:51																																																																																																																																																					
Air Blank	0.000	09:51																																																																																																																																																					
Control Test Stats																																																																																																																																																							
Average	0.2013																																																																																																																																																						
Std Dev	0.0006																																																																																																																																																						
Rel Std Dev(%)	0.2868																																																																																																																																																						
Test	g/210L	Time																																																																																																																																																					
Air Blank	0.000	09:25																																																																																																																																																					
Control Test	0.082	09:25																																																																																																																																																					
Air Blank	0.000	09:26																																																																																																																																																					
Control Test	0.081	09:26																																																																																																																																																					
Air Blank	0.000	09:27																																																																																																																																																					
Control Test	0.081	09:27																																																																																																																																																					
Air Blank	0.000	09:27																																																																																																																																																					
Control Test Stats																																																																																																																																																							
Average	0.0813																																																																																																																																																						
Std Dev	0.0006																																																																																																																																																						
Rel Std Dev(%)	0.7099																																																																																																																																																						
Operator's Signature		Operator's Signature		Operator's Signature		Operator's Signature																																																																																																																																																	

Comments:

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: HILLSBOROUGH CO SO
Time of Inspection: 11:39

Date of Inspection: 01/27/2022

Serial Number: 80-007479
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202010A Exp: 10/05/2022	0.08g/210L Test (g/210L) Lot#:202010B Exp: 10/05/2022	0.20g/210L Test (g/210L) Lot#:202010D Exp: 10/06/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG026705 Exp: 09/23/2022
0.000	0.050	0.080	0.201	0.081
0.000	0.050	0.080	0.201	0.081
0.000	0.050	0.080	0.201	0.081
0.000	0.050	0.080	0.201	0.081
0.000	0.049	0.080	0.201	0.081
0.000	0.050	0.080	0.201	0.080
0.000	0.050	0.080	0.200	0.080
0.000	0.050	0.080	0.201	0.081
0.000	0.050	0.080	0.201	0.081
0.000	0.050	0.080	0.201	0.081


Standard Deviations	0.0003	0.0000	0.0003	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0002 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 TAYLOR D GUTSCHOW
Signature and Printed Name

01/27/2022
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-007479, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-007479</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>HILLSBOROUGH CO SO</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>01/27/2022</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>11:39</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.


01/27/2022 Date
TAYLOR D GUTSCHOW,

Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

INSTRUMENT PROCESSING SHEET

Agency Hillsborough County Sheriff's Office

S/N 80-007479

Florida Department of
Law Enforcement

Date In 7/14/2022 DI Completion Date 7/25/2022

☐ Ship ☒ P/U ☐ H/D ☐ CMI ☐ EE

Intake		Quality Checks		Flow Calibration																
By DERR _____ Date 7/25/2022		By DERR _____ Date 7/25/2022		By _____ Date _____																
☒ Annual ☐ Registration ☒ Return from CMI / EE Visual Inspection: ☒ Case ☒ Keyboard ☒ Feet ☒ Ports ☒ Handle ☒ Dry Gas Shelf ☒ Breath Tube ☒ Screws Tight Other Equipment/ Accessories: ☒ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: _____ _____ _____ _____ _____ _____ _____ _____ _____ _____		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>179</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP101</u> 32 mm <u>0.167</u> (.139 - .169) 36 mm <u>0.183</u> (.156 - .190) 53 mm <u>0.253</u> (.228 - .278) 103 mm <u>0.511</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28199</u> ☒ Stability Checks		Flow Column # _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																
		<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>SD3963</td> <td>202201C 01/11/2024</td> </tr> <tr> <td>0.080</td> <td>SD3967</td> <td>202201D 01/18/2024</td> </tr> <tr> <td>0.200</td> <td>SD3968</td> <td>202201E 01/18/2024</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG115904 06/08/2023</td> </tr> </tbody> </table>		Simulator	Serial #	Lot #/Exp	0.050	SD3963	202201C 01/11/2024	0.080	SD3967	202201D 01/18/2024	0.200	SD3968	202201E 01/18/2024	0.080 DGS	N/A	AG115904 06/08/2023	Maintenance By _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____ _____ _____ _____ _____ _____	
Simulator	Serial #	Lot #/Exp																		
0.050	SD3963	202201C 01/11/2024																		
0.080	SD3967	202201D 01/18/2024																		
0.200	SD3968	202201E 01/18/2024																		
0.080 DGS	N/A	AG115904 06/08/2023																		

Calibration Adjustment				By <u>DERR</u>	
Barometric Pressure Gauge <u>1018</u> ID # <u>68639</u>				Department Inspection	
				Barometric Pressure ID# <u>26932</u>	
				Gauge <u>1018</u> Instrument <u>1018</u>	
				Mouth Alcohol Solution Lot # <u>2021-D</u>	
				Acetone Stock Solution Lot # <u>2021-C</u>	
Simulator		Serial #	Lot #	Expiration	
0.000		MP5095	N/A	N/A	
0.040		MP5096	21070	03/01/2023	
0.100		MP5098	21080	03/08/2023	
0.200		MP5100	20510	12/03/2022	
0.300		MP5101	21030	02/02/2023	
0.080 DGS		N/A	00521080A2	02/05/2023	
☒ Post Calibration Adjustment Stability Checks					
Simulator		Serial #	Lot #	Expiration	
0.050		SD3963	202201C	01/11/2024	
0.080		SD3967	202201D	01/18/2024	
0.200		MP5094	202201E	01/18/2024	
0.080 DGS		N/A	AG115904	06/08/2023	
Notes/Suggested Service: _____ Instrument was calibrated to bring values closer to nominal. Simulator SD3968 has a leak, replaced with MP5094. DERR _____ _____ _____ _____					
Attachments					
☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☒ Calibration Adjustment			☒ Post-Stability Checks ☐ Flow Calibration ☐ Form 40 ☐ Other _____		
☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use					
Israel Soto Digitally signed by Israel Soto Date: 2022.07.26 07:30:20 +04'00'		 2022.07.26 12:07:5 2:04'00"			
Tech Review / Date			Admin Review / Date		



Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

Calibration Certificate

This is to certify the calibration of Intoxilyzer 8000 serial number 80-007479, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-007479</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>HILLSBOROUGH CO SO</u>	0.050 g/210 L	0.004
Calibration Date:	<u>07/25/2022</u>	0.080 g/210 L	0.004
Calibration Time:	<u>15:03</u>	0.200 g/210 L	0.007
		0.080 g/210 L Dry Gas Control	0.005

All results are reported in g/210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/IEC 17025 standards.
This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

07/25/2022

Date


DAVID E REYES-RIVERA,

Department Inspector

Service • Integrity • Respect • Quality

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: HILLSBOROUGH CO SO

Time of Inspection: 15:03

Date of Inspection: 07/25/2022

Serial Number: 80-007479

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG115904 Exp: 06/08/2023
0.000	0.048	0.079	0.197	0.080
0.000	0.048	0.079	0.197	0.080
0.000	0.048	0.079	0.198	0.080
0.000	0.048	0.079	0.198	0.079
0.000	0.048	0.079	0.198	0.078
0.000	0.048	0.079	0.198	0.078
0.000	0.048	0.079	0.198	0.079
0.000	0.048	0.079	0.198	0.079
0.000	0.048	0.082	0.198	0.079
0.000	0.048	0.080	0.198	0.079

Standard Deviations	0.0000	0.0009	0.0004	0.0007
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0005 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

David E Reyes-Rivera Signature and Printed Name

07/25/2022
Date

Type of Test	Serial Number	Agency	Date	Performed By
Post Stabilities	80-007479	Hillsborough County Sheriff's Office	7/25/2022	DERR <i>[Signature]</i>

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																																
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083																																																																																																																																																
HILLSBOROUGH CO SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-007479 07/25/2022 Software: 8100.27 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>12:20</td></tr><tr><td>Control Test</td><td>0.048</td><td>12:21</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:21</td></tr><tr><td>Control Test</td><td>0.048</td><td>12:22</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:22</td></tr><tr><td>Control Test</td><td>0.049</td><td>12:23</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:24</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0483</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>1.1945</td><td></td></tr></tbody></table> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	12:20	Control Test	0.048	12:21	Air Blank	0.000	12:21	Control Test	0.048	12:22	Air Blank	0.000	12:22	Control Test	0.049	12:23	Air Blank	0.000	12:24	Control Test Stats			Average	0.0483		Std Dev	0.0006		Rel Std Dev(%)	1.1945		HILLSBOROUGH CO SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-007479 07/25/2022 Software: 8100.27 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.010</td><td>12:25</td></tr><tr><td>Control Test</td><td>0.079</td><td>12:25</td></tr><tr><td>Air Blank</td><td>0.010</td><td>12:26</td></tr><tr><td>Control Test</td><td>0.078</td><td>12:27</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:27</td></tr><tr><td>Control Test</td><td>0.079</td><td>12:28</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:28</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0787</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.7339</td><td></td></tr></tbody></table> Operator's Signature	Test	g/210L	Time	Air Blank	0.010	12:25	Control Test	0.079	12:25	Air Blank	0.010	12:26	Control Test	0.078	12:27	Air Blank	0.000	12:27	Control Test	0.079	12:28	Air Blank	0.000	12:28	Control Test Stats			Average	0.0787		Std Dev	0.0006		Rel Std Dev(%)	0.7339		HILLSBOROUGH CO SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-007479 07/25/2022 Software: 8100.27 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>12:30</td></tr><tr><td>Control Test</td><td>0.199</td><td>12:30</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:31</td></tr><tr><td>Control Test</td><td>0.198</td><td>12:32</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:32</td></tr><tr><td>Control Test</td><td>0.198</td><td>12:33</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:33</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.1983</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.2911</td><td></td></tr></tbody></table> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	12:30	Control Test	0.199	12:30	Air Blank	0.000	12:31	Control Test	0.198	12:32	Air Blank	0.000	12:32	Control Test	0.198	12:33	Air Blank	0.000	12:33	Control Test Stats			Average	0.1983		Std Dev	0.0006		Rel Std Dev(%)	0.2911		HILLSBOROUGH CO SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-007479 07/25/2022 Software: 8100.27 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>12:35</td></tr><tr><td>Control Test</td><td>0.078</td><td>12:35</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:36</td></tr><tr><td>Control Test</td><td>0.079</td><td>12:36</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:37</td></tr><tr><td>Control Test</td><td>0.079</td><td>12:37</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:37</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0787</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.7339</td><td></td></tr></tbody></table> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	12:35	Control Test	0.078	12:35	Air Blank	0.000	12:36	Control Test	0.079	12:36	Air Blank	0.000	12:37	Control Test	0.079	12:37	Air Blank	0.000	12:37	Control Test Stats			Average	0.0787		Std Dev	0.0006		Rel Std Dev(%)	0.7339	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:20																																																																																																																																																	
Control Test	0.048	12:21																																																																																																																																																	
Air Blank	0.000	12:21																																																																																																																																																	
Control Test	0.048	12:22																																																																																																																																																	
Air Blank	0.000	12:22																																																																																																																																																	
Control Test	0.049	12:23																																																																																																																																																	
Air Blank	0.000	12:24																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0483																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	1.1945																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.010	12:25																																																																																																																																																	
Control Test	0.079	12:25																																																																																																																																																	
Air Blank	0.010	12:26																																																																																																																																																	
Control Test	0.078	12:27																																																																																																																																																	
Air Blank	0.000	12:27																																																																																																																																																	
Control Test	0.079	12:28																																																																																																																																																	
Air Blank	0.000	12:28																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0787																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7339																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:30																																																																																																																																																	
Control Test	0.199	12:30																																																																																																																																																	
Air Blank	0.000	12:31																																																																																																																																																	
Control Test	0.198	12:32																																																																																																																																																	
Air Blank	0.000	12:32																																																																																																																																																	
Control Test	0.198	12:33																																																																																																																																																	
Air Blank	0.000	12:33																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1983																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.2911																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:35																																																																																																																																																	
Control Test	0.078	12:35																																																																																																																																																	
Air Blank	0.000	12:36																																																																																																																																																	
Control Test	0.079	12:36																																																																																																																																																	
Air Blank	0.000	12:37																																																																																																																																																	
Control Test	0.079	12:37																																																																																																																																																	
Air Blank	0.000	12:37																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0787																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7339																																																																																																																																																		

<<<< CHANNEL 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 1.4030 (0.0020)
Sample #2 = 1.3770 (0.0210)
Sample #3 = 1.3800 (0.0170)
Sample #4 = 1.3950 (0.0100)
Avg % Abs = 1.3840 (0.0160)
STD DEV = 0.0096 (0.0056)
REL STD DEV = 0.697 (34.799)

HILLSBOROUGH CO SO
Intoxilizer - Alcohol Analyzer
Model 8000
07/25/2022
SN 80-007479
11:24:54

Auto Calibration
Max Power Res Value = 77
Auto Range Res Value = 60
Sol Value = 0.100 g/210L ***
Fit value = 0.4762 mg/l %ZZZ
Samples Taken = 4, Discarded = 1
3um Io = 12706, 9um Io = 12944
<<<< CHANNEL 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 1.8580 (-0.0120)
Sample #2 = 1.8290 (0.0280)
Sample #3 = 1.8210 (0.0360)
Sample #4 = 1.8030 (0.0380)
Avg % Abs = 1.8177 (0.0340)
STD DEV = 0.0133 (0.0053)
REL STD DEV = 0.733 (15.563)

<<<< CHANNEL 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 3.3160 (0.0000)
Sample #2 = 3.2860 (0.0280)
Sample #3 = 3.2980 (0.0310)
Sample #4 = 3.2850 (0.0290)
Avg % Abs = 3.2897 (0.0293)
STD DEV = 0.0072 (0.0015)
REL STD DEV = 0.220 (5.207)
Sol Value = 0.200 g/210L ***
Fit value = 0.9524 mg/l %ZZZ
Samples Taken = 4, Discarded = 1
3um Io = 12701, 9um Io = 12942
<<<< CHANNEL 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 3.5200 (-0.0040)
Sample #2 = 3.5230 (-0.0040)
Sample #3 = 3.5050 (0.0020)
Sample #4 = 3.5440 (0.0060)
Avg % Abs = 3.5240 (0.0080)
STD DEV = 0.0195 (0.0131)
REL STD DEV = 0.554 (163.936)

Sol Value = 0.040 g/210L ***
Fit value = 0.1905 mg/l %ZZZ
Samples Taken = 4, Discarded = 1
3um Io = 12707, 9um Io = 12946
<<<< CHANNEL 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.7780 (-0.0050)
Sample #2 = 0.7660 (-0.0030)
Sample #3 = 0.7690 (0.0020)
Sample #4 = 0.7800 (-0.0090)
Avg % Abs = 0.7717 (-0.0033)
STD DEV = 0.0074 (0.0055)
REL STD DEV = 0.955 (165.227)

<<<< CHANNEL 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 6.3470 (0.0120)
Sample #2 = 6.3350 (0.0310)
Sample #3 = 6.3050 (0.0610)
Sample #4 = 6.3380 (0.0380)
Avg % Abs = 6.3260 (0.0433)
STD DEV = 0.0182 (0.0157)
REL STD DEV = 0.288 (36.219)

Sol Value = 0.300 g/210L ***
Fit value = 1.4286 mg/l %ZZZ
Samples Taken = 4, Discarded = 1
3um Io = 12699, 9um Io = 12939
<<<< CHANNEL 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 5.1680 (-0.0030)
Sample #2 = 5.1930 (-0.0020)
Sample #3 = 5.1760 (0.0290)
Sample #4 = 5.1910 (0.0080)
Avg % Abs = 5.1867 (0.0117)
STD DEV = 0.0093 (0.0158)
REL STD DEV = 0.179 (135.617)

<<<< CHANNEL 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 9.2270 (-0.0070)
Sample #2 = 9.2140 (0.0210)
Sample #3 = 9.1980 (0.0370)
Sample #4 = 9.2090 (0.0310)
Avg % Abs = 9.2070 (0.0297)
STD DEV = 0.0082 (0.0081)
REL STD DEV = 0.089 (27.246)

Optical Calibration	
SN:	80-007479
Agency:	Hillsborough CSO
Date:	7/25/2022
Quadratic Fit:	+/- 0.002g/210L
By:	DERR <i>DM</i>

***** AUTO CAL DATA *****
<<<< CHANNEL 1 >>>>
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.081
Std Dev = 0.02 Rel Std Dev = 28.05
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.772
Std Dev = 0.01 Rel Std Dev = 0.96
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 1.818
Std Dev = 0.01 Rel Std Dev = 0.73
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 3.524
Std Dev = 0.02 Rel Std Dev = 0.55
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 5.187
Std Dev = 0.01 Rel Std Dev = 0.18
Zero Order Coef = -207.90
First Order Coef = 2701.87
Second Order Coef = 17.69
Standard Deviation = 11.646366

<<<< CHANNEL 2 >>>>
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.088
Std Dev = 0.02 Rel Std Dev = 18.29
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 1.384
Std Dev = 0.01 Rel Std Dev = 0.70
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 3.290
Std Dev = 0.01 Rel Std Dev = 0.22
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 6.326
Std Dev = 0.02 Rel Std Dev = 0.29
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 9.207
Std Dev = 0.01 Rel Std Dev = 0.19
Zero Order Coef = -121.37
First Order Coef = 1439.61
Second Order Coef = 13.57
Standard Deviation = 5.429244

Solution Stats Quadratic Fit Chan 1
Act Fit Residual
g/210L g/210L g/210L
0.000 0.000 -0.0003
0.040 0.040 0.0004
0.100 0.100 0.0000
0.200 0.200 -0.0002
0.300 0.300 0.0001

Solution Stats Quadratic Fit Chan 2
Act Fit Residual
g/210L g/210L g/210L
0.000 0.000 -0.0001
0.040 0.040 0.0002
0.100 0.100 0.0000
0.200 0.200 -0.0001
0.300 0.300 0.0000

Sol Value = 0.080 g/210L ***
Fit value = 0.3810 mg/l %ZZZ
Samples Taken = 4, Discarded = 1
***** CHANNEL 1 *****
Sample #1 = 3208.00
Sample #2 = 3183.00
Sample #3 = 3242.00
Sample #4 = 3205.00
Average Result = 3210.0000
STD DEV = 29.8161
REL STD DEV = 0.929
***** CHANNEL 2 *****
Sample #1 = 3419.00
Sample #2 = 3403.00
Sample #3 = 3440.00
Sample #4 = 3406.00
Average Result = 3416.3333
STD DEV = 20.5508
REL STD DEV = 0.602
***** CHANNEL 1 *****
Dry Gas H2O Adjust Results *****
Barometric Pressure = 1018
3 um H2O Adjust (mg/l x 10,000) = 599
9 um H2O Adjust (mg/l x 10,000) = 393
***** AUTO CAL PASS *****

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities Continued	80-007479	Hillsborough County Sheriff's Office	7/25/2022	DERR

☐	☐	☒	0.077 to 0.083																																				
HILLSBOROUGH CO SO Intoxilyzer - Alcohol Analyzer Model 8000 07/25/2022 Software: 8100.27 SN 80-007479 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>10:41</td></tr> <tr><td>Control Test</td><td>0.199</td><td>10:41</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:42</td></tr> <tr><td>Control Test</td><td>0.199</td><td>10:43</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:43</td></tr> <tr><td>Control Test</td><td>0.199</td><td>10:44</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:44</td></tr> <tr><td>Control Test</td><td>0.199</td><td>10:44</td></tr> <tr><td>Average</td><td>0.1990</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> Operator's Signature Simulator was changed to MP5094, there is a leak on SD3968. DERR 7/25/22				Test	g/210L	Time	Air Blank	0.000	10:41	Control Test	0.199	10:41	Air Blank	0.000	10:42	Control Test	0.199	10:43	Air Blank	0.000	10:43	Control Test	0.199	10:44	Air Blank	0.000	10:44	Control Test	0.199	10:44	Average	0.1990		Std Dev	0.0000		Rel Std Dev(%)	0.0000	
Test	g/210L	Time																																					
Air Blank	0.000	10:41																																					
Control Test	0.199	10:41																																					
Air Blank	0.000	10:42																																					
Control Test	0.199	10:43																																					
Air Blank	0.000	10:43																																					
Control Test	0.199	10:44																																					
Air Blank	0.000	10:44																																					
Control Test	0.199	10:44																																					
Average	0.1990																																						
Std Dev	0.0000																																						
Rel Std Dev(%)	0.0000																																						

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-007479	Hillsborough County Sheriff's Office	7/25/2022	DERR <i>[Signature]</i>

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																																
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083																																																																																																																																																
HILLSBOROUGH CO SO Intoxilyzer - Alcohol Analyzer Model 8000 07/25/2022 Software: 8100.27 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>09:41</td></tr><tr><td>Control Test</td><td>0.047</td><td>09:42</td></tr><tr><td>Air Blank</td><td>0.000</td><td>09:42</td></tr><tr><td>Control Test</td><td>0.048</td><td>09:43</td></tr><tr><td>Air Blank</td><td>0.000</td><td>09:43</td></tr><tr><td>Control Test</td><td>0.048</td><td>09:44</td></tr><tr><td>Air Blank</td><td>0.000</td><td>09:45</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0477</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>1.2112</td><td></td></tr></tbody></table> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	09:41	Control Test	0.047	09:42	Air Blank	0.000	09:42	Control Test	0.048	09:43	Air Blank	0.000	09:43	Control Test	0.048	09:44	Air Blank	0.000	09:45	Control Test Stats			Average	0.0477		Std Dev	0.0006		Rel Std Dev(%)	1.2112		HILLSBOROUGH CO SO Intoxilyzer - Alcohol Analyzer Model 8000 07/25/2022 Software: 8100.27 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>09:46</td></tr><tr><td>Control Test</td><td>0.078</td><td>09:47</td></tr><tr><td>Air Blank</td><td>0.000</td><td>09:47</td></tr><tr><td>Control Test</td><td>0.077</td><td>09:48</td></tr><tr><td>Air Blank</td><td>0.000</td><td>09:49</td></tr><tr><td>Control Test</td><td>0.078</td><td>09:49</td></tr><tr><td>Air Blank</td><td>0.000</td><td>09:50</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0777</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.7434</td><td></td></tr></tbody></table> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	09:46	Control Test	0.078	09:47	Air Blank	0.000	09:47	Control Test	0.077	09:48	Air Blank	0.000	09:49	Control Test	0.078	09:49	Air Blank	0.000	09:50	Control Test Stats			Average	0.0777		Std Dev	0.0006		Rel Std Dev(%)	0.7434		HILLSBOROUGH CO SO Intoxilyzer - Alcohol Analyzer Model 8000 07/25/2022 Software: 8100.27 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>09:51</td></tr><tr><td>Control Test</td><td>0.193</td><td>09:52</td></tr><tr><td>Air Blank</td><td>0.000</td><td>09:52</td></tr><tr><td>Control Test</td><td>0.192</td><td>09:53</td></tr><tr><td>Air Blank</td><td>0.000</td><td>09:54</td></tr><tr><td>Control Test</td><td>0.193</td><td>09:54</td></tr><tr><td>Air Blank</td><td>0.000</td><td>09:55</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.1927</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.2997</td><td></td></tr></tbody></table> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	09:51	Control Test	0.193	09:52	Air Blank	0.000	09:52	Control Test	0.192	09:53	Air Blank	0.000	09:54	Control Test	0.193	09:54	Air Blank	0.000	09:55	Control Test Stats			Average	0.1927		Std Dev	0.0006		Rel Std Dev(%)	0.2997		HILLSBOROUGH CO SO Intoxilyzer - Alcohol Analyzer Model 8000 07/25/2022 Software: 8100.27 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>09:58</td></tr><tr><td>Control Test</td><td>0.083</td><td>09:58</td></tr><tr><td>Air Blank</td><td>0.000</td><td>09:59</td></tr><tr><td>Control Test</td><td>0.082</td><td>09:59</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:00</td></tr><tr><td>Control Test</td><td>0.083</td><td>10:00</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:01</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0827</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.6984</td><td></td></tr></tbody></table> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	09:58	Control Test	0.083	09:58	Air Blank	0.000	09:59	Control Test	0.082	09:59	Air Blank	0.000	10:00	Control Test	0.083	10:00	Air Blank	0.000	10:01	Control Test Stats			Average	0.0827		Std Dev	0.0006		Rel Std Dev(%)	0.6984	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	09:41																																																																																																																																																	
Control Test	0.047	09:42																																																																																																																																																	
Air Blank	0.000	09:42																																																																																																																																																	
Control Test	0.048	09:43																																																																																																																																																	
Air Blank	0.000	09:43																																																																																																																																																	
Control Test	0.048	09:44																																																																																																																																																	
Air Blank	0.000	09:45																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0477																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	1.2112																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	09:46																																																																																																																																																	
Control Test	0.078	09:47																																																																																																																																																	
Air Blank	0.000	09:47																																																																																																																																																	
Control Test	0.077	09:48																																																																																																																																																	
Air Blank	0.000	09:49																																																																																																																																																	
Control Test	0.078	09:49																																																																																																																																																	
Air Blank	0.000	09:50																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0777																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7434																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	09:51																																																																																																																																																	
Control Test	0.193	09:52																																																																																																																																																	
Air Blank	0.000	09:52																																																																																																																																																	
Control Test	0.192	09:53																																																																																																																																																	
Air Blank	0.000	09:54																																																																																																																																																	
Control Test	0.193	09:54																																																																																																																																																	
Air Blank	0.000	09:55																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1927																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.2997																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	09:58																																																																																																																																																	
Control Test	0.083	09:58																																																																																																																																																	
Air Blank	0.000	09:59																																																																																																																																																	
Control Test	0.082	09:59																																																																																																																																																	
Air Blank	0.000	10:00																																																																																																																																																	
Control Test	0.083	10:00																																																																																																																																																	
Air Blank	0.000	10:01																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0827																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.6984																																																																																																																																																		