

INSTRUMENT PROCESSING SHEET

Agency Orange City Police Department

S/N 80-007491


Florida Department of
Law Enforcement

Date In 6/9/2022

DI Completion Date 6/16/2022

☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake By <u>DERR</u> ☐ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☐ Case ☐ Handle ☐ Keyboard ☐ Dry Gas Shelf ☐ Feet ☐ Breath Tube ☐ Ports ☐ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes: _____ _____ _____ _____ _____ _____ _____ _____ 	Quality Checks By <u>DERR</u> Date <u>6/16/2022</u> ☐ Breath Tube Screen ☐ Replace External O-Rings ☐ Instrument Set Up Verified ☐ R-Value <u>202</u> ☐ Flow Verification (L/s) Flow Column # <u>ATP101</u> 32 mm <u>0.156</u> (.139 - .169) 36 mm <u>0.171</u> (.156 - .190) 53 mm <u>0.238</u> (.228 - .278) 103 mm <u>0.500</u> (.447 - .547) ☐ Barometric Pressure Check Gauge ID # <u>28199</u> ☐ Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> <tr> <td>0.050</td> <td rowspan="2">MP4863</td> <td>202201C 01/11/2024</td> </tr> <tr> <td>0.080</td> <td>202201D 01/18/2024</td> </tr> <tr> <td>0.200</td> <td rowspan="2">MP5097</td> <td>202201E 01/18/2024</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A AG115904 06/08/2023</td> </tr> </table>	Simulator	Serial #	Lot #/Exp	0.050	MP4863	202201C 01/11/2024	0.080	202201D 01/18/2024	0.200	MP5097	202201E 01/18/2024	0.080 DGS	N/A AG115904 06/08/2023	Flow Calibration By _____ Date _____ Flow Column # _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547) Maintenance By _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____ _____ _____ _____ 																																																		
Simulator	Serial #	Lot #/Exp																																																															
0.050	MP4863	202201C 01/11/2024																																																															
0.080		202201D 01/18/2024																																																															
0.200	MP5097	202201E 01/18/2024																																																															
0.080 DGS		N/A AG115904 06/08/2023																																																															
Calibration Adjustment By _____ Barometric Pressure Gauge _____ ID # _____ <table border="1" style="width:100%; border-collapse: collapse;"> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </table> ☐ Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </table> Notes/Suggested Service: _____ _____ _____ _____ _____ _____ _____ 		Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Department Inspection By <u>DERR</u> Barometric Pressure ID# <u>26932</u> Gauge <u>1017</u> Instrument <u>1019</u> Mouth Alcohol Solution Lot # <u>2021-D</u> Acetone Stock Solution Lot # <u>2021-C</u> <table border="1" style="width:100%; border-collapse: collapse;"> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> <tr> <td>0.000</td> <td>SD1014</td> </tr> <tr> <td>Interferent</td> <td>SD1017</td> </tr> <tr> <td>0.050</td> <td>MP4863</td> </tr> <tr> <td>0.080</td> <td>MP4864</td> </tr> <tr> <td>0.200</td> <td>MP5097</td> </tr> </table> Attachments <table style="width:100%;"> <tr> <td style="width:50%; vertical-align: top;"> ☐ Form 41 ☐ Stability Checks ☐ Calibration Certificate ☐ Calibration Adjustment </td> <td style="width:50%; vertical-align: top;"> ☐ Post-Stability Checks ☐ Flow Calibration ☐ Form 40 ☐ Other _____ </td> </tr> </table> ☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use 2022.06. Taylor Gutschow Digitally signed by Taylor Gutschow Date: 2022.06.16 10:30:07 -0400 16 14:14:59 04 00 Tech Review / Date _____ Admin Review / Date _____		Simulator	Serial Number	0.000	SD1014	Interferent	SD1017	0.050	MP4863	0.080	MP4864	0.200	MP5097	☐ Form 41 ☐ Stability Checks ☐ Calibration Certificate ☐ Calibration Adjustment	☐ Post-Stability Checks ☐ Flow Calibration ☐ Form 40 ☐ Other _____
Simulator	Serial #	Lot #	Expiration																																																														
0.000		N/A	N/A																																																														
0.040																																																																	
0.100																																																																	
0.200																																																																	
0.300																																																																	
0.080 DGS	N/A																																																																
Simulator	Serial #	Lot #	Expiration																																																														
0.050																																																																	
0.080																																																																	
0.200																																																																	
0.080 DGS	N/A																																																																
Simulator	Serial Number																																																																
0.000	SD1014																																																																
Interferent	SD1017																																																																
0.050	MP4863																																																																
0.080	MP4864																																																																
0.200	MP5097																																																																
☐ Form 41 ☐ Stability Checks ☐ Calibration Certificate ☐ Calibration Adjustment	☐ Post-Stability Checks ☐ Flow Calibration ☐ Form 40 ☐ Other _____																																																																



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-007491, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-007491</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>ORANGE CITY P.D.</u>	0.050 g/210 L	0.004
Calibration Date:	<u>06/16/2022</u>	0.080 g/210 L	0.004
Calibration Time:	<u>09:19</u>	0.200 g/210 L	0.007
		0.080 g/210 L Dry Gas Control	0.005

All results are reported in g/210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

06/16/2022

Date

DAVID E REYES-RIVERA,

Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Page 1 of 1

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: ORANGE CITY P.D.

Time of Inspection: 09:19

Date of Inspection: 06/16/2022

Serial Number: 80-007491

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG115904 Exp: 06/08/2023
0.000	0.049	0.080	0.201	0.079
0.000	0.048	0.079	0.201	0.079
0.000	0.047	0.079	0.202	0.079
0.000	0.047	0.079	0.201	0.079
0.000	0.047	0.079	0.201	0.080
0.000	0.046	0.079	0.202	0.079
0.000	0.046	0.079	0.201	0.079
0.000	0.046	0.080	0.202	0.079
0.000	0.046	0.080	0.202	0.079
0.000	0.045	0.079	0.202	0.079

Standard Deviations	0.0011	0.0004	0.0005	0.0003
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0005 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

David E Reyes-Rivera DAVID E REYES-RIVERA
Signature and Printed Name

06/16/2022
Date

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-007491	Orange City Police Department	6/16/2022	DERR

0.05g/210L 0.047 to 0.053	0.08g/210L 0.077 to 0.083	0.20g/210L 0.194 to 0.206	DGS 0.08g/210L 0.077 to 0.083																																																																																																																																																
ORANGE CITY P.D. Intoxilyzer - Alcohol Analyzer Model 8000 06/16/2022 Software: 8100.27	ORANGE CITY P.D. Intoxilyzer - Alcohol Analyzer Model 8000 06/16/2022 Software: 8100.27	ORANGE CITY P.D. Intoxilyzer - Alcohol Analyzer Model 8000 06/16/2022 Software: 8100.27	ORANGE CITY P.D. Intoxilyzer - Alcohol Analyzer Model 8000 06/16/2022 Software: 8100.27																																																																																																																																																
<table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>06:48</td></tr><tr><td>Control Test</td><td>0.050</td><td>06:49</td></tr><tr><td>Air Blank</td><td>0.000</td><td>06:49</td></tr><tr><td>Control Test</td><td>0.049</td><td>06:50</td></tr><tr><td>Air Blank</td><td>0.000</td><td>06:50</td></tr><tr><td>Control Test</td><td>0.049</td><td>06:51</td></tr><tr><td>Air Blank</td><td>0.000</td><td>06:52</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0493</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>1.1703</td><td></td></tr></table>	Test	g/210L	Time	Air Blank	0.000	06:48	Control Test	0.050	06:49	Air Blank	0.000	06:49	Control Test	0.049	06:50	Air Blank	0.000	06:50	Control Test	0.049	06:51	Air Blank	0.000	06:52	Control Test Stats			Average	0.0493		Std Dev	0.0006		Rel Std Dev(%)	1.1703		<table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>06:53</td></tr><tr><td>Control Test</td><td>0.079</td><td>06:54</td></tr><tr><td>Air Blank</td><td>0.000</td><td>06:54</td></tr><tr><td>Control Test</td><td>0.079</td><td>06:55</td></tr><tr><td>Air Blank</td><td>0.000</td><td>06:56</td></tr><tr><td>Control Test</td><td>0.080</td><td>06:56</td></tr><tr><td>Air Blank</td><td>0.000</td><td>06:57</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0793</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.7277</td><td></td></tr></table>	Test	g/210L	Time	Air Blank	0.000	06:53	Control Test	0.079	06:54	Air Blank	0.000	06:54	Control Test	0.079	06:55	Air Blank	0.000	06:56	Control Test	0.080	06:56	Air Blank	0.000	06:57	Control Test Stats			Average	0.0793		Std Dev	0.0006		Rel Std Dev(%)	0.7277		<table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>06:58</td></tr><tr><td>Control Test</td><td>0.202</td><td>06:58</td></tr><tr><td>Air Blank</td><td>0.000</td><td>06:59</td></tr><tr><td>Control Test</td><td>0.200</td><td>07:00</td></tr><tr><td>Air Blank</td><td>0.000</td><td>07:00</td></tr><tr><td>Control Test</td><td>0.202</td><td>07:01</td></tr><tr><td>Air Blank</td><td>0.000</td><td>07:02</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.2013</td><td></td></tr><tr><td>Std Dev</td><td>0.0012</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.5735</td><td></td></tr></table>	Test	g/210L	Time	Air Blank	0.000	06:58	Control Test	0.202	06:58	Air Blank	0.000	06:59	Control Test	0.200	07:00	Air Blank	0.000	07:00	Control Test	0.202	07:01	Air Blank	0.000	07:02	Control Test Stats			Average	0.2013		Std Dev	0.0012		Rel Std Dev(%)	0.5735		<table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>07:02</td></tr><tr><td>Control Test</td><td>0.079</td><td>07:03</td></tr><tr><td>Air Blank</td><td>0.000</td><td>07:03</td></tr><tr><td>Control Test</td><td>0.079</td><td>07:03</td></tr><tr><td>Air Blank</td><td>0.000</td><td>07:04</td></tr><tr><td>Control Test</td><td>0.079</td><td>07:04</td></tr><tr><td>Air Blank</td><td>0.000</td><td>07:05</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0790</td><td></td></tr><tr><td>Std Dev</td><td>0.0000</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr></table>	Test	g/210L	Time	Air Blank	0.000	07:02	Control Test	0.079	07:03	Air Blank	0.000	07:03	Control Test	0.079	07:03	Air Blank	0.000	07:04	Control Test	0.079	07:04	Air Blank	0.000	07:05	Control Test Stats			Average	0.0790		Std Dev	0.0000		Rel Std Dev(%)	0.0000	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	06:48																																																																																																																																																	
Control Test	0.050	06:49																																																																																																																																																	
Air Blank	0.000	06:49																																																																																																																																																	
Control Test	0.049	06:50																																																																																																																																																	
Air Blank	0.000	06:50																																																																																																																																																	
Control Test	0.049	06:51																																																																																																																																																	
Air Blank	0.000	06:52																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0493																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	1.1703																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	06:53																																																																																																																																																	
Control Test	0.079	06:54																																																																																																																																																	
Air Blank	0.000	06:54																																																																																																																																																	
Control Test	0.079	06:55																																																																																																																																																	
Air Blank	0.000	06:56																																																																																																																																																	
Control Test	0.080	06:56																																																																																																																																																	
Air Blank	0.000	06:57																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0793																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7277																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	06:58																																																																																																																																																	
Control Test	0.202	06:58																																																																																																																																																	
Air Blank	0.000	06:59																																																																																																																																																	
Control Test	0.200	07:00																																																																																																																																																	
Air Blank	0.000	07:00																																																																																																																																																	
Control Test	0.202	07:01																																																																																																																																																	
Air Blank	0.000	07:02																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.2013																																																																																																																																																		
Std Dev	0.0012																																																																																																																																																		
Rel Std Dev(%)	0.5735																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	07:02																																																																																																																																																	
Control Test	0.079	07:03																																																																																																																																																	
Air Blank	0.000	07:03																																																																																																																																																	
Control Test	0.079	07:03																																																																																																																																																	
Air Blank	0.000	07:04																																																																																																																																																	
Control Test	0.079	07:04																																																																																																																																																	
Air Blank	0.000	07:05																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0790																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Operator's Signature	Operator's Signature	Operator's Signature	Operator's Signature																																																																																																																																																