



INSTRUMENT PROCESSING SHEET

Agency Highlands CSOS/N 80-007507Florida Department of
Law EnforcementDate In 12/15/2021 DI Completion Date 1/11/2022☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By TDG	Quality Checks	By TDG	Date 1/11/2022	Flow Calibration	By	Date	
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☒ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: AI reports Purge Fails and Ambient Fails after the MA part of the inspection. Instrument is recently registered (TDG)		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>235</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP101</u> 32 mm <u>0.164</u> (.139 - .169) 36 mm <u>0.179</u> (.156 - .190) 53 mm <u>0.246</u> (.228 - .278) 103 mm <u>0.507</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>68639</u> ☒ Stability Checks			Flow Column # _____ ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)			
					Maintenance By _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____			

Calibration Adjustment	By	Department Inspection	By TDG																																																												
Barometric Pressure Gauge _____ ID # _____ <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr> <tr><td>0.040</td><td></td><td></td><td></td></tr> <tr><td>0.100</td><td></td><td></td><td></td></tr> <tr><td>0.200</td><td></td><td></td><td></td></tr> <tr><td>0.300</td><td></td><td></td><td></td></tr> <tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr> </tbody> </table> ☐ Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr><td>0.050</td><td></td><td></td><td></td></tr> <tr><td>0.080</td><td></td><td></td><td></td></tr> <tr><td>0.200</td><td></td><td></td><td></td></tr> <tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr> </tbody> </table>	Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A				Barometric Pressure ID# <u>28199</u> Gauge <u>1025</u> Instrument <u>1022</u> Mouth Alcohol Solution Lot # <u>2021-B</u> Acetone Stock Solution Lot # <u>2020-A</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr><td>0.000</td><td>SD3963</td></tr> <tr><td>Interferent</td><td>SD1017</td></tr> <tr><td>0.050</td><td>MP5092</td></tr> <tr><td>0.080</td><td>MP4864</td></tr> <tr><td>0.200</td><td>MP5094</td></tr> </tbody> </table>	Simulator	Serial Number	0.000	SD3963	Interferent	SD1017	0.050	MP5092	0.080	MP4864	0.200	MP5094	
Simulator	Serial #	Lot #	Expiration																																																												
0.000		N/A	N/A																																																												
0.040																																																															
0.100																																																															
0.200																																																															
0.300																																																															
0.080 DGS	N/A																																																														
Simulator	Serial #	Lot #	Expiration																																																												
0.050																																																															
0.080																																																															
0.200																																																															
0.080 DGS	N/A																																																														
Simulator	Serial Number																																																														
0.000	SD3963																																																														
Interferent	SD1017																																																														
0.050	MP5092																																																														
0.080	MP4864																																																														
0.200	MP5094																																																														

Attachments	Notes/Suggested Service
☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Calibration Adjustment ☐ Post-Stability Checks ☐ Flow Calibration ☐ Form 40 ☐ Other _____	Notes/Suggested Service: Unable to reproduce the Purge Fail and Ambient Fail mentioned by the AI. Suggest adding a fan to promote air flow, as well as distancing from the instrument after providing a sample of mouth alcohol. If exception messages persist, will schedule a field visit to investigate further. (TDG)

☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use	David Eliezer Reyes-Rivera Tech Review / Date _____ Admin Review / Date _____
---	--

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-00 7507	Highlands CSO	1/11/2022	TDG <i>ML</i>

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083

HIGHLANDS COUNTY SO Intoxilyzer - Alcotest Analyzer Model 8000 01/11/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 10:33 Control Test 0.049 10:34 Air Blank 0.000 10:35 Control Test 0.048 10:35 Air Blank 0.000 10:36 Control Test 0.048 10:37 Air Blank 0.000 10:37 Control Test Stats Average 0.0483 Std Dev 0.0006 Rel Std Dev(%) 1.1945	HIGHLANDS COUNTY SO Intoxilyzer - Alcotest Analyzer Model 8000 01/11/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 10:40 Control Test 0.080 10:41 Air Blank 0.000 10:42 Control Test 0.079 10:42 Air Blank 0.000 10:43 Control Test 0.079 10:44 Air Blank 0.000 10:45 Control Test Stats Average 0.0793 Std Dev 0.0006 Rel Std Dev(%) 0.7277	HIGHLANDS COUNTY SO Intoxilyzer - Alcotest Analyzer Model 8000 01/11/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 10:48 Control Test 0.199 10:49 Air Blank 0.000 10:50 Control Test 0.198 10:50 Air Blank 0.000 10:51 Control Test 0.198 10:52 Air Blank 0.000 10:53 Control Test Stats Average 0.1983 Std Dev 0.0006 Rel Std Dev(%) 0.2911	HIGHLANDS COUNTY SO Intoxilyzer - Alcotest Analyzer Model 8000 01/11/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 10:54 Control Test 0.078 10:54 Air Blank 0.000 10:55 Control Test 0.078 10:55 Air Blank 0.000 10:56 Control Test 0.078 10:56 Air Blank 0.000 10:57 Control Test Stats Average 0.0780 Std Dev 0.0000 Rel Std Dev(%) 0.0000
<i>ML</i> Operator's Signature	<i>ML</i> Operator's Signature	<i>ML</i> Operator's Signature	<i>ML</i> Operator's Signature

Comments:

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: HIGHLANDS COUNTY SO
Time of Inspection: 13:23

Date of Inspection: 01/11/2022

Serial Number: 80-007507
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202010A Exp: 10/05/2022	0.08g/210L Test (g/210L) Lot#:202010B Exp: 10/05/2022	0.20g/210L Test (g/210L) Lot#:202010D Exp: 10/06/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG115904 Exp: 06/08/2023
0.000	0.050	0.079	0.199	0.079
0.000	0.049	0.080	0.199	0.080
0.000	0.049	0.080	0.199	0.079
0.000	0.050	0.080	0.199	0.079
0.000	0.049	0.081	0.199	0.079
0.000	0.050	0.080	0.199	0.079
0.000	0.050	0.081	0.200	0.079
0.000	0.050	0.081	0.199	0.079
0.000	0.050	0.081	0.200	0.079
0.000	0.051	0.081	0.200	0.079

Standard Deviations	0.0006	0.0006	0.0004	0.0003
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

01/11/2022
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-007507, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-007507</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>HIGHLANDS COUNTY SO</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>01/11/2022</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>13:23</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

01/11/2022

Date




TAYLOR D GUTSCHOW,

Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-007507, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-007507</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>HIGHLANDS COUNTY SO</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>05/09/2022</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>08:44</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

05/09/2022

Date

DAVID E REYES-RIVERA,

Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: HIGHLANDS COUNTY SO
Time of Inspection: 08:44

Date of Inspection: 05/09/2022

Serial Number: 80-007507
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG115904 Exp: 06/08/2023
0.000	0.049	0.079	0.200	0.080
0.000	0.048	0.079	0.200	0.080
0.000	0.049	0.079	0.200	0.080
0.000	0.049	0.079	0.200	0.080
0.000	0.049	0.080	0.200	0.080
0.000	0.049	0.078	0.200	0.080
0.000	0.049	0.079	0.200	0.080
0.000	0.049	0.080	0.200	0.080
0.000	0.049	0.079	0.200	0.080
0.000	0.049	0.080	0.201	0.080

Standard Deviations	0.0003	0.0006	0.0003	0.0000
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


DAVID E REYES-RIVERA

 Signature and Printed Name

05/09/2022
 Date

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-007507	Highlands County Sheriff's Office	5/9/2022	DERR <i>[Signature]</i>

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																																
0.047 to 0.053 ☒ HIGHLANDS COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-007507 05/09/2022 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>06:58</td></tr> <tr><td>Control Test</td><td>0.049</td><td>06:59</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:00</td></tr> <tr><td>Control Test</td><td>0.049</td><td>07:00</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:01</td></tr> <tr><td>Control Test</td><td>0.049</td><td>07:02</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:02</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0490</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> <i>[Signature]</i> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	06:58	Control Test	0.049	06:59	Air Blank	0.000	07:00	Control Test	0.049	07:00	Air Blank	0.000	07:01	Control Test	0.049	07:02	Air Blank	0.000	07:02	Control Test Stats			Average	0.0490		Std Dev	0.0000		Rel Std Dev(%)	0.0000		0.077 to 0.083 ☒ HIGHLANDS COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-007507 05/09/2022 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>07:03</td></tr> <tr><td>Control Test</td><td>0.079</td><td>07:04</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:04</td></tr> <tr><td>Control Test</td><td>0.079</td><td>07:05</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:06</td></tr> <tr><td>Control Test</td><td>0.079</td><td>07:06</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:07</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0790</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> <i>[Signature]</i> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	07:03	Control Test	0.079	07:04	Air Blank	0.000	07:04	Control Test	0.079	07:05	Air Blank	0.000	07:06	Control Test	0.079	07:06	Air Blank	0.000	07:07	Control Test Stats			Average	0.0790		Std Dev	0.0000		Rel Std Dev(%)	0.0000		0.194 to 0.206 ☒ HIGHLANDS COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-007507 05/09/2022 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>07:08</td></tr> <tr><td>Control Test</td><td>0.201</td><td>07:09</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:10</td></tr> <tr><td>Control Test</td><td>0.201</td><td>07:10</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:11</td></tr> <tr><td>Control Test</td><td>0.201</td><td>07:12</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:12</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.2007</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.2877</td><td></td></tr> </tbody> </table> <i>[Signature]</i> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	07:08	Control Test	0.201	07:09	Air Blank	0.000	07:10	Control Test	0.201	07:10	Air Blank	0.000	07:11	Control Test	0.201	07:12	Air Blank	0.000	07:12	Control Test Stats			Average	0.2007		Std Dev	0.0006		Rel Std Dev(%)	0.2877		0.077 to 0.083 ☒ HIGHLANDS COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-007507 05/09/2022 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>07:14</td></tr> <tr><td>Control Test</td><td>0.079</td><td>07:14</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:15</td></tr> <tr><td>Control Test</td><td>0.079</td><td>07:15</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:15</td></tr> <tr><td>Control Test</td><td>0.080</td><td>07:16</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:16</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0793</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7277</td><td></td></tr> </tbody> </table> <i>[Signature]</i> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	07:14	Control Test	0.079	07:14	Air Blank	0.000	07:15	Control Test	0.079	07:15	Air Blank	0.000	07:15	Control Test	0.080	07:16	Air Blank	0.000	07:16	Control Test Stats			Average	0.0793		Std Dev	0.0006		Rel Std Dev(%)	0.7277	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	06:58																																																																																																																																																	
Control Test	0.049	06:59																																																																																																																																																	
Air Blank	0.000	07:00																																																																																																																																																	
Control Test	0.049	07:00																																																																																																																																																	
Air Blank	0.000	07:01																																																																																																																																																	
Control Test	0.049	07:02																																																																																																																																																	
Air Blank	0.000	07:02																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0490																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	07:03																																																																																																																																																	
Control Test	0.079	07:04																																																																																																																																																	
Air Blank	0.000	07:04																																																																																																																																																	
Control Test	0.079	07:05																																																																																																																																																	
Air Blank	0.000	07:06																																																																																																																																																	
Control Test	0.079	07:06																																																																																																																																																	
Air Blank	0.000	07:07																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0790																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	07:08																																																																																																																																																	
Control Test	0.201	07:09																																																																																																																																																	
Air Blank	0.000	07:10																																																																																																																																																	
Control Test	0.201	07:10																																																																																																																																																	
Air Blank	0.000	07:11																																																																																																																																																	
Control Test	0.201	07:12																																																																																																																																																	
Air Blank	0.000	07:12																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.2007																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.2877																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	07:14																																																																																																																																																	
Control Test	0.079	07:14																																																																																																																																																	
Air Blank	0.000	07:15																																																																																																																																																	
Control Test	0.079	07:15																																																																																																																																																	
Air Blank	0.000	07:15																																																																																																																																																	
Control Test	0.080	07:16																																																																																																																																																	
Air Blank	0.000	07:16																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0793																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7277																																																																																																																																																		