



Agency Desoto Co. SO

S/N 80-007508

Date In 12/21/2021 DI Completion Date 1/3/2022

☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

FDLE/ATP Form 48 January 2021
Issuing Authority: Alcohol Testing Program

Page 1 of 1

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: DESOTO COUNTY S.O.
Time of Inspection: 17:06

Date of Inspection: 01/03/2022

Serial Number: 80-007508
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202010A Exp: 10/05/2022	0.08g/210L Test (g/210L) Lot#:202010B Exp: 10/05/2022	0.20g/210L Test (g/210L) Lot#:202010D Exp: 10/06/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG113403 Exp: 05/14/2023
0.000	0.048	0.078	0.198	0.000 / 0.080
0.000	0.048	0.079	0.198	0.000 / 0.080
0.000	0.049	0.078	0.198	0.000 / 0.080
0.000	0.048	0.079	0.198	0.080 / 0.080
0.000	0.048	0.079	0.198	0.080 / 0.080
0.000	0.049	0.079	0.198	0.080 / 0.080
0.000	0.048	0.078	0.199	0.080 / 0.079
0.000	0.048	0.079	0.198	0.080 / 0.080
0.000	0.049	0.079	0.198	0.080 / 0.080
0.000	0.048	0.079	0.198	0.080 / 0.080

Standard Deviations	0.0004	0.0004	0.0003	0.0386 / 0.0003
---------------------	--------	--------	--------	-----------------


Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

08: Control Outside Tolerance, FORGOT TO PLUG IN HOSE..

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



 Signature and Printed Name

RICHARD A WILLIAMS

01/03/2022
 Date

DESOTO COUNTY S.O.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007508
01/03/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:15
Control Test	0.047	14:15
Air Blank	0.000	14:16
Control Test	0.048	14:17
Air Blank	0.000	14:17
Control Test	0.048	14:18
Air Blank	0.000	14:18
Control Test Stats		
Average	0.0477	
Std Dev	0.0006	
Rel Std Dev(%)	1.2112	

05

PAW

Operator's Signature

DESOTO COUNTY S.O.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007508
01/03/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:20
Control Test	0.078	14:21
Air Blank	0.000	14:21
Control Test	0.079	14:22
Air Blank	0.000	14:22
Control Test	0.078	14:23
Air Blank	0.000	14:24
Control Test Stats		
Average	0.0783	
Std Dev	0.0006	
Rel Std Dev(%)	0.7370	

08
wet

PAW

Operator's Signature

DESOTO COUNTY S.O.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007508
01/03/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:25
Control Test	0.197	14:26
Air Blank	0.000	14:27
Control Test	0.198	14:27
Air Blank	0.000	14:28
Control Test	0.198	14:29
Air Blank	0.000	14:29
Control Test Stats		
Average	0.1977	
Std Dev	0.0006	
Rel Std Dev(%)	0.2921	

20

PAW

Operator's Signature

DESOTO COUNTY S.O.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007508
01/03/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:34
Control Test	0.080	14:34
Air Blank	0.000	14:35
Control Test	0.080	14:35
Air Blank	0.000	14:36
Control Test	0.079	14:36
Air Blank	0.000	14:37
Control Test Stats		
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	

08
DGS

PAW

Operator's Signature

Stability Checks
80-007508
01/03/22



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2729 Fort Knox Blvd.
Bldg. 2, Suite 1300
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-007508, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-007508</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>DESOTO COUNTY S.O.</u>	0.050 g/ 210 L	0.005
Calibration Date:	<u>01/03/2022</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>17:06</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Richard A

Digitally signed by
Richard A Williams

Williams

Date: 2022.01.04

14:06:45 -05:00

Date

01/03/2022

Date

RICHARD A WILLIAMS,

Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Agency Desoto County SO

S/N 80-007508

Date In 03/18/2022 DI Completion Date 04-05-2022

☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Calibration Adjustment

By _____

Barometric Pressure Gauge _____ ID #

☐ Post Calibration Adjustment Stability Checks

Notes/Suggested Service:

Department Inspection

By IS

Barometric Pressure ID# 28421

Gauge 1009	Instrument 1006
------------	-----------------

Mouth Alcohol Solution Lot # 2021-D

Acetone Stock Solution Lot # 2021-C

Attachments

- | | |
|---|--|
| ☒ Form 41 | ☐ Post-Stability Checks |
| ☒ Stability Checks | ☐ Flow Calibration |
| ☒ Calibration Certificate | ☒ Form 40 |
| ☐ Calibration Adjustment | ☐ Other |

- ☒ Instrument Complies with Chapter 11D-8, FAC
 - ☐ Instrument Does Not Comply with Chapter 11D-8, FAC
-
- ☒ Return to/Place into Evidentiary Use
 - ☐ Remain Out of Evidentiary Use
-
- ☒ Conduct an Agency Inspection Before Evidentiary Use

Taylor
Gutschow

Digitally signed by Taylor
Gutschow
Date: 2022.04.05 13:50:05 -04'00'

Tech Review / Date

Admin Review / Date

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: DESOTO COUNTY S.O.

Time of Inspection: 10:34

Date of Inspection: 04/05/2022

Serial Number: 80-007508

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG113403 Exp: 05/14/2023
0.000	0.050	0.080	0.200	0.080
0.000	0.049	0.079	0.200	0.080
0.000	0.049	0.080	0.200	0.080
0.000	0.049	0.080	0.200	0.080
0.000	0.049	0.080	0.201	0.080
0.000	0.049	0.080	0.200	0.080
0.000	0.050	0.079	0.200	0.079
0.000	0.049	0.080	0.200	0.079
0.000	0.050	0.080	0.200	0.080
0.000	0.049	0.080	0.200	0.080

Standard Deviations	0.0004	0.0004	0.0003	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Israel Soto

ISRAEL SOTO

Signature and Printed Name

04/05/2022

Date

stability checks

DESOTO COUNTY S.O.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007508
04/04/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:35
Control Test	0.048	14:35
Air Blank	0.000	14:36
Control Test	0.048	14:37
Air Blank	0.000	14:37
Control Test	0.049	14:38
Air Blank	0.000	14:38
Control Test Stats		
Average	0.0483	
Std Dev	0.0006	
Rel Std Dev(%)	1.1945	

Operator's Signature

DESOTO COUNTY S.O.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007508
04/04/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:40
Control Test	0.078	14:41
Air Blank	0.000	14:41
Control Test	0.079	14:42
Air Blank	0.000	14:43
Control Test	0.079	14:43
Air Blank	0.000	14:44
Control Test Stats		
Average	0.0787	
Std Dev	0.0006	
Rel Std Dev(%)	0.7339	

Operator's Signature

DESOTO COUNTY S.O.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007508
04/04/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:47
Control Test	0.199	14:48
Air Blank	0.000	14:49
Control Test	0.199	14:49
Air Blank	0.000	14:50
Control Test	0.198	14:51
Air Blank	0.000	14:51
Control Test Stats		
Average	0.1987	
Std Dev	0.0006	
Rel Std Dev(%)	0.2906	

Operator's Signature

DESOTO COUNTY S.O.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007508
04/04/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:52
Control Test	0.080	14:52
Air Blank	0.000	14:53
Control Test	0.079	14:53
Air Blank	0.000	14:54
Control Test	0.079	14:54
Air Blank	0.000	14:55
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

Operator's Signature



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road.
Suite B1032
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-007508, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-007508</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>DESOTO COUNTY S.O.</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>04/05/2022</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>10:34</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Israel Soto
Soto
Date: 2022.04.05
12:46:09 -04'00'

04/05/2022

Date

ISRAEL SOTO,

Department Inspector

FDLE/ATP Form 69 March 2022

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: DESOTO COUNTY S.O.

Time of Inspection: 14:02

Date of Inspection: 04/04/2022

Serial Number: 80-007508

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Simulators Used: _____

Remarks:

BYPASSED AI TO OPERATE INSTRUMENT

sd
n/A compliance not determined

The above instrument complies (☒) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Israel Soto

ISRAEL SOTO

Signature and Printed Name

04/04/2022

Date

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-007508	Pitt CSO Mr. Desoto CSO	06/09/2022	TDG MK

0.05g/210L		0.08g/210L		0.20g/210L		DGS 0.08g/210L	
0.047 to 0.053		0.077 to 0.083		0.194 to 0.206		0.077 to 0.083	
✓	✓	✓	✓	✓	✓	✓	✓
DESOTO COUNTY S.O. Intoxilyzer - Alcohol Analyzer Model 8000 06/09/2022 Software: 8100.27 SN 80-007508		DESOTO COUNTY S.O. Intoxilyzer - Alcohol Analyzer Model 8000 06/09/2022 Software: 8100.27 SN 80-007508		DESOTO COUNTY S.O. Intoxilyzer - Alcohol Analyzer Model 8000 06/09/2022 Software: 8100.27 SN 80-007508		DESOTO COUNTY S.O. Intoxilyzer - Alcohol Analyzer Model 8000 06/09/2022 Software: 8100.27 SN 80-007508	
Test g/210L Time		Test g/210L Time		Test g/210L Time		Test g/210L Time	
Air Blank 0.000 12:29		Air Blank 0.000 12:38		Air Blank 0.000 12:45		Air Blank 0.000 12:50	
Control Test 0.048 12:30		Control Test 0.077 12:39		Control Test 0.197 12:45		Control Test 0.079 12:50	
Air Blank 0.000 12:30		Air Blank 0.000 12:39		Air Blank 0.000 12:46		Air Blank 0.000 12:51	
Control Test 0.048 12:31		Control Test 0.077 12:40		Control Test 0.198 12:47		Control Test 0.078 12:51	
Air Blank 0.000 12:31		Air Blank 0.000 12:41		Air Blank 0.000 12:47		Air Blank 0.000 12:52	
Control Test 0.048 12:32		Control Test 0.077 12:41		Control Test 0.198 12:48		Control Test 0.079 12:52	
Air Blank 0.000 12:33		Air Blank 0.000 12:42		Air Blank 0.000 12:48		Air Blank 0.000 12:53	
Control Test Stats		Control Test Stats		Control Test Stats		Control Test Stats	
Average 0.0480		Average 0.0770		Average 0.1977		Average 0.0787	
Std Dev 0.0000		Std Dev 0.0000		Std Dev 0.0006		Std Dev 0.0006	
Rel Std Dev(%) 0.0000		Rel Std Dev(%) 0.0000		Rel Std Dev(%) 0.2921		Rel Std Dev(%) 0.7339	
Operator's Signature TDG		Operator's Signature TDG		Operator's Signature MK		Operator's Signature MK	

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: DESOTO COUNTY S.O.
Time of Inspection: 15:03

Date of Inspection: 06/09/2022

Serial Number: 80-007508
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG115904 Exp: 06/08/2023
0.000	0.047	0.077	0.197	0.079
0.000	0.048	0.077	0.198	0.079
0.000	0.048	0.077	0.198	0.079
0.000	0.048	0.077	0.197	0.079
0.000	0.048	0.078	0.198	0.079
0.000	0.048	0.077	0.198	0.079
0.000	0.048	0.078	0.198	0.079
0.000	0.048	0.077	0.198	0.079
0.000	0.049	0.078	0.199	0.079
0.000	0.049	0.078	0.199	0.079

Standard Deviations	0.0005	0.0005	0.0006	0.0000
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

06/09/2022
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-007508, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-007508</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>DESOTO COUNTY S.O.</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>06/09/2022</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>15:03</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

06/09/2022

Date

TAYLOR D GUTSCHOW,

Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality